

LVP-1000 Telemetry Transmitter User Guide

This guide provides information on the intended use of the DSI LVP-1000 Telemetry Transmitters. It also contains information relating to the requirements of 1999/5/EC (R&TTE Directive), FCC compliance and transmitter disposal. Please read this guide carefully and retain for future reference.

Intended Use:

The DSI LVP-1000 Telemetry Transmitter is intended for measuring physiologic parameters in laboratory animals. This device measures left ventricular pressure and temperature and should be placed subcutaneously, no more than 1.5 inches from the surface of the skin.

Declaration of Conformity:

These transmitters conform to the requirements of 1999/5/EC (R&TTE Directive). The declaration of conformity may be viewed on the Internet at www.transomamedical.com.

Use of these devices may be subject to licensing in the following countries. For licensing options please contact the national spectrum management authorities

Austria	Belgium	Denmark
France	Greece	Iceland
Ireland	Italy	Liechtenstein
Luxembourg	The Netherlands	Norway
Portugal	Spain	Switzerland

FCC Statement:

TX FCC ID MHALVP916 , IC:5681A-LVP916

These devices comply with part 15 of the FCC rules and Industry Canada RSS 210. Operation is subject to the following two conditions: (1) This device may not cause harmful interference and (2) this device must accept any interference that may cause undesired operation.

Caution:

Changes or modifications not expressly approved by the party responsible for compliance could void the users authority to operate the equipment.

Transmitter Disposal

Transmitters contain a primary battery cell. The battery cell may be removed by destructively dismantling the transmitter. Recycle the battery cell and remaining electronics in accordance with applicable federal, state, and local regulations.

