



Certificate of Quality

Catalog Number 226-30-18

Sorbent: XAD[®]-2 coated with Naphthylisothiocyanate
Tube Holder: Type A

**SKC is an ISO 9001 registered
manufacturing company.**

For compliance sampling, use these tubes as specified in a validated sampling method. It is the user's responsibility, employing a suitable method, to establish appropriate safety and health practices and to determine the applicability of regulatory limitations before use. The user should adjust the sampling procedures to specific conditions and should test tubes to ensure that the desired results will be obtained.

Buyer shall not be entitled to recover from SKC Inc. any consequential damages, damages to property, damages for loss of use, loss of time, loss of profits, or income or other incidental damages. Nor shall buyer be entitled to recover from SKC Inc. any consequential damages resulting from defect of the instrument including, but not limited to, any recovery under section 402A of the Restatement, Second of Torts.

The SKC Certificate of Quality assures that these sorbent tubes have been manufactured to meet or exceed the requirements for sorbent purity, sample collection and retention, storage stability, and desorption efficiency for the following chemical hazards and analytical methods published by NIOSH, OSHA or EPA:

<i>1,2- Diaminoethane</i>	<i>NIOSH 2540</i>
<i>Diethylenetriamine</i>	<i>OSHA 60</i>
<i>Ethylenediamine</i>	<i>NIOSH 2540 & OSHA 60</i>
<i>Triethylenetetramine</i>	<i>OSHA 60</i>

Form #37334 Rev 0903