



Certificate of Compliance

Certificate: 2225157 (LR 44340C)

Master Contract: 162029

Project: 2525517

Date Issued: August 1, 2012

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA

Attention: Thomas Smith

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cindy Johnson

Issued by: Cindy Johnson, AScT.

PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Digital Multimeters, Models 27 II, 28 II, 28 II SI, battery powered by 3 AA Alkaline batteries, NEDA 15A IEC LR6.

Notes:

- The above models are Equipment Class 2, Pollution Degree 2, Measurement Category: CAT III 1000V, CAT IV 600V, 10 A.
- Environmental Conditions: Normal: 2000 m max, Humidity: 0% to 95% (0°C to 35°C), 0% to 70% (35°C to 55°C), Temperature: -15°C to 55°C, -40°C for 20 minutes when taken from 20°C. 10 A continuous up to 35 °C. <20 minutes on, 5 minutes off at 35 °C to 55 °C. >10 A to 20 A for 30 seconds maximum, 5 minutes off.



Certificate: 2225157 (LR 44340C)

Master Contract: 162029

Project: 2525517

Date Issued: August 1, 2012

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-030-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

UL Std. No. 61010-1 (3rd ed.) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-030 (1st ed.) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

Reference Standards

IEC 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

IEC 61010-2-030:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits