

Manual Supplement

Manual Title: BP Pump 2 Operator Supplement Issue: **3**
Part Number: 2196592 Issue Date: 4/13
Print Date: June 2007 Page Count: 1
Revision/Date:

This supplement contains information
necessary to ensure the accuracy of
the above manual.

Change #1, 55935

On page 1-11, following **Pressure Generation** add:

Dynamic NIBP Simulation Repeatability.....±2 mmHg (at maximal pulse size independent of device under test)

Change #2, 56150, 56170

On page 1-11, under **Pressure Measurement**, change the High-Accuracy Version to the following:

High-Accuracy Version.....±0.7 mmHg (0.09 kPa)

Change #3, 63302, 65378

On page 1-4, delete the entire **General Safety Considerations** section.

On page 1-4, in the **Symbols** Table delete the **ETL** Symbol and add:

	Conforms to relevant North American Safety Standards.
	Conforms to relevant South Korean EMC Standards.

On page 1-10, following **Specifications**, delete **Mains Voltage**, **Environmental Conditions** and add the new section:

General Specifications

Power Line 100 - 240 V ac 50/60 Hz, 60 VA
Operating Temperature 15 °C to 40 °C
Storage Temperature -20 °C to +65 °C
Relative Humidity 90 % max
Operating Altitude 2000m
Safety IEC 61010-1: CAT II 300 V, Pollution Degree 2

Electromagnetic Environment IEC 61326-1: Basic

Electromagnetic Compatibility Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment)^[1]

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.