

Manual Supplement

Manual Title: 481 Operators
Print Date: April 2011
Revision/Date:

Supplement Issue: 1
Issue Date: 11/14
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 481
CD Rev. & Date: 1, 6/2011
CD PN: 4026903

Change #1, 381

On page 3, replace Batteries with and add the following to the **Specifications**:

- Batteries** Two 9 volt batteries (NEDA 1604A or 6LR61 or 6AM6) provide over 200 hours continuous operation. Three lithium cells provide chamber bias voltage of 63 V (10 year life expectancy)
- Altitude** 2000 m
For indoor use
- IP Rating** IEC 60529: IP40
- Safety** IEC 61010-1: Pollution Degree 2

Electromagnetic Compatibility (EMC) .. IEC 61326-1, Electromagnetic Environment: Portable

IEC CISPR 11: Group 1, Class A .- Group 1 ISM equipment: group 1 contains all ISM equipment in which there is intentionally generated and/or used conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself. Class A equipment *is* equipment suitable for use in all establishments other than domestic and those directly connected to a low voltage power-supply network which supplies buildings used for domestic purposes.

Korea (KCC) - Class A Equipment (Industrial Broadcasting & Communication Equipment) This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.

On page 13, replace table 2 with:

Table 2. Replaceable Parts List

Part Number	Description
4026926	FLUKE-481-8002,Top Case Decal
4026971	FLUKE-481-8004,Keypad
4026998	FLUKE-481-2002,Handle
3567776	Handle, Foam Grip, Black
3267644	Check Source, Model 450UCS
3569067	Wrist Strap
4400081	Cover Battery Box
3568830	Grill
614487	Battery, 9 V, Alkaline