

# **Certificate of Compliance**

Certificate: 2686208 Master Contract: 162029

**Issued to:** Fluke Corporation

6920 Seaway Blvd Everett, WA 98203

USA

**Attention: Joe Swanzy** 

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cíndy Johnson

**Issued by:** Cindy Johnson, AScT.

### **PRODUCTS**

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

**CLASS 3631 86** - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to

**US Standards** 

Pressure Calibrators and Modules, Models 719 Pro rated 12 Vdc - 8 AA batteries (IEC LR06), Model 721 rated 6Vdc - 4 AA batteries (IEC LR06) and Pressure Module Accessories Models 750Pxx and 750Rxx (where xx denotes the type and pressure rating).

Page: 1

## Notes:

- 1. The above models are Equipment Class II, Pollution Degree 2.
- 2. Mode of operation: Continuous.
- 3. Environmental Conditions: -10 °C to +50 °C, 2000 m max.

DQD 507 Rev. 2012-05-22



Certificate: 2686208 Master Contract: 162029

#### CONDITIONS OF ACCEPTABILITY

(1) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.

(2) Equipment is not to be used with flammable liquids.

## **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-030-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-030 (1st Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

#### Reference Standards:

IEC 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

IEC 61010-2-030:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

DQD 507 Rev. 2012-05-22 Page: 2