

USER GUIDE

RECEIVER-IN-EAR HEARING INSTRUMENT



ReSound



rediscover hearing

Hearing instrument type designations for models included in this user guide are: **SY312**, FCC ID: X26SY312, IC: 6941C-SY312; **SY312e**, FCC ID: X26SY312e, IC: 6941C-SY312e; and **MRIE**, FCC ID: X26MRIE, IC: 6941C-MRIE. **VE312**, FCC ID: X26VE312; IC: 6941C-VE312, Luxor. Please see page 6 for list of models referring to all types.

Statement:

This device complies with Part 15 of the FCC Rules and IC rules.

Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

Note: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules and ICES-003 of the IC rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures: Reorient or relocate the receiving antenna.

- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from the one in which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

Changes or modifications can void the user's authority to operate the equipment.

Intended use

Generic air-conduction hearing instruments are wearable sound-amplifying devices intended to compensate for impaired hearing. The fundamental operating principle of hearing instruments is to receive, amplify, and transfer sound to the ear drum of a hearing impaired person.

List of countries:

Products without wireless functionality are intended for worldwide sales.

Products with wireless functionality are intended for sale in countries within European Economic Area as well as Switzerland.

The products are in compliance with the following regulatory requirements:

- In EU: the device conforms to the Essential Requirements according to Annex I of Council Directive 93/42/EEC for medical devices (MDD) and essential requirements and other relevant provisions of Directive 1999/5/EC (R&TTE).
- The declaration of conformity may be consulted at www.resound.com
- In US: FCC CFR 47 Part 15, subpart C.
- In Canada: these hearing instruments are certified under the rules of IC
- Other identified applicable international regulatory requirements in countries outside EU and US. Please refer to local country requirements for these areas.

- Cet appareil numérique de la classe B est conforme à la norme NMB-003 du Canada.
- L'exploitation est autorisée aux deux conditions suivantes: (1) l'appareil ne doit pas produire de brouillage, et (2) l'utilisateur de l'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.
- Japanese Radio Law and Japanese Telecommunications Business Law Compliance.
This device is granted pursuant to the Japanese Radio Law (電波法) and the Japanese Telecommunications Business Law (電気通信事業法) This device should not be modified (otherwise the granted designation number will become invalid)

Introduction

Congratulations on the purchase of your new hearing instruments. ReSound's innovative sound technology and design, combined with the customised device programming selected by your hearing care professional, will make hearing a more enjoyable experience. Hearing instruments will enable you to hear sounds that you may not have heard in years because of your hearing loss. Practice and a positive attitude are important in learning to use hearing instruments. Your ReSound instruments have been adjusted according to your individual hearing loss and needs. Some people adjust quickly to wearing hearing instruments in their ears and hearing new sounds; other people may need more time.

Please read this manual carefully in order to wholly benefit from the use of your hearing instruments. With proper care, maintenance, and usage, your hearing instruments will aid you in better communication for many years. Ask your hearing care professional if you have any questions.

ReSound is a registered trade mark of ReSound A/S

Hearing instrument model: _____

- Model 60: Battery size 10
 Model 61: Battery size 312
 Model 62: Battery size 312

Receiver tube length: _____

Dome size: _____

Left serial number: _____

Right serial number: _____

Micro receiver-in-the-ear (RIE) hearing instruments with size 10A battery are available in the following variants:

**AL960-DR, AL760-DR, AL560-DR
ALT960-DR
VO960-DR, VO760-DR, VO560-DR,
VOT960-DR, VOT760-DR**

Mini receiver-in-the-ear (RIE) hearing instruments of type SY312 with FCC ID X26SY312, IC: 6941C-SY312 and size 312 battery are available in the following variants:

**AL961-DRW, AL761-DRW
AL561-DRW, AL461-DRW
ALT961-DRW**

Mini receiver-in-the-ear (RIE) hearing instruments of type SY312e with FCC ID X26SY312e, IC: 6941C-SY312e and size 312 battery are available in the following variants:

**VO961-DRW, VO761-DRW, VO561-DRW,
VOT961-DRW, VOT761-DRW**

Mini Receiver In-the-Ear (RIE) hearing instruments of type VE312 with FCC ID: X26VE312, IC: 6941C-VE312 and size 312 battery are available in the following variants:

LN961-DRW, LN761-DRW

Receiver-in-the-ear (RIE) hearing instruments of type MRIE with FCC ID X26MRIE, IC: 6941C-MRIE and size 312 battery are available in the following variants:




**AL962-DVIRW, AL762-DVIRW
AL562-DVIRW, AL462-DVIRW
ALT962-DVIRW, ALT762-DVIRW
VO962-DRW, VO762-DRW, VO562-DRW,
VOT962-DRW, VOT762-DRW**

The identification number for the MRIE, SY312, SY312e, and VE312 device models can be found at location “15” as indicated in the illustrations on page 9.



Contents

Introduction	4	Replacing the wax filter (receiver in-the-ear models).	27
Description of the hearing instruments	8	How to apply domes	28
Getting started	10	Using ReSound hearing instruments with smart phone apps	29
On/Off function	10	General precautions.	30
SmartStart	10	General warnings.	30
Inserting/Replacing the battery	10	Tinnitus Sound Generator (TSG) module	32
Low battery indicator	11	The scientific concepts that form the basis for the device	34
Sports lock.	12	Significant physical characteristics	34
Inserting/removing hearing instruments	12	Prescription use of this TSG hearing instrument	35
Program and Multi-Function buttons.	14	Important notice for prospective sound generator users.	35
Flight mode	17	TSG warning to hearing healthcare professionals.	36
Telephone use	19	Battery warning Information.	37
Listen to radio or TV.	20	Hearing instrument expectations.	37
Cellular phones	20	Warning to hearing aid dispensers	38
PhoneNow	20	Important notice for prospective hearing aid users.	38
Telecoil (optional)	22	Children with hearing loss	39
Tele-loop systems	23	Troubleshooting guide	40
Direct audio input (optional)	23	Technical data	44
Connecting/Disconnecting audio boots	23	Warranty and repairs	54
Care and maintenance.	24	Temperature test, transport and storage information	54
Daily maintenance	25		
The receiver tube	26		
Cleaning the receiver tubes and domes	26		
Cleaning RIE moulds	26		



1. Receiver tube
2. Receiver Open Dome
3. Receiver Tulip Dome
4. Receiver Power Dome
5. RIE mould
6. Sports lock
7. Battery compartment
8. Receiver
9. Push button
10. Volume Control (optional)
11. S receiver tube
12. NP receiver tube
13. HP receiver tube
14. Left/right indicator
15. Model, and serial number
16. Direct audio input
17. UP Receiver/Mould



(Identification number for the SY312e and SY312 devices)




ReSound LiNX™ device



Getting started

On/Off function

1. When the battery door is closed, the hearing instrument turns on, and the default program will be activated.
2. To turn off the hearing instrument, open the battery door. Use your fingernail to pull it open.




SmartStart

Hearing instruments can be turned on once you have placed them on your ears. If you prefer to turn them on just prior to placing them on your ear, your hearing care professional can activate a function called SmartStart. This function will delay the time in which the hearing instruments turn on by ten seconds after the battery compartment is closed. With SmartStart, a beep will be heard for each second of the delay period.

Inserting/Replacing the battery

1. Open the battery door completely by using your fingernail. Remove the used battery if present.
2. Prepare the new battery (please refer to page 7 for information on appropriate battery type/size for your hearing instrument). Remove the protective foil and wait 2 minutes before inserting the battery into the hearing instrument to allow activation of the battery.
3. Insert the new battery with the positive side in the correct position.
4. Gently close the battery door.



Tip:

1. Always use new Zinc-Air batteries that have a minimum remaining shelf life of one year.
2. Whenever the hearing instruments are not in use, remember to turn them off to avoid unnecessary battery consumption.

Low battery indicator

Your hearing care professional can set your hearing instrument to give an acoustical indication when the battery is reaching its end of life. The hearing instrument will reduce amplification and emit a melody if battery power gets too low. This signal will recur every five minutes until the hearing instrument automatically switches off. It is recommended that you keep spare batteries on hand.

Low battery indicator (instruments paired with accessories only)

Active usage of the ReSound Unite accessories (Remote Control, Phone Clip, TV Streamer and Mini Microphone) requires more battery power from the hearing instruments than when these are working on their own meaning that battery life is highly dependent on the amount of wireless accessory usage. When the battery in the hearing instrument has depleted to a level at which use of the ReSound Unite TV Streamer, Phone Clip and Mini Microphone cannot be supported, the hearing instrument will play two sets of descending tones. After this, your hearing instrument and ReSound Unite Remote Control will continue to work as usual, but you will not be able to use your ReSound Unite TV Streamer, Phone Clip and Mini Microphone. At some point the battery level will not support the remote control either and you will once again hear the descending tones. The hearing instruments will continue to work as usual. Once a new battery is inserted, full operation of the accessories will resume.



Sports lock

The sports lock will be applied or adjusted by your hearing care professional.

Inserting/Removing hearing instruments

Insertion (custom RIE and UP moulds)

1. Hold the RIE mould between your thumb and index finger and position its sound outlet in your ear canal.
2. Slide the RIE mould all the way into your ear with a gentle, twisting movement.
3. Move the RIE mould up and down and gently press to ensure it is positioned correctly in the ear. Opening and closing your mouth can ease insertion.
4. Make sure the hearing instrument is seated behind the ear.



By experimenting, an easier method may be discovered. With proper insertion, hearing instruments should fit snugly but comfortably. If hearing instruments cause irritation of the ears, contact your hearing care professional.




Never attempt to modify the shape of the hearing instrument, RIE moulds, or tubing yourself.



Tip: It may be helpful to pull the top of your ear back with your opposite hand during insertion to open the ear canal.

Insertion (domes)

1. Hold the receiver tube where it bends, and gently place the dome into the ear canal. Push the dome far enough into the ear canal so that the receiver tube lies flush with the side of the head.
2. It is important that the tube and the dome fit correctly into your ear.
3. When the dome is placed correctly, you should not be able to see the receiver tube sticking out when facing a mirror.




Note: You should never attempt to bend or modify the shape of the receiver tube.

Removal (RIE moulds)

1. Grasp the removal string and pull the RIE or UP mould outward.
2. Consult your hearing care professional if you have difficulties removing the hearing instrument.


Removal (domes)

1. Hold the receiver tube with your thumb and forefinger and remove the tube.
2. Consult your hearing care professional if you have difficulties removing the hearing instrument.




Program and Multi-Function buttons

If you have a hearing aid with a program button or multi-function button, this will allow you to use up to four different listening programs, each of them suitable for certain situations.



Program button (61 models only)



Multi-function button (62 models only)

Your hearing care professional can fill out the following table for you.

Programme	Description of when to use
1	
2	
3	
4	

When using the program or multi-function buttons to switch programs, each press will move the instrument to the next program. For example, if it was in program 1 it will switch to program 2, if it was in program 2 it will switch to program 3 etc.

When you close the battery door and switch the instrument on, it will start in program 1. Press the program or multi-function buttons if you want to move to a different listening program.

If you have two hearing instruments with the synchronization function enabled, program changes to one instrument will automatically repeat in the second instrument. When a program change is made in one instrument, you will hear the same amount of confirmation beeps in the second instrument.

Your instrument has a fully automatic volume control. Therefore, it should not be necessary to control the volume manually.

However, in addition to controlling listening programs, the multi-function button (62 models only) provides you with the ability to adjust the amplification to your liking.

If you have two hearing instruments with the synchronization function enabled and you have the Multifunction Button set for volume control functionality, volume adjustments to one instrument will automatically repeat in the second instrument. When a volume control adjustment is made in one instrument, you will hear a confirmation beep. A beep in the second instrument will follow.

The multi-function button is designed to change the volume or listening programs of the hearing instrument, based on different ways it is pressed.

If necessary, your hearing care practitioner can change these settings and fill in the following table to indicate the new settings:

Multi-function button action	Default setting	New setting
Short press up	Increases volume	
Short press down	Decreases volume	
Long press up (3 seconds)	Changes programs	
Long press down (3 seconds)	Activates streaming	



Flight mode* (Wireless models only)

When boarding a flight or entering an area where RF transmitters are prohibited, wireless functionality must be deactivated, as it is not allowed to radiate radio signals during flights or in otherwise restricted areas.

For Alera wireless hearing instruments follow the following steps to enter and leave flight mode:

It is possible to disable wireless operation by opening and closing the battery compartment of the hearing instrument while at the same time pressing the push button.


When disabled manually, wireless operation may be re-enabled by opening and closing the battery compartment normally, (i.e. without at the same time pressing the push button).

For Verso wireless hearing instruments follow the following steps to enter and leave flight mode:

It is possible to disable wireless operation by opening and closing the battery compartment three times within a ten second period (open-close, open-close, open-close). Your instruments will now be in flight mode.

If the hearing instrument is in flight mode, the hearing instrument must have been operating in flight mode for at least 10 seconds before attempting to enable wireless again. It is possible to re-enable wireless operation by opening and closing the battery door once. 10 seconds after this operation is completed, wireless operation will begin again.

**For wireless models only*

 Note: It is important to wait an additional 15 seconds after wireless function resumes before opening and closing the battery compartment again for any reason. If the battery compartment is opened and closed during this 15 second window, flight mode will resume.

Telephone use

If your hearing instruments are fit with a receiver open dome or receiver tulip dome, you can probably use the telephone as you normally would by holding it up to your ear canal opening. If your hearing instruments are fit with a receiver power dome or RIE or UP mould, finding the optimal position for holding a telephone while using a hearing instrument may require practice for some individuals, and one or more of the following suggestions may be helpful.

1. Hold the telephone as you would normally.
2. Hold the telephone towards the top of the ear (closer to where the microphones are located).
3. If whistling occurs, it may take a brief moment of holding the telephone in the same position before the hearing instrument adapts and reduces the feedback.
4. Any whistling may also be decreased by holding the telephone slightly away from the ear.
5. Depending on your individual needs, your hearing care professional may activate a programme specifically for telephone use.



Listen to radio or TV

When listening to the TV or the radio, start out by listening to news commentators since they usually speak clearly, then try other programmes.

If you find it difficult to listen to TV or radio, your hearing care professional will be able to give you advice on available accessories to enhance your listening capabilities for TV and radio.

Cellular phones

Your hearing instrument is designed to comply with the most stringent Standards of International Electro-magnetic Compatibility. However, not all cell phones are hearing instrument compatible. The varying degree of disturbance can be due to the nature of your particular cellular phone or of your wireless telephony service provider.

If you find it difficult to obtain a good result while using your cellular phone, your hearing care professional will be able to give you advice on available accessories to enhance listening capabilities.


PhoneNow

The PhoneNow function, allows your hearing instrument to automatically switch to your telephone programme when a telephone receiver is raised to the ear. When the telephone receiver is removed from the ear, the hearing instrument automatically returns to the previous listening programme.

Placement of PhoneNow magnets

Place PhoneNow magnet on your telephone receiver to allow operation of the PhoneNow function. In order to place PhoneNow magnet properly:

1. Clean the telephone receiver thoroughly.
2. Hold the telephone vertically, in a position similar to when making a telephone call.
3. Place the magnets just below the telephone receiver. Make sure not to cover the microphone openings. If necessary, move the magnet to another position to improve ease of use and comfort while speaking.
4. If you are not satisfied with the strength of PhoneNow, you can reposition the PhoneNow magnet or add additional PhoneNow magnets.



PhoneNow usage

Telephones can be used in a normal manner. A short melody will indicate that the PhoneNow feature has automatically switched the hearing instrument to your telephone programme. Initially, you may need to move the telephone receiver slightly to find the best position for reliable PhoneNow activation and good hearing on the telephone.

If you have two hearing instruments with the synchronization function enabled, the volume of hearing instrument on the non-phone ear will be turned down.



Only use recommended cleaning agent to clean the telephone prior to placing the magnet on the phone in order to obtain best possible adherence.



PhoneNow precautions

1. Keep magnets out of reach of pets, children and mentally challenged persons. If a magnet is swallowed, please seek advice from a medical practitioner.
2. The magnet may affect some medical devices or electronic systems. The manufacturer of any magnetically sensitive devices (e.g. pacemakers) should advise you regarding appropriate safety precautions when using your hearing instrument and magnet in close proximity to the medical device or electronic system in question. If the manufacturer cannot issue a statement, we recommend keeping the magnet or a telephone equipped with the magnet 30 cm (12") away from magnetically sensitive devices (e.g. pacemakers).
3. High distortion during dialing or phoning may mean that the magnet is not in the optimal position relative to the telephone receiver. To avoid the issue, please move the magnet to another place on the telephone receiver.
4. Only use magnets supplied by ReSound.

Telecoil (optional)

If equipped, a telecoil can be activated by your hearing care professional and accessed through one of the additional programmes. A telecoil picks up a telephone's magnetic signal and converts it to sound. An optional telephone programme may help to improve speech understanding on the telephone. When using a telecoil programme, the receiver of the telephone may need to be held closer to the hearing instrument. The handset of the telephone may need to be moved to slightly different positions in order to find the best reception.

Tele-loop systems

Many places, such as theatres, houses of worship, and schools are equipped with tele-loop systems. When using a telecoil programme with tele-loop systems, sound is picked up directly and may improve speech understanding. If there is no sound from the hearing instruments in a tele-loop system and with a telecoil programme activated, the tele-loop system may not be turned on or is not operating correctly. If a facility is not equipped with a tele-loop system, sitting as close as possible to the front may be helpful.


Direct audio input (optional)

Use of direct audio input (DAI), which enables a direct connection of the hearing instruments to items such as television, radio, and remote microphones, may increase speech understanding for some individuals. The sound source is connected to the hearing instruments by a cable or a wireless FM system to the audio boot. This accessory connects to the bottom of the hearing instruments, and once properly clicked into place, the hearing instruments switch to DAI automatically.

Connecting/Disconnecting audio boots

Connecting audio boots


1. Align the tip of the audio boot with the groove just above the battery compartment and below the model number.



- Once in place, move the boot in the direction of the battery compartment.
- Gently click the audio boot onto the hearing instrument.

Disconnecting audio boots

- Press and hold the button on the front side of the audio boot.
- Gently remove the audio boot from the hearing instrument.



i Care and maintenance


Your hearing instrument is protected by a layer of protective, hydrophobic nanocoat material. Please follow the following instructions to prolong the durability of your hearing instruments:

- Keep your hearing instrument clean and dry. Wipe the case with a soft cloth or tissue after use to remove grease or moisture. Do not use water or solvents, as these can damage the hearing instrument(s).
- Never immerse hearing instruments in water or other liquids, as liquids may cause permanent damage to the hearing instruments.

- Avoid rough handling of hearing instruments or dropping them on hard surfaces or floors.
- Do not leave hearing instruments in or near direct heat or sunlight, such as in a hot, parked car, as excessive heat can cause damage or deform the casing.
- Do not wear your instrument while showering, swimming, in heavy rain or in a moist atmosphere such as a steam bath or sauna.
- If your instrument does get wet, or if it has been exposed to high humidity or perspiration, it should be left to dry out overnight with the battery out and the battery compartment open. It is also a good idea to put the instrument and battery in a sealed container together with a drying agent (desiccator) overnight. Do not use the instrument until it is completely dry. Consult your hearing care professional as to which drying agent to use.
- Remove your hearing instrument when applying such things as cosmetics, perfume, aftershave, hair spray, and suntan lotion. These might get into the instrument and cause damage.

i Daily maintenance

It is important to keep your hearing instrument clean and dry. On a daily basis, clean the hearing instruments using a soft cloth or tissue.



The receiver tube


The receiver tube contains the wiring to the receiver which delivers the sound to the ear canal. It is important that the receiver tube and the receiver dome/RIE mold fits correctly in your ear. If the receiver tube or the receiver dome/RIE mould irritates your ear in any way and prevents you from wearing your hearing instrument, please contact your hearing care professional. You should never attempt to modify the shape of the receiver tube yourself. The receiver tube and the receiver dome/RIE mould should be cleaned regularly. Please see instructions in the next section.

Cleaning the receiver tubes and domes


The receiver tube and the receiver dome should be cleaned regularly. Use a damp cloth to clean the receiver tube and receiver dome on the outside. Do not use water when you are cleaning the receiver tubes or the receiver domes. This process is also used to clean the UP receiver mould. Please see instruction on page 30 or 31 for how to change the wax guard filter.

Cleaning RIE or moulds

1. Separate the mould from the receiver tube.
2. Clean the RIE mould using a mild soap, and rinse with lukewarm water.
3. After cleaning, dry RIE moulds thoroughly and remove any residual water and debris from the tubing utilising an air bulb and wire loop.



Changing wax guard for receiver tube



For NP receivers:

1. Clean any debris from the old waxguard.
2. Insert the wand into the old waxguard.
3. Twist the wand with the waxguard in a clockwise direction to ensure it is attached to the wand.
4. Pull the wand and waxguard away from the tube/mould.
5. Insert the old waxguard into the center of the HF3 wheel.
6. Dispose of the old waxguard by drawing the wand to the narrow end of the center disposal area.
7. Insert the empty wand into a new waxguard on the HF3 wheel.
8. Pull the new waxguard attached to the wand away from the HF3 wheel.
9. Insert the wand into the receiver tube/mould.
10. Twist the wand to release the new waxguard onto the receiver tube/mould.

For S and HP₂ receivers:


1. To remove the old wax guard, insert the removal side of the wax guard tool into the used wax guard so that the shaft of the tool is touching the rim of the wax guard. Slowly pull the wax guard straight out.
2. To insert the new wax guard, gently press the replacement side of the wax guard tool straight into the hole of the sound outlet until the outer ring lies flush with the outside of the receiver. Pull the tool straight out -the new wax guard will remain in place.

How to apply domes

It is recommended that your hearing care professional change domes, as incorrect dome replacement could result in injury.

ReSound standard domes


1. Push the new dome over the receiver.
2. Make sure that the new dome is properly and securely mounted.



ReSound tulip domes

Tulip domes are mounted in a similar manner to standard domes, but a few extra steps are required. Tulip domes consist of two “petals”. It is important to note that the largest petal is the outermost petal. To ensure this:

1. Push the largest petal away from the receiver tube using a finger. This bends the petal forward.
2. Then push the largest “petal” back, and it will be placed on top of the smaller petal.



Using ReSound hearing instruments with smart phone apps

i Intended use of smart phone apps:

GN ReSound smart phone apps are intended to be used with GN ReSound wireless hearing aids. GN ReSound smart phone apps send and receive signals from the GN ReSound wireless hearing aids via smart phones for which the apps have been developed.

Use with smart phone apps:

- Notifications of app updates should not be disabled, and it is recommended that the user installs all updates to ensure that the app will function correctly and will be kept up to date.
- The app must only be used with GNR devices for which it is intended, and GNR take no responsibility if the app is used with other devices.



General precautions

1. Special care should be exercised in selecting and fitting a hearing instrument(s) who's maximum sound pressure level exceeds 132 dB SPL (with an IEC 60711:1981 occluded ear simulator), because there may be a risk of impairing the remaining hearing of the hearing instrument user
2. Do not leave hearing instruments in the sun, near an open fire, or in a hot, parked car.
3. Do not wear hearing instruments while showering, swimming, in heavy rain, or in a moist atmosphere such as a steam bath or sauna.
4. Should the hearing instrument become moist, remove the battery and place the hearing instrument in a closed container with a drying agent. Your hearing care professional can provide options for drying containers or kits.
5. Remove the hearing instruments when applying items such as cosmetics, perfume, after-shave, hair spray, and suntan lotion.
6. When wireless function is activated, the device uses low-powered digitally coded transmissions in order to communicate with other wireless devices. Although unlikely, nearby electronic devices may be affected. In that case, move the hearing instrument away from the affected electronic device.
7. When using wireless functionality and the devices are affected by electromagnetic interference, move away from the source.
8. Use only original GN ReSound consumables e.g. tubes and domes.
Never attempt to modify the shape of the hearing instrument, ear-moulds, or tubing yourself.
9. Do only connect ReSound hearing instruments to ReSound accessories intended and qualified to be used with ReSound hearing instruments.



General warnings

1. Consult a hearing care professional if you discover a foreign object in your ear canal, if you experience skin irritation, or if excessive ear wax accumulates with the use of the hearing instrument.

2. Different types of radiation, from e.g. NMR, MRI, or CT scanners, may damage hearing instruments. It is recommended not to wear hearing instruments during these or other similar procedures. Other types of radiation, such as burglar alarms, room surveillance systems, radio equipment, mobile telephones, contain less energy and will not damage hearing instruments. However, they have the potential to momentarily affect the sound quality or temporarily create strange sounds from hearing instruments.
3. Do not wear hearing instruments in mines, oil fields, or other explosive areas unless those areas are certified for hearing instrument use.
4. Do not allow others to use your hearing instruments. This may cause damage to the hearing instruments or to the hearing of the other individual.
5. Instrument usage by children or mentally challenged persons should be supervised at all times to ensure their safety. The hearing instrument contains small parts that could be swallowed by children. Please be mindful not to leave children unsupervised with this hearing instrument.
6. Hearing instruments should be used only as prescribed by your hearing care professional. Incorrect use may result in hearing loss.
7. External devices connected to the electrical input must be safe according to the requirements of IEC 60601-1-1, IEC 60065, or IEC 60950-1, as appropriate (wired connection, f.ex. HI-PRO), SpeedLink).
8. If device is broken, do not use.
9. Be careful when boarding flights, to remember to deactivate the wireless functionality.
Turn off your wireless functionality by using the flight mode in areas where radio frequency emission is prohibited.



Note:

- * ReSound wireless devices operate in the frequency range of 2.4 GHz - 2.48 GHz.
- * ReSound wireless devices include a RF transmitter that operates in the range of 2.4 GHz - 2.48 GHz.
- * For use of wireless functionality only use ReSound Unite accessories. For further guidance regarding e.g. pairing, please refer to the user guide of the relevant ReSound Unite accessory.

Tinnitus Sound Generator (TSG) module

Intended use for the TSG module

Your ReSound hearing instruments may also include the Tinnitus Sound Generator function, a tool for generating sounds to be used in tinnitus management programmes to relieve suffering from tinnitus.



TSG warnings

- Sound generators can be dangerous if improperly used.
- Sound generators should be used only as advised by your doctor, audiologist, or hearing healthcare professional.
- Sound generators are not toys and should be kept out of reach of anyone who might cause themselves injury (especially children and pets).

User instructions for the TSG module

Description of the device

The Tinnitus Sound Generator (TSG) Module is a software tool that generates sounds to be used in tinnitus management programmes to relieve suffering from tinnitus.

Explanation of how the device functions

The TSG module is a frequency and amplitude shaped white-noise generator. Noise signal level and frequency characteristics can be adjusted to the specific therapeutic needs as determined by your doctor, audiologist or hearing healthcare professional.

Your doctor, audiologist or hearing healthcare professional can modulate the generated noise with the purpose of making it more pleasant. The noise can then resemble, for example, crushing waves on a shore. Modulation level and speed can also be configured to your likes and needs.

If your tinnitus troubles you only in quiet environments, your doctor, audiologist or hearing healthcare professional can set the TSG Module so that it becomes audible exclusively in such surroundings. The overall sound level can be adjusted via an optional volume control. Your doctor, audiologist or hearing healthcare professional will review with you the need for having such a control.

TSG volume control

The sound generator is set to a specific loudness level by the hearing healthcare professional. When switching the sound generator on, the volume will have this optimal setting. Therefore, it might not be necessary to control the volume (loudness) manually. However, the volume control provides the ability to adjust the volume, or amount of stimulus, to the liking of the user.



Caution

- The volume control is an optional feature in the TSG module used for adjusting the sound generator output level. To prevent unintended usage by pediatric or physically or mentally challenged users, the volume control must, if enabled, be configured to only provide a decrease of the sound generator output level.



TSG precautions

- Should the user develop any side effects from using the sound generator, such as dizziness, nausea, headaches, perceived decrease in auditory function or increase in tinnitus perception, the user should discontinue use of the sound generator and seek medical evaluation.
- Children and physically or mentally challenged users will require guardian supervision while wearing the TSG hearing instrument.

The scientific concepts that form the basis for the device

The TSG Module provides sound enrichment with the aim of surrounding the tinnitus sound with a neutral sound which is easily ignored. Sound enrichment is an important component of most approaches to tinnitus management, such as Tinnitus Retraining Therapy (TRT). To assist habituation to tinnitus, this needs to be audible. The ideal level of the TSG module, therefore, should be set so that it starts to blend with the tinnitus, and so that you can hear both your tinnitus as well as the sound used.

In a majority of instances, the TSG module can also be set to mask the tinnitus sound, so to provide temporary relief by introducing a more pleasant and controllable sound source.

Significant physical characteristics

Audio signal technology

Digital

Available sounds

White noise signal which can be shaped with the following configurations:

High-pass filter	Low-pass filter
500 Hz	2000 Hz
750 Hz	3000 Hz
1000 Hz	4000 Hz
1500 Hz	5000 Hz
2000 Hz	6000 Hz

The white noise signal can be modulated in amplitude with an attenuation depth of up to 14dB.

Prescription use of this TSG hearing instrument

The TSG module should be used as prescribed by your doctor, audiologist or hearing healthcare professional. In order to avoid permanent hearing damages, the maximum daily usage depends on the level of the generated sound.

Should you develop any side effects from using the sound generator, such as dizziness, nausea, headaches, perceived decrease in auditory function or increase in tinnitus perception, you should discontinue use of sound generator and seek medical evaluation.

The target population is primarily the adult population over 18 years of age. This product may also be used with children 5 years of age or older. However, children and physically or mentally challenged users will require training by a doctor, audiologist, hearing healthcare professional or the guardian for the insertion and removal of the hearing instrument containing the TSG module.

Important notice for prospective sound generator users

A tinnitus masker is an electronic device intended to generate noise of sufficient intensity and bandwidth to mask internal noises. It is also used as an aid in hearing external noises and speech.

Good health practice requires that a person with a tinnitus condition have a medical evaluation by a licensed physician (preferably a physician who specializes in diseases of the ear) before using a sound generator. Licensed physicians who specialize in diseases of the ear are often referred to as otolaryngologists, otologists or otorhinolaryngologists.


The purpose of medical evaluation is to assure that all medically treatable conditions that may affect tinnitus are identified and treated before the sound generator instrument is used. The sound generator instrument is

a tool to generate sounds to be used with appropriate counselling and/or in a tinnitus management program to relieve patients suffering from tinnitus.

TSG warning to hearing healthcare professionals

A hearing healthcare professional should advise a prospective sound generator user to consult promptly with a licensed physician (preferably an ear specialist) before getting a sound generator if the hearing healthcare professional determines through inquiry, actual observation, or review of any other available information concerning the prospective user that the prospective user has any of the following conditions:

- (i) Visible congenital or traumatic deformity of the ear.
- (ii) History of active drainage from the ear within the previous 90 days.
- (iii) History of sudden or rapidly progressive hearing loss within the previous 90 days.
- (iv) Acute or chronic dizziness.
- (v) Unilateral hearing loss of sudden or recent onset within the previous 90 days.
- (vi) Audiometric air-bone gap equal to or greater than 15dB at 500 hertz (Hz), 1000 Hz, and 2000 Hz.
- (vii) Visible evidence of significant cerumen accumulation or a foreign body in the ear canal.
- (viii) Pain or discomfort in the ear.

 **CAUTION:** The maximum output of the sound generator falls into the range that can cause hearing loss according to OSHA regulations. The user should not use the sound generator for more than eight (8) hours a day when this is set below 90dB SPL. Above that level, the device should not be used for more than two (2) hours per day. In no case should the sound generator be worn at uncomfortable levels.

Battery warning information

Batteries, although very small, contain dangerous substances, and should be disposed of carefully. This is for the safety of you and the environment. Please note:

1. Do not attempt to recharge batteries (Zinc Air) which are not specifically designated as rechargeable because they may leak or explode.
2. DO NOT attempt to dispose of batteries by burning them. Used batteries are harmful to the environment. Please dispose of them according to local regulations or return them to your hearing care practitioner.
3. DO NOT place batteries in your mouth. Consult a physician immediately if a battery has been swallowed, as they can be harmful to your health.
4. Keep batteries away from pets, children and mentally challenged persons.
5. Remove the batteries to prevent leakage when the hearing instruments are not in use for an extended period of time.

Hearing instrument expectations

A hearing instrument will not restore normal hearing and will not prevent or improve a hearing impairment resulting from organic conditions. Consistent use of the hearing instrument is recommended. In most cases, infrequent use does not permit you to attain full benefit from it.

The use of a hearing instrument is only part of hearing rehabilitation and may need to be supplemented by auditory training and instructions in lip-reading.



Warning to hearing aid dispensers (US Only)

A hearing aid dispenser should advise a prospective hearing aid user to consult promptly with a licensed physician (preferably an ear specialist) before dispensing a hearing aid if the hearing aid dispenser determines through inquiry, actual observation, or review of any other available information concerning the prospective user, that the prospective user has any of the following conditions:

- (i) Visible congenital or traumatic deformity of the ear.
- (ii) History of active drainage from the ear within the previous 90 days.
- (iii) History of sudden or rapidly progressive hearing loss within the previous 90 days.
- (iv) Acute or chronic dizziness.
- (v) Unilateral hearing loss of sudden or recent onset within the previous 90 days.
- (vi) Audiometric air-bone gap equal to or greater than 15 decibels at 500 hertz (Hz), 1,000 Hz, and 2,000 Hz.
- (vii) Visible evidence of significant cerumen accumulation or a foreign body in the ear canal.
- (viii) Pain or discomfort in the ear.

Important notice for prospective hearing aid users (US Only)

Good health practice requires that a person with a hearing loss have a medical evaluation by a licensed physician (preferably a physician who specializes in diseases of the ear) before purchasing a hearing aid. Licensed physicians who specialize in diseases of the ear are often referred to as otolaryngologists, otologists or otorhinolaryngologists. The purpose of medical evaluation is to assure that all medically treatable conditions that may affect hearing are identified and treated before the hearing aid is purchased.

Following the medical evaluation, the physician will give you a written statement that states that your hearing loss has been medically evaluated and that you may be considered a candidate for a hearing aid. The physician will refer you to an audiologist or a hearing aid dispenser, as appropriate, for a hearing aid evaluation. The audiologist or hearing aid dispenser will conduct a hearing aid evaluation to assess your ability to hear with and without a hearing aid. The hearing aid evaluation will enable the audiologist or dispenser to select and fit a hearing aid to your individual needs. If you have reservations about your ability to adapt to amplification, you should inquire about the availability of a trial-rental or purchase-option program. Many hearing aid dispensers now offer programs that permit you to wear a hearing aid for a period of time for a nominal fee after which you may decide if you want to purchase the hearing aid.

Federal law restricts the sale of hearing aids to those individuals who have obtained a medical evaluation from a licensed physician. Federal law permits a fully informed adult to sign a waiver statement declining the medical evaluation for religious or personal beliefs that preclude consultation with a physician. The exercise of such a waiver is not in your best health interest and its use is strongly discouraged.

Children with hearing loss (US Only)

In addition to seeing a physician for a medical evaluation, a child with a hearing loss should be directed to an audiologist for evaluation and rehabilitation since hearing loss may cause problems in language development and the educational and social growth of a child. An audiologist is qualified by training and experience to assist in the evaluation and rehabilitation of a child with a hearing loss.

Troubleshooting Guide

SYMPTOM	CAUSE	POSSIBLE REMEDY
No sound	Not turned on	Turn on by closing the battery door
	Dead battery	Replace battery
	Battery door will not close	Insert battery properly
	Blocked RIE mould or tube	Clean RIE mould or tube
	Blocked wax filter	Replace wax filter or consult your hearing care professional
Not loud enough	Incorrect RIE mould placement	Reinsert RIE mould
	Blocked RIE mould or dome	Clean RIE mould, replace dome, replace filter
	Change in hearing sensitivity	Consult your hearing care professional
	Excessive ear wax	Consult your hearing care professional
	Volume set too low	Consult your hearing care professional

Troubleshooting Guide

SYMPTOM	CAUSE	POSSIBLE REMEDY
Excessive whistling / feedback	Incorrect RIE mould placement	Re-insert RIE mould carefully
	Incorrect dome placement	Re-insert dome
	Excessive ear wax	Consult your hearing care professional
	Feedback control may need adjustment	Consult your hearing care professional
	RIE mould tubing worn or damaged	Consult your hearing care professional
	Hearing instrument settings not optimal	Consult your hearing care professional
Sound distorted / not clear	Weak battery	Replace battery
	Improper RIE mould or dome fit	Consult your hearing care professional
	Hearing instrument damaged	Consult your hearing care professional
	Hearing instrument settings not optimal	Consult your hearing care professional
Wireless does not work	Possible Root Cause - Device is in flight mode	For Alera devices with push button: Open and close the battery compartment. For Alera devices without push button: Open and close the battery door twice within 10 seconds For all Verso devices: Open and close the battery compartment once. Wireless will reactivate 10 seconds later. (If Root Cause is device in flight mode)

If there are any other problems not mentioned in this guide, please contact your hearing care professional.



Technical data

Micro and Mini RIE—NP receiver

Models: AL960-DR, AL760-DR, AL560-DR, AL961-DRW, AL761-DRW, AL561-DRW, ALT960-DR, ALT961-DRW

		Open	Closed	
Reference test gain (60 dB SPL input)	HFA	30	32	dB
Full-on gain (50 dB SPL Input)	Max	47	50	dB
	HFA	41	42	dB
Maximum output (90 dB SPL input)	Max	114	114	dB SPL
	HFA	108	108	dB SPL
Total harmonic distortion	500 Hz	0.8	0.8	%
	800 Hz	0.8	0.9	%
	1600 Hz	0.7	0.8	%
Equivalent input noise (w/o noise reduction)		24	25	dB SPL
Frequency range (DIN 45605)		100–6790	100–6720	Hz
Current drain (in test mode)	ALx60	1.1	1.1	mA
	ALx61	1.2	1.2	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





— Open configuration
— Closed configuration

Micro and Mini RIE—HP receiver

Models: AL960-DR, AL760-DR, AL560-DR, AL961-DRW, AL761-DRW, AL561-DRW, ALT960-DR, ALT961-DRW

		Open	Closed	
Reference test gain (60 dB SPL input)	HFA	35	35	dB
Full-on gain (50 dB SPL Input)	Max	57	58	dB
	HFA	49	49	dB
Maximum output (90 dB SPL input)	Max	117	117	dB SPL
	HFA	112	112	dB SPL
Total harmonic distortion	500 Hz	0.8	0.7	%
	800 Hz	1.1	1.0	%
	1600 Hz	0.9	0.8	%
Equivalent input noise (w/o noise reduction)		26	26	dB SPL
Frequency range (DIN 45605)		100–7150	100–7140	Hz
Current drain (in test mode)	ALx60	1.1	1.1	mA
	ALx61	1.2	1.2	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





— Open configuration
— Closed configuration

RIE—NP receiver

Models: AL962-DVIRW, AL762-DVIRW, AL562-DVIRW, ALT962-DVIRW, ALT762-DVIRW

		Open	Closed	
Reference test gain (60 dB SPL input)	HFA	30	32	dB
Full-on gain (50 dB SPL Input)	Max	47	50	dB
	HFA	41	42	dB
Maximum output (90 dB SPL input)	Max	114	114	dB SPL
	HFA	108	108	dB SPL
Total harmonic distortion	500 Hz	0.8	0.8	%
	800 Hz	0.8	0.9	%
	1600 Hz	0.7	0.8	%
Telecoil sensitivity (SPLIV @ 31.6 mA/m)		90	91	dB SPL
Equivalent input noise (w/o noise reduction)		24	25	dB SPL
Frequency range (DIN 45605)		100–6790	100–6720	Hz
Current drain (in test mode)		1.2	1.2	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





— Open configuration
— Closed configuration

RIE—HP receiver

Models: AL962-DVIRW, AL762-DVIRW, AL562-DVIRW, ALT962-DVIRW, ALT762-DVIRW

		Open	Closed	
Reference test gain (60 dB SPL input)	HFA	35	35	dB
Full-on gain (50 dB SPL Input)	Max	57	58	dB
	HFA	49	49	dB
Maximum output (90 dB SPL input)	Max	117	117	dB SPL
	HFA	112	112	dB SPL
Total harmonic distortion	500 Hz	0.8	0.7	%
	800 Hz	1.1	1.0	%
	1600 Hz	0.9	0.8	%
Telecoil sensitivity (SPLIV @ 31.6 mA/m)		96	96	dB SPL
Equivalent input noise (w/o noise reduction)		26	26	dB SPL
Frequency range (DIN 45605)		100–7150	100–7140	Hz
Current drain (in test mode)		1.2	1.2	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





— Open configuration
— Closed configuration

Micro and Mini RIE—S receiver

Models: VO960-DR, VO760-DR, VO560-DR,
VO961-DRW, VO761-DRW, VO561-DRW
VOT960-DR, VOT760-DR, VOT961-DRW, VOT761-DRW

		Open	Closed	
Reference test gain (60 dB SPL input)	HFA	31	31	dB
Full-on gain (50 dB SPL Input)	Max	46	50	dB
	HFA	41	42	dB
Maximum output (90 dB SPL input)	Max	113	113	dB SPL
	HFA	108	108	dB SPL
Total harmonic distortion	500 Hz	0.5	0.3	%
	800 Hz	0.5	0.6	%
	1600 Hz	0.8	1.0	%
Equivalent input noise (w/o noise reduction)		23	24	dB SPL
Frequency range (DIN 45605)		100–7110	100–7100	Hz
Current drain (in test mode)	VOx60	1.1	1.1	mA
	VOx61	1.1	1.1	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





— Open configuration
- - Closed configuration

Micro and Mini RIE—NP receiver

Models: VO960-DR, VO760-DR, VO560-DR,
VO961-DRW, VO761-DRW, VO561-DRW
VOT960-DR, VOT760-DR, VOT961-DRW, VOT761-DRW

		Open	Closed	
Reference test gain (60 dB SPL input)	HFA	32	33	dB
Full-on gain (50 dB SPL Input)	Max	47	51	dB
	HFA	42	43	dB
Maximum output (90 dB SPL input)	Max	114	114	dB SPL
	HFA	109	109	dB SPL
Total harmonic distortion	500 Hz	0.8	0.9	%
	800 Hz	0.9	0.8	%
	1600 Hz	0.9	1.0	%
Equivalent input noise (w/o noise reduction)		22	23	dB SPL
Frequency range (DIN 45605)		100–6840	100–6780	Hz
Current drain (in test mode)	VOx60	1.1	1.1	mA
	VOx61	1.1	1.1	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





— Open configuration
- - Closed configuration

Micro and Mini RIE—HP₂ receiver

Models: VO960-DR, VO760-DR, VO560-DR,
VO961-DRW, VO761-DRW, VO561-DRW
VOT960-DR, VOT760-DR, VOT961-DRW, VOT761-DRW

		Open	Closed	
Reference test gain (60 dB SPL input)	HFA	36	36	dB
Full-on gain (50 dB SPL Input)	Max	59	59	dB
	HFA	50	51	dB
Maximum output (90 dB SPL input)	Max	118	119	dB SPL
	HFA	114	114	dB SPL
Total harmonic distortion	500 Hz	1.4	1.2	%
	800 Hz	1.4	1.6	%
	1600 Hz	1.1	1.0	%
Equivalent input noise (w/o noise reduction)		23	23	dB SPL
Frequency range (DIN 45605)		100–6790	100–6710	Hz
Current drain (in test mode)	VOx60	1.1	1.1	mA
	VOx61	1.1	1.1	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





— Open configuration
— Closed configuration

RIE—S receiver

Models: VO962-DRW, VO762-DRW, VO562-DRW
VOT962-DRW, VOT762-DRW

		Open	Closed	
Reference test gain (60 dB SPL input)	HFA	31	31	dB
Full-on gain (50 dB SPL Input)	Max	46	50	dB
	HFA	41	42	dB
Maximum output (90 dB SPL input)	Max	113	113	dB SPL
	HFA	108	108	dB SPL
Total harmonic distortion	500 Hz	0.5	0.4	%
	800 Hz	0.5	0.5	%
	1600 Hz	0.9	1.0	%
Telecoil sensitivity (SPLIV @ 31.6 mA/m)		91	92	dB SPL
Equivalent input noise (w/o noise reduction)		23	24	dB SPL
Frequency range (DIN 45605)		100–7110	100–7100	Hz
Current drain (in test mode)		1.2	1.2	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





— Open configuration
— Closed configuration

RIE—NP receiver

Models: VO962-DRW, VO762-DRW, VO562-DRW
VOT962-DRW, VOT762-DRW

		Open	Closed	
Reference test gain (60 dB SPL input)	HFA	30	33	dB
Full-on gain (50 dB SPL Input)	Max	46	50	dB
	HFA	41	43	dB
Maximum output (90 dB SPL input)	Max	113	114	dB SPL
	HFA	108	109	dB SPL
Total harmonic distortion	500 Hz	0.5	0.9	%
	800 Hz	0.5	1.0	%
	1600 Hz	0.9	1.1	%
Telecoil sensitivity (SPLIV @ 31.6 mA/m)		91	93	dB SPL
Equivalent input noise (w/o noise reduction)		23	24	dB SPL
Frequency range (DIN 45605)		100–7110	100–6770	Hz
Current drain (in test mode)		1.2	1.2	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





— Open configuration
— Closed configuration

RIE—HP₂ receiver

Models: VO962-DRW, VO762-DRW, VO562-DRW
VOT962-DRW, VOT762-DRW

		Open	Closed	
Reference test gain (60 dB SPL input)	HFA	36	36	dB
Full-on gain (50 dB SPL Input)	Max	59	59	dB
	HFA	50	51	dB
Maximum output (90 dB SPL input)	Max	118	119	dB SPL
	HFA	114	114	dB SPL
Total harmonic distortion	500 Hz	1.4	1.2	%
	800 Hz	1.4	1.6	%
	1600 Hz	1.1	1.0	%
Telecoil sensitivity (SPLIV @ 31.6 mA/m)		96	98	dB SPL
Equivalent input noise (w/o noise reduction)		23	23	dB SPL
Frequency range (DIN 45605)		100–6790	100–6710	Hz
Current drain (in test mode)		1.2	1.2	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





— Open configuration
— Closed configuration

Mini RIE—S receiver

Models: LN961-DRW, LN761-DRW

Reference test gain (60 dB SPL input)	HFA	32	dB
Full-on gain (50 dB SPL input)	Max. HFA	50 43	dB
Maximum output (90 dB SPL input)	Max. HFA	114 109	dB SPL
Total harmonic distortion	500 Hz 800 Hz 1600 Hz	0,5 0,6 0,8	%
Equivalent input noise		24	dB SPL
Frequency range (DIN 45605/ANSI)		100- 6930	Hz
Current Drain		1,2	mA

Data in accordance with ANSI S3.22-2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





Mini RIE—NP receiver

Models: LN961-DRW, LN761-DRW

Reference test gain (60 dB SPL input)	HFA	33	dB
Full-on gain (50 dB SPL input)	Max. HFA	55 48	dB
Maximum output (90 dB SPL input)	Max. HFA	115 110	dB SPL
Total harmonic distortion	500 Hz 800 Hz 1600 Hz	1,2 1,5 1,6	%
Equivalent input noise		23	dB SPL
Frequency range (DIN 45605/ANSI)		100- 6470	Hz
Current Drain		1,3	mA

Data in accordance with ANSI S3.22-2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





Mini RIE—HP receiver

Models: LN961-DRW, LN761-DRW

Reference test gain (60 dB SPL input)	HFA	38	dB
Full-on gain (50 dB SPL input)	Max. HFA	65 55	dB
Maximum output (90 dB SPL input)	Max. HFA	118 115	dB SPL
Total harmonic distortion	500 Hz 800 Hz 1600 Hz	1,5 2,4 1,5	%
Equivalent input noise		24	dB SPL
Frequency range (DIN 45605/ANSI)		100- 6300	Hz
Current Drain		1,2	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler





Mini RIE—UP receiver

Models: LN961-DRW, LN761-DRW

Reference test gain (60 dB SPL input)	HFA	46	dB
Full-on gain (50 dB SPL input)	Max. HFA	76 66	dB
Maximum output (90 dB SPL input)	Max. HFA	128 123	dB SPL
Total harmonic distortion	500 Hz 800 Hz 1600 Hz	2,1 2,3 0,6	%
Equivalent input noise		24	dB SPL
Frequency range (DIN 45605/ANSI)		100- 5690	Hz
Current Drain		1,2	mA

Data in accordance with ANSI S3.22–2003 and IEC 60118-7; Supply Voltage 1.3 V, 2cc coupler



Warranty and repairs

ReSound provides a warranty on hearing instruments in the event of defects in workmanship or materials, as described in applicable warranty documentation. In its service policy, ReSound pledges to secure functionality at least equivalent to the original hearing instrument. As a signatory to the United Nations Global Compact initiative, ReSound is committed to doing this in line with environment-friendly best practices. Hearing instruments therefore, at ReSound's discretion, may be replaced by new products or products manufactured from new or serviceable used parts, or repaired using new or refurbished replacement parts. The warranty period of hearing instruments is designated on your warranty card, which is provided by your hearing care professional.

For hearing instruments that require service, please contact your hearing care professional for assistance. ReSound hearing instruments that malfunction must be repaired by a qualified technician. Do not attempt to open the case of hearing instruments, as this will invalidate the warranty.

Temperature test, transport and storage information

GN ReSound Hearing Instruments are subjected to various tests in temperature and damp heating cycling between -25 C and +70C according to internal and industry standards. During transport or storage, the temperature should not exceed the limit values of -20C to 60C and relative humidity of 90% RH, non condensing (for limited time). The air pressure between 500 and 1100 hPa is appropriate.

Be aware of information marked with the warning symbol



WARNING points out a situation that could lead to serious injuries,
CAUTION indicates a situation that could lead to minor and moderate injuries”



Advice and tips on how to handle your hearing instrument better.



Equipment includes RF transmitter



Product is a Type B applied part



“Made for iPhone” means that an electronic accessory has been designed to connect specifically to iPhone and has been certified by the developer to meet Apple performance standards.

Apple is not responsible for the operation of this device or its compliance with safety and regulatory standards. Please note that the use of this accessory with iPhone may affect wireless performance.



Please ask your local hearing care professional concerning disposal of your hearing instrument

Worldwide Headquarters

ReSound A/S
Lautrupbjerg 7
DK-2750 Ballerup, Denmark
Tel.: +45 45 75 11 11
Fax: +45 45 75 11 19
www.resound.com

CVR no. 55082715

United Kingdom

GN ReSound Ltd.
Kirtlington Business Centre
Portway
Kirtlington
Oxon OX5 3JA
Tel.: +44 1869 352 800
Fax: +44 1869 343 466
www.gnresound.co.uk

Australia

GN ReSound Pty. Ltd.
Unit R1 Regents Park Estate
391 Park Road
Regents Park NSW 2143
Tel.: (free) 1800 658 955
Fax: +61 2 9743 7472
www.gnresound.com.au

New Zealand

GN ReSound (NZ) Ltd.
12 Parkway Drive
Mairangi Bay
Auckland
Tel.: (free) 0800 900 126
Fax: (free) 0800 007 695
www.gnresound.co.nz



Any issues relating to the EU Medical Device Directive 93/42/EEC, or Council Directive 1999/5/EC on Radio Equipment and Telecommunications terminal equipment should be directed to ReSound A/S

ReSound

rediscover hearing