


Front

- A** Wearer's name
- B** Wear location
- C** Department
- D** Account number
- E** Wearer number
- F** Read button
- G** Serial number
- H** Clip
- I** Light indicator



Back

Instadose®+ Dosimeter

Instadose® radiation monitoring technology has revolutionized the way radiation exposure is captured, read and reported.

Instadose®+ provides faster, easier exposure feedback with on-demand (manual) and automatic calendar-enabled dose reads using the InstaLink™ Hotspot, or the Instadose® Mobile App. Best of all, you keep your dosimeter – No more collecting and redistributing new dosimeters every wear period!

Care for Your Dosimeter

Extreme temperatures (high or low) may permanently damage internal components. Do not leave outside or in hot car. If dosimeter gets too hot or cold, transmission is not possible until it cools down and recovers to room temperature.



Do not allow dosimeter to overheat.



Do not allow dosimeter to freeze.



Do not immerse dosimeter in water.

Capturing Dose Reads

- 1** To communicate and transmit dose data, you must be within range of a wireless transmission source (Instadose® App, or the InstaLink™ Hotspot).




Instadose®
Mobile App



InstaLink™
Hotspot

- 2** Push and hold the button on the back of the dosimeter for at least 4 seconds, until you see a solid green light flash. Within 1-minute, a second green light will flash to confirm transmission.



For more information, visit www.instadose.com

FCC Compliance Statement

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

- (1) this device may not cause harmful interference, and
- (2) this device must accept any interference received, including interference that may cause undesired operation.

CAUTION: The grantee is not responsible for any changes or modifications not expressly approved by the party responsible for compliance. Such modifications could void the user's authority to operate the equipment.

NOTE: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy, and if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

Canadian Compliance Statement

This device contains license-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada license exempt RSS(s). Operation is subject to the following two conditions:

- (1) This device may not cause interference.
- (2) This device must accept any interference, including interference that may cause undesired operation of the device.

L'émetteur/récepteur exempt de licence contenu dans le présent appareil est conforme aux CNR d'Innovation, Sciences et Développement économique

Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes:

- 1) L'appareil ne doit pas produire de brouillage;
- 2) L'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

Mirion Technologies, Inc.

PRODUCT NAME: Instadose+
PRODUCT MODEL: v1.2
RATING: 3V ~ 150mA
FCC ID: 2A42N-INSTADOSE
IC: 21847-INSTADOSE

This device complies with part 15 of the FCC Rules.
Operation is subject to the following two conditions:
(1) This device may not cause harmful interference, and
(2) This device must accept any interference received,
including interference that may cause undesired operation.

