

✓ Where People Make the Difference

March 5, 2008

AmericanTCB 6731 Whittier Avenue McLean VA 22101

Attn: Director of Certification

Dear Sir or Madam

We, Nonin Medical, Inc, are requesting Limited Modular Approval (LMA) on the SABER and hereby declare that the SABER will only be integrated into Nonin Medical, Inc products. Therefore, Nonin Medical, Inc will retain full control of final installation

No manual exists for this product since Nonin Medical, Inc is the integrator. However, the following regulatory notices will be included in the end product manuals.

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

This device contains FCC ID: RG9SABER / IC: 4743A-SABER.

CAUTION: Changes or modifications not expressly approved by Nonin Medical, Inc could void the user's authority to operate the equipment.

This device is intended only for OEM integrators under the following conditions:

- 1) The transmitter module may not be co-located with any other transmitter or antenna.
- 2) The Module is approved using the FCC 'unlicensed modular transmitter approval' method. Therefore the module must only be used with the originally approved antennas.

"IMPORTANT NOTE: To comply with FCC RF exposure compliance requirements, the antenna used for this transmitter not be co-located or operating in conjunction with any other antenna or transmitter."

The end user should NOT be provided any instructions on how to remove or install the device.

In addition, the wording "Contains FCC ID: RG9SABER / IC: 4743A-SABER" will be included on the label of the device in which the SABER is installed.

Sincerely



✓ Where People Make the Difference

Harold Rudnick Sr. Quality Engineer 763-553-9968

763-553-7807

harold.rudnick@nonin.com