















Specifications are subject to change without notice.  
© 2016 Koninklijke Philips N.V.  
All rights reserved

Philips Personal Health Solutions  
High Tech Campus 37  
5656 AE Eindhoven, The Netherlands  
4222.100.5746.1 (1/2016)





English

## Introduction

Congratulations on your purchase and welcome to Philips! To fully benefit from the support that Philips offers, register your product at [www.philips.com/welcome](http://www.philips.com/welcome).

### Intended use

The product is intended for measuring the temperature of the human body. The device is intended for all ages in a home environment.

### General

This Philips ear thermometer has been developed for accurate and fast human body temperature measurements in the ear in the adjusted mode. Research indicates that the ear is an ideal site for taking the body temperature. The body core temperature is regulated by the hypothalamus (Fig. 2). The hypothalamus (1) shares the same blood supply as the tympanic membrane (2). Measurement results can be transmitted via Bluetooth® Smart to the Philips HealthSuite health app for charting and tracking purposes.

### General description (Fig. 1)

- 1 Protection cap
- 2 Infrared sensor
- 3 Probe tip
- 4 History icon
- 5 Bluetooth® Smart icon
- 6 Battery symbol
- 7 Temperature display
- 8 Power button
- 9 Temperature light ring
- 10 °F/°C selection button
- 11 Battery compartment
- 12 Battery cover

## IMPORTANT SAFEGUARDS

### Warning: READ ALL INSTRUCTIONS BEFORE USING

When using battery-operated products, especially when children are present, basic safety precautions should always be followed, including the following:

#### Warning



- Do not measure the body temperature at the ear site if the ear is inflamed or infected.
- Stop using the device in the occurrence of any pain. It is likely to damage the ear canal.
- If the measured body temperature exceeds 100.4°F / 38°C and the person whose temperature is being measured feels uncomfortable, contact your physician.
- It is recommended not to use the device if you suffer from ear disease, including otitis externa and tympanitis.
- Keep the device out of the reach of children and pets to avoid inhalation or swallowing of small parts. Children may not be able to use the device according to the instructions in this user manual. It is not a toy.
- Do not throw disposable batteries into fire. Batteries may explode.
- Do not put the device in a wet ear canal after swimming or bathing. This may cause injury to the ear canal.
- Do not use the device if it is broken or damaged. Using a broken or damaged device may cause injury.
- Do not use the thermometer under temperature extremes (below 59°F / 15°C or over 95°F / 35°C) or humidity extremes (below 20%RH or over 80%RH). If you do, this may cause inaccurate measurements.
- Do not expose the thermometer to temperature extremes (below 14°F / -10°C or over 131°F / 55°C) or humidity extremes (below 20%RH or over 90%RH). If you do, this may cause inaccurate measurements.
- Do not store the device in extreme temperature (below -4°F / -20°C or over 140°F / 60°C) or in extreme levels of humidity (below 15%RH or over 85%RH). If you do, this may cause inaccurate measurements.

#### Caution



- Only use the device as described in this user manual. Do not use it for any other purpose.

- The device is not intended as a substitute for consulting a doctor. Measurement results are for reference only.
- When the device is used to measure the temperature of a child, it has to be operated by an adult. Adults can measure their own temperature.
- If you clean the ear, wait 5-10 minutes before you use the device.
- Earwax in the ear canal may cause inaccurate measurements. Make sure the ear canal is clean before you use the device.
- Do not take apart, repair or change any part of the device at any time, except for replacing batteries.
- Keep the infrared sensor dry, clean and undamaged to ensure accurate measurements.
- Do not touch or blow at the infrared sensor. A dirty infrared sensor may cause inaccurate measurements.
- Never clean the device with an abrasive cleaner, thinner, benzene or immerse the device in water or other liquids.
- When the temperature of the storage area differs from the temperature of the measuring area, wait at least 30 minutes for the device to acclimatize before you use the device.
- If you do not intend to use the device for a long period of time, it is advised to remove the batteries before storing. This is to prevent possible damage due to leakage of the batteries. If the batteries leak, remove them carefully and replace with new batteries.

## SAVE THESE INSTRUCTIONS




### Compliance with standards

- The device meets the relevant standards for this type of Class IIa medical device for home use.
- This Philips device complies with all applicable standards and regulations regarding exposure to electromagnetic fields and complies with IEC 60601-1-2
- Skin contact parts (ABS, TPU, metal) comply with ISO 10993-5 and ISO 10993-10. Cytotoxicity, sensitization and irritation test results meet the requirements.
- Based on the current science and technology, other potential allergic reactions are unknown.

Clinical accuracy:

Subject age group	Operation mode	Measuring site	Reference body site	Clinical bias ( $\Delta$ cb)	Limits of agreement (LA)	Clinical repeatability ( $\sigma$ )
A1	Adjusted mode	Ear	Body core temperature	-0.020	0.203	0.07
A2				-0.033	0.195	0.09
B				-0.014	0.198	0.08
C				-0.016	0.199	0.08

### Display

Symbol	Description	Explanation
	History icon	Display shows last measurement result when switching on the device.
	Bluetooth® Smart icon	The device uses Bluetooth® Smart for communication.
	Battery symbol	Battery symbol indicating status of battery.

### Battery status indications

#### Battery full

To indicate that the battery is fully charged the battery symbol shows a full battery (Fig. 3).

#### Battery low

To indicate that the battery is low, the battery symbol shows a nearly empty battery (Fig. 4). When using 2x 1.5V AAA, Alkaline batteries, approximately 500 measurements can be realized (73.4 °F ± 9 °F, 50 ± 20% RH).

### Preparing for use

The batteries (2x 1.5V AAA) are included with the device. Remove the battery strip from the battery compartment before first use. (Fig. 15)

Slide the battery cover back until it locks into place (Fig. 5).

### Pairing the thermometer to your mobile device

Your Philips ear thermometer is equipped with Bluetooth® Smart. Download the Philips HealthSuite health app from the App Store or Google Play. Use the search term: Philips HealthSuite health app. The Philips health app is available for iOS® 8.0+ and Android™ 4.4+.

- 1 Download the Philips HealthSuite health app on your mobile device, start the Setup wizard and follow the steps to create a user profile and add the thermometer.
- 2 **Make sure the Philips health app is active and Bluetooth on your mobile device is on when pairing is in progress.**
- 3 Press the power button once, to switch on the thermometer (Fig. 6).
  - The app identifies the thermometer and requests to pair.
  - The Bluetooth icon lights up when the thermometer is connected to your mobile device.
- 4 The thermometer switches off automatically after 30 seconds of no activity.
  - When the thermometer has paired successfully with your mobile device, your personal measurement results are automatically transmitted to your mobile device via Bluetooth.

Note: Only when the Philips HealthSuite health app is active, your personal health data can be transmitted.

## Cleaning the infrared sensor

To achieve accurate measurements, it is important to check if the infrared sensor is clean. Use a cotton swab to clean the infrared sensor. See chapter Cleaning and storage for instructions. After cleaning, wait approximately 5-10 minutes before you use the device.

## Cleaning the ear

Earwax in the ear canal may cause inaccurate measurements. Make sure the ear canal is clean before you use the device.

Carefully clean your ear canal. If you clean your ear, wait for 5-10 minutes before you use the device.

## Resetting the thermometer

If you press the °F/°C selection button on the side of the thermometer for longer than 10 seconds, all measurements are deleted from the device and all settings return to the factory values.

## Measuring temperature

### Tips for proper measurement

External factors can influence your body temperature. In the following cases, wait at least 30 minutes before taking your temperature:

- When you have been lying on your ear
- When you had your ears covered
- When you have been exposed to very low or very high temperatures
- When you have been swimming or bathing
- If you wear earplugs or hearing aids, remove them first

We do not advise you to use the device in the following circumstances as this may cause inaccurate measurements:

- When the temperature of the storage area differs from the temperature of the measuring area. In this case, wait at least 30 minutes before you use the device.

### Body temperature

Normal body temperature is a range. A person's normal temperature range tends to decrease with age. The following table shows normal temperature ranges by age.

The range of normal body temperature varies from person to person and can be influenced by many factors such as time of day, level of activity and emotions.

Age	Temperature in °Fahrenheit	Temperature in °Celsius
0-2 years	97.5 - 100.4 °F	36.4 - 38.0 °C
3-10 years	97.0 - 100.0 °F	36.1 - 37.8 °C
11-65 years	96.6 - 99.7 °F	35.9 - 37.6 °C
> 65 years	96.4 - 99.5 °F	35.8 - 37.5 °C

Source: Chamberlain, J.M., et al., Determination of Normal Ear Temperature with an Infrared Emission Detection Thermometer, Annals of Emergency Medicine, January 1995, Vol. 25, pp. 15-20.

### Performing a measurement

Note: Before measuring, make sure that the probe tip of the device is clean.

- 1 Remove the protection cap before you use the device (Fig. 7).
- 2 Press the power button to switch on the device.
  - The display shows the last measurement result as well as the history icon.
- 3 Hold the outer ear and gently pull it towards the rear of the head to straighten the ear canal.
  - Children under the age of 1: hold the outer ear and gently pull it straight back. (Fig. 8)
  - Children aged 1 year or older: hold the outer ear and gently pull the ear up and back (Fig. 9).
- 4 Insert the probe tip very gently and slowly into the ear canal (Fig. 10).
  - Make sure the probe tip points straight towards the eardrum when you hold it.
  - Always insert the probe tip into the same ear, in the same direction and at the same depth.

Note: A temperature measurement taken in the right ear may differ from a measurement taken in the left ear.

- 5 Press the power button to start the measurement. (Fig. 11)
  - When the measurement is done, you hear 2 beeps.
- 6 Remove the probe tip from the ear canal (Fig. 12).
  - The display shows the measured temperature and the temperature light ring around the power button lights up red, if the measured temperature is above 100.4 °F / 38.0 °C.

Note: The temperature display on the device shows the measurement result for 30 seconds or until you perform a new measurement.

### Temperature light ring

This device is equipped with a temperature light ring around the power button (Fig. 13).

After the measurement, the ring lights up red, depending on the measured temperature.

- When the temperature light ring lights up red, the measured temperature is between 100.4 °F (38.0 °C) and 109.2 °F (42.9 °C).

### Transmit and display personal measurement results in the app

Note: Your personal measurement data is only stored and displayed in the Philips HealthSuite health app.

- 1 Start up the Philips HealthSuite health app and switch on Bluetooth on your mobile device.
- 2 The Bluetooth symbol will light up on the display of the thermometer to indicate it is connected to your mobile device.

- Once successfully connected, the measurement results will be automatically transmitted to your mobile device via Bluetooth.
- If the data transmission is successful, the measurement results are displayed in the Philips HealthSuite health app.
- The time and date of each measurement is shown in the Philips HealthSuite health app.

## Removing/inserting batteries

When the batteries are almost empty, the display shows a flashing battery symbol and the text "bAtt" (Fig. 14).

When the battery low symbol appears on the display (Fig. 4) or if the device does not function at all, replace the batteries as soon as possible. The device works on 2x 1.5V AAA Alkaline batteries.

Data will be lost when the batteries are completely empty or are taken out of the thermometer.

- 1 Slide the battery cover downwards to open the battery compartment and remove the empty batteries (Fig. 15).
- 2 Insert two new batteries in the battery compartment according to the polarity indications ( + and - ) marked inside the compartment. Press down the batteries until they click into place (Fig. 16).
- 3 Slide the battery cover back until it locks into place (Fig. 5).

## Setting the measurement unit

You can select either Fahrenheit (°F) or Celsius (°C) for measuring temperature.

- 1 Switch on the device.
- 2 Press and hold the °F/°C selection button for 3 seconds to change the measurement unit (Fig. 17).

## Cleaning and storage

**⚠ Caution: Do not expose the device to extreme temperatures (below -4 °F/-20 °C or over 140 °F/60 °C), extreme levels of humidity (below 15%RH or over 85%RH), direct sunlight or shock. This may result in inaccurate measurements.**

Use a cotton swab to clean the infrared sensor. For hygienic reasons, the device should be cleaned after each use

- 1 Moisten the cotton swab with a few drops of alcohol and wipe the surface of the infrared sensor gently with the swab (Fig. 18).
- 2 Immediately wipe the surface dry with a clean cotton swab.
- 3 Use a soft dry cloth to clean the body of the device (Fig. 19).
- 4 Store the device in a clean, dry place at room temperature.

## Calibration

The device has been calibrated when it was manufactured. If the device is used according to the instructions, periodic recalibration is not required. If you question the accuracy of the measurement at any time, call 1-844-531-6861 for assistance.

## Accessories

Philips accessories may be purchased at a store near you, or on our website

**[www.philips.com/store](http://www.philips.com/store)**.

This device has no other user-serviceable parts.

For assistance call 1-844-531-6861.

## Specifications

Power supply	2 x AAA non-rechargeable alkaline batteries
Temperature measuring range	90.3 °F - 109.2 °F / 32.4 °C - 42.9 °C
In-ear measurement accuracy	±0.4 °F (±0.2 °C) within the range of 90.3 °F - 109.2 °F / 32.4 °C - 42.9 °C
Dimensions	5.3" (L) x 1.5" (W) x 1.1" (H) mm
Weight	About 0.12 lb / 54.6 g (without battery)
Operating conditions	50.0 °F - 104.0 °F (10.0 °C - 40.0 °C) with a relative humidity of 15% - 85%, 86 kPa to 106 kPa
Storage and transport conditions	-4 °F - 140 °F (-20 °C - 60 °C) with a relative humidity of 15% - 85%, 86 kPa to 106 kPa

## Disposal

This product may contain lead and mercury. Disposal of these materials may be regulated due to environmental considerations. For disposal or recycling information, please contact your local authorities or visit **[www.recycle.philips.com](http://www.recycle.philips.com)**.

This product contains batteries:

- Dispose of batteries properly. Do not incinerate. Batteries may explode if overheated.
- Do not wrap in metal or aluminum foil. Tape the waste battery terminals before discarding.
- It is suggested that you contact your local town or city to determine proper battery redemption site(s) in your area.
- Battery must be removed from the device before discarding.

## Troubleshooting

This chapter summarizes the most common problems you could encounter with the device. If you are unable to solve the problem with the information below, visit **[www.philips.com/support](http://www.philips.com/support)** for a list of frequently asked questions or call 1-844-531-6861 for assistance.

Problem	Possible cause	Solution
The device does not respond or resets automatically when I pull out the battery strip.	The batteries are empty.	Replace the batteries.

Problem	Possible cause	Solution
	The poles of the batteries point in the wrong direction.	Remove the batteries and reinsert them properly.
	The batteries are not making proper contact.	Remove the batteries and reinsert them properly.
	The measured temperature is lower than 90.3 °F / 32.4 °C or higher than 109.2 °F / 42.9 °C.	Follow the instructions in the user manual for proper measurement.
	Temperature is out of temperature measuring range.	Use the thermometer in the range of operating conditions.
The thermometer seems to be inaccurate or the measurement results seem questionable.	The infrared sensor is not clean enough.	Clean the infrared sensor with a cotton swab according to user manual.
	You are not measuring temperature in the correct way.	Make sure that you have read the user manual and know how to use the device properly.
	Your room temperature is too low or too high.	Use your thermometer at room temperatures between 50.0 °F and 104.0 °F (10.0 °C and 40.0 °C).
	You are using the device outdoors.	The device is only intended for indoor use.
	You have held the device in your hand too long.	Put the device on the table in the room where the measurement is taking place and let it cool down first.
The battery symbol flashes on the display, or the display is blank.	The batteries are empty.	Replace the empty batteries with new ones.
The display shows Err0	Self-test error	Remove and replace the batteries. If the error persists, call 1-844-531-6861 for assistance.
The display shows Err2	Stabilization error	Wait 30 minutes and measure again.
The display shows Err3	Battery level low	Replace the empty batteries with new ones.

## Assistance

For assistance, visit our website:

**[www.philips.com/support](http://www.philips.com/support)** or call toll free **1-844-531-6861**

Online information is available 24 hours a day, 7 days a week.

## Full Two-Year Warranty

Philips Electronics North America Corporation warrants each new Philips product, Model DL8740, against defects in materials or workmanship for a period of two years from the date of purchase and agrees to repair or replace any defective product without charge.

**IMPORTANT:** This warranty does not cover damage resulting from accident, misuse or abuse, lack of reasonable care, the affixing of any attachment not provided with the product or loss of parts or subjecting the product to any but the specified battery.\*

**NO RESPONSIBILITY IS ASSUMED FOR ANY SPECIAL, INCIDENTAL OR CONSEQUENTIAL DAMAGES.**

In order to obtain warranty service, simply go to [www.philips.com/support](http://www.philips.com/support) or call toll-free 1-844-531-6861. It is suggested that for your protection you return shipments of product by insured mail, insurance prepaid. Damage occurring during shipment is not covered by this warranty.

**NOTE:** No other warranty, written or oral, is authorized by Philips Electronics North America Corporation. This warranty gives you specific legal rights, and you may also have other rights which vary from state to state. Some states do not allow the exclusion or limitation of incidental or consequential damages, so the above exclusion and limitations may not apply to you.

\* Read enclosed instructions carefully.

In U.S. only manufactured for: Philips Personal Health Solutions, High tech Campus 37, 5656 AE Eindhoven, The Netherlands.

PHILIPS and Philips Shield are Registered Trademarks of Koninklijke Philips N.V.

## Bluetooth wordmark

The Bluetooth® Smart wordmark and logos are registered trademarks owned by Bluetooth SIG, Inc. and any use of such marks by Philips is under license.

## App Store and iPhone

App Store and iPhone are trademarks of Apple Inc., registered in the U.S. and other countries. App Store is a service mark of Apple Inc.

## Google Play and Android

Google Play and Android are trademarks of Google Inc.

## EMC Guidance

- The ear thermometer needs special precautions regarding EMC and needs to be installed and put into service according to the EMC information provided in this document.
- Wireless communications equipment such as wireless home network devices, mobile phones, cordless telephones and their base stations, walkie-talkies can affect this equipment and should be kept at least a distance  $d = 3.3$  m away from the equipment.

---

Note: As indicated in IEC 60601-1-2:2007 for ME equipment, a typical cell phone with a maximum output power of 2 W yields  $d = 3.3$  m at an immunity level of 3V/m.

---

## FCC Compliance information

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

### Radio interference

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation.

If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment to an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

### RF Radiation exposure statement

This equipment complies with FCC RF radiation exposure limits set forth for an uncontrolled environment. For handheld/body-worn operation, this equipment has been tested and meets the FCC RF exposure guidelines. This transmitter must not be co-located or operating in conjunction with any other antenna or transmitter. Use of other accessories may not ensure compliance with FCC RF guidelines.

Do not attempt to repair or modify this equipment. Any repairs or alterations made by the user to the equipment may void the warranty and compliance of the equipment. Changes or modifications made to this equipment not expressly approved by Philips may void the FCC authorization to operate this equipment. For assistance visit our website [www.philips.com/support](http://www.philips.com/support) or call toll-free 1-844-531-6861.

## Electromagnetic emissions and immunity

The device is approved according to EMC safety standard IEC 60601-1-2. It is designed to be used in typical domestic environments.

### Table 1 Guidance and manufacturer's declaration – electromagnetic emissions – for all ME equipment and ME systems

Guidance and manufacturer's declaration – electromagnetic emissions

The device is intended for use in the electromagnetic environment specified below. The customer or the user of the device should assure that it is used in such an environment.

Emissions test	Compliance	Electromagnetic environment – guidance
RF emissions CISPR 11	Group 1	The device must emit electromagnetic energy in order to perform its intended function. Nearby electronic equipment may be affected.
RF emissions CISPR 11	Class B	
Harmonic emissions IEC 61000-3-2	Not applicable	
Voltage fluctuations/flicker emissions IEC 61000-3-3	Not applicable	

### Guidance and manufacturer's declaration – electromagnetic immunity – for all ME equipment and ME systems

Guidance and manufacturer's declaration – electromagnetic immunity

The device is intended for use in the electromagnetic environment specified below. The customer or the user of the device should assure that it is used in such an environment.

Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Electrostatic discharge (ESD) IEC 61000-4-2	±6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Power frequency (50/60-Hz) magnetic field IEC 61000-4-8	3A/m	3A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.

**Table 4 Guidance and manufacturer's declaration – electromagnetic immunity – for ME equipment and ME systems that are not life supporting**

Guidance and manufacturer's declaration – electromagnetic immunity. The device is intended for use in the electromagnetic environment specified below. The customer or the user of the device should assure that it is used in such an environment.

IMMUNITY test	IEC 60601 TEST LEVEL	Compliance level
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 Vrms
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2.5 GHz	3 V/m

**Electromagnetic environment - guidance**

Portable and mobile RF communications equipment should be used no closer to any part of the device, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.

Recommended separation distance:

$$d = 1.167 \sqrt{P}$$

$$d = 1.167 \sqrt{P} \text{ 80 MHz to 800MHz}$$

$$d = 2.333 \sqrt{P} \text{ 800 MHz to 2.5 GHz}$$

where  $P$  is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and  $d$  is the recommended separation distance in metres (m).

Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey (a), should be less than the compliance level in each frequency range (b).

Interference may occur in the vicinity of equipment marked with the following symbol:



NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies.

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

(a) Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the device is used exceeds the applicable RF compliance level above, the device should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the device.

(b) Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3V/m.

**Table 6 Recommended separation distances between portable and mobile RF communications equipment and the ME equipment or ME system – for ME equipment and ME systems that are not life supporting**

Recommended separation distances between portable and mobile RF communications equipment and the device.

The device is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the device can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the device as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter (W)	Separation distance according to frequency of transmitter (m)		
	150 kHz to 80 MHz	80 MHz to 800 MHz	800 MHz to 2.5 GHz
	$d = 1.167 \sqrt{P}$	$d = 1.167 \sqrt{P}$	$d = 2.333$
0.01	0.117	0.117	0.233
0.1	0.369	0.369	0.738
1	1.167	1.167	2.333
10	3.690	3.690	7.378



### Separation distance according to frequency of transmitter (m)

100	11.67	11.67	23.33
-----	-------	-------	-------

For transmitters rated at a maximum output power not listed above, the recommended separation distance  $d$  in metres (m) can be estimated using the equation applicable to the frequency of the transmitter, where  $P$  is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1 At 80MHz and 800MHz, the separation distance for the higher frequency range applies.

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

### Explanation of symbols

The warning signs and symbols are essential to ensure that you use this product safely and correctly and to protect you and others from injury. Below you find the meaning of the warning signs and symbols on the label and in the user manual.



Read the user manual before you use the device.



This symbol means that the part of the device that comes into physical contact with the user (also known as the applied part, probe tip) is of type BF (Body Floating) according to IEC 60601-1.



Indicates the manufacturer.



Symbol for the 'Bluetooth Smart mark'. The device uses Bluetooth for communication.



Indicates the manufacturer's serial number so that a specific medical device can be identified.



Indicates manufacturer's catalog number of the device.



Indicates caution. The user should consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself.

IP22

IP22: The first number 2: Protected against solid foreign objects of 12,5 mm  $\Phi$  and greater. The second number: Protected against vertically falling water drops when enclosure is tilted up to 15°. Vertically falling drops shall have no harmful effects when the enclosure is tilted at any angle up to 15° on either side of the vertical.



Indicates the storage and transportation temperature limits to which the medical device can be safely exposed: -4°F - 140°F / -20°C to 60°C.



Indicates the storage and transportation relative humidity limits to which the medical device can be safely exposed: 15% - 85%



Symbol for the 2 year Philips warranty.



This symbol is a certification mark. It is used on electronic products manufactured or sold in the United States. It confirms that the electromagnetic interference from the device is lower than the limits approved by the FCC.