Passport[™] ITE Guide



Table of Contents

| Your Passport™ Hearing Instruments | 1 |
|---|----|
| Practical Solutions to Everyday Problems | 2 |
| Getting the Most Out of Your Passport™ Hearing Instruments | 3 |
| Using the Passport™ ITE Guide | 4 |
| Turning Your Passport™ Hearing Instruments On and Off | 6 |
| Inserting and Removing ITE and CIC Hearing Instruments | 7 |
| Feedback-Free Listening | 8 |
| Replacing the Battery | 9 |
| Operating Instructions ITE and CIC | 11 |
| Caring for Your Passport™ Hearing Instruments | 22 |
| Cleaning Your Passport™ Hearing Instruments | 23 |
| Warnings | 24 |
| Troubleshooting Guide | 26 |
| Warning to Hearing Instrument Dispensers | 28 |

Your Passport™ Hearing Instruments

| Hearing Healthcare Professional: | |
|---|--|
| | |
| | |
| Telephone: | |
| Model: | |
| | |
| Serial Number: | |
| Replacement Batteries: ☐ Size 10 ☐ Size 312 ☐ Size 13 | |
| Warranty: | |
| Program 1 is the Automatic Program | |
| Program 2 is the Manual Program for: | |
| Program 3 is the Manual Program for: | |
| Program 4 is the Manual Program for: | |
| Date of Purchase: | |

1

Practical Solutions to Everyday Problems

Congratulations on choosing Passport™ ITE (In-the-Ear) hearing instruments. For over 40 years, Unitron Hearing has been committed to making life better for people with hearing loss. This means a commitment to developing high-quality hearing solutions that incorporate special features to solve the everyday problems and concerns you have with hearing loss and hearing instruments.

Passport is a premium hearing instrument that delivers superior sound quality, performance, and extraordinary results. Passport ITE hearing instruments can be purchased with Unifi Wireless option which gives access to Bluetooth accessories. Passport's automatic program identifies your listening environment and automatically determines hearing instrument settings that are most effective for optimal listening as your listening needs change throughout the day. Three additional manual programs give you added flexibility to meet your particular listening needs. Passport offers you the ability to adjust your personal comfort and clarity settings, giving you greater control of your listening experience. Passport will also learn these adjustments, leading to a more personalized hearing instrument.

Getting the Most Out of Your Passport[™] Hearing Instruments

Adjusting to your new hearing instruments will take some time. In the beginning it is important that you do not use the hearing instruments for longer than is comfortable. Depending on your previous experience with hearing instruments, a few hours a day may be enough and then you can increase wearing time gradually. Once you have become accustomed to your hearing instruments, you should wear them all day everyday since frequent use will help you adapt to your hearing instruments and enjoy their full benefits. The quicker you get used to the everyday sounds around you, the less you will notice that you are wearing hearing instruments.


Using the Passport™ ITE Guide

Refer to the table of contents for a complete listing of the topics covered in this guide.

Use the diagram on the next page to identify some of the components on your hearing instruments. Because each instrument is custom-designed and molded to fit an individual ear, the physical appearance and exact positioning on your hearing instruments may vary slightly from those shown. Have your hearing healthcare professional place a check beside the diagram that best describes your ITE style.

| my nearing instrument | nas (check all that apply) |
|-----------------------|------------------------------|
| ☐ SmartFocus | ☐ Program Button |
| ☐ Volume Control | ☐ Push Button Volume Control |

☐ Full Shell or Full Shell Power



☐ Half Shell or Half Shell Power



☐ Canal or Power Canal



☐ Mini Canal or Power Mini Canal



☐ Completely-in-the-Canal (CIC) or Power CIC



Legend

- 1 Microphone
- 2 Program Button or Push Button Volume Control (depending on your customized fitting)
- 3 Volume Control or SmartFocus (depending on your customized fitting)
- 4 Battery Door
- 5 Vent or IntelliVent
- 6 Removal Handle

Turning Your Passport™ Hearing Instruments On and Off

To turn your hearing instruments on, close the battery door. Note: If the start up delay is activated, your hearing instruments will not turn on for 10-15 seconds after closing the battery door.

To turn your hearing instruments off, open the battery door or remove the battery.

Inserting and Removing ITE and CIC Hearing Instruments

ITE and CIC hearing instruments are color-coded red for the right ear and blue for the left ear. The color is indicated either on the shell or on a label on the hearing instrument.

To insert ITEs and CICs:

- Hold the hearing instrument between your thumb and index finger with the battery door away from your ear.
- If your hearing instrument is a CIC, insert by holding the hearing instrument with the removal handle at the bottom.
- Gently insert the canal portion of the hearing instrument into your ear using your index finger to push back and tuck the hearing instrument completely into place.

To remove ITEs and CICs:

- If your hearing instrument is a CIC, gently pull on the removal handle. To remove ITE hearing instruments, grasp the instrument with your thumb and index finger.
- 2. Move your jaw up and down or apply pressure to the back of your ear to help loosen the instrument.

Note: Never use the volume control, program button or the battery door to remove your hearing instruments.

Feedback-Free Listening

Your Passport hearing devices have an optional start up delay that can be activated by your hearing healthcare professional. If the start up delay is activated, your hearing devices will not turn on for 10-15 seconds after you close the battery door. The hearing devices return to the start up delay position each time you turn your hearing devices on. This allows you to insert your hearing devices without experiencing whistling.


If you wear hearing devices now, you may have experienced whistling from your hearing devices when you talk, chew, use the telephone or hug someone. Passport's feedback technology dramatically reduces this uncomfortable whistling before you or anyone else can perceive it.


Replacing the Battery

- 1. Gently swing out the battery compartment with your fingernail.
- Grasp the battery with your thumb and index finger and remove.
- Insert the new battery into the battery compartment with the plus (+) sign on the battery facing the same way as the plus (+) sign on the edge of the battery door. This will ensure that the battery door closes properly.

Note: If the battery is inserted incorrectly, the door will not close.

4. Close the battery door.





Low Battery Warning

When you hear two long beeps, your hearing instruments are warning you that their batteries are low. You will hear the warning approximately every 30 minutes until you change the batteries or the batteries

die. After the first warning, you may experience some reduction in sound quality. This is normal and can be remedied by inserting fresh batteries in the hearing devices. If you prefer, your hearing healthcare professional can change the pitch and loudness of the low battery beep or turn it off entirely.

Caring for Batteries

- Always discard batteries carefully.
- To prolong battery life, remember to turn your hearing instruments off when not in use.
- Remove the batteries and keep the battery door open while hearing instruments are not in use. This will allow internal moisture to evaporate.

Warnings

- Never leave hearing instruments or batteries where small children and pets can reach them.
- Never put hearing instruments or batteries in your mouth. If a hearing instrument or battery is swallowed, call a physician immediately.

Operating Instructions ITE and CIC

Changing Programs on Your Passport™ Hearing Instruments

Your Passport hearing instruments have an automatic program and up to three additional manual programs. Passport's automatic program will satisfy most of your listening needs by adjusting to different listening environments without requiring manual adjustments. Up to three additional manual programs can be individually tailored by your hearing healthcare professional to meet your particular listening needs.

Your hearing instruments may come equipped with a push button which has been set for switching between programs or adjusting your volume. This push button could also be disabled by your hearing healthcare professional. If you have an active push button program control, each time you push the button, you will move to a new program.

Your Passport hearing instruments may also come with an optional remote control which allows you to switch between different listening programs. Ask your hearing healthcare professional for more information on the remote control.

duol ink

By pressing the push button on only one hearing

instrument, you can adjust both instruments simultaneously if this feature has been enabled by your hearing healthcare professional.

Program Beeps

Your hearing instruments beep to indicate which program you are in (i.e. one beep for program 1, two beeps for program 2, etc.) If you prefer, your hearing healthcare professional can adjust the pitch and loudness of the program beep or turn it off entirely.

| Program 1 (e.g. Automatic Program) | \$ | 1 beep |
|------------------------------------|--|--------------|
| Program 2 (e.g. Group/Party Noise) | 1-1 | 2 beeps |
| Program 3 (e.g. Easy-t/Telephone) | 1-1-1 | 3 beeps |
| Program 4 (e.g. Music) | 1-1-1-1 | 4 beeps |
| Easy-t/Telephone | 717 | short melody |
| Easy-DAI | 111 | short melody |
| duoLink | beeps the same as feature being adjusted | |

Adjusting the Volume

The volume on your hearing instruments adjusts automatically in response to the loud or quiet sounds around you. Your hearing instruments are set to your ideal volume by default. Your ideal volume is indicated by 1 beep. If you have a push button volume control or a volume control rotary dial, you can further adjust the volume.

Your Passport hearing instruments will learn your volume control adjustments while the hearing instruments are set to the automatic program, if this feature has been enabled by your hearing healthcare professional. Over time, the hearing instruments will adjust the default volume settings in the automatic program to the levels that you typically use.

Your Passport hearing instruments may also come with an optional remote control which allows you to adjust your volume levels. Ask your hearing healthcare professional for more information on the remote control.

Push Button Volume Control

If your push button has been configured as a volume control, you can adjust the volume level by pushing the button. Note that your volume control may be disabled by your hearing healthcare professional and, therefore, could be inactive.

As you change the volume level, your hearing instrument will beep. Please refer to the chart below to learn more about the different volume beeps.

| Volume Setting | Beeps | ; |
|--------------------|-------|------------------------------|
| Ideal volume level | 1 | 1 beep |
| Increased loudness | 11 | 1 beep + 1 high-pitched beep |
| Decreased loudness | 11 | 1 beep + 1 low-pitched beep |

Your hearing healthcare professional can help you identify your different push button volume control settings. If you prefer, your hearing healthcare professional can adjust the pitch and loudness of the push button volume control beeps or turn them off entirely.

Volume Control

As you change the volume level on your rotary or scroll volume control, your hearing instruments will beep.

Please refer to the chart below to learn more about the different volume beeps.

| Volume Setting | Beep | |
|----------------------|------|---------|
| Ideal volume level | 1 | 1 beep |
| Maximum volume level | 11 | 2 beeps |
| Minimum volume level | 11 | 2 beeps |

Your hearing healthcare professional can help you identify your different volume control settings. If you prefer, your hearing healthcare professional can adjust the pitch and loudness of the volume control beeps or turn them off entirely.

Rotary Volume Control

If the rotary dial has been configured as a volume control, rotate the volume control slowly forward

towards your nose to increase the volume or slowly backward away from your nose to decrease the volume. Note that your volume control may be disabled by your hearing healthcare professional and, therefore, could be inactive.

Scroll Volume Control

If your hearing instruments have a scroll volume control, turn the volume control slowly upward towards the ceiling to increase the volume, or slowly downward towards the floor to decrease the volume. Note that your volume control may be disabled by your hearing healthcare professional and, therefore, could be inactive.

Adjusting the SmartFocus

With a Passport hearing instrument, you can also manage the levels of speech and background noise in your listening environment via the SmartFocus. This control provides an additional sound refinement option beyond a traditional volume control. If your hearing healthcare professional has enabled the SmartFocus, you can adjust your comfort-clarity levels using the rotary dial on your hearing instruments.

Your Passport hearing instruments will learn your comfort-clarity settings while the hearing instruments are set to the automatic program, if this feature has been enabled by your hearing healthcare professional.

Over time, the hearing instruments will adjust the preprogrammed comfort-clarity settings to your preferred levels.

Your Passport hearing instruments may also come with an optional remote control which allows you to adjust your SmartFocus comfort and clarity levels. Ask your hearing healthcare professional for more information on the remote control.

SmartFocus

As you change the SmartFocus comfort and clarity level, your hearing instruments will beep. Please refer to the chart below to learn more about the different comfort-clarity beeps.

| SmartFocus Setting | Beep | |
|---------------------------------|------|---------|
| Maximum sound clarity level | 11 | 2 beeps |
| Maximum listening comfort level | 11 | 2 beeps |
| Ideal comfort and clarity level | 1 | 1 beep |

Note: As the SmartFocus is adjusted toward the midpoint of the control range, the effect on sound is minimized.

Your hearing healthcare professional can help you identify your different SmartFocus settings. If you prefer, your hearing healthcare professional can adjust the pitch and loudness of the SmartFocus beeps or turn them off entirely.

Rotary SmartFocus

If the rotary dial on your hearing instruments has been configured as a SmartFocus, rotate the dial slowly forward towards your nose to increase clarity of sounds, such as speech. For greater comfort in noisy listening situations, slowly rotate the dial backwards away from your nose.

Scroll SmartFocus

If the scroll wheel on your hearing instruments has been configured as a SmartFocus, turn the scroll slowly upward towards the ceiling to increase clarity of sounds in front of you, such as speech. For greater overall listening comfort in noisy listening situations, turn the scroll wheel slowly downward towards the floor.

Listening in Windy Environments

Passport's wind noise manager will engage automatically based on whether wind conditions are moderate or high. When the wind noise manager is engaged, sounds such as speech may become quieter because the wind noise manager is working to reduce the loud noise produced by the wind. When you are no longer in a windy environment, the wind noise manager will not be active and desirable sounds, such as speech, will once again become louder.

Listening in Quiet and Noisy Environments

Your Passport hearing instruments may have a directional microphone system to meet your listening needs in different environments. The directional system focuses on sounds in front of you (i.e., speech) while reducing sounds from the sides or behind you (i.e., noise). The directional system can be set to track moving noise sources and adapt to changing noise levels so that background noise is reduced. Your hearing healthcare professional can tell you how your directional system has been customized for you and which listening programs have the directional microphones activated.

In addition, Passport contains antiShockTM technology that identifies and minimizes sudden impulse sounds that many hearing device wearers find irritatingly loud such as slamming doors or clattering dishes. This technology is designed to increase listening comfort in adverse listening situations without impacting sound quality or your ability to understand conversations.

Using Passport™ with the Telephone

Passport has the ability, depending on hearing loss, vent size and style of hearing device to provide feedback-free phone use without program changes. For many wearers this means, when the phone rings, all you have to do is pick up the telephone and hold it

to your ear normally. In some situations when using a cell phone, you may experience digital interference that sounds like static, buzzing or beeping. If you experience interference, increase the distance between your hearing instrument and the phone receiver.

Easy-t for the Telephone or Cell/Mobile Phone

Passport can also come equipped with an optional easy-t (automatic telephone switch) that can help you listen on the telephone. Easy-t automatically switches your hearing instrument into a telephone listening mode with hearing instrument compatible phones. You will hear a short melody to indicate you are in the telephone (easy-t) program. If your phone is hearing instrument compatible, it will have a magnetic coil and easy-t will activate automatically when the telephone is held to the ear. Once the telephone is removed from the ear, the hearing instrument will switch back to the normal listening mode. Since the location and strength of the magnetic coil varies among phone manufacturers, it may be necessary to move the telephone receiver slightly to find the best reception. If the hearing instrument does not switch to telephone program automatically when the telephone receiver is placed in proximity, the magnet for easy-t hearing instruments should be attached to the telephone receiver. The magnet is designed to strengthen the

magnetic field at the ear piece of hearing instrument compatible telephones.

Depending on the programming provided by your Hearing Healthcare Professional, easy-t may also be equipped to automatically adjust the hearing instrument settings on your non-phone ear to assist with telephone situations (eg. reduce loudness and resulting interference of environmental sounds in non-phone ear).

To affix the easy-t magnet:

- 1. Clean the telephone receiver.
- Hold the magnet near the "listening end" of your telephone receiver and release it (Figure 1). The magnet will flip to the appropriate side and seek the optimal position on the telephone receiver.
- 3. Place the double-sided tape in this optimal position on the telephone receiver (Figure 2) and attach the magnet to the tape (Figure 3).



Figure 1



Figure 2



Figure 3

Warnings

- Be sure the magnet is securely affixed to the telephone.
- Keep loose magnets out of reach of children and pets.
- If magnet falls into your ear, contact your hearing healthcare professional.
- If magnet is swallowed, contact your physician immediately.
- The magnet may affect some medical devices or electronic systems. Always keep the magnet (or the telephone equipped with the magnet) at least 30 cm (12") away from pacemakers, credit cards, floppy disks or other magnetically sensitive devices.
- Too high distortion during dialing or phoning may mean that the phone handset is stressed by the magnet. To avoid any damage, please move the magnet to another place on the telephone receiver.

uDirect

Your Passport hearing devices may come with an optional uDirect™, a connectivity device between your hearing instruments and Bluetooth-enabled cell phones, or a standard audio jack. Ask your hearing healthcare professional for more information on uDirect.

Your Passport hearing instruments may also come with an optional uDirect accessory. uDirect is a connectivity device that is worn around your neck, and provides wireless transmission of various sound sources directly to your hearing instruments, thereby helping to reduce impact of background noise on the signals you are hearing. uDirect provides easy handsfree access to clear audio signals from Bluetoothenabled devices such as cell phones or to other audio sources using a standard audio jack plug-in.

Caring for Your Passport™ Hearing Instruments

- Open the battery door when not in use.
- Always remove your hearing instruments when using hair care products. The hearing instruments can become clogged and cease to function properly.
- Do not wear your hearing devices in the bath or shower or immerse them in water.
- If your hearing instruments do become wet, do not attempt to dry them in an oven or microwave. Do not adjust any controls. Open the battery doors immediately, and allow your hearing instruments to dry naturally for 24 hours.
- Protect your hearing instruments from excessive heat (hair dryer, vehicle glove box or dashboard).
- Regular use of a dehumidifier, such as a Dri-Aid kit, can help prevent corrosion and prolong the life of your hearing instruments. See your hearing healthcare professional for more information.
- Do not drop your hearing instruments or knock them against hard surfaces.

Cleaning Your Passport™ Hearing Instruments

Ear wax is natural and common. Ensuring your hearing instruments are free of ear wax is an important step in your daily cleaning and maintenance routine.

- Using the brush that was provided with your hearing instruments, clean ear wax from your hearing instruments everyday.
- Never use alcohol to clean your hearing instruments.
- Do not use sharp tools to dislodge ear wax.
 Sticking household items into your hearing instruments can seriously damage them.
- Talk to your hearing healthcare professional about regular appointments to have the wax removed from your hearing instruments.

Warnings

- Hearing instruments should only be used as directed by your physician or hearing healthcare professional.
- Hearing instruments will not restore normal hearing and will not prevent or improve a hearing impairment resulting from organic conditions.
- Do not use your hearing instruments in explosion hazard areas.
- Allergic reactions to hearing instruments are unlikely. However, if you experience itching, redness, soreness, inflammation or a burning sensation in or around your ears, inform your hearing healthcare professional and contact your physician.
- In the unlikely case that any parts remain in the ear canal after the removal of the hearing instrument, contact a physician immediately.
- Remove your hearing instruments for CT and MRI scans or for other electromagnetic procedures.
- Special care should be exercised in wearing hearing instruments when maximum sound pressure levels exceed 132 decibels. There may be a risk of impairing your remaining hearing. Speak with your hearing healthcare professional to ensure the maximum output of your hearing instruments is

25

suitable for your particular hearing loss.

Precautions

- The use of hearing instruments is only part of hearing rehabilitation; auditory training and lip reading instruction may be required as well.
- In most cases, infrequent use of hearing instruments does not provide full benefit. Once you have become accustomed to your hearing instruments, wear your hearing instruments everyday all day.
- Your hearing instruments use the most modern components to provide the best possible sound quality in every listening situation. However, communication devices such as digital cell phones can create interference (a buzzing sound) in hearing instruments. If you experience interference from a cell phone being used close by, you can minimize this interference in a number of ways. Switch your hearing instruments to another program, turn your head in a different direction or locate the cell phone and move away from it.

Labeling

For most ITE hearing instruments, the year of manufacture is located above the serial number. The first two digits indicate the year of manufacture.

Troubleshooting Guide

| CAUSE | POSSIBLE REMEDY |
|---|--|
| No sound | |
| Not turned on Low/dead battery Poor battery contact Battery upside down Hearing instruments blocked with ear wax | Turn on Replace battery Consult your hearing healthcare professional Insert battery plus (+) side up Refer to section "Cleaning Your Passport Hearing Instruments". Consult your hearing healthcare professional. |
| Not loud enough | |
| Low volume Low battery Hearing instruments not inserted properly Change in hearing Hearing instruments blocked with ear wax | Turn up volume; see hearing healthcare professional for models without a manual volume control or if problem persists. Replace battery Reinsert carefully. See "Inserting and Removing ITE and CIC Hearing Instruments". Consult your hearing healthcare professional Refer to section "Cleaning Your Passport Hearing Instruments". Consult your hearing healthcare professional. |
| Intermittent | |
| Low batteryDirty battery contact | Replace battery Consult your hearing healthcare professional |
| Two long beeps | |
| • Low battery | Replace battery |

| CAUSE | POSSIBLE REMEDY |
|--|---|
| Whistling | |
| Hearing instruments not inserted properly Volume too loud Hand/clothing near ear Poorly fitting hearing instruments | Remove and reinsert Turn down volume Remove hand/clothing from ear Consult your hearing healthcare professional |
| Not clear, distorted | |
| Poorly fitting hearing instruments Hearing instruments blocked with ear wax Change in hearing Low battery | Consult your hearing healthcare professional Refer to section "Cleaning Your Passport Hearing Instruments". Consult your hearing healthcare professional. Consult your hearing healthcare professional Replace battery |
| Weak on the telephone | |
| Low volume Telephone not positioned properly Hearing instrument requires adjustment | Turn up volume Move telephone around ear for clearer signal. See "Using Passport with the Telephone" and "Easy-t for the Telephone or Cell/Mobile Phone". Consult your hearing healthcare professional |

For any problems not listed in the guide, contact your hearing healthcare professional. If you do not have a hearing healthcare professional, please contact the nearest office listed on the back page of this booklet.

Warning to Hearing Instrument Dispensers

A hearing instrument dispenser should advise a prospective hearing instrument user to consult promptly with a licensed physician (preferably an ear specialist) before dispensing a hearing instrument if the hearing instrument dispenser determines through inquiry, actual observation, or review of any other available information concerning the prospective user, that the prospective user has any of the following conditions: (i) Visible congenital or traumatic deformity of the ear. (ii) History of active drainage from the ear within the previous 90 days. (iii) History of sudden or rapidly progressive hearing loss within the previous 90 days. (iv) Acute or chronic dizziness. (v) Unilateral hearing loss of sudden or recent onset within the previous 90 days. (vi) Audiometric air-bone gap equal to or greater than 15 decibels at 500 hertz (Hz), 1,000 Hz, and 2,000 Hz. (vii) Visible evidence of significant cerumen accumulation or a foreign body in the ear canal. (viii) Pain or discomfort in the ear. Special care should be exercised in selecting and fitting a hearing instrument whose maximum sound pressure level exceeds 132 decibels because there may be risk of impairing the remaining hearing of the hearing instrument user. [This provision is required only for those hearing instruments with a maximum sound pressure capability greater than 132 decibels (dB).]

Important Notice for Prospective Hearing Instrument Users

Good health practice requires that a person with a hearing loss have a medical evaluation by a licensed physician (preferably a physician who specializes in diseases of the ear) before purchasing a hearing instrument.

Licensed physicians who specialize in diseases of the ear are often referred to as otolaryngologists, otologists or otorhinolaryngologists. The purpose of medical evaluation is to assure that all medically treatable conditions that may affect hearing are identified and treated before the hearing instrument is purchased. Following the medical evaluation, the physician will give you a written statement that states that your hearing loss has been medically evaluated and that you may be considered a candidate for a hearing instrument. The physician will refer you to an audiologist or a hearing instrument dispenser, as appropriate, for a hearing instrument evaluation. The audiologist or hearing instrument dispenser will conduct a hearing instrument evaluation to assess your ability to hear with and without a hearing instrument. The hearing instrument evaluation will enable the audiologist or dispenser to select and fit a hearing instrument to your individual needs. If you have reservations about your ability to adapt to

amplification, you should inquire about the availability of a trial-rental or purchase-option program. Many hearing instrument dispensers now offer programs that permit you to wear a hearing instrument for a period of time for a nominal fee after which you may decide if you want to purchase the hearing instrument. Federal law restricts the sale of hearing instruments to those individuals who have obtained a medical evaluation from a licensed physician. Federal law permits a fully informed adult to sign a waiver statement declining the medical evaluation for religious or personal beliefs that preclude consultation with a physician. The exercise of such a waiver is not in your best health interest and its use is strongly discouraged.

Children With Hearing Loss

In addition to seeing a physician for a medical evaluation, a child with a hearing loss should be directed to an audiologist for evaluation and rehabilitation since hearing loss may cause problems in language development and the educational and social growth of a child. An audiologist is qualified by training and experience to assist in the evaluation and rehabilitation of a child with a hearing loss.

Notices

Notice 1

This instrument is certified under:

FCC ID: VMY-UWITE IC: 2756A-UWITE

This device complies with Part 15 of the FCC Rules and with RSS-210 of Industry Canada. Operation is subject to the following two conditions:

- 1) this device may not cause harmful interference, and
- 2) this device must accept any interference received, including interference that may cause undesired operation. Changes or modifications made to this equipment not expressly approved by Phonak may void the FCC authorization to operate this equipment.

Notice 2

This Class B digital apparatus complies with Canadian ICES-003.

Notice 3

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna
- Increase the separation between the equipment and receiver
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected
- Consult the dealer or an experienced radio/TV technician for help

Unitron Distributors

CORPORATE OFFICE/INTERNATIONAL

20 Beasley Drive, P.O. Box 9017, Kitchener, ON N2G 4X1 Canada

AUSTRALIA

Level 2, Norwest Quay, 21 Solent Circuit, Baulkham Hills, NSW, NSW 2153

BELGIUM

Baron de Vironlaan, 60 b-1700 Dilbeek

CANADA

20 Beasley Drive, P.O. Box 9017, Kitchener, ON N2G 4X1

CHINA

No. 200 Suhong Road, Export Processing Zone 4D, Suzhou Industrial Park 215021

DENMARK

Nitivej 10, DK-2000 Frederiksberg

EUROPEAN REPRESENTATIVE

Daimlerstrasse 22, 70736 Fellbach-Oeffingen, Germany

FRANCE

5, rue Maryse Bastié - BP 15 69671 Bron Cedex, France

NETHERLANDS

Archimedesbaan 19, P.O. Box 1214, 3430 BE Nieuwegein

NEW ZEALAND

10/215 Rosedale Rd., M277 Private Bag, 300987, Albany, Auckland

NORWAY

Brugata 14 0186 Oslo, Norway

SOUTH AFRICA

First Floor – Selborne House, Fourways Golf Park, Roos Street, Fourways, Johannesburg

SPAIN

Urb. El Palmeral Bl. IX, 17-27 Alicante, Spain -- 03008

SWEDEN

Förmansvägen 2, 4 tr SE-11743, Stockholm, Sweden

UNITED KINGDOM

St. George House, Cygnet Court, Centre Park, Warrington, Cheshire WA1 1PD

U.S.A.

Suite A, 2300 Berkshire Lane North, Plymouth, MN 55441

Manufacturer

UNITRON HEARING LTD.

20 Beasley Drive, P.O. Box 9017, Kitchener, ON N2G 4X1 Canada





DISTRIBUTOR

www.unitron.com

