

Rhein Tech Laboratories, Inc.
360 Herndon Parkway
Suite 1400
Herndon, VA 20170
<http://www.rheintech.com>

Client: Vocollect, Inc.
Model #: AVA-100-01
Standards: FCC 15.247 & RSS-210
ID's: MQOH1AVA-10001/2570A-H1AVA101
Report #: 2007241

Appendix J: Manual

Please refer to the following page.

Version and Compliance Information: Model Number H1 AVA 100-01

Note: This document contains information about Vocollect Healthcare Systems AccuNurse[®] Voice Assistant[™] H1 device.

- [Version](#)
- [Regulatory Policy Compliance](#)
AccuNurse Voice Assistant H1 Device Federal Communications Commission Compliance
- [Specific Absorption Rate \(SAR\) Level – VHS H1 Device](#)

Version

For version information, please consult the **About** page in the AccuNurse Web application.

Software	AccuNurse Version
AccuNurse Voice Assistant H1 Software	Consult the release notes.

Regulatory Policy Compliance

AccuNurse Voice Assistant H1 Federal Communications Commission Compliance

This device complies with [Part 15 \(b\) of the Federal Communications Commission \(FCC\) Rules](#).



Specific Absorption Rate (SAR) Level – VHS H1 Device

The AccuNurse Voice Assistant H1 with the SDC MCF10G radio has been SAR tested. The maximum SAR value measured was 0.484 W/kg.

This product has been tested to the following standards:

	Standard
United States/Federal Communications Commission	FCC Part 15, Class B: Code of Federal Regulations, Title 47 Telecommunication Part 15-Radio Frequency Devices
Canada/Industry Canada	RSS-210: Low-power License-exempt Radio Communication Devices (All Frequency Bands) Category 1 Equipment

The H1 product may contain one of these radio devices. See device label.

Card Manufacturer and P/N	AccuNurse Voice Assistant H1 FCC ID #	AccuNurse Voice Assistant H1 Industry Canada ID #
Summit SDC-MCF10G	MQOH1AVA-10001	2570A-H1AVA101

The VHS H1 product is separately approved for:

1. FCC Part 15 Subpart C
2. Industry Canada RSS210 and RSS139

The VHS H1 is nominally a Class B digital device, pursuant to Part 15 of the FCC Rules

Caution: Exposure to Radio Frequency Radiation

The AccuNurse Voice Assistant H1 device contains an internal low-power radio. The radiated output power of the radio is far below the FCC radio frequency exposure limits. Nevertheless, the device shall be used in such a manner that the potential for human contact with the radio antenna during normal operation is minimized. The device should not be used if the case is open or if the internal antenna is exposed. The device should be worn in accordance with the instructions for this device.

VHS devices are designed to be compliant with the rules and regulations in the locations into which they are sold and are labeled as required. VHS devices are type approved and do not require the user to obtain license or authorization before using them. Changes or modifications not expressly approved by Vocollect Healthcare Systems, Inc., could void the user's authority to operate the equipment.

This Class B digital apparatus complies with Canadian ICES-003.
Cet appareil numérique de la classes B est conforme à la norme NMB-003 du Canada.

Warning: The AccuNurse Voice Assistant H1 device is a class B product. In a domestic environment, this product may cause radio interference in which case the user may be required to take adequate measures.

Part 15 (b) of the Federal Communications Commission (FCC) Rules

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

Note: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced

Made in the U.S.A.

Vocollect Healthcare Systems, Inc.

Pittsburgh, PA