WIDEX **DREAM**



USER INSTRUCTIONS THE DREAM™ SERIES

D-CIC/D-CIC-TR model Completely-in-canal





YOUR WIDEX HEARING AID

(To be filled out by the hearing care professional)

Vent:	☐ Short	☐ Full length	☐ None
☐ DRE	AM440 AM330 AM220 AM110		
Progra	ms:		
☐ Mast	ter		Audibility Extender
☐ Musi	ic		Audibility Extender
□ TV			Audibility Extender
□ Com	fort		Audibility Extender
☐ Phor	ne		Audibility Extender
□ Zen			Audibility Extender
☐ Mast	ter + Zen		Audibility Extender
☐ Mast	ter + Phone		Audibility Extender
□ Zen-	+		Audibility Extender
	Date	Vous hooving	care professional
	Date	rour nearing	care professional

CONTENTS

SYMBOLS	. 4
THE HEARING AID	. 5
Acoustic indicators	. 8
The battery	. 9
Inserting the battery	. 9
Low battery indication	
Turning the hearing aid on and off	12
Right/left identification	13
Positioning the hearing aid	14
Removing the hearing aid	15
Volume adjustment	16
Lost partner alarm	. 17
Programs	18
Switching between the listening programs	21
Zen+	21
CLEANING	22
The hearing aid	23
The sound outlet	25
The relief vent	26
The microphone opening	
NANOCARE WAX GUARD	28
Changing the wax guard	
ACCESSORIES	
IN CASE OF MALFUNCTION	
CARING FOR YOUR HEARING AID	
WARNINGS	37
ADVICE	40
Adapting to your hearing aids	41
REGULATORY INFORMATION	
ADDITIONAL PROGRAM FORMS	48

SYMBOLS

The following symbols will be used throughout the manual:



Messages with this heading indicate serious adverse reactions, potential safety hazards and inadequate performance of device.




Messages with this heading indicate/include information regarding any special care to be exercised.



Not for general waste.

THE HEARING AID

- 1. Microphone opening
- 2. Battery drawer (on/off function)
- 3. Nail grip
- 4. Sound outlet
- 5. NanoCare wax guard
- 6. Extraction cord
- 7. Relief vent (short)
- 8. Relief vent (full length)



On page 2 of this booklet your hearing care professional can indicate the type of vent used in your hearing aid, if any.



This booklet contains important information and instructions. Please read this booklet carefully before you start using the hearing aid.

NOTE

Your hearing aid and accessories may not look exactly as illustrated in this booklet. We also reserve the right to make any changes considered necessary.

Intended use

The hearing aids are intended as air conduction amplification devices to be used in everyday listening environments. The hearing aids may be provided with the Zen program intended to provide a relaxing sound background (i.e. music/noise source) for adults who desire to listen to such a background in quiet.

Indications for use

The devices are indicated for individuals with a range of hearing loss severity from minimal (10 dB HL) to moderately severe (76 dB HL) and all hearing loss configurations.

They are to be programmed by licensed hearing care professionals (audiologists, hearing aid specialists, oto-laryngologists) who are trained in hearing (re)habilitation.

Description of device

The hearing aid uses a proprietary wireless technology, WidexLink, to enable communication between the left and the right hearing aids, as well as between the hearing aids and the DEX accessories.

Your hearing aid may be provided with a listening program called Zen. It makes musical tones (and sometimes a rushing noise) in the background. These sounds are shaped according to your hearing loss.

Acoustic indicators

The hearing aid may be set to produce a signal to indicate the use of certain functions. The signal may be a spoken message or tones. The signal can also be deactivated.

Functions	Default settings	Other settings
Adjusting volume via remote control	Tone	Off
Confirming RC program key use	Clicking sound	Off
Changing program via remote control	Message	Tones/off
Starting up the hearing aid	Message	Tone/off
Warning about low battery	Message	4 tones/off
Warning about lost partner	Off	Message
Service reminder	Off	Message

Note: "Warning about lost partner" can only be made available in DREAM440.

The battery

We recommend **zinc-air batteries**. Use a **size 10** battery for the hearing aid.

To obtain replacement batteries, please consult your hearing care professional. It is important to take note of the expiry date and the recommendations on the battery pack regarding disposal of used batteries. Otherwise there is a risk that the battery life may be reduced.




Inserting the battery

Before inserting a new battery into the hearing aid, remember to remove the adhesive tab. Once the tab has been removed, the battery will start functioning after a few seconds.



CAUTION


Do not use batteries if there is a sticky residue from the tab or other unwanted substance, as this can cause the hearing aid to malfunction.



Use the nail grip to open the battery drawer. Do not press the battery drawer beyond the open position.

Place the battery in the drawer so that the small plus (+) sign on the battery is visible when you hold the hearing aid as shown.

You can use the battery magnet provided to steer the battery into place.



If the battery drawer does not close easily, the battery is incorrectly inserted.

When changing battery, it is a good idea to hold the hearing aid over a table.

Low battery indication

An acoustic indicator will sound when the battery is nearly exhausted, unless this function has been disabled (see page 8). We recommend that you always have a spare battery with you.



WARNING

Never leave an exhausted battery in the hearing aid. Exhausted batteries may leak, damaging the hearing aid.



WARNING

Your hearing aid may stop functioning, for instance if the battery is exhausted. You should be aware of this possibility, in particular when you are in traffic or are otherwise dependent on warning signals.

Turning the hearing aid on and off

The battery drawer of the hearing aid also functions as the on/off switch.



Close the battery drawer to turn on the hearing aid. An acoustic indicator will indicate that the hearing aid has been switched on unless this function has been disabled.




Open the battery drawer to turn off the hearing aid.

NOTE: Another way to verify that the hearing aid is turned on is to cup it in your hand. If it is on, it will whistle.

Please remember to turn off the hearing aid when it is not in use. Remove the battery if the hearing aid will not be used for several days.

Right/left identification



Your hearing aids will be marked with a red Widex logo for the right ear and a blue logo for the left ear, or they may have a red or a blue shell, respectively. The arrow shows the position of the logo.

Positioning the hearing aid



Close the battery drawer completely.



Hold your hearing aid by the extraction cord with your thumb and forefinger. Slowly guide the hearing aid into your ear canal until you feel resistance.



Let go of the extraction cord and use the tip of your forefinger to gently push the hearing aid into your ear until it rests comfortably in your canal.



To help position the hearing aid properly you could pull the outer ear backwards and upwards with the opposite hand.

Removing the hearing aid



Gently pull the extraction cord with your thumb and forefinger.

If it is difficult, you can try to move the hearing aid carefully from side to side while pulling the extraction cord. It may also help to pull the outer ear backwards and upwards with the opposite hand.

Volume adjustment

The hearing aid volume is adjusted automatically in accordance with your sound environment.

If you have a remote control, you can also adjust the volume manually or mute your hearing aid. Any adjustment of the volume setting will be cancelled when your hearing aid is turned off, or when you change program.

Depending on hearing aid features and setting, any change you make will affect both hearing aids.

To alter the increments in which the volume is increased or decreased, consult your hearing care professional.

If you have a remote control, please also refer to the user instructions for this.



If the volume in the hearing aid is generally too loud or too weak, or the reproduced sounds are distorted, or if you would like any further information, consult your hearing care professional.

To mute the hearing aid with a remote control:

Keep pressing the volume down key on the remote control after the long beep-tone has sounded and until it stops. Pressing one of the volume keys briefly will bring back the sound.

NOTE: Your hearing aid has been set to gradually increase the amplification provided to allow you to acclimatize yourself to the sound through the hearing aid. Therefore you may notice an increase in level after a period of time.

Lost partner alarm

(Note: Only applicable if this feature is available and has been activated by your hearing care professional).

If one of your hearing aids falls out of your ear, and/ or the battery is exhausted, a tone signal and a spoken message will sound two times in the other hearing aid.

Programs

By default your hearing aid has one program. However, with a remote control a number of additional programs, plus the special Zen program called Zen+ can be made available.

Please also refer to the user instructions for your remote control.

Standard
For listening to music
For listening to the TV
Attenuates background noise
For listening to the telephone
Generates different kinds of tones or noise
The Master program in one ear and Zen or Phone in the other
Special Zen program with up to three different Zen styles

Your hearing aid has a function called the Audibility Extender. This feature makes high frequency sounds audible, and can be set to operate for one or more or all available programs.

If your needs change over time, the combination of listening programs can be altered.

Zen program

Your hearing aid may be provided with a unique optional listening program called Zen. It makes musical tones (and sometimes a rushing noise) in the background. The Zen program may be used alone (without amplification) in quiet when you are not required to hear surrounding sounds. Or, it may be used with amplification so both the surrounding sounds and the generated sounds (fractal tones and noise) are heard together.



CAUTION

Use of the different Zen programs may interfere with hearing surrounding sounds including speech. The programs should not be used when hearing such sounds is important. Switch the hearing aid to a non-Zen program in those situations.



CAUTION

If you perceive decrease in loudness, tolerance of sounds, speech not as clear, or worsening tinnitus, contact your hearing care professional.

Benefits

The Zen program may provide a relaxing listening background for some people. When the Zen program is used in a tinnitus management program, its wearer may experience some relief from tinnitus.

Indications for use

The Zen program is intended to provide a relaxing sound background (i.e., music/noise source) for adults who desire to listen to such a background in quiet. It may be used as a sound therapy tool in a tinnitus treatment program that is prescribed by a hearing care professional (audiologists, hearing aid specialists, otolaryngologists) who is trained in tinnitus management.

Directions for use

Because of the unique ways in which Zen is programmed in your hearing aid, please follow the recommendations of your hearing care professionals as to how to use the program, when to use the program and/or how long to use the program.

Switching between the listening programs

If you have a remote control, you can change programs with a simple push of the program key. Each time you change program, an acoustic indicator will sound unless this function has been disabled.

Program 1: Message or one short beep Program 2: Message or two short beeps Program 3: Message or three short beeps

Program 4: Message or one long and one short beepProgram 5: Message or one long and two short beeps

Zen+: Message or tone

Depending on hearing aid features and setting, any change you make will affect both hearing aids.

Zen+

This program is accessed via a long press (more than 1 second) on the remote control program key. A short press then allows you to cycle through the available Zen styles.

You can exit Zen+ by pressing and holding the program key for more than 1 second.

CLEANING

The following cleaning accessories are available for cleaning the hearing aid. Ask your hearing care professional which accessories you need.

- 1. Brush
- 2. Cloth
- 3. Wax removing tool with a short and a long end
- 4. Long wax removing tool



Contact your hearing care professional if you need additional supplies of cleaning accessories.

The hearing aid

It is very important for the hearing aid's performance that it is kept free from earwax and dirt.



Clean your hearing aid after use with the soft cloth to keep it dry and clean of earwax or dirt.



WARNING

Never clean the hearing aid with water or cleaning solutions, as this may cause it to malfunction.




When the hearing aid is not in use, the battery drawer should be left open, so as to ventilate the hearing aid and allow it to dry.


We recommend that every time you take the hearing aid out of your ear, you inspect the sound outlet, the microphone opening in the battery drawer and the relief vent opening to ensure that no wax or dirt has gathered there.

Below is an example of how the sound outlet, relief vent and microphone opening may be positioned in a hearing aid provided with a short vent and a full length vent, respectively.

- 1. Sound outlet
- 2. Relief vent (short)
- 3. Microphone opening



- 1. Sound outlet
- 2. Relief vent (full length, outlet)
- 3. Relief vent (full length, inlet)
- 4. Microphone opening



The sound outlet

It is important that the sound outlet is not blocked by earwax



Remove any visible earwax from around the sound outlet with the small brush or cloth. Do not attempt to clean the wax guard.

If the sound outlet is still blocked, you should change the wax guard (see pages 28-31).



WARNING

If you cannot clean the sound outlet completely, you should contact your hearing care professional. Do not insert anything into the sound outlet, as this could damage the hearing aid.

The relief vent

If earwax has accumulated in and/or around the relief vent opening, do as follows:



Short vent:

Clean any earwax from the vent as far as the long end of wax removing tool no. 3 can reach.



Full length vent:

Clean any earwax from the vent by passing the long wax removing tool (no. 4) all the way through the vent.



WARNING

It is important for the performance of the hearing aid that the relief vent is never blocked. Make it a habit to clean the relief vent every day! In rare cases so much earwax can accumulate in the relief vent that the hearing aid sound is affected. In that case you should see your hearing care professional.

The microphone opening

If dirt or earwax has gathered around the microphone opening on the hearing aid front, do as follows:

Open the battery drawer and remove the battery.

Turn the hearing aid so the open battery drawer is facing downwards.



Guide the short end of wax removing tool no. 3 through the microphone opening on the battery drawer from the inside as illustrated.



If you suspect that earwax or dirt has migrated through the microphone opening and into the actual microphone inside the hearing aid, contact your hearing care professional. Do not insert anything into the actual microphone.

NANOCARE WAX GUARD

The NanoCare wax guard helps to protect the hearing aid against earwax. Always use NanoCare wax guards with your hearing aid. Otherwise the warranty for your hearing aid will be void.

The wax guard consists of the following parts:

- 1. Holder
- 2. Removal hook
- 3. Wax guard






The wax guard is disposable and for one-time use only. Do **not** attempt to clean or reuse a wax guard. This could damage your hearing aid and/or cause the wax guard to become dislodged in your ear.

Changing the wax guard

Insert the removal hook into the used wax guard placed in the sound outlet. Pull the wax guard straight out.

- 1. Vent opening
- 2. Sound outlet
- 3. Used wax guard






Turn the holder around so that the new wax guard is in position to be inserted.



Insert the holder into the opening of the sound outlet and gently press the wax guard into the opening.



Then pull the holder straight out.

The new wax guard will automatically detach from the holder. Press the hearing aid gently against a flat surface to ensure that the wax guard remains in place.



Discard the holder with the used wax guard.

The frequency with which the wax guard should be changed varies from one individual to another.

If you have any questions regarding your wax guard, please contact your hearing care professional.



WARNING

If the wax guard fits loosely, discard it and insert another. If the wax guard becomes dislodged from the hearing aid while in the ear canal, please contact your physician. Do not try to remove the wax guard from your ear canal yourself.

ACCESSORIES

A variety of assistive listening devices are available for your hearing aid.

RC-DEX	remote control
TV-DEX	for listening to TV and audio
M-DEX	for mobile phone applications and remote control
PHONE-DEX	for easy landline use
FM+DEX	the FM+DEX is a high-quality streaming device designed specifically for Widex hearing aids

For help in determining whether you could benefit from a remote control or other accessories, please consult your hearing care professional.

IN CASE OF MALFUNCTION

The following pages contain some quick advice on what to do if your hearing aid stops working or performs unsatisfactorily. If the problems persist, contact your hearing care professional for assistance.

Problem	Possible cause	Solution
The hearing aid is com- pletely silent	The hearing aid is not turned on	Make sure the battery drawer is closed
	The battery does not work	Insert a new battery in the hearing aid battery drawer
	The sound outlet is blocked	See page 25
	The microphone opening is blocked	See page 27
The hearing aid volume is	The sound outlet is blocked	See page 25
not powerful enough	Your ear is blocked by earwax	Contact your ENT doctor/physician
	Your hearing may have changed	Contact your hearing care professional
The hearing aid whistles	Your ear is blocked by earwax	Contact your ENT doctor/physician
continuously	The hearing aid is not correctly placed in your ear canal	Take the hearing aid out and reinsert it
	The hearing aid fits loosely in your ear canal	Contact your hearing care professional

Problem	Possible cause	Solution
The hearing aid is uncomfortable to wear	The hearing aid is not placed correctly in your ear canal	Consult your hearing care professional on how to place the hearing aid correctly in your ear canal.
	The hearing aid does not fit your ear prop- erly	Contact your hearing care professional
	Your outer ear or ear canal is sore	It is important that first-time users only wear the hearing aid for short periods of time at the beginning. The skin in the ear canal is very thin and may swell a bit because of the pressure from the hearing aid. Take 2-3 weeks to gradually get used to the hearing aid. If the soreness continues, contact your hearing care professional
Your two hear- ing aids are not working in synchrony	The connection between the hearing aids is lost	Turn the hearing aids off and on again

Problem	Possible cause	Solution
The hearing aids do not respond with a corresponding change in volume or program to the DEX	a. The DEX is used beyond the transmission range b. Strong electromagnetic interference in the vicinity c. The DEX and the DREAM hearing aids are not matched	a. Move the DEX closer to the DREAM hearing aids b. Move away from known source of EM interference c. Check with hearing care professional to make sure DEX is matched with DREAM hearing aids
The wearer hears "interrupted" speech (on and off) from the hearing aids or no speech (muted) from the transmitting ear.	a. The battery in one of the hearing aids has expired b. Strong electromagnetic interference in the vicinity	a. Replace battery in one or both hearing aids b. Move away from known sources of interference

CARING FOR YOUR HEARING AID

The hearing aid is a valuable object and should be treated with care. Here are some things you can do to prolong the life of your hearing aid:



CAUTION

- Turn off your hearing aid when it is not in use. Remove the battery if the hearing aid will not be used for several days.
- When the hearing aid is not in use, keep it in its case in a cool, dry location out of reach of children and pets.
- Do not expose the hearing aid to extreme temperatures or high humidity. Make sure to dry the hearing aid thoroughly after heavy perspiration such as that which may occur during intense physical activity, e.g. playing sports.
- Avoid dropping your hearing aid perform cleaning and battery changes while holding the hearing aid above a soft surface.
- Do not wear your hearing aid while in the shower or swimming, or when using a hair dryer, perfume, hair and body sprays or gels such as suntanning lotions or creams.
- In environments with high humidity, a drying kit may be used daily to reduce the amount of moisture inside the hearing aid.
 See the instructions enclosed with the drying kit.

WARNINGS



Hearing aids and batteries can be dangerous if swallowed or used improperly. Swallowing or improper use can result in severe injury or even fatalities. In case of ingestion, contact a physician immediately.

- Keep hearing aids and their parts, accessories and batteries out of reach of children and anyone else who might swallow such items or otherwise cause injury to themselves. Do not change batteries in front of them and do not let them see where you keep your battery supply. Discard used batteries carefully.
- Batteries are very small and can easily be mistaken for pills or the like. Never put a battery or hearing aid in your mouth for any reason as you may risk swallowing it.

M WARNING

- Risk of explosion if battery is replaced by an incorrect type or recharged. Dispose of used batteries according to the instructions
- Never allow others to wear your hearing aid, as this could cause permanent damage to their hearing.
- When selecting a listening program, please remember that there are situations in which it is particularly important to be able to hear the surrounding sounds (e.g. traffic, warning signals).
- The hearing aid is made of modern non-allergenic materials. Nonetheless, in rare cases skin irritation may occur. If you notice skin irritation in or around your ear or ear canal, contact your hearing care professional.
- Please note that when using any type of hearing aid, you must allow regular ventilation of the ear. If the ear is not adequately ventilated, there may be a slightly increased risk of infection or disease in the ear canal. We therefore recommend that you remove the hearing aid and ear-set from your ear periodically to allow the ear canal to be ventilated. Make sure that you clean and inspect your hearing aid and ear-set as required. If an ear infection or disease occurs, you should seek medical attention and contact your hearing care professional for advice on how to disinfect the various hearing aid parts. Do not under any circumstances use alcohol, chlorine or similar substances for this purpose.
- Do not use Widex hearing aids in mines or other areas with explosive gases.



• Do not wear your hearing aid during radiation, X-rays, MRIs, CT or other medical treatments and scans. The emissions from these procedures as well as from other types of radiation, such as that in a microwave oven, can damage your hearing aid. Radiation from, for example, room surveillance equipment, burglar alarms and mobile phones is weaker and will not damage the hearing aid, but may create audible interference



- Your hearing aid has been tested for interference according to international standards. Nevertheless, it is possible that unforeseen interference may occur in the hearing aid due to electromagnetic radiation from other products such as alarm systems, room surveillance equipment and mobile phones.
- Although your hearing aid has been designed to comply with the most stringent international electromagnetic compatibility standards, the possibility cannot be excluded that it may cause interference with other equipment, such as medical devices.
- Never try to open or repair the hearing aid yourself.

ADVICE

NOTE

- The hearing aid will not restore normal hearing and will not prevent or improve a hearing loss resulting from organic conditions. However, the hearing aid can help you to make the best possible use of your remaining hearing ability. You should also bear in mind that it can take time to get used to a new hearing aid and new sounds.
- In most cases, using the hearing aid infrequently will not permit you to gain full benefit from it.
- The use of a hearing aid is only part of hearing habilitation and may need to be supplemented by auditory training and instruction in lipreading.
- The use of hearing aids increases the risk of accumulation of earwax. Contact your physician/ENT doctor if you suspect that a plug of earwax has accumulated in your ear. Earwax may not only reduce your own hearing but also the effect of the hearing aid considerably. It is a good idea to ask your physician to clean your ears a couple of times a year.

Adapting to your hearing aids

Congratulations for taking the first steps toward restoring your communication skills with the purchase of state-of-the-art hearing aids from Widex.

Because many of the sounds you may now hear have not been fully audible without hearing aids, we urge you to take a few minutes to perform the following simple activities in the order listed. Do these first in quiet settings, then try them in noisier surroundings.

Activity 1: First, listen to a friend or relative speak to you with your eyes closed and your hearing aids off. Now, put your hearing aids on and continue to listen. Notice the difference? The purpose of this activity is not to hear every word, but just to recognize the difference in amplified speech from a familiar source.

Activity 2: Listen to your friend or relative speak to you with your hearing aids on, but with your eyes closed. Now continue to listen with your eyes open. The goal is to demonstrate how helpful it is to receive information from your eyes to supplement your ears.

Activity 3: Purchase two copies of today's newspaper. Ask a friend or relative to read a story aloud while you read along with the same story. Alternatively, read it aloud and listen to your own voice. It is also helpful to watch captioned TV.

- We hear in our brain, not in our ears. It will take your brain some time to fully adapt to the new sounds you are hearing. It is not unusual, for example, to hear certain sounds, such as your own footsteps, or a newspaper rustling, or the refrigerator humming, that people with normal hearing take for granted and thus don't consciously process. Be patient and allow yourself a few weeks to adapt. Your brain will learn to ignore them if they are not important. Of course, if sounds are too loud or uncomfortable for you, contact your hearing care professional.
- Conversational speech has many redundant cues.
 Worrying about a word you may have missed will
 likely lead to additional missed words. Focus on the
 essence of the conversation. It may help to ask someone you are comfortable with to subtly alert you
 about the overall topic when listening in groups.
- Before going to a restaurant, call ahead and request a seat away from the kitchen or bar area. Remember that in noisy places, even normal hearing people may struggle to hear.

For more activities, training materials, and helpful communication hints please go to www.widex.com.

REGULATORY INFORMATION

FCC ID: TTY-DCIC IC: 5676B-DCIC

Federal Communications Commission Statement

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

- (1) This device may not cause harmful interference, and
- (2) this device must accept any interference received, including interference that may cause undesired operation.

NOTF:

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

NOTF:

This equipment complies with FCC radiation exposure limits set forth for an uncontrolled environment. This transmitter must not be co-located or operating in conjunction with any other antenna or transmitter.

Changes or modifications to the equipment not expressly approved by Widex could void the user's authority to operate the equipment.

Industry Canada Statement / Déclaration d'industrie Canada

Under Industry Canada regulations, this radio transmitter may only operate using an antenna of a type and maximum (or lesser) gain approved for the transmitter by Industry Canada.

To reduce potential radio interference to other users, the antenna type and its gain should be so chosen that the equivalent isotropically radiated power (e.i.r.p.) is not more than that necessary for successful communication.

This device complies with Industry Canada licence-exempt RSS standard(s). Operation is subject to the following two conditions:

- (1) this device may not cause interference, and
- (2) this device must accept any interference, including interference that may cause undesired operation of the device.

Conformément à la réglementation d'Industrie Canada, le présent émetteur radio peut fonctionner avec une antenne d'un type et d'un gain maximal (ou inférieur) approuvé pour l'émetteur par Industrie Canada.

Dans le but de réduire les risques de brouillage radioélectrique à l'intention des autres utilisateurs, il faut choisir le type d'antenne et son gain de sorte que la puissance isotrope rayonnée équivalente (p.i.r.e.) ne dépasse pas l'intensité nécessaire à l'établissement d'une communication satisfaisante.

Le présent appareil est conforme aux CNR d'Industrie Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes :

- (1) l'appareil ne doit pas produire de brouillage, et
- (2) l'utilisateur de l'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

CE 0459

Hereby, Widex A/S declares that this D-CIC and D-CIC-TR are in compliance with the essential requirements and other relevant provisions of Directive 1999/5/EC.

A copy of the Declaration of Conformity can be found at: http://www.widex.com





Hearing aids, accessories and batteries should not be disposed of with normal household waste. Please consult your national Widex distributor for advice on how to dispose of these items.

ADDITIONAL PROGRAM FORMS

Programs available in your hearing aid:

Program 1:
Program 2:
Program 3:
Program 4:
Program 5:
Special program:

Programs available in your hearing aid:

Program 1:
Program 2:
Program 3:
Program 4:
Program 5:
Special program:

Manufacturer

CE 0459

WIDEX A/S • Nymoellevej 6, DK-3540 Lynge Denmark • www.widex.com





Printed by HTO / 2012-04 9 514 0216 001 #01