

**Instruction Manual**

PN 51-Xmt-C/rev.E

January 2011

# **Model Solu Comp™ Xmt-C**

Two-Wire Contacting Conductivity Transmitter



# **ESSENTIAL INSTRUCTIONS**

## **READ THIS PAGE BEFORE PROCEEDING!**

Rosemount Analytical designs, manufactures, and tests its products to meet many national and international standards. Because these instruments are sophisticated technical products, you must properly install, use, and maintain them to ensure they continue to operate within their normal specifications. The following instructions must be adhered to and integrated into your safety program when installing, using, and maintaining Rosemount Analytical products. Failure to follow the proper instructions may cause any one of the following situations to occur: Loss of life; personal injury; property damage; damage to this instrument; and warranty invalidation.

- Read all instructions prior to installing, operating, and servicing the product. If this Instruction Manual is not the correct manual, telephone 1-800-654-7768 and the requested manual will be provided. Save this Instruction Manual for future reference.
- If you do not understand any of the instructions, contact your Rosemount representative for clarification.
- Follow all warnings, cautions, and instructions marked on and supplied with the product.
- Inform and educate your personnel in the proper installation, operation, and maintenance of the product.
- Install your equipment as specified in the Installation Instructions of the appropriate Instruction Manual and per applicable local and national codes. Connect all products to the proper electrical and pressure sources.
- To ensure proper performance, use qualified personnel to install, operate, update, program, and maintain the product.
- When replacement parts are required, ensure that qualified people use replacement parts specified by Rosemount. Unauthorized parts and procedures can affect the product's performance and place the safe operation of your process at risk. Look alike substitutions may result in fire, electrical hazards, or improper operation.
- Ensure that all equipment doors are closed and protective covers are in place, except when maintenance is being performed by qualified persons, to prevent electrical shock and personal injury.

### **NOTICE**

If a Model 375 Universal Hart® Communicator is used with these transmitters, the software within the Model 375 may require modification. If a software modification is required, please contact your local Emerson Process Management Service Group or National Response Center at 1-800-654-7768.

## **About This Document**

This manual contains instructions for installation and operation of the Model Xmt-C Two-Wire Contacting Conductivity Transmitter. The following list provides notes concerning all revisions of this document.

<b><u>Rev. Level</u></b>	<b><u>Date</u></b>	<b><u>Notes</u></b>
A	3/05	This is the initial release of the product manual. The manual has been reformatted to reflect the Emerson documentation style and updated to reflect any changes in the product offering. This manual contains information on HART Smart and FOUNDATION Fieldbus versions of Model Solu Comp Xmt-C.
B	9/05	Add Foundation fieldbus agency approvals and FISCO version.
C	2/06	Revised section 1.0, page 1, and the case specification on page 2. Added new drawings of FF and FI on section 4.0, pages 24-35.
D	12/10	Removed mention of patent, "Liquid Division" on address, and updated DNV logo.
E	1/11	Updated DWG no 9241581-00 pg 22 from rev A to rev B.

#### **Emerson Process Management**

2400 Barranca Parkway  
Irvine, CA 92606 USA  
Tel: (949) 757-8500  
Fax: (949) 474-7250  
<http://www.raihome.com>

# QUICK START GUIDE

## FOR MODEL SOLU COMP Xmt-C TRANSMITTER

1. Refer to page 11 for installation instructions.
2. Wire conductivity sensor to the transmitter. Refer to the sensor instruction sheet for details.
3. Once connections are secure and verified, apply DC power to the transmitter.
4. When the transmitter is powered up for the first time, **Quick Start** screens appear. Using **Quick Start** is easy.
  - a. A blinking field shows the position of the cursor.
  - b. Use the ◀ or ▶ key to move the cursor left or right. Use the ▲ or ▼ key to move the cursor up or down or to increase or decrease the value of a digit. Use the ▲ or ▼ key to move the decimal point.
  - c. Press ENTER to store a setting. Press EXIT to leave without storing changes. Pressing EXIT also returns the display to the previous screen.

English	Français
Español	>>

5. Choose the desired language. Choose >> to show more choices.

Measure?	Cond
Resistivity	>>

6. Choose measurement: **Conductivity, Resistivity, Total Dissolved Solids, or Custom.**

Measure?	
TDS	Custom >>

Cell Constant?
1.0000/cm

7. Enter the cell constant. See label attached to sensor.

ROSEMOUNT ANALYTICAL	ENDURANCE™ CONDUCTIVITY SENSOR		
	MODEL/PN: 404-11-16-50-54	REV B C99	PT1000
	MAX TEMP: 140°F AT 20 psig	CAL CONST= 450	K= .01
	MAX PRESS: 100 psig AT 77°F	CELL CONST= .00950/cm	

If there is no cell constant on the label, calculate it from the equation:

$$\text{cell const} = K \left( \frac{500 + \text{cal const}}{1000} \right)$$

Temperature in?	
°C	°F

8. Choose temperature units: °C or °F.

Enter Data Pts	
Ref Temp	Slope

9. If you selected **Custom**, you must enter the appropriate conductivity and concentration data points. From the main display, press MENU. Choose **Program** followed by **Measurement** and **Custom**. The screen shown at left appears. Select **Enter Data Pts**. Follow the prompts and enter the display units, the number of data points, and enter the concentration and conductivity data points. Enter the reference temperature and the temperature coefficient (slope). Once the analyzer has been configured, press EXIT. For a guide to the program menu, see the menu trees on pages 5 & 6.

Concentration	
Units?	% PPM none

10. To change output settings, to scale the 4-20 mA output, to change measurement-related settings from the default values, and to set security codes, press MENU. Select Program and follow the prompts. Refer to the menu tree on pages 5 & 6.

11. To return the transmitter to default settings, choose **ResetAnalyzer** in the Program menu.



# MODEL XMT-C TWO-WIRE CONDUCTIVITY TRANSMITTER

## TABLE OF CONTENTS

Section	Title	Page
<b>1.0</b>	<b>DESCRIPTION AND SPECIFICATIONS .....</b>	<b>1</b>
1.1	Features and Applications.....	1
1.2	Specifications .....	2
1.3	Hazardous Location Approval .....	4
1.4	Menu Tree for Model XMT-C-HT.....	5
1.5	Menu Tree for Model XMT-C-FF .....	6
1.6	HART Communications.....	7
1.7	FOUNDATION Fieldbus .....	7
1.8	Asset Management Solutions .....	8
1.9	Ordering Information .....	10
1.10	Accessories .....	10
<b>2.0</b>	<b>INSTALLATION .....</b>	<b>11</b>
2.1	Unpacking and Inspection.....	11
2.2	Installation .....	11
<b>3.0</b>	<b>WIRING.....</b>	<b>15</b>
3.1	Power Supply / Current Loop — Model XMT-C-HT .....	15
3.2	Power Supply Wiring for Model XMT-C-FF .....	16
3.2	Sensor Wiring .....	17
<b>4.0</b>	<b>INTRINSICALLY SAFE INSTALLATION.....</b>	<b>18</b>
<b>5.0</b>	<b>DISPLAY AND OPERATION .....</b>	<b>36</b>
5.1	Display .....	36
5.2	Keypad.....	36
5.3	Programming and Calibrating the Model Xmt — Tutorial.....	37
5.4	Menu Trees .....	38
5.5	Diagnostic Messages.....	38
5.6	Security .....	41
5.7	Using Hold .....	41
<b>6.0</b>	<b>OPERATION WITH MODEL 375.....</b>	<b>42</b>
6.1	Note on Model 375 HART and Foundation Fieldbus Communicator.....	42
6.2	Connecting the HART and Foundation Fieldbus Communicator .....	42
6.3	Operation .....	43
<b>7.0</b>	<b>CALIBRATION — TEMPERATURE.....</b>	<b>47</b>
7.1	Introduction .....	47
7.2	Calibrating Temperature.....	47
<b>8.0</b>	<b>CALIBRATION — CONDUCTIVITY .....</b>	<b>48</b>
8.1	Introduction .....	48
8.2	Entering the Cell Constant .....	49
8.3	Zeroing the Instrument.....	50
8.4	Calibrating the Sensor in a Conductivity Standard.....	51
8.5	Calibrating the Sensor to a Laboratory Instrument .....	52

## TABLE OF CONTENTS CONT'D

<b>9.0</b>	<b>PROGRAMMING THE TRANSMITTER.....</b>	<b>53</b>
9.1	General .....	53
9.2	Changing Start-up Settings .....	53
9.3	Configuring and Ranging the Output .....	54
9.4	Choosing and Configuring the Analytical Measurement .....	57
9.5	Choosing Temperature Units & Manual or Automatic Temperature Compensation ..	58
9.6	Setting a Security Code .....	59
9.7	Making HART-related Settings.....	60
9.8	Resetting Factory Calibration and Factory Default Settings .....	60
9.9	Selecting a Default Screen and Screen Contrast .....	61
<b>10.0</b>	<b>MAINTENANCE .....</b>	<b>62</b>
10.1	Overview .....	62
10.2	Replacement Parts .....	62
<b>11.0</b>	<b>THEORY OF OPERATION .....</b>	<b>63</b>
11.1	Conductivity / Resistivity / % Concentration.....	63
11.2	Temperature Correction .....	63
<b>12.0</b>	<b>THEORY — REMOTE COMMUNICATIONS.....</b>	<b>65</b>
12.1	Overview of HART Communications.....	65
12.2	HART Interface Devices.....	65
12.2	Asset Management Solutions .....	66
<b>13.0</b>	<b>RETURN OF MATERIAL.....</b>	<b>67</b>

## LIST OF FIGURES

Number	Title	Page
1-1	Menu Tree — XMT-C-HT .....	5
1-2	Menu Tree — XMT-C-FF .....	6
1-3	Configuring Model XMT Transmitter with FOUNDATION Fieldbus .....	7
1-4	HART and FOUNDATION Fieldbus Communicators .....	8
1-5	AMS Main Menu Tools .....	9
2-1	Removing the Knockouts .....	11
2-2	Power Supply / Current Loop Wiring .....	11
2-3	Panel Mount Installation .....	12
2-4	Pipe Mount Installation .....	13
2-5	Surface Mount Installation .....	14
3-1	Load/Power Supply Requirements .....	15
3-2	Power Supply / Current Loop Wiring .....	15
3-3	Typical Fieldbus Network Electrical Wiring Configuration .....	16
3-4	Loop Power and Sensor Wiring .....	16
4-1	FM Intrinsically Safe Label for Model XMT-C-HT .....	18
4-2	FM Intrinsically Safe Installation for Model XMT-C-HT .....	19
4-3	CSA Intrinsically Safe Label for Model XMT-C-HT .....	20
4-4	CSA Intrinsically Safe Installation for Model XMT-C-HT .....	21
4-5	ATEX Intrinsically Safe Label for Model XMT-C-HT .....	22
4-6	ATEX Intrinsically Safe Installation for Model XMT-C-HT .....	23
4-7	FM Intrinsically Safe Label for Model XMT-C-FF .....	24
4-8	FM Intrinsically Safe Installation for Model XMT-C-FF .....	25
4-9	CSA Intrinsically Safe Label for Model XMT-C-FF .....	26
4-10	CSA Intrinsically Safe Installation for Model XMT-C-FF .....	27
4-11	ATEX Intrinsically Safe Label for Model XMT-C-FF .....	28
4-12	ATEX Intrinsically Safe Installation for Model XMT-C-FF .....	29
4-13	FM Intrinsically Safe Label for Model XMT-C-FI .....	30
4-14	FM Intrinsically Safe Installation for Model XMT-C-FI .....	31
4-15	CSA Intrinsically Safe Label for Model XMT-C-FI .....	32
4-16	CSA Intrinsically Safe Installation for Model XMT-C-FI .....	33
4-17	ATEX Intrinsically Safe Label for Model XMT-C-FI .....	34
4-18	ATEX Intrinsically Safe Installation for Model XMT-C-FI .....	35
5-1	Displays During Normal Operation .....	36
5-2	Solu Comp Xmt Keypad .....	36
5-3	Menu Tree for Model XMT-C-HT .....	39
5-4	Menu Tree for Model XMT-C-FF .....	40
6-1	Connecting the Model 375 Communicator .....	42
6-2	XMT-C-HT HART / Model 375 Menu Tree .....	44
12-1	HART Communicators .....	65
12-2	AMS Main Menu Tools .....	66





## SECTION 1.0 DESCRIPTION AND SPECIFICATIONS

### Model Xmt Family of Two-wire Transmitters

- CHOICE OF COMMUNICATION PROTOCOLS: HART® or FOUNDATION® Fieldbus
- CLEAR, EASY-TO-READ two-line display shows commissioning menus and process measurement displays in English
- SIMPLE TO USE MENU STRUCTURE
- CHOICE OF PANEL OR PIPE/SURFACE MOUNTING
- NON-VOLATILE MEMORY retains program settings and calibration data during power failures
- SIX LOCAL LANGUAGES - English, French, German, Italian, Spanish and Portuguese



#### 1.1 FEATURES AND APPLICATIONS

The Solu Comp Model Xmt family of transmitters can be used to measure pH, ORP, conductivity (using either contacting or toroidal sensors), resistivity, oxygen (ppm and ppb level), free chlorine, total chlorine, monochloramine and ozone in a variety of process liquids. The Xmt is compatible with most Rosemount Analytical sensors. See the Specification sections for details.

The transmitter has a rugged, weatherproof, corrosion-resistant enclosure (NEMA 4X and IP65). The panel mount version fits standard ½ DIN panel cutouts, and its shallow depth is ideally suited for easy mounting in cabinet-type enclosures. A panel mount gasket is included to maintain the weather rating of the panel. Surface/pipe mount enclosure includes self-tapping screws for surface mounting. A pipe mounting accessory kit is available for mounting to a 2-inch pipe.

The transmitter has a two-line 16-character display. Menu screens for calibrating and registering choices are simple and intuitive. Plain language prompts guide the user through the procedures. There are no service codes to enter before gaining access to menus.

Two digital communication protocols are available: HART (model option -HT) and FOUNDATION fieldbus (model option -FF or -FI). Digital communications allow access to AMS (Asset Management Solutions). Use AMS to set up and configure the transmitter, read process variables, and troubleshoot problems from a personal computer or host anywhere in the plant.

The seven-button membrane-type keypad allows local programming and calibrating of the transmitter. The HART Model 375 communicator can also be used for programming and calibrating the transmitter.

The Model Xmt-C Transmitter with the appropriate sensor

measures dissolved oxygen (ppm and ppb level), free chlorine, total chlorine, monochloramine, and ozone in water and aqueous solutions. The transmitter is compatible with Rosemount Analytical 499A amperometric sensors for oxygen, chlorine, monochloramine, and ozone; and with Hx438, Bx438, and Gx448 steam-sterilizable oxygen sensors.

For free chlorine measurements, both automatic and manual pH correction are available. pH correction is necessary because amperometric free chlorine sensors respond only to hypochlorous acid, not free chlorine, which is the sum of hypochlorous acid and hypochlorite ion. To measure free chlorine, most competing instruments require an acidified sample. Acid lowers the pH and converts hypochlorite ion to hypochlorous acid. The Model Xmt-C eliminates the need for messy and expensive sample conditioning by measuring the sample pH and using it to correct the chlorine sensor signal. If the pH is relatively constant, a fixed pH correction can be used, and the pH measurement is not necessary. If the pH is greater than 7.0 and fluctuates more than about 0.2 units, continuous measurement of pH and automatic pH correction is necessary. See Specifications section for recommended pH sensors. Corrections are valid to pH 9.5.

The transmitter fully compensates oxygen, ozone, free chlorine, total chlorine, and monochloramine readings for changes in membrane permeability caused by temperature changes.

For pH measurements — pH is available with free chlorine only — the Xmt-C features automatic buffer recognition and stabilization check. Buffer pH and temperature data for commonly used buffers are stored in the transmitter. Glass impedance diagnostics warn the user of an aging or failed pH sensor.

## 1.2 SPECIFICATIONS

### 1.2.1 GENERAL SPECIFICATIONS

**Case:** ABS (panel mount), polycarbonate (pipe/wall mount); NEMA 4X/CSA 4 (IP65)

**Dimensions**

**Panel (code -10):** 6.10 x 6.10 x 3.72 in. (155 x 155 x 94.5 mm)

**Surface/Pipe (code -11):** 6.23 x 6.23 x 3.23 in. (158 x 158 x 82 mm); see page 15 for dimensions of pipe mounting bracket.

**Conduit openings:** Accepts PG13.5 or 1/2 in. conduit fittings

**Ambient Temperature:** 32 to 122°F (0 to 50°C). Some degradation of display above 50°C.

**Storage Temperature:** -4 to 158°F (-20 to 70°C)

**Relative Humidity:** 10 to 90% (non-condensing)


**Weight/Shipping Weight:** 2 lb/3 lb (1 kg/1.5 kg)

**Display:** Two line, 16-character display. Character height: 4.8 mm; first line shows process variable, second line shows process temperature and output current. Fault and warning messages, when triggered, alternate with temperature and output readings.

During calibration and programming, messages, prompts, and editable values appear on the two-line display.

**Temperature resolution:** 0.1°C ( $\leq 99.9^\circ\text{C}$ );  
1°C ( $\geq 100^\circ\text{C}$ )

**Hazardous Location Approval:** For details, see specifications for the measurement of interest.


**RFI/EMI:** EN-61326 

*Solu Comp is a trademark of Rosemount Analytical.*

*Xmt is a trademark of Rosemount Analytical.*

*HART is a registered trademark of the HART Communication Foundation.*

*FOUNDATION is a registered trademark of Fieldbus Foundation.*



### DIGITAL COMMUNICATIONS:

#### HART —

**Power & Load Requirements:** Supply voltage at the transmitter terminals should be at least 12 Vdc. Power supply voltage should cover the voltage drop on the cable plus the external load resistor required for HART communications (250  $\Omega$  minimum). Minimum power supply voltage is 12 Vdc. Maximum power supply voltage is 42.4 Vdc. The graph shows the supply voltage required to maintain 12 Vdc (upper line) and 30 Vdc (lower line) at the transmitter terminals when the current is 22 mA.

**Analog Output:** Two-wire, 4-20 mA output with superimposed HART digital signal. Fully scalable over the operating range of the sensor.

**Output accuracy:**  $\pm 0.05$  mA

#### FOUNDATION fieldbus —

**Power & Load Requirements:** A power supply voltage of 9-32 Vdc at 13 mA is required.

**Fieldbus Intrinsically Safe COnccept/FISCO-compliant versions of Model Xmt Foundation Fieldbus transmitters are available.**

**1.2.2 FUNCTIONAL SPECIFICATIONS**

**Automatic Temperature Compensation:**

3-wire Pt 100 or Pt 1000 RTD

Conductivity: 0 to 200°C (32 to 392°F)

Resistivity: 0 to 100°C (32 to 212°F)

Low Conductivity: 0 to 100°C (32 to 212°F)

**Diagnostics:** The internal diagnostics can detect:

- Calibration Error
- Temperature Slope Error
- High Temperature Warning
- Low Temperature Warning
- ROM Failure
- Zero Error
- CPU Failure
- Input Warning

Once one of the above is diagnosed, the Xmt-C will display a message describing the problem.

**Digital Communications:**

**HART:** PV, SV, and TV assignable to measurement (conductivity, resistivity, or concentration), temperature, and raw conductivity. Raw conductivity is measured conductivity before temperature correction.

**Fieldbus:** Three AI blocks assignable to measurement (conductivity, resistivity, or concentration), temperature, and raw conductivity. Raw conductivity is measured conductivity before temperature correction. Execution time 75 msec. One PID block; execution time 150 msec. Device type: 4084. Device revision: 1. Certified to ITK 4.5.

**1.2.3 TRANSMITTER SPECIFICATIONS @ 25°C**

**Measured Range:** 0-20,000 µS/cm

**Accuracy:** ± 0.7% of reading and ± 0.002 µS/cm

**Repeatability:** ± 0.25% of reading

**Temperature Accuracy:** ± 0.2°C between 0 and 50°C; ± 0.5°C above 50°C (excludes inaccuracies in sensor)

**Temperature Compensation:** Slope 0-5%/°C, ultra-pure water, cation conductivity, or raw (uncompensated) conductivity.

**Compatible RTD:** 100Ω or 1000Ω with automatic recognition

**Ambient Temperature Coefficient:** ± 0.05% of reading/°C

**Maximum Cable Length:** 200 ft (61 m)

**1.2.4 LOOP SPECIFICATIONS**

**Accuracy:** under controlled laboratory conditions at 25°C (77°F) with perfectly calibrated ENDURANCE sensor of appropriate cell constant:

Cell Constant	Range	Loop accuracy
0.01/cm	up to 50 µS/cm	±0.7% of reading ±0.002 µS/cm
0.1/cm	0.4 to 50 µS/cm 50 to 200 µS/cm	±0.7% of reading ±2% of reading
1.0/cm	4 to 5000 µS/cm 5000 to 20,000 µS/cm	±0.7% of reading ±2% of reading

**Calibration:** Calibrate against previously calibrated standard sensor and analyzer, or calibrate against solution of known conductivity.

**1.2.5 SENSOR SELECTION GUIDELINES**

Cell Constant	Suggested Conductivity Range
0.01/cm	up to 50 µS/cm
0.1/cm	0.4 to 500 µS/cm
1.0/cm	4 to 20,000 µS/cm

**Note:** The conductivity values shown in the above chart are for UNCOMPENSATED (or RAW) conductivity at 25°C. Maximum range values will vary due to temperature compensation selection, process temperature, and other process conditions.

**RECOMMENDED SENSORS:**

- Model 140 Retractable Conductivity
- Model 141 Insertion High Conductivity
- Model 142 Insertion Low Conductivity
- Model 150 Insertion/Submersion Conductivity
- Model 400/VP Screw-In Low Conductivity
- Model 401 Screw-In High Conductivity
- Model 402/VP Retractable Conductivity
- Model 403/VP Sanitary Conductivity
- Model 404 Low Flow Conductivity

### 1.3 HAZARDOUS LOCATION APPROVALS

#### Intrinsic Safety:



Class I, II, III, Div. 1  
Groups A-G  
T4 Tamb = 50°C



Class I, II, III, Div. 1  
Groups A-G  
T4 Tamb = 50°C

ATEX

CE 1180



II 1 G  
Baseefa04ATEX0214X  
EEx ia IIC T4  
Tamb = 0°C to 50°C

#### Non-Incendive:




Class I, Div. 2, Groups A-D  
Dust Ignition Proof  
Class II & III, Div. 1, Groups E-G  
NEMA 4/4X Enclosure



Class I, Div. 2, Groups A-D  
Dust Ignition Proof  
Class II & III, Div. 1, Groups E-G  
NEMA 4/4X Enclosure  
T4 Tamb = 50°C


1.4 MENU TREE FOR MODEL XMT-C-HT

FIGURE 1-1. MENU TREE FOR MODEL SOLU COMP XMT-C-HT TRANSMITTER



1.5 MENU TREE FOR MODEL XMT-C-FF

FIGURE 1-2. MENU TREE FOR MODEL SOLU COMP XMT-C-FF TRANSMITTER



### 1.6 HART COMMUNICATIONS

#### 1.6.1 OVERVIEW OF HART COMMUNICATION

HART (highway addressable remote transducer) is a digital communication system in which two frequencies are superimposed on the 4 to 20 mA output signal from the transmitter. A 1200 Hz sine wave represents the digit 1, and a 2400 Hz sine wave represents the digit 0. Because the average value of a sine wave is zero, the digital signal adds no dc component to the analog signal. HART permits digital communication while retaining the analog signal for process control.

The HART protocol, originally developed by Fisher-Rosemount, is now overseen by the independent HART Communication Foundation. The Foundation ensures that all HART devices can communicate with one another. For more information about HART communications, call the HART Communication Foundation at (512) 794-0369. The internet address is <http://www.hartcomm.org>.

#### 1.6.2 HART INTERFACE DEVICES

The Model 375 HART Communicator is a hand-held device that provides a common link to all HART SMART instruments and allows access to AMS (Asset Management Solutions). Use the HART communicator to set up and control the XMT-C-HT and to read measured variables. Press ON to display the on-line menu. All setup menus are available through this menu.

HART communicators allow the user to view measurement data (conductivity, TDS, resistivity, and temperature), program the transmitter, and download information from the transmitter for transfer to a computer for analysis. Downloaded information can also be sent to another HART transmitter. Either a hand-held communicator, such as the Rosemount Model 375, or a computer can be used. HART interface devices operate from any wiring termination point in the 4 - 20 mA loop. A minimum load of 250 ohms must be present between the transmitter and the power supply. See Figure 1-4.

If your communicator does not recognize the Model XMT-C transmitter, the device description library may need updating. Call the manufacturer of your HART communication device for updates.

### 1.7 FOUNDATION FIELDBUS

Figure 1-3 shows a Xmt-C-FF being used to measure conductivity. The figure also shows three ways in which Fieldbus communication can be used to read process variables and configure the transmitter.





FIGURE 1-3. CONFIGURING MODEL XMT-C TRANSMITTER WITH FOUNDATION FIELDBUS



**FIGURE 1-4. HART and FOUNDATION™ Fieldbus Communicators.**

Both the Rosemount Model 375 (or 275) and a computer can be used to communicate with a HART transmitter. The 250 ohm load (minimum) must be present between the transmitter and the power supply.

## 1.8 ASSET MANAGEMENT SOLUTIONS

Asset Management Solutions (AMS) is software that helps plant personnel better monitor the performance of analytical instruments, pressure and temperature transmitters, and control valves. Continuous monitoring means maintenance personnel can anticipate equipment failures and plan preventative measures before costly breakdown maintenance is required.


AMS uses remote monitoring. The operator, sitting at a computer, can view measurement data, change program settings, read diagnostic and warning messages, and retrieve historical data from any HART-compatible device, including the Model XMT-C transmitter. Although AMS allows access to the basic functions of any HART compatible device, Rosemount Analytical has developed additional software for that allows access to all features of the Model Xmt-C transmitter.

AMS can play a central role in plant quality assurance and quality control. Using AMS Audit Trail, plant operators can track calibration frequency and results as well as warnings and diagnostic messages. The information is available to Audit Trail whether calibrations were done using the infrared remote transmitter, the Model 375 HART communicator, or AMS software.

AMS operates in Windows 2000, NT, and XP operating systems. See Figure 1-5 for a sample screen. AMS communicates through a HART-compatible modem with any HART transmitters, including those from other manufacturers. AMS is also compatible with FOUNDATION Fieldbus, which allows future upgrades to Fieldbus instruments.

Rosemount Analytical AMS windows provide access to all transmitter measurement and configuration variables. The user can read raw data, final data, and program settings and can reconfigure the transmitter from anywhere in the plant.





## 1.9 ORDERING INFORMATION

The **Solu Comp Model Xmt Two-Wire Transmitter** is intended for conductivity and resistivity measurements using contacting conductivity sensors.

MODEL Xmt SMART TWO-WIRE MICROPROCESSOR TRANSMITTER	
CODE	REQUIRED SELECTION
P	pH/ORP
CODE	REQUIRED SELECTION
HT	Analog 4-20 mA output with superimposed HART digital signal
FF	Foundation fieldbus digital output
FI	Foundation fieldbus digital output with FISCO
CODE	REQUIRED SELECTION
10	Panel mounting enclosure
11	Pipe/Surface mounting enclosure (pipe mounting requires accessory kit PN 23820-00)
CODE	AGENCY APPROVALS
60	No approval
67	FM approved intrinsically safe and non-incendive (when used with appropriate sensor and safety barrier)
69	CSA approved intrinsically safe and non-incendive (when used with appropriate sensor and safety barrier)
73	ATEX approved intrinsically safe (when used with appropriate sensor and safety barrier)
<b>Xmt-P-HT-10-67      EXAMPLE</b>	

## 1.10 ACCESSORIES

**POWER SUPPLY:** Use the Model 515 Power Supply to provide dc loop power to the transmitter. The Model 515 provides two isolated sources at 24Vdc and 200 mA each. For more information refer to product data sheet 71-515.

**ALARM MODULE:** The Model 230A alarm Module receives the 4-20 mA signal from the XMT-C-HT transmitter and activates two alarm relays. High/high, low/low, and high/low are available. Hysteresis (deadband) is also adjustable. For more information, refer to product data sheet 71-230A.

**HART COMMUNICATOR:** The Model 375 HART communicator allows the user to view measurement values as well as to program and configure the transmitter. The Model 375 attaches to any wiring terminal across the output loop. A minimum 250  $\Omega$  load must be between the power supply and transmitter. Order the Model 375 communicator from Emerson Process Management. Call (800) 999-9307.

### ACCESSORIES

MODEL/PN	DESCRIPTION
515	DC loop power supply (see product data sheet 71-515)
230A	Alarm module (see product data sheet 71-230A)
23820-00	2-in. pipe mounting kit
9240048-00	Stainless steel tag, specify marking
23554-00	Gland fittings PG 13.5, 5 per package

## SECTION 2.0 INSTALLATION

### 2.1 Unpacking and Inspection

### 2.2 Installation

#### 2.1 UNPACKING AND INSPECTION

Inspect the shipping container. If it is damaged, contact the shipper immediately for instructions. Save the box. If there is no apparent damage, remove the transmitter. Be sure all items shown on the packing list are present. If items are missing, immediately notify Rosemount Analytical.

Save the shipping container and packaging. They can be reused if it is later necessary to return the transmitter to the factory.

#### 2.2 INSTALLATION

1. Although the transmitter is suitable for outdoor use, do not install it in direct sunlight or in areas of extreme temperatures.
2. Install the transmitter in an area where vibrations and electromagnetic and radio frequency interference are minimized or absent.
3. Keep the transmitter and sensor wiring at least one foot from high voltage conductors. Be sure there is easy access to the transmitter.
4. The transmitter is suitable for panel (Figure 2-3), pipe (Figure 2-4), or surface (Figure 2-5) mounting.
5. The transmitter case has two 1/2-inch (PG13.5) conduit openings and either three or four 1/2-inch knockouts. The panel mount XMT-C-HT has four knockouts. The pipe/surface mount transmitter has three knockouts\*. One conduit opening is for the power/output cable; the other opening is for the sensor cable.

Figure 1 shows how to remove a knockout. The knockout grooves are on the outside of the case. Place the screwdriver blade on the inside of the case and align it approximately along the groove. Rap the screwdriver sharply with a hammer until the groove cracks. Move the screwdriver to an uncracked portion of the groove and continue the process until the knockout falls out. Use a small knife to remove the flash from the inside of the hole.

6. Use weathertight cable glands to keep moisture out to the transmitter. If conduit is used, plug and seal the connections at the transmitter housing to prevent moisture from getting inside the instrument.
7. To reduce the likelihood of stress on wiring connections, do not remove the hinged front panel (-11 models) from the base during wiring installation. Allow sufficient wire leads to avoid stress on conductors.

\*NEMA plug may be supplied instead of knockout for pipe/surface version.




FIGURE 2-1. Removing the Knockouts






FIGURE 2-2. Power Supply/Current Loop Wiring


Panel Mounting.



Pipe Mounting.



Surface Mounting.



**FIGURE 2-5. Surface Mount Installation**

*The front panel is hinged at the bottom. The panel swings down for access to the wiring terminals.*

## SECTION 3.0 WIRING

### 3.1 POWER SUPPLY/CURRENT LOOP — MODEL XMT-C-HT

#### 3.1.1 Power Supply and Load Requirements.

Refer to Figure 3-1.

The supply voltage must be at least 12.0 Vdc at the transmitter terminals. The power supply must be able to cover the voltage drop on the cable as well as the load resistor (250 Ω minimum) required for HART communications. The maximum power supply voltage is 42.0 Vdc. For intrinsically safe installations, the maximum power supply voltage is 30.0 Vdc. The graph shows load and power supply requirements. The upper line is the power supply voltage needed to provide 12 Vdc at the transmitter terminals for a 22 mA current. The lower line is the power supply voltage needed to provide 30 Vdc for a 22 mA current.




FIGURE 3-1. Load/Power Supply Requirements

The power supply must provide a surge current during the first 80 milliseconds of startup. The maximum current is about 24 mA.

For digital communications, the load must be at least 250 ohms. To supply the 12.0 Vdc lift off voltage at the transmitter, the power supply voltage must be at least 17.5 Vdc.

#### 3.1.2 Power Supply-Current Loop Wiring.

For general purpose areas, wire power as shown in Figure 3-2. For hazardous areas, please see hazardous area installation drawings.

Run the power/signal wiring through the opening nearest TB-2.

For optimum EMI/RFI protection . . .

1. Use shielded power/signal cable and ground the shield at the power supply.
2. Use a metal cable gland and be sure the shield makes good electrical contact with the gland.
3. Use the metal backing plate when attaching the gland to transmitter enclosure.

The power/signal cable can also be enclosed in an earth-grounded metal conduit.

Do not run power supply/signal wiring in the same conduit or cable tray with AC power lines or with relay actuated signal cables. Keep power supply/signal wiring at least 6 ft (2 m) away from heavy electrical equipment.




FIGURE 3-2. Power Supply/Current Loop Wiring

### 3.2 POWER SUPPLY WIRING FOR MODEL XMT-C-FF


**3.2.1 Power Supply Wiring.** For general purpose areas, wire power as shown in Figure 3-4. For hazardous areas, please see hazardous area installation drawings. Refer to Figure 3-3 and Figure 3-4.

Run the power/signal wiring through the opening nearest TB2. Use shielded cable and ground the shield at the power supply. To ground the transmitter, attach the shield to TB2-3.


**NOTE**

For optimum EMI/RFI immunity, the power supply/output cable should be shielded and enclosed in an earth-grounded metal conduit.

Do not run power supply/signal wiring in the same conduit or cable tray with AC power lines or with relay actuated signal cables. Keep power supply/signal wiring at least 6 ft (2 m) away from heavy electrical equipment.



**FIGURE 3-3. Typical Fieldbus Network Electrical Wiring Configuration**



**FIGURE 3-4. Loop Power and Sensor Wiring**



### 3.3 SENSOR WIRING

Keep sensor wiring separate from power wiring. For best EMI/RFI protection, use shielded output signal cable in an earth-grounded metal conduit. See Figure 3-4. Refer to the Instruction Sheet provided with each sensor for specific wiring instructions.

#### 3.1.1 WIRING THROUGH A JUNCTION BOX

The sensor can be wired to the analyzer through a remote junction box (PN 23550-00). Wire the extension cable and sensor cable point-to-point. See Figure 3-4. Refer to the sensor instruction manual for more details.

Factory-terminated (PN 23294-05) and unterminated (PN 9200276) connecting cable are available. The use of factory-terminated cable is strongly recommended. To prepare unterminated cable for use, follow the instructions in the sensor instruction manual.

For maximum EMI/RFI protection, the outer braid of the sensor cable should be connected to the outer braided shield of the extension cable. At the instrument, connect the outer braid of the extension cable to earth ground.

#### 3.1.2 SENSOR SELECTION

All Rosemount Analytical contacting conductivity sensors with PT100 RTD or PT1000 RTD are compatible with the Model Xmt-C transmitter. Refer to the Instruction Sheet provided with each sensor for specific wiring instructions.

Choose a contacting conductivity sensor that is appropriate for your process conditions and range of conductivity measurement.

**TABLE 3-1. Model Xmt-C Sensor Selection**

Cell Constant	Range
0.01/cm	0 to 50 $\mu\text{S/cm}$
0.1/cm	1 to 500 $\mu\text{S/cm}$
1.0/cm	10 to 20,000 $\mu\text{S/cm}$

# SECTION 4.0 INTRINSICALLY SAFE INSTALLATION

## INTRINSICALLY SAFE INSTALLATIONS FOR MODEL XMT-C-HT

For FM Intrinsically Safe Label, see Figure 4-1.  
For FM Intrinsically Safe Installation, see Figure 4-2.  
For CSA Intrinsically Safe Label, see Figure 4-3.

For CSA Intrinsically Safe Installation, see Figure 4-4.  
For ATEX Intrinsically Safe Label, see Figure 4-5.  
For ATEX Intrinsically Safe Installation, see Figure 4-6.




FIGURE 4-1. FM Intrinsically Safe Label for Model Xmt-C-HT




FIGURE 4-2. FM Intrinsically Safe Installation for Model Xmt-C-HT









FIGURE 4-5. ATEX Intrinsically Safe Label for Model Xmt-C-HT





1. ARTWORK IS SHEET 2 OF 2.

NOTES: UNLESS OTHERWISE SPECIFIED

4. NO CHANGE WITHOUT FM APPROVAL.

3. ALL ALPHA AND NUMERIC CHARACTERS ON LABEL TO BE BLACK HELVETICA MEDIUM. BACKGROUND TO BE WHITE.

2. MATERIAL: 3M SCOTCHCAL #3650-10 (WHITE VINYL FACESTOCK) OR POLYESTER, (.002 REFERENCE THICKNESS CLEAR MATTE MYLAR OVERLAMINATE .002-.005 FINISH THICKNESS. PRESSURE SENSITIVE ADHESIVE, FAR SIDE AND SPLIT LINER) OR (INTERMEC PN L7211210, 2 MIL GLOSS WHITE POLYESTER WITH PRESSURE SENSITIVE ACRYLIC ADHESIVE. NOMENCLATURE TO BE PRINTED USING INTERMEC SUPER PREMIUM BLACK THERMAL TRANSFER RIBBON) SEE BLANK LABEL PN 9241406-01.

FIGURE 4-7. FM Intrinsically Safe Label for Model Xmt-C-FF








FIGURE 4-9. CSA Intrinsically Safe Label for Model Xmt-C-FF





FIGURE 4-10. CSA Intrinsically Safe Installation for Model Xmt-C-FF





**WARNING-** SUBSTITUTION OF COMPONENTS MAY IMPAIR INTRINSIC SAFETY OR SUITABILITY FOR DIVISION 2.

**WARNING-** TO PREVENT IGNITION OF FLAMMABLE OR COMBUSTIBLE ATMOSPHERE, DISCONNECT POWER BEFORE SERVICING.

**TABLE I**

GAS GROUPS	Ca (uF)	La (uH)
IIC	13	800
IIB	240	300
IIA	240	300

**TABLE II**

OUTPUT PARAMETERS	MODEL XMT-C-FF TB1-1 THRU T2
Io	7.16V
Ib	221mA
Po	280mW
CI	8.8 1/2 F
LI	0mH

**TABLE III**

XMT-C-FF ENTITY PARAMETERS

MODEL NO.	Vmax (Vdc)	Imax (mA)	Pmax (W)	CI (nF)	LI (mH)
XMT-C-FF	30	300	1.3	0.4	0

ENTITY PARAMETERS: REMOTE TRANSMITTER INTERFACE

MODEL NO.	Vmax IN: Vdc	Imax IN: ma	Pmax IN: W	CI (uF)	LI (mH)	Voc max OUT: Vdc	Isc max OUT: uA
375	30	200	1.0	0.0	0.0	1.9	32

Baseefa Certified Product  
 No modifications permitted  
 without the approval of  
 the Authorized Person  
 Related Drawing

RELEASE DATE	6-30-05	ECONO.	9065	REV.	A
THIS DWG CONVERTED TO	SOLID EDGE				
DATE	10/8/04	BY	J. FLOCK	REV.	1
DATE	08/04	BY	J. FLOCK	REV.	2
DATE	08/04	BY	J. FLOCK	REV.	3
DATE	08/04	BY	J. FLOCK	REV.	4
DATE	08/04	BY	J. FLOCK	REV.	5
DATE	08/04	BY	J. FLOCK	REV.	6
DATE	08/04	BY	J. FLOCK	REV.	7
DATE	08/04	BY	J. FLOCK	REV.	8
DATE	08/04	BY	J. FLOCK	REV.	9
DATE	08/04	BY	J. FLOCK	REV.	10
DATE	08/04	BY	J. FLOCK	REV.	11
DATE	08/04	BY	J. FLOCK	REV.	12
DATE	08/04	BY	J. FLOCK	REV.	13
DATE	08/04	BY	J. FLOCK	REV.	14
DATE	08/04	BY	J. FLOCK	REV.	15
DATE	08/04	BY	J. FLOCK	REV.	16
DATE	08/04	BY	J. FLOCK	REV.	17
DATE	08/04	BY	J. FLOCK	REV.	18
DATE	08/04	BY	J. FLOCK	REV.	19
DATE	08/04	BY	J. FLOCK	REV.	20
DATE	08/04	BY	J. FLOCK	REV.	21
DATE	08/04	BY	J. FLOCK	REV.	22
DATE	08/04	BY	J. FLOCK	REV.	23
DATE	08/04	BY	J. FLOCK	REV.	24
DATE	08/04	BY	J. FLOCK	REV.	25
DATE	08/04	BY	J. FLOCK	REV.	26
DATE	08/04	BY	J. FLOCK	REV.	27
DATE	08/04	BY	J. FLOCK	REV.	28
DATE	08/04	BY	J. FLOCK	REV.	29
DATE	08/04	BY	J. FLOCK	REV.	30
DATE	08/04	BY	J. FLOCK	REV.	31
DATE	08/04	BY	J. FLOCK	REV.	32
DATE	08/04	BY	J. FLOCK	REV.	33
DATE	08/04	BY	J. FLOCK	REV.	34
DATE	08/04	BY	J. FLOCK	REV.	35
DATE	08/04	BY	J. FLOCK	REV.	36
DATE	08/04	BY	J. FLOCK	REV.	37
DATE	08/04	BY	J. FLOCK	REV.	38
DATE	08/04	BY	J. FLOCK	REV.	39
DATE	08/04	BY	J. FLOCK	REV.	40
DATE	08/04	BY	J. FLOCK	REV.	41
DATE	08/04	BY	J. FLOCK	REV.	42
DATE	08/04	BY	J. FLOCK	REV.	43
DATE	08/04	BY	J. FLOCK	REV.	44
DATE	08/04	BY	J. FLOCK	REV.	45
DATE	08/04	BY	J. FLOCK	REV.	46
DATE	08/04	BY	J. FLOCK	REV.	47
DATE	08/04	BY	J. FLOCK	REV.	48
DATE	08/04	BY	J. FLOCK	REV.	49
DATE	08/04	BY	J. FLOCK	REV.	50
DATE	08/04	BY	J. FLOCK	REV.	51
DATE	08/04	BY	J. FLOCK	REV.	52
DATE	08/04	BY	J. FLOCK	REV.	53
DATE	08/04	BY	J. FLOCK	REV.	54
DATE	08/04	BY	J. FLOCK	REV.	55
DATE	08/04	BY	J. FLOCK	REV.	56
DATE	08/04	BY	J. FLOCK	REV.	57
DATE	08/04	BY	J. FLOCK	REV.	58
DATE	08/04	BY	J. FLOCK	REV.	59
DATE	08/04	BY	J. FLOCK	REV.	60
DATE	08/04	BY	J. FLOCK	REV.	61
DATE	08/04	BY	J. FLOCK	REV.	62
DATE	08/04	BY	J. FLOCK	REV.	63
DATE	08/04	BY	J. FLOCK	REV.	64
DATE	08/04	BY	J. FLOCK	REV.	65
DATE	08/04	BY	J. FLOCK	REV.	66
DATE	08/04	BY	J. FLOCK	REV.	67
DATE	08/04	BY	J. FLOCK	REV.	68
DATE	08/04	BY	J. FLOCK	REV.	69
DATE	08/04	BY	J. FLOCK	REV.	70
DATE	08/04	BY	J. FLOCK	REV.	71
DATE	08/04	BY	J. FLOCK	REV.	72
DATE	08/04	BY	J. FLOCK	REV.	73
DATE	08/04	BY	J. FLOCK	REV.	74
DATE	08/04	BY	J. FLOCK	REV.	75
DATE	08/04	BY	J. FLOCK	REV.	76
DATE	08/04	BY	J. FLOCK	REV.	77
DATE	08/04	BY	J. FLOCK	REV.	78
DATE	08/04	BY	J. FLOCK	REV.	79
DATE	08/04	BY	J. FLOCK	REV.	80
DATE	08/04	BY	J. FLOCK	REV.	81
DATE	08/04	BY	J. FLOCK	REV.	82
DATE	08/04	BY	J. FLOCK	REV.	83
DATE	08/04	BY	J. FLOCK	REV.	84
DATE	08/04	BY	J. FLOCK	REV.	85
DATE	08/04	BY	J. FLOCK	REV.	86
DATE	08/04	BY	J. FLOCK	REV.	87
DATE	08/04	BY	J. FLOCK	REV.	88
DATE	08/04	BY	J. FLOCK	REV.	89
DATE	08/04	BY	J. FLOCK	REV.	90
DATE	08/04	BY	J. FLOCK	REV.	91
DATE	08/04	BY	J. FLOCK	REV.	92
DATE	08/04	BY	J. FLOCK	REV.	93
DATE	08/04	BY	J. FLOCK	REV.	94
DATE	08/04	BY	J. FLOCK	REV.	95
DATE	08/04	BY	J. FLOCK	REV.	96
DATE	08/04	BY	J. FLOCK	REV.	97
DATE	08/04	BY	J. FLOCK	REV.	98
DATE	08/04	BY	J. FLOCK	REV.	99
DATE	08/04	BY	J. FLOCK	REV.	100

**FIGURE 4-12. ATEX Intrinsically Safe Installation for Model Xmt-C-FF**









FIGURE 4-15. CSA Intrinsically Safe Label for Model Xmt-C-FI







4. NO CHANGE WITHOUT Baseefa APPROVAL.

3. ALL ALPHA AND NUMERIC CHARACTERS ON LABEL TO BE BLACK HELVETICA MEDIUM. BACKGROUND TO BE WHITE.

2 MATERIAL: 3M SCOTCHCAL #3650-10 (WHITE VINYL FACESTOCK) OR POLYESTER, (.002 REFERENCE THICKNESS CLEAR MATTE MYLAR OVERLAMINATE, .002-.005 FINISH THICKNESS, PRESSURE SENSITIVE ADHESIVE, FARSIDE AND SPLIT LINER).

1. ARTWORK IS SHEET 2 OF 2.

NOTES: UNLESS OTHERWISE SPECIFIED

FIGURE 4-17. ATEX Intrinsically Safe Label for Model Xmt-C-FI




## SECTION 5.0 DISPLAY AND OPERATION

### 5.1. DISPLAY

The Model Xmt-C has a two-line display. Generally, the user can program the transmitter to show one of four displays. If the transmitter has been configured to measure total dissolved solids, resistivity, or percent conductivity, similar displays are available. Figure 5-1 shows the displays available for conductivity.

The transmitter has information screens that supplement the data in the main display. Press ▼ to view the information screens. The first information screen shows the type of measurement being made (conductivity, resistivity, TDS, custom). **The last information screen is the software version number.**

During calibration and programming, key presses cause different displays to appear. The displays are self-explanatory and guide the user step-by-step through the procedure.




**FIGURE 5-1. Displays During Normal Operation**

Screen A shows the conductivity reading, the temperature, and the output current generated by the transmitter. Screen B shows the same information as Screen A except conductivity is shown as percent of scale. Screen C shows conductivity as parts per million total dissolved solids. Screen D displays conductivity as resistivity in mega-ohms per cm (MΩ-cm).

### 5.2 KEYPAD

Figure 5-2 shows the Solu Comp Xmt keypad.



**FIGURE 5-2. Solu Comp Xmt Keypad**

Four arrow keys move the cursor around the screen. A blinking word or numeral show the position of the cursor. The arrow keys are also used to change the value of a numeral. Pressing ENTER stores numbers and settings and moves the display to the next screen. Pressing EXIT returns to the previous screen without storing changes. Pressing MENU always causes the main menu screen to appear. Pressing MENU followed by EXIT causes the main display to appear.

### 5.3 PROGRAMMING AND CALIBRATING THE MODEL XMT - TUTORIAL

Setting up and calibrating the Model Xmt is easy. The following tutorial describes how to move around in the programming menus. For practice, the tutorial also describes how to assign values to the 4 and 20 mA output.

<b>Calibrate</b>	Hold
Program	Display

Calibrate	Hold
<b>Program</b>	Display

<b>Output</b>	Temp
Measurement	>>

<b>Security</b>	HART
ResetAnalyzer	>>

Output?	<b>Test</b>
Configure	Range

Output Range?	
4mA	0.000µS/cm

Output Range?	
20mA	2000µS/cm

Output?	Test
Configure	<b>Range</b>

1. If the menu screen (shown at the left) is not already showing, press MENU. **Calibrate** is blinking, which means the cursor is on **Calibrate**.
2. To assign values to the current output, the **Program** sub-menu must be open. Press ▼. The cursor moves to **Program** (**Program** blinking.) Press ENTER. Pressing ENTER opens the **Program** sub-menu.
3. The **Program** sub-menu permits the user to configure and assign values to the 4-20 mA output, to test and trim the output, to change the type of measurement from what was selected during Quick Start, to set manual or automatic temperature correction for membrane permeability, and to set security codes. When the sub-menu opens, **Output** is blinking, which means the cursor is on **Output**. Press ▼ or ► (or any arrow key) to move the cursor around the display. Move the cursor to >> and press ENTER to cause a second screen with more program items to appear. There are three screens in the **Program** sub-menu. Pressing >> and ENTER in the third screen cause the display to return to the first screen (**Output, Temp, Measurement**).
4. For practice, assign values to the 4 and 20 mA output. Move the cursor to **Output** and press ENTER.
5. The screen shown at left appears. **Test** is blinking. Move the cursor to **Range** and press ENTER.
6. The screen shown at left appears. + is blinking, which means the cursor is on +.
  - a. To toggle between + and - press ▲ or ▼.
  - b. To move from one digit to the next, press ◀ or ▶.
  - c. To increase or decrease the value of a digit, press ▲ or ▼.
  - d. To move the decimal point, press ◀ or ▶ until the cursor is on the decimal point. Press ▲ to move the decimal to the right. Press ▼ to move the decimal point to the left.
  - e. Press ENTER to store the number.
7. The screen shown at left appears. Use this screen to assign a full scale value to the 20 mA output. Use the arrow keys to change the number to the desired value. Press ENTER to store the setting.
8. The screen shown at left appears. To configure the output or to test the output, move the cursor to the appropriate place and press ENTER.
9. To return to the main menu, press MENU. To return to the main display, press MENU then EXIT, or press EXIT repeatedly until the main display appears. To return to the previous display, press EXIT.

#### NOTE

To store values or settings, press ENTER before pressing EXIT.

## 5.4 MENU TREES

The Model Xmt-C-HT transmitter has four menus: CALIBRATE, PROGRAM, HOLD, and DISPLAY. Under the Calibrate and Program menus are several sub-menus. For example, under CALIBRATE, the sub-menus are **Conductivity** and **Temperature**. Under each sub-menu are prompts. Under PROGRAM, the sub-menus for Xmt-C-HT are **Output**, **Temp**, **Measurement**, **Security**, **HART**, and **Reset Analyzer**. The HOLD menu (HART only) enables or disables the 4-20 mA outputs. The DISPLAY menu allows the user to configure the main display information fields and to adjust the LCD display contrast. Figure 5-3 shows the complete menu tree for Model Xmt-C-HT. Figure 5-4 shows the complete menu tree for Model Xmt-C-FF.

## 5.5 DIAGNOSTIC MESSAGES

Whenever a warning or fault limit has been exceeded, the transmitter displays diagnostic messages to aid in troubleshooting. "Fault" or "Warn" appears in the main display to alert the user of an adverse condition. The display alternates between the regular display and the Fault or Warning message. If more than one warning or fault message has been generated, the messages appear alternately.

FIGURE 5-3. MENU TREE FOR MODEL SOLU COMP XMT-C-HT TRANSMITTER





FIGURE 5-4. MENU TREE FOR MODEL SOLU COMP XMT-C-FF TRANSMITTER







## SECTION 6.0 OPERATION WITH MODEL 375

### 6.1 Note on Model 375 HART and Foundation Fieldbus Communicator

The Model 375 HART Communicator is a product of Emerson Process Management, Rosemount Inc. This section contains selected information on using the Model 375 with the Rosemount Analytical Model XMT-C-HT Transmitter and Model XMT-C-FF Transmitter. For complete information on the Model 375 Communicator, see the Model 375 instruction manual. For technical support on the Model 375 Communicator, call Rosemount Inc. at (800) 999-9307 within the United States. Support is available worldwide on the internet at <http://rosemount.com>.

### 6.2 Connecting the HART and Foundation Fieldbus Communicator

Figure 6-1 shows how the Model 275 or 375 Communicator connects to the output lines from the Model XMT-C Transmitter.





FIGURE 6-1. Connecting the Model 375 Communicator

## 6.3 Operation

### 6.3.1 Off-line and On-line Operation

The Model 375 Communicator features off-line and on-line communications. On-line means the communicator is connected to the transmitter in the usual fashion. While the communicator is on line, the operator can view measurement data, change program settings, and read diagnostic messages. Off-line means the communicator is not connected to the transmitter. When the communicator is off line, the operator can still program settings into the communicator. Later, after the communicator has been connected to a transmitter, the operator can transfer the programmed settings to the transmitter. Off-line operation permits settings common to several transmitters to be easily stored in all of them.

### 6.3.2 Making HART related settings from the keypad

Calibrate	Hold
<b>Program</b>	Display

1. Press MENU. The main menu screen appears. Choose **Program**.

Output	Temp
Measurement	>>

2. Choose >>.

Security	<b>HART</b>
	>>


3. Choose **HART**.

<b>DevID</b>	PollAddr
Burst	Preamble

4. To display the device ID, choose **DevID**. To change the polling address, choose **PollAddr**. To make burst mode settings, choose **Burst**. To change the preamble count, choose **Preamble**.

### 6.3.3 Menu Tree

The menu tree for the Model 275 and Model 375 HART communicators are on the following pages.



## Detailed setup

FIGURE 6-2. XMT-C-HT HART/Model 375 Menu Tree (2 of 3)

## Sensors

## Main sensor

PV Type [Conductivity, Resistivity, 0-12% NaOH, 0-15% HCl, 0-25% H2SO4, 96-99.7% H2SO4, 0-20% NaCl, TDS, Custom]

PV Snsr unit [uS/cm, uS/m, mS/cm, mS/m, Mohm-cm, %, ppm, \_]

Cond unit [uS/cm, uS/m, mS/cm, mS/m] \*\*\*\*

Define curve \*\*\*\*

View custom points \*\*\*\*

Cell constant

Temp comp type [Linear, Neutral salt, Cation, None/Off] \*..

Temp slope \*\*\*

Ref temp \*\*\* AND \* , \*\*\*\*

PV sensor type

Sensor information

LSL

USL

Min span

## Temperature

ATC [On, Off]

Man temp

Temp unit [°C, °F]

Temp snsr [RTD PT100, RTD PT1000, Manual]

## Diag override

PV>display limit [ON, OFF]

EE write Error [ON, OFF]

EE chksum Error [ON, OFF]

EE buf overflow [ON, OFF]

% out of range [ON, OFF]

Need zero cal [ON, OFF]

Temperature high [ON, OFF]

Temperature low [ON, OFF]

Sense line open [ON, OFF]

Need factory cal [ON, OFF]

Input Overrange [ON, OFF]

A2D Read Error [ON, OFF]

RTDohm overrange [ON, OFF]

RTD open [ON, OFF]

## Signal condition

LRV

URV

AO Damp

% rng

Xfer fnctn

AO1 lo end point

AO1 hi end pt

## Output condition

Analog output

AO1

AO Alm typ

Fixed

Fault mode [Fixed, Live]

Fault

Loop test

D/A trim

FIGURE 6-2. XMT-C-HT HART/Model 375 Menu Tree (3 of 3)

HART output	
PV is Process	
SV is Temperature	
TV is Raw process	
Poll addr	
Burst option [PV, %range/current, Process vars/crnt]	
Burst mode [Off, On]	
Num req preams	
Num resp preams	
Device information	
Distributor	
Model	
Dev id	
Tag	
Date	
Physicl signl code	
Write protect	
Snsr text	
Descriptor	
Message	
Revision #'s	
Universal rev	
Fld dev rev	
Software rev	
Hardware rev	
Local Display	
AO LOI units [mA, %]	
LOI cfg code	
LOI cal code	
Load Default Conf.	
Review	
PV	
PV AO	
PV LRV	
PV URV	
-----	
Notes:	
* Can be Cond, Res, NaOH, HCl, H2SO4, NaCl, TDS or Custom	
** Valid only when PV Type = NaOH, HCl, 0-25% H2SO4, or NaCl	
*** Valid only when Temp comp type = Linear	
**** Valid only when PV Type = Custom	
*. Valid only when PV Sensor Type = Toroidal	
*.. Valid only when PV Type = Conductivity or Resistivity. For contacting sensors, the item list is [Linear, Neutral salt, Cation, None/off]. For toroidal sensors, the item list is [Linear, None/off].	
*... Valid only when PV Sensor Type = Contacting	
/* Xmt CT done */	

# SECTION 7.0

## CALIBRATION — TEMPERATURE

### 7.1 INTRODUCTION

The Calibrate Menu allows the user to calibrate the conductivity readings and temperature response of the sensor.

### 7.2 CALIBRATING TEMPERATURE

#### 7.2.1 Purpose

This procedure is used to ensure an accurate temperature measurement by the temperature sensor. It enables the instrument to display process temperature accurately as well as to compensate for the effect of temperature on the conductivity reading when the temperature in your process changes. The following steps should be performed with the sensor in the process or in a grab sample near the operating temperature of the process.

#### 7.2.2 Procedure

1. Check the instrument temperature reading (main display) to make sure the sensor has acclimated to the process temperature. Compare the instrument temperature to a calibrated temperature reading device. Proceed to the next step if the reading requires adjustment.

<b>Calibrate</b>	Hold
Program	Display

2. Press MENU. The menu screen appears. Choose **Calibrate**.

Cal?	Conductivity
	<b>Temperature</b>

3. Choose **Temperature**.

Live	25.0°C
Cal	+025.0°C

4. To calibrate the temperature, change the number in the second line to match the temperature measured **in the process**. Press ENTER.
5. Press MENU then EXIT to return to the main display.

## SECTION 8.0 CALIBRATION — CONDUCTIVITY

The following procedures are described in this section:

- Entering the cell constant (Section 8.1)
- Zeroing the transmitter (Section 8.2)
- Calibrating the sensor in a Conductivity Standard (Section 8.3)
- Calibrating the sensor to a Conductivity Meter (Section 8.4)
- Entering the temperature slope (Section 8.5)

### INTRODUCTION

Calibration is the process of adjusting or standardizing the transmitter to a lab test (such as free acid titration) or a calibrated laboratory instrument, or standardizing to some known reference (such as a commercial chemical standard). Calibration ensures that the transmitter reads an accurate, and therefore, repeatable reading of conductivity and temperature. This section contains procedures for the first time use and for routine calibration of the Model Xmt-C transmitter.

Since conductivity measurements are affected by temperature, the Model Xmt-C reads the temperature at the probe and compensates for the changing temperature by referencing all conductivity measurements to 25°C (77°F).

To ensure the transmitter's accuracy, it is important to perform all the calibration procedures provided in this section if you are:

- installing this unit for the first time
- changing or replacing a probe
- troubleshooting

After the initial calibration, the accuracy of the conductivity reading should be checked periodically against some known standard of conductivity and temperature.

Entering the cell constant, zeroing the instrument, and calibrating the sensor should be done when first commissioning the Xmt and when changing the conductivity probe. These operations should be conducted with the conductivity probe wired to the transmitter with full length of extension cable (if any) for best results.



## 8.1 ENTERING THE CELL CONSTANT

The cell constant should be entered:

- When the unit is installed for the first time
- When the probe is replaced
- During troubleshooting

All cell constants can be located on the cable label of the conductivity probe.

```
Cal?      Conductivity
          Temperature
```

```
Cal?      InProcess
Meter     Zero     >>
```

```
Enter Cell Const
Temp Slope >>
```

```
Cell Constant?
          1.0000/cm
```

1. Press MENU. The main menu appears. Choose **Calibrate**. Press ENTER.
2. Choose **Conductivity**. The screen at the left will appear. Press ENTER.
3. Scroll right with the right arrow key. The screen at the left will appear.
4. Choose **Enter Cell Constant**. The screen at the left will appear.
5. Enter the actual cell constant as printed on the sensor tag attached to the sensor cable.

### NOTE

**For sensors that show a "cal constant" on the label, the actual cell constant can be calculated by adding 500 to the cal constant, multiply this value by the nominal cell constant, then divide the result by 1000.**

If there is no cell constant on the label, calculate it from the equation:

$$\text{cell const} = K \left( \frac{500 + \text{cal const}}{1000} \right)$$

6. Press ENTER. All conductivity readings will reference this sensor-specific cell constant.
7. Press EXIT once.

### NOTE

The cell constant you have entered is changed after the Standardizing Conductivity procedure is performed. **For inductive sensors and contacting sensors that only show nominal cell constants, do not change it back to the value as shown on the probe label.**

## 8.2 ZEROING THE INSTRUMENT

This procedure is used to compensate for small offsets to the conductivity signal that are present even when there is no conductivity to be measured. This procedure is affected by the length of extension cable and should always be repeated if any changes in extension cable or sensor have been made. **Electrically connect the conductivity probe as it will actually be used and place the measuring portion of the probe in air.**

Complete the following procedure. **Verify that the sensor is actually in air.** If the displayed value is not very close to zero, then press ENTER. The transmitter will establish a new zero.

Calibrate	Hold
Program	Display

1. If continuing from the previous procedure, the screen on the left will appear.

Cal?	<b>Conductivity</b>
	Temperature

2. Choose **Conductivity**. Press ENTER.

Ca?	InProcess
Meter	<b>Zero</b> >>

3. Choose **Zero**. Press ENTER. The screens at the left will appear briefly during zeroing.

Live	1.000 $\mu$ S
Zeroing	<b>Wait</b>

After a few seconds, the display will return to a value of 0  $\mu$ S/cm and may then change slightly. A slight variation from zero is to be expected, and the procedure may be repeated several times, if necessary. A successful zero is indicated with a message of "Sensor Zero Done".

Live	0.000 $\mu$ S
Sensor Zero Done	

An unsuccessful zero will result if the conductivity reading is more than 1000  $\mu$ S/cm or if the reading is too unstable. The "Zero offset error" message indicates the reading is too high for the zero routine. If repeated attempts do not result in an acceptable zero, there is a good chance that there is a wiring problem.

4. After electronic zeroing, the Conductivity menu screen will appear.

### 8.3 CALIBRATING THE SENSOR IN A CONDUCTIVITY STANDARD

This procedure is used to check and correct the conductivity reading of the Model Xmt-C to ensure that the reading is accurate. This is done by submerging the probe in the sample of known conductivity, then adjusting the displayed value, if necessary, to correspond to the conductivity value of the sample.

This procedure must always be done after cleaning the probe. The temperature reading must also be checked and standardized if necessary, prior to performing this procedure.

**Important:** If you are submerging the probe in the commercial conductivity standard solution, follow steps 1 through 3 below. If you are leaving the probe submerged in a chemical solution bath and checking conductivity against a laboratory instrument, complete the meter cal procedure in Section 8.4.

1. Be sure that the probe has been cleaned of heavy deposits of dirt, oils, or chemical residue.
2. Commercial standards are referenced to a known temperature, for example, 4000 micromhos at 25°C (77°F). As the temperature of the standard changes, the conductivity will change. Therefore it is recommended that this procedure be performed at a temperature between 22 and 28 °C. **Be sure the probe has reached a stable temperature before standardizing.**
3. Pour the standard into a clean container. Submerge the clean probe in the standard. Place the probe so that a minimum of 1 in. of liquid surrounds the probe. Do not allow the probe to be closer than 1 in. to the sides or bottom of the container. Shake the probe slightly to eliminate any trapped air bubbles. Observe the displayed conductivity to determine if the sensor needs to be moved.

Alternatively, the probe can be calibrated while inserted in the process. This is done by adjusting the Model Xmt-C conductivity reading to the known conductivity of the process water which is close to the probe to be calibrated.

Live	10.00µS/cm
Cal	10.00µS/cm

Cal in Progress.
Please wait.

Updated	Cell
Const:	1.0013/cm

- a. Choose **In Process**. Press ENTER. The screen at the left will appear.
- b. Using the arrow keys, enter the actual conductivity of the conductivity standard in µS/cm. Press ENTER. The screens at the left will appear during this standardization process.

The conductivity reading in the display will change to the new value and the cell constant or cell factor will be recalculated. The cell factor can be viewed in the information screens.

**If too large an adjustment is attempted, the transmitter will display "calibration error" and no change will be made.**

## 8.4 CALIBRATING THE SENSOR TO A LABORATORY INSTRUMENT

This procedure is used to check the and correct the conductivity reading of the Model Xmt-C using a laboratory conductivity instrument. This is done by submerging the conductivity probe in a bath and measuring the conductivity of a grab sample of the same bath water with a separate laboratory instrument. The Model Xmt-C reading is then adjusted to match the conductivity reading of the lab instrument.

```
Use precision
resistors only
```

```
Live          50.00kΩ
Input         50.00kΩ
```

1. Choose **Meter**. Press ENTER. The screens at the left will appear.
2. Take a grab sample that is as close to the sensor as possible.
3. Using a calibrated laboratory instrument **with automatic temperature compensation**, determine the conductivity of the process or grab sample (as close to actual process temperature as possible). Continue with the following step if an adjustment is needed.
4. Using the arrow keys, enter the conductivity that appears on the laboratory instrument. Press ENTER. The screens at the left will appear during this standardization process.

```
Cal in progress.
Please wait.
```

## 8.5 CALIBRATING TEMPERATURE SLOPE

Conductivity is temperature-compensated per a constant linear slope of 0-5%/°C. The temperature slope currently being used by the transmitter is displayed. If this value is acceptable, press EXIT. 2%/°C is a good value for natural waters. To change the temperature slope, complete the following procedure.

```
Cal?          InProcess
Meter         Zero      >>
```

```
Enter Cell Const
Temp Slope    >>
```

```
Temperature
Slope?        2.00%/°C
```

1. If continuing from the previous steps, the screen on the left will appear.
2. Scroll right until the screen at the left appears. Choose **Temp Slope**. Press ENTER.
3. Enter the percent change of conductivity ( $\mu\text{S}/\text{cm}$ ) per degrees Centigrade. Press ENTER. All conductivity readings will be temperature compensated according to the linear slope constant that you have manually entered.

## SECTION 9.0 PROGRAMMING THE TRANSMITTER

### 9.1 GENERAL

This section describes how to program the transmitter using the keypad.

1. Configure and assign values to the 4-20 mA output [-HT version only] (Section 9.3)
2. Test and trim the current output [-HT version only] (Section 9.3)
3. Select the measurement to be made [conductivity, resistivity, or TDS] (Section 9.4)
4. Choose temperature units and automatic or manual temperature mode (Section 9.5)
5. Set a security code (Section 9.6)
6. Resetting factory default settings (Section 9.8)
7. Selecting a default display screen and adjusting screen contrast (Section 9.9)

### 9.2 CHANGING START-UP SETTINGS

When the Solu Comp Xmt is powered up for the first time, startup screens appear. The screens prompt the user to enter the measurement being made, to enter the cell constant, and to select temperature units. If incorrect settings were entered at startup, enter the correct settings now. To change the measurement, refer to Section 9.4.

## 9.3 CONFIGURING AND RANGING THE OUTPUT (-HT version only)

### 9.3.1 Purpose

1. Configuring an output means
  - a. displaying the output reading in units of mA or percent of full scale.
  - b. changing the time constant for output dampening.
  - c. assigning the value the output current will take if the transmitter detects a fault in itself or the sensor.
2. Ranging the output means assigning values to the 4 mA and 20 mA outputs.
3. Testing an output means entering a test value from the keypad to check the operation of recorders or transmitters.
4. Trimming an output means calibrating the 4 and 20 mA current outputs against a referee milliammeter.

### 9.3.2 Definitions

1. CURRENT OUTPUT. The transmitter provides a continuous 4-20 mA output current directly proportional to the pH of the sample.
2. FAULT. The transmitter continuously monitors itself and the sensor for faults. If the transmitter detects a fault, the 4-20 mA output can be programmed to go to a fixed value or it can be programmed to continue to display the live current reading. In any event Fault appears intermittently in the second line of the display.
3. DAMPEN. Output dampening smooths out noisy readings. But it also increases the response time of the output. To estimate the time (in minutes) required for the output to reach 95% of the final reading following a step change, divide the setting by 20. Thus, a setting of 140 means that, following a step change, the output takes about seven minutes to reach 95% of final reading. The output dampen setting does not affect the response time of the process display. The maximum setting is 255.
4. TEST. The transmitter can be programmed to generate a test current.

9.3.3 Procedure: Configuring the Output

Calibrate	Hold
<b>Program</b>	Display

1. Press MENU. The menu screen appears. Choose **Program**.

<b>Output</b>	Temp
Measurement <sup>a</sup>	>>

2. Choose **Output**.

Output?	Test
<b>Configure</b>	Range

3. Choose **Configure**.

Configure?	<b>Fault</b>
mA/%	Damping

4. Choose **Fault**.

Set to value?	
<b>Fixed</b>	Live

5. Choose **Fixed** or **Live**.

Current Output	
if Fault:22.00mA	

6. If you chose **Fixed**, the screen at left appears. Use the arrow keys to change the fault current to the desired value. The limits are 4.00 to 22.00 mA. If you chose **Live**, there are no settings to make.

Configure?	Fault
mA/%	Damping

7. The screen at left appears. Choose **mA/%**.

Display Output?	
mA	percent

8. Choose **mA** or **percent**. Percent means the display will show percent of full scale reading.

Configure?	Fault
mA/%	<b>Damping</b>

9. The screen at left appears. Choose **Damping**.

Damping?	000-255
	000 sec

10. Use the arrow keys to change the blinking display to the desired time constant.

7.3.4 Procedure: Ranging the output

Calibrate	Hold
<b>Program</b>	Display

1. From the main display, press MENU. The menu screen appears. Choose **Program**.

<b>Output</b>	Temp
Measurement <sup>a</sup>	>>

2. Choose **Output**.

Output?	Test
Configure	<b>Range</b>

3. Choose **Range**.

Output range?	
4mA	0.000µS/cm

4. Assign a value to the 4 mA output and press ENTER. Then assign a value to the 20 mA output. Press ENTER. Use the arrow keys to change the flashing display to the desired value.

**9.3.5 Procedure: Testing the output**

Calibrate	Hold
<b>Program</b>	Display

1. From the main display, press MENU. The menu screen appears. Choose **Program**.

<b>Output</b>	Temp
Measurement <sup>a</sup>	>>

2. Choose **Output**.

Output?	<b>Test</b>
Configure	Range

3. Choose **Test**.

<b>Test Output</b>
Trim Output

4. Choose **Test Output**.

Current Output
for Test:12.00mA

5. Use the arrow keys to change the displayed current to the desired value. Press ENTER. The output will change to the value just entered.

6. To return to normal operation, press EXIT. The output will return to the value determined by the process variable.

7. To return to the main display, press MENU then EXIT.

**9.3.6 Procedure: Trimming the output**

1. Connect an accurate milliammeter in series with the current output.

Calibrate	Hold
<b>Program</b>	Display

2. Press MENU. The menu screen appears. Choose **Program**.

<b>Output</b>	Temp
Measurement	>>

3. Choose **Output**.

Output?	<b>Test</b>
Configure	Range

4. Choose **Test**.

Test Output
<b>Trim Output</b>

5. Choose **Trim Output**.

Meter reading:
04.00mA

6. The output goes to 4.00 mA. If the milliammeter does not read 4.00 mA, use the arrow keys to change the display to match the current measured by the milliammeter. Press ENTER.

Meter reading:
20.00mA

7. The output goes to 20.00 mA. If the milliammeter does not read 20.00 mA, use the arrow keys to change the display to match the current measured by the milliammeter. Press ENTER.

Trim Complete
---------------

8. To return to the main display, press MENU then EXIT.



## 9.4 CHOOSING AND CONFIGURING THE ANALYTICAL MEASUREMENT

### 9.4.1 Purpose

This section describes how to do the following:

1. Configure the transmitter to measure conductivity, resistivity, or TDS.
2. If resistivity was selected, one of the following temperature corrections must be selected:
  - a. Neutral salt
  - b. Slope
  - c. Cation conductivity
  - d. Raw/no temp correction

### 9.4.3 Procedure: Measurement.

To choose a menu item, move the cursor to the item and press ENTER.  
To store a number or setting, press ENTER.

Calibrate                      Hold  
**Program**                      Display

1. Press MENU. The main menu screen appears. Choose **Program**.

Outputs                      Temp  
**Measurement**                      >>

2. Choose **Measurement**.

Measure?                      **Cond**  
Resistivity                      >>

3. Choose **Conductivity, Resistivity, TDS, or Custom**.

If you chose **Conductivity** or **Resistivity**, do steps 5 through 7.

If you chose **TDS**, no further measurement programming is required.

If you chose **Custom**, see Quick Start, step 8.

Measure?  
TDS                      Custom                      >>

Unit?                      (**µS, mS**)/cm  
(µS, mS, S)/m

4. For **Conductivity**, select measurement units: (µS, mS)/cm or (µS, mS, S)/m

Temp Correction?  
Slope                      Cation                      >>

5. Choose **slope, cation, raw/none, or neutral salt** for temperature correction.

Temp Correction?  
Raw/None                      >>

Temp Correction?  
**NeutralSalt**                      >>

Temperature  
Slope?                      **2.00%**/°C

6. If you chose **Slope**, enter the linear conductivity change in percent per degrees (C or F). Press ENTER.

7. To return to the main display, press MENU followed by EXIT.

## 9.5 CHOOSING TEMPERATURE UNITS AND MANUAL OR AUTOMATIC TEMPERATURE COMPENSATION

### 9.5.1 Purpose

This section describes how to do the following:

1. Choose temperature display units (°C or °F).
2. Choose automatic or manual temperature compensation.
3. Enter a temperature for manual temperature compensation

### 9.5.2 Definitions

1. **AUTOMATIC TEMPERATURE COMPENSATION.** The analyzer uses a temperature-dependent factor to convert raw conductivity to temperature-compensated conductivity. In automatic temperature compensation, the analyzer measures the temperature and automatically calculates the correct conversion factor unless **Slope** (user-entered) or **Raw/None** was selected. For maximum accuracy, use automatic temperature compensation.
2. **MANUAL TEMPERATURE COMPENSATION.** In manual temperature compensation, the analyzer converts raw conductivity to temperature-compensated conductivity using the temperature entered by the user. It does not use the actual process temperature. Do **NOT** use manual temperature compensation unless the process temperature varies no more than about  $\pm 2^{\circ}\text{C}$  or the pH is between 6 and 8. Manual temperature compensation is useful if the sensor temperature element has failed and a replacement sensor is not available. If manual temperature correction is selected, the display will not show the measured temperature. It will show the manually entered value.

### 9.5.3 Procedure: Temperature.

To choose a menu item, move the cursor to the item and press ENTER.

To store a number or setting, press ENTER.

Calibrate	Hold
<b>Program</b>	Display

1. Press MENU. The main menu screen appears. Choose **Program**.

Outputs	<b>Temp</b>
Measurement	>>

2. Choose **Temp**.

Config Temp?	
<b>°C/F</b>	Live/Manual

3. Choose **°C/F** to change temperature units. Choose **Live/Manual** to turn on (Live) or turn off (Manual) automatic temperature compensation.
  - a. If **°C/F** is chosen, select **°C** or **°F** in the next screen.
  - b. If **Live/Manual** is chosen, select **Live** or **Manual** in the next screen.
  - c. If **Manual** is chosen, enter the temperature in the next screen. The temperature entered in this step will be used in all subsequent measurements, no matter what the process temperature is.

## 9.6 SETTING A SECURITY CODE

### 9.6.1 Purpose

This section describes how to set a security code. There are three levels of security:

- A user can view the default display and information screens only.
- A user has access to the calibration and hold menus only.
- A user has access to all menus.

The security code is a three-digit number. The table shows what happens when security codes are assigned to **Calib** (calibration) and **Config** (configure). In the table XXX and YYY are the assigned security codes. To bypass security, enter 555.

Code assignments		What happens
Calib	Config	
000	XXX	User enters XXX and has access to all menus.
XXX	YYY	User enters XXX and has access to calibration and hold menus only. User enters YYY and has access to all menus.
XXX	000	User needs no security code to have access to all menus.
000	000	User needs no security code to have access to all menus.

### 9.6.2 Procedure: Setting a security code

```
Calibrate          Hold
Program           Display
```

```
Outputs          Temp
Measurement      >>
```

```
Security          HART
Reset Analyzer    >>
```

```
Lock?
Calib            Config
```

- Press MENU. The menu screen appears. Choose **Program**.
- Choose **>>**.
- Choose **Security**.
- Choose **Calib** or **Config**.
  - If you chose **Calib**, enter a three-digit security code.
  - If you chose **Config**, enter a three-digit security code.
- To return to the main display, press MENU the EXIT.

## 9.7 MAKING HART RELATED SETTINGS

For more information refer to Section 6.0.

## 9.8 RESETTING FACTORY CALIBRATION AND FACTORY DEFAULT SETTINGS

### 9.8.1 Purpose

This section describes how to install factory calibration and default values. The process also clears all fault messages and returns the display to the first quick start screen.

### 9.8.2 Procedure: Installing default settings

Calibrate	Hold
<b>Program</b>	Display

1. Press MENU. The menu screen appears. Choose **Program**.

Outputs	Temp
Measurement	>>

2. Choose >>.

Security	HART
<b>Reset Analyzer</b>	>>

3. Choose **ResetTransmitter**.

Load factory		
settings?	<b>Yes</b>	No

4. Choose **Yes** or **No**. Choosing **Yes** clears previous settings and calibrations and returns the transmitter to the first quick start screen.

## 9.9 SELECTING A DEFAULT SCREEN AND SCREEN CONTRAST

### 9.9.1 Purpose

This section describes how to do the following:

1. Set a default screen. The default screen is the screen shown during normal operation. The Solu Comp Xmt allows the user to choose from a number of screens. Which screens are available depends on the measurement the transmitter is making.
2. Change the screen contrast.

### 9.9.2 Procedure: Choosing a display screen.

```
Calibrate           Hold
Program            Display
```

```
Default Display
Display Contrast
```

1. Press MENU. The menu screen appears. Choose **Display**.
2. Choose **Default Display**.
3. Press ↓ until the desired screen appears. Press ENTER. One of two display screen layouts can be selected for any measurement selected.
4. The display returns to the screen in step 2. Press MENU then EXIT to return to the main display.

### 9.9.3 Procedure: Changing screen contrast.

```
Calibrate           Hold
Program            Display
```

```
Default Display
Display Contrast
```

```
Display contrast
Lighter            Darker
```

1. Press MENU. The menu screen appears. Choose **Display**.
2. Choose **Display Contrast**.
3. To increase the contrast, select **darker**. Press ENTER. Each key press increases the contrast. To reduce the contrast, select **lighter**, Press ENTER. Each key press decreases the contrast.
4. To return to the main display, press MENU then EXIT.

#### NOTE:

Screen contrast can also be adjusted from the main display. Press MENU and ↑ at the same time to increase contrast. Press MENU and ↓ at the same time to decrease contrast. Repeatedly pressing the arrow key increases or reduces the contrast.

## SECTION 10.0 MAINTENANCE

### 10.1 OVERVIEW

The Solu Comp Xmt needs little routine maintenance. The calibration of the analyzer and sensor should be checked periodically. To recalibrate the sensor and analyzer, refer to sections 7 through 9.

### 10.2 REPLACEMENT PARTS

Only a few components of the analyzer are replaceable. Refer to the tables below. Circuit boards, display, and enclosure are not replaceable.

#### REPLACEMENT PARTS FOR SOLU COMP XMT (PANEL MOUNT VERSION)

PART NUMBER	DESCRIPTION	SHIPPING WEIGHT
23823-00	Panel mounting kit, includes four brackets and four set screws	1 lb/0.5 kg
33654-00	Gasket, front, for panel mount version	1 lb/0.5 kg
33658-00	Gasket, rear cover, for panel mount version	1 lb/0.5 kg

#### REPLACEMENT PARTS FOR SOLU COMP XMT (PIPE/SURFACE MOUNT VERSION)

PART NUMBER	DESCRIPTION	SHIPPING WEIGHT
33655-00	Gasket for pipe/surface mount version	1 lb/0.5 kg
23833-00	Surface mount kit, consists of four self tapping screws and four O-rings	1 lb/0.5 kg

## SECTION 11.0 THEORY OF OPERATION

### 11.1 CONDUCTIVITY / RESISTIVITY / % CONCENTRATION

Liquids can only conduct electrical currents when they contain particles that carry charges. These particles are called ions, and they are produced when acids, bases, and salts are mixed with water. The conductivity of a substance determines how well it can carry electrical currents and is used to indicate the concentration of acids, bases, and salts in water.

Conductance is the reciprocal of resistance. The traditional unit of conductance is mho, a term representing the reciprocal of ohm. Recently, the unit siemen has replaced the mho, but the amount of conductance is exactly the same. Liquid water has relatively low conductivity, so measurements are expressed in millisiemens (.001 siemen) or microsiemens (.000001 siemen), and abbreviated as mS or  $\mu$ S, respectively.

The Model Xmt-C conductivity transmitter is a device used to measure conductivity in most chemical processes. Conductivity is a function of ion concentration, ionic charge, and ion mobility. Ions in water conduct current when an electrical potential is applied across electrodes immersed in the solution.

Model Xmt-C conductivity transmitter uses conductivity probes with electrodes (contacting). Model Xmt-T is designed for use with inductive (toroidal) probes for measurement of high conductivity. Generally, contacting (electrode) probes are used for conductivity below 200 microsiemens, such as water rinses in metal finishing or ultra-pure boiler water applications. The electrode design is more sensitive for low level measurement and these water solutions tend to be non-fouling.

For % concentration measurement, the Model Xmt-C transmitter uses the measured temperature and absolute conductivity and applies specific algorithms that have been developed for each of the substances available in the instrument.

### 11.2 TEMPERATURE CORRECTION

The conductivity of an electrolyte solution depends strongly on temperature. To allow comparison among measurements made at different temperatures, conductivity values are usually converted to the value at 25°C. The Model Xmt-C performs the correction automatically following one of three temperature correction algorithms.

1. User-programmable linear temperature coefficient (slope)
2. Ultra-pure water
3. Cation conductivity (dilute hydrochloric acid)

Temperature correction can also be turned off. If temperature correction is off, the Model Xmt-C displays the raw or non-temperature corrected conductivity. Temperature corrections apply whether the measurement is in conductivity or resistivity units.

**DEFINITIONS**

1. **LINEAR TEMPERATURE COEFFICIENT OR TEMPERATURE SLOPE.** The change in the conductivity of most electrolyte solutions having conductivity greater than about 5 mS/cm at 25°C can be expressed by the following equation:

$$C_{25} = \frac{C_t}{1 + a(t - 25)}$$

In the equation,  $C_{25}$  is the conductivity at 25°C,  $C_t$  is the conductivity at  $t^\circ\text{C}$ , and  $a$  is the linear temperature coefficient. The linear temperature coefficient, sometimes called the temperature slope, has units of  $\%/^\circ\text{C}$ . In the equation, the temperature coefficient is expressed as a decimal fraction. The linear temperature coefficient depends to some extent on both the temperature and the concentration of the salt solution. The temperature coefficient also varies from salt to salt.

For maximum accuracy, the temperature coefficient must be appropriate for the salt or salts in solution, their concentration, and the temperature. Frequently the relationship must be determined by experiment. Fortunately, for most dilute neutral electrolyte solutions, a linear temperature coefficient of  $2.00\%/^\circ\text{C}$  (0.0200) works reasonably well. The table below gives typical ranges for different electrolytes.

	Slope ( $\%/^\circ\text{C}$ )
Neutral salts	1.8 - 3.0
Acids	1.0 - 1.6
Bases	1.8 - 2.2
High purity water	Use standard correction

Temperature compensated conductivity measurements are important in the power industry. The table lists temperature slopes for different types of treatment chemicals. The slopes apply across the range of concentrations typically encountered.

	Slope ( $\%/^\circ\text{C}$ )
Condensate treated with ammonia	2.00
Boiler water treated with phosphate/caustic	2.00

- CATION TEMPERATURE CORRECTION.** Cation conductivity, sometimes called acid conductance, is used in steam power plants to measure salt contamination in the boiler feedwater and steam. The Model Xmt-C automatically corrects for the variation in the conductivity of extremely dilute hydrochloric acid with temperature and displays cation conductivity measurements. Cation conductivity temperature also applies to semiconductor etch rinse baths, which contain trace amounts of acids.
- RAW.** Raw conductivity is the conductivity of the sample at the measurement temperature.



## SECTION 12.0

# THEORY - REMOTE COMMUNICATIONS

### 12.1 OVERVIEW OF HART COMMUNICATION


HART (highway addressable remote transducer) is a digital communication system in which two frequencies are superimposed on the 4 to 20 mA output signal from the transmitter. A 1200 Hz sine wave represents the digit 1, and a 2400 Hz sine wave represents the digit 0. Because the average value of a sine wave is zero, the digital signal adds no dc component to the analog signal. HART permits digital communication while retaining the analog signal for process control.

The HART protocol, originally developed by Fisher-Rosemount, is now overseen by the independent HART Communication Foundation. The Foundation ensures that all HART devices can communicate with one another. For more information about HART communications, call the HART Communication Foundation at (512) 794-0369. The internet address is <http://www.hartcomm.org>.

### 12.2 HART INTERFACE DEVICES

HART communicators allow the user to view measurement data (conductivity readings and temperature), program the transmitter, and download information from the transmitter for transfer to a computer for analysis. Downloaded information can also be sent to another HART transmitter. Either a hand-held communicator, such as the Rosemount Model 275, or a computer can be used. HART interface devices operate from any wiring termination point in the 4 - 20 mA loop. A minimum load of 250 ohms must be present between the transmitter and the power supply. See Figure 12-1.

If your communicator does not recognize the Model XMT-C transmitter, the device description library may need updating. Call the manufacturer of your HART communication device for updates.



### 12.3 ASSET MANAGEMENT SOLUTIONS


Asset Management Solutions (AMS) is software that helps plant personnel better monitor the performance of analytical instruments, pressure and temperature transmitters, and control valves. Continuous monitoring means maintenance personnel can anticipate equipment failures and plan preventative measures before costly breakdown maintenance is required.

AMS uses remote monitoring. The operator, sitting at a computer, can view measurement data, change program settings, read diagnostic and warning messages, and retrieve historical data from any HART-compatible device, including the Model XMT-C transmitter. Although AMS allows access to the basic functions of any HART compatible device, Rosemount Analytical has developed additional software for that allows access to all features of the Model XMT-C transmitter.

AMS can play a central role in plant quality assurance and quality control. Using AMS Audit Trail, plant operators can track calibration frequency and results as well as warnings and diagnostic messages. The information is available to Audit Trail whether calibrations were done using the infrared remote controller, the Model 375 or 275 HART communicator, or AMS software.

AMS operates in Windows 95. See Figure 12-2 for a sample screen. AMS communicates through a HART-compatible modem with any HART transmitters, including those from other manufacturers. AMS is also compatible with FOUNDATION™ Fieldbus, which allows future upgrades to Fieldbus instruments.

For more information about AMS, including upgrades, renewals, and training, call Fisher-Rosemount Systems, Inc. at (612) 895-2000.



## SECTION 13.0 RETURN OF MATERIAL

### 13.1 GENERAL.

To expedite the repair and return of instruments, proper communication between the customer and the factory is important. Call 1-949-757-8500 for a Return Materials Authorization (RMA) number.

### 13.2 WARRANTY REPAIR.

The following is the procedure for returning instruments still under warranty:

1. Call Rosemount Analytical for authorization.
2. To verify warranty, supply the factory sales order number or the original purchase order number. In the case of individual parts or sub-assemblies, the serial number on the unit must be supplied.
3. Carefully package the materials and enclose your "Letter of Transmittal" (see Warranty). If possible, pack the materials in the same manner as they were received.
4. Send the package prepaid to:

Emerson Process Management  
Liquid Division  
2400 Barranca Parkway  
Irvine, CA 92606

Attn: Factory Repair

RMA No. \_\_\_\_\_

Mark the package: Returned for Repair

Model No. \_\_\_\_\_

### 13.3 NON-WARRANTY REPAIR.

The following is the procedure for returning for repair instruments that are no longer under warranty:

1. Call Rosemount Analytical for authorization.
2. Supply the purchase order number, and make sure to provide the name and telephone number of the individual to be contacted should additional information be needed.
3. Do Steps 3 and 4 of Section 13.2.

#### NOTE

Consult the factory for additional information regarding service or repair.



## **WARRANTY**

Goods and part(s) (excluding consumables) manufactured by Seller are warranted to be free from defects in workmanship and material under normal use and service for a period of twelve (12) months from the date of shipment by Seller. Consumables, pH electrodes, membranes, liquid junctions, electrolyte, O-rings, etc. are warranted to be free from defects in workmanship and material under normal use and service for a period of ninety (90) days from date of shipment by Seller. Goods, part(s) and consumables proven by Seller to be defective in workmanship and / or material shall be replaced or repaired, free of charge, F.O.B. Seller's factory provided that the goods, parts(s), or consumables are returned to Seller's designated factory, transportation charges prepaid, within the twelve (12) month period of warranty in the case of goods and part(s), and in the case of consumables, within the ninety (90) day period of warranty. This warranty shall be in effect for replacement or repaired goods, part(s) and consumables for the remaining portion of the period of the twelve (12) month warranty in the case of goods and part(s) and the remaining portion of the ninety (90) day warranty in the case of consumables. A defect in goods, part(s) and consumables of the commercial unit shall not operate to condemn such commercial unit when such goods, parts(s) or consumables are capable of being renewed, repaired or replaced.

The Seller shall not be liable to the Buyer, or to any other person, for the loss or damage, directly or indirectly, arising from the use of the equipment or goods, from breach of any warranty or from any other cause. All other warranties, expressed or implied are hereby excluded.

IN CONSIDERATION OF THE STATED PURCHASE PRICE OF THE GOODS, SELLER GRANTS ONLY THE ABOVE STATED EXPRESS WARRANTY. NO OTHER WARRANTIES ARE GRANTED INCLUDING, BUT NOT LIMITED TO, EXPRESS AND IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE.

## **RETURN OF MATERIAL**

Material returned for repair, whether in or out of warranty, should be shipped prepaid to:

**Emerson Process Management  
Liquid Division  
2400 Barranca Parkway  
Irvine, CA 92606**

The shipping container should be marked:

Return for Repair

Model \_\_\_\_\_

The returned material should be accompanied by a letter of transmittal which should include the following information (make a copy of the "Return of Materials Request" found on the last page of the Manual and provide the following thereon):

1. Location type of service, and length of time of service of the device.
2. Description of the faulty operation of the device and the circumstances of the failure.
3. Name and telephone number of the person to contact if there are questions about the returned material.
4. Statement as to whether warranty or non-warranty service is requested.
5. Complete shipping instructions for return of the material.

Adherence to these procedures will expedite handling of the returned material and will prevent unnecessary additional charges for inspection and testing to determine the problem with the device.

If the material is returned for out-of-warranty repairs, a purchase order for repairs should be enclosed.



*The right people,  
the right answers,  
right now.*

ROSEMOUNT ANALYTICAL  
CUSTOMER SUPPORT CENTER  
1-800-854-8257



Credit Cards for U.S. Purchases Only.



ON-LINE ORDERING NOW AVAILABLE ON OUR WEB SITE  
<http://www.raihome.com>

*Specifications subject to change without notice.*

**Emerson Process Management**

2400 Barranca Parkway  
Irvine, CA 92606 USA  
Tel: (949) 757-8500  
Fax: (949) 474-7250

<http://www.raihome.com>

