



CERTIFICATE OF ANALYSIS
Complies with ISO Guide 34, ISO Guide 31,
ISO Guide 35, and ISO 9001
TRACEABLE® CERTIFIED REFERENCE MATERIAL



This certificate indicates traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

Certificate No.: 4580-8151717 **Amended Certificate**
Reference Certificate 4580-7887064

Description: Conductivity Standard 149,640 $\mu\text{S/cm}$

Catalog Number: 39193 **Lot:** CC15187

Certificate Date: August 16, 2016 **Expiration Date:** August 16, 2017

Certified Value: 149,640 $\mu\text{S/cm}$ $U = \pm 520 \mu\text{S/cm (k=2)}$ at 25°C

Derived Values: 149640 micromho/cm, 6.68 ohm-cm, 99760 PPM D.S.

Certification measurements are performed under ISO Guide 34, A2LA accreditation no. 1750.02 and are traceable to recognized national and international standards via an unbroken chain of comparisons. Electrical conductance is the reciprocal of electrical impedance. The International System of units (SI), derived unit of conductance, is Siemens (S), also referred to as (mhos) the reciprocal of ohms. The certified value is expressed in microsiemens per centimeter ($\mu\text{S/cm}$).

MEASUREMENT: Ten (10) 100 ml samples were measured from this lot. The conductivity of each sample was derived from a measurement of the impedance of the solution using a conductivity meter and calibrated cell. The cell and sample were temperature controlled by submersion in a water bath at 25°C $\pm 0.015^\circ\text{C}$.

UNCERTAINTY: The certified value is given as the average of the measured samples. The reported expanded uncertainty (U) is determined from the measurement variation from sample to sample, change due to shelf life, and from the uncertainty of the measurement process. The value of uncertainty is multiplied by $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%. Uncertainty is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM).

METHOD: The certified reference material is prepared and analyzed by Control Company. The certified reference material consists of a mixture of a dilute solution of less than 14% (by mass) potassium chloride (KCL), of less than 2% (by mass) propanol, and of less than 88% (by mass) deionized water in equilibrium with atmospheric carbon dioxide. Mixing was performed by circulation utilizing a proprietary method.


Aaron Judice, Technical Manager


Nicol Rodriguez, Quality Manager

Traceability: Standards and Equipment Used

Description	Serial Number	Due Date	Traceable Reference
Conductivity Probe (4W)/ Meter	10124-F02	9/09/16	TC26-7501964
Digital Thermometer	140073819	2/03/17	4000-7392222
Calibration Bath TC-337	B5C477		

Laboratory environment conditions: 26.0°C 49%RH 1,013mb/hPa

CONTROL COMPANY 12554 Old Galveston RD Suite B230 Webster TX 77598 USA
Tel: (281) 482 1714 Fax: (281) 482 9448 sales@control3.com www.control3.com

Control Company is an ISO Guide 34:2009 Certified Reference Material (CRM) Producer Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.02). This certificate fulfills the requirements of ISO Guide 31:2000 (Reference Materials - Contents of Certificates and Labels), ISO Guide 34:2009 "Quality System Guidelines for the Production of Reference Materials", and ISO Guide 35:2006 "Certification of Reference Materials - General and Statistical Principles". Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.01). Control Company is ISO 9001:2008 certified by DNV GL (Certificate No. CERT-01805-2006-AQ-HOU-RvA). TRACEABLE® is a registered trademark of Control 3 Inc.



CERTIFICATE OF ANALYSIS
Complies with ISO Guide 34, ISO Guide 31,
ISO Guide 35, and ISO 9001
TRACEABLE® CERTIFIED REFERENCE MATERIAL



PACKAGING: This material is available in both a 460 ml bottle and a 100 ml One-Shot™.

INTENDED USE: The certified reference material is intended for the calibration of conductivity cell constants, for conductivity measurement, for the validation of analytical methods, and for the preparation of working reference standards.

INSTRUCTIONS FOR USE: The certified reference material should be open for the minimum time. Rinse the cell in a small amount of the certified reference material and discard. The recommended sample size for measurement is 100 ml. Discard the standard after use and under the following circumstances: if the expiration date is past due, four months after opening, or if any color, turbidity, or visible microbiological growth become evident. Standards which have been opened are not protected from growth. Do not return used solution to this standard. Contaminates and evaporation have a significant effect on conductivity. Keep the standard closed. Keep the standard stored at a stable temperature. Select a standard as near as possible to that of the unknown solution to be measured. Do not standardize at 10,000 uS and then measure unknowns at 100 uS. Reference any accompanying instructions shipped with this product.

Temperature has a significant effect on conductivity. For measurements at a temperature other than 25°C, refer to the temperature correction table provided. This product should be used as near as possible to 25°C.

HOMOGENEITY: Ten (10) 100 ml samples were selected for analytical control. Results from different samples showed no statistically significant differences, nor was there any correlation between values obtained and the bottling sequence. Bottle-to-bottle (One-Shot™ to One-Shot™) variations of the samples measured are included as a part of the calculated measurement uncertainty stated on page 1 of this certificate. A minimum sample size of 100 ml should be used to maintain the certified value and the associated statement of uncertainty. This standard as formulated is considered infinitely soluble.

STABILITY, SHELF LIFE: The expiration date stated on page 1 indicates the period of time which the certified reference material in a properly packaged, unopened, unused, and stored under environmentally controlled and monitored conditions remains within the specified uncertainty range.

EXPIRATION DATE: The date after which a certified reference material should be discarded.

STORAGE: Store below 40°C and above 0°C.

SHIPPING: Ship below 50°C and above 0°C.

MAINTENANCE OF CERTIFICATION: Control Company monitors representative samples from this lot over the period of its certification. If a change occurs that affects the certification before the expiration date, Control Company posts amended certificates at www.control3.com/crmupdate.htm.

MSDS INFORMATION: Please refer to the Material Safety Data sheet for information regarding this certified reference material at www.control3.com (Search MSDS). Use only the first four digits of the certificate number to locate the MSDS.

QUALITY STANDARD DOCUMENTATION:

ISO Guide 34:2009 General Requirements for the Competence of Reference Material Producers, accredited A2LA Certificate Number 1750.02.

ISO Guide 31:2000 Reference Materials - Contents of Certificates and Labels.

ISO Guide 35:2006 Certification of Reference Materials - General and Statistical Principals.

ISO/IEC 17025:2005 General Requirements for the Competence of Testing and Calibration Laboratories, accredited A2LA Certificate Number 1750.01.

ANSI/NCSL Z540-1:1994 Calibration Laboratories and Measuring and Test Equipment-General Requirements.

ISO 9001:2008 Quality Management System Registration - DNV GL Certificate Number CERT-01805-2006-AQ-HOU-RvA

SUPPORTED METHODS: This certified reference material meets test requirements for Federal, State, and local agencies, CAP, CLSI, ACS, and CLIA. Traceable® Certified Reference Material complies with and is essential for use in these official methods: AOAC 973.40, EPA 120.1, Standard Method 2510 (APHA, AWWA, WEF), ISO 7888, DIN 38404, ASTM D1125, USGS I-1780, USP 645, OIML R56, IUPAC, and for A2LA/NVLAP accreditations /ISO 9000 certifications. Material may be used to calibrate all conductivity meters and to determine all conductivity cell constants.

CONTROL COMPANY 12554 Old Galveston RD Suite B230 Webster TX 77598 USA
Tel: (281) 482 1714 Fax: (281) 482 9448 sales@control3.com www.control3.com

Control Company is an ISO Guide 34:2009 Certified Reference Material (CRM) Producer Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.02). This certificate fulfills the requirements of ISO Guide 31:2000 (Reference Materials - Contents of Certificates and Labels), ISO Guide 34:2009 "Quality System Guidelines for the Production of Reference Materials", and ISO Guide 35:2006 "Certification of Reference Materials - General and Statistical Principals". Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.01). Control Company is ISO 9001:2008 certified by DNV GL (Certificate No. CERT-01805-2006-AQ-HOU-RvA). TRACEABLE® is a registered trademark of Control 3 Inc.



CERTIFICATE OF ANALYSIS
Complies with ISO Guide 34, ISO Guide 31,
ISO Guide 35, and ISO 9001
TRACEABLE® CERTIFIED REFERENCE MATERIAL



Temperature Correction Information

1.707 %

If your conductivity meter allows you to set a temperature coefficient (temperature correction) then the underlined number shown above is the best approximation for this specific analysis for this specific Traceable® Certified Reference Material. For more precise measurements use the chart. Use the chart below only for making absolute measurements. That is, measurements without any automatic temperature correction (temperature coefficient set to 0). The chart below displays derived values.

Using a thermometer, measure the temperature of this certified reference material. Shown on the chart is temperature (in the far left column) in whole degrees. Shown across the top row is temperature in tenths of a degree. Locate the measured temperature in whole numbers on the far left column, and then follow across the row to the temperature in tenths of a degree. At the intersection is the certified reference material value at that specific temperature. Standardize your meter using that value. Example: Measured temperature is 25.0°C. Find 25°C in the far left column, find the row 0.0°C. Where 25°C and 0.0°C intersect, read the value in microsiemens/cm.

Temperature Correction Chart in microsiemens/cm

	0.0°C	0.1°C	0.2°C	0.3°C	0.4°C	0.5°C	0.6°C	0.7°C	0.8°C	0.9°C
15°C	124420	124668	124916	125164	125412	125661	125909	126158	126406	126655
16°C	126904	127153	127402	127651	127900	128149	128399	128648	128898	129147
17°C	129397	129647	129897	130147	130397	130647	130897	131148	131398	131649
18°C	131899	132150	132401	132652	132903	133154	133405	133657	133908	134159
19°C	134411	134663	134914	135166	135418	135670	135922	136175	136427	136679
20°C	136932	137185	137437	137690	137943	138196	138449	138702	138955	139209
21°C	139462	139716	139969	140223	140477	140731	140985	141239	141493	141747
22°C	142002	142256	142511	142765	143020	143275	143530	143785	144040	144295
23°C	144550	144806	145061	145317	145572	145828	146084	146340	146596	146852
24°C	147108	147365	147621	147878	148134	148391	148648	148904	149161	149418
25°C	149640	149933	150190	150448	150705	150963	151220	151478	151736	151994
26°C	152252	152510	152768	153027	153285	153544	153802	154061	154320	154579
27°C	154838	155097	155356	155615	155875	156134	156394	156653	156913	157173
28°C	157433	157693	157953	158213	158473	158734	158994	159255	159515	159776
29°C	160037	160298	160559	160820	161081	161343	161604	161866	162127	162389
30°C	162651	162912	163174	163436	163699	163961	164223	164486	164748	165011
31°C	165273	165536	165799	166062	166325	166588	166852	167115	167378	167642
32°C	167905	168169	168433	168697	168961	169225	169489	169753	170018	170282
33°C	170547	170811	171076	171341	171606	171871	172136	172401	172666	172932
34°C	173197	173463	173729	173994	174260	174526	174792	175058	175324	175591
35°C	175857	176124	176390	176657	176924	177191	177457	177725	177992	178259

CONTROL COMPANY 12554 Old Galveston RD Suite B230 Webster TX 77598 USA
Tel: (281) 482 1714 Fax: (281) 482 9448 sales@control3.com www.control3.com

Control Company is an ISO Guide 34:2009 Certified Reference Material (CRM) Producer Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.02). This certificate fulfills the requirements of ISO Guide 31:2000 (Reference Materials - Contents of Certificates and Labels), ISO Guide 34:2009 "Quality System Guidelines for the Production of Reference Materials", and ISO Guide 35:2006 "Certification of Reference Materials - General and Statistical Principles". Control Company is an ISO/IEC 17025:2005 Calibration Laboratory Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.01). Control Company is ISO 9001:2008 certified by DNV GL (Certificate No. CERT-01805-2006-AQ-HOU-RvA). TRACEABLE® is a registered trademark of Control 3 Inc.