

K131482

NOV 22 2013

510(k) Summary	
Trade Name:	SOFIA™ Distal Access Catheter
Generic Name:	Percutaneous Catheter
Classification:	Class II, 21 CFR 870.1250 (DQY), 21 CFR 870.1200 (DQO)
Submitted By:	MicroVention, Inc 1311 Valencia Avenue Tustin, California U.S.A.
Contact:	Naomi Gong
Date:	2013 November 13
Predicate Device:	Chaperon Guiding Catheter (K082385) Headway Duo Microcatheter (K120917)

Device Description:

The SOFIA Distal Access Catheter is a single-lumen, flexible catheter designed with coil and braid reinforcement. The distal segment is steam-shapeable and it has a hydrophilic coating for navigation through the vasculature. The radiopaque marker is located at the distal end of the catheter for visualization under fluoroscopy. An introducer sheath and shaping mandrel are also provided.

Indications For Use:

The SOFIA Distal Access Catheter is indicated for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate introduction of diagnostic or therapeutic devices. It is not intended for use in coronary arteries.

Technological Comparison:

	Chaperon Guiding Catheter	SOFIA Distal Access Catheter	Headway Duo Microcatheter
Intended Use	Intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.	Intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.	Intended for general intravascular use, including the peripheral and coronary vasculature for the infusion of diagnostic agents, such as contrast media, and therapeutic agents, such as embolization materials. Intended for neurovascular use for the infusion of diagnostic agents, such as contrast media, and therapeutic agents that have been cleared or approved for use in the neurovasculature and are compatible with the inner diameter of the microcatheter.
Material Catheter Body	Outer layer of polyester elastomer; stainless steel braid; inner layer of PTFE (polytetrafluoroethylene).	Outer layer of polyurethane elastomer (Polyblend and Pellethane), polyether block amide (Pebax) and polyamide (Grilamid); inner layer of stainless steel braid/coil, PTFE and polyolefin elastomer	Outer layer of polyether block amide (Pebax) and polyamide (Grilamid); inner layer of stainless steel braid/coil and PTFE
Marker	Tungsten	Platinum/Iridium	Platinum/Iridium
Hub	Nylon	Nylon	Nylon
Strain Relief	Polyester elastomer	Polyurethane	Pebax
Introducer	Not applicable	Pebax	Pebax
Shaping Mandrel	Not applicable	Stainless steel	Stainless steel

	Chaperon Guiding Catheter	SOFIA Distal Access Catheter	Headway Duo Microcatheter
Catheter size	5 F (outer catheter)	5 F	1.6 - 2.1 F
ID	0.059 inch (1.5 mm)	0.055 inch (1.4 mm)	0.0165 inch (0.42 mm)
OD	0.068 inch (1.7 mm)	0.068 inch (1.7 mm)	0.023 – 0.0275 inch (0.58 – 0.70 mm)
Effective Length	95 cm (outer) 117 cm (inner)	125 cm	157 and 168 cm
Coating	Hydrophilic coating (Terumo proprietary coating)	Hydrophilic coating (Hydak [®] – same)	Hydrophilic coating (Hydak [®] – same)
Tip Configuration	Preshaped	Steam shapeable by user	Steam shapeable by user
Guidewire Compatibility	0.035 inch or 0.038 inch	0.035 inch or 0.038 inch	0.014 inch or smaller
Accessories	N/A	Introducer sheath and shaping mandrel	Introducer sheath and shaping mandrel .
Method of Supply	Sterile and single use	Sterile and single use	Sterile and single use
Sterilization Method	Ethylene Oxide	Ethylene Oxide	Ethylene Oxide
Packaging Configuration	Catheter placed on packaging card that is inserted into Tyvek [®] pouch. Pouch and IFU placed in bleached sulfate carton box.	Catheter placed into a HDPE dispenser tube. Dispenser tube, introducer and shaping mandrel placed on a polyethylene packaging card that is inserted into a Tyvek [®] pouch. Pouch and IFU placed in bleached sulfate carton box.	Catheter placed into a polyethylene dispenser hoop. Introducer and shaping mandrel placed on polyethylene packaging card. Dispenser hoop and packaging card inserted into Tyvek [®] pouch. Pouch and IFU placed in bleached sulfate carton box.

Verification and Test Summary:

Bench Testing		
Test	Results	Conclusions
Simulated Use	Test articles achieved a rating ≥ 3 for preparation/ease of assembly, introducer sheath interaction, introducer peel away, tracking with guidewire/microcatheter, microcatheter/guidewire lockup, lubricity and durability of hydrophilic coating, microcatheter/guidewire removal, removal/aspiration of clot, mechanical clot retriever and stent delivery with no particles	Device performs as intended under simulated use conditions
Equipment Interface	Test articles compatible with 0.035-inch and 0.038-inch guidewires, 6F or larger guide catheter/guiding sheath, common RHVs using insertion tool, stopcocks and ≤ 0.027 -inch microcatheters	Device compatible with recommended accessories commonly used in intravascular procedures
Dimensional and Physical Attributes	Test articles met the specified dimensional requirements for catheter OD, catheter ID, overall working length, length of distal section, length of distal tip to marker band and total length of hub/strain relief	Device met established dimensional and physical specifications
Kink Resistance	No kinks at 1 cm, 4 cm, 12 cm and 25 cm from distal tip when wrapped around 0.025-inch and 0.030-inch pin gauges No kinks noted during simulated use testing	Device resistant to kinking around small radii turns
Tip Shapeability	Tip angle of test article equivalent to competitive devices after steam shaping using mandrel with an angle of approximately 90°	Shapeability of distal tip after steam shaping equivalent to competitive devices
Radio Detectability	Distal marker band visible under fluoroscopy	Device radiopacity equivalent to or better than predicate and competitive devices

Bench Testing		
Test	Results	Conclusions
Gauging (ISO 594-2)	Gauging pin and hub align in limit planes	Device hub meets the requirements of ISO 594-2
Separation Force (ISO 594-2)	Mating parts separation force greater than 25 N	Device hub meets the requirements of ISO 594-2
Unscrewing Torque (ISO 594-2)	Test article luer remains attached after applying an unscrewing torque not less than 0.02 Nm for a minimum of 10 seconds	Device hub meets the requirements of ISO 594-2
Stress Cracking (ISO 594-2)	No stress cracks on test article hub	Device hub meets the requirements of ISO 594-2
Ease of Assembly (ISO 594-2)	Components fit together securely with no resistance observed between test article luer and reference fitting	Device hub meets the requirements of ISO 594-2
Resistance to Overriding (ISO 594-2)	Test article luer does not override reference fitting threads	Device hub meets the requirements of ISO 594-2
Durability/Lubricity of Hydrophilic Coating	Test article achieved a rating of ≥ 3 during simulated use testing for coating durability and lubricity.	Device tracks easily with no coating cracking or separation
Catheter Stiffness	Device stiffness equivalent to predicate and competitive devices	Device tracks in tortuous anatomy while advancing to target site
Torque Strength	No catheter breakage after 50 rotations	Device torque strength same as predicate device
Catheter Flexural Fatigue	No flexural fatigue following repeated bending during simulated use testing and repeated hoop stress following pressure and air aspiration testing	Device integrity suitable for intended clinical use
Surface Contamination	Test article free from surface contaminants from uncured coating surface particulates $> 0.02 \text{ mm}^2$, embedded particulates Distal tip smooth and tapered PTFE inner layer not delaminated	Device integrity suitable for intended clinical use
Force at Break (Distal and Hub)	Catheter force at break $\geq 2.25 \text{ lbf}$ for distal section and hub/catheter junction	Tensile strength test results equivalent to predicate and competitive devices

Bench Testing		
Test	Results	Conclusions
Flow Rate	Flow rate at 100 psi and 300 psi with diagnostic agents (e.g., saline, contrast media) equivalent to or better than competitive devices	Device meets specified requirements for delivery of diagnostic agents
Static Burst Pressure	No damage of pressurized catheter at 46 psi	Device integrity suitable for intended clinical use and met requirements of ISO 10555-1
Fluid Leakage at > 46 psi	No liquid leakage from hub and catheter shaft at 46 psi for 30 seconds	Device integrity suitable for intended clinical use and met requirements of ISO 10555-1
Air Leakage	No air leakage at hub into syringe for 15 seconds	Device integrity suitable for intended clinical use and met requirements of ISO 10555-1
Dynamic Burst	Test articles did not burst at or below 300 psi	Device met labeled maximum infusion pressure of 300 psi
Particulate Test	Less than 25 particles greater than 10 microns per ml volume and less than 3 particles less than 25 microns per ml volume No particles greater than 70 microns	Device met specifications for maximum allowable particles

Biocompatibility		
Test	Results	Conclusions
Cytotoxicity – MEM Elution Assay (ISO 10993-5)	Cell culture treated with test article exhibited slight reactivity (Grade 1)	Non-cytotoxic
Sensitization/Irritation – Kligman Maximization Test (ISO 10993-10)	Extracts of the test article elicited no reaction at the challenge (0% sensitization) following the induction phase (Grade 1).	Weak allergic potential or sensitizing capacity
Sensitization/Irritation – Intracutaneous Injection Test (ISO 10993-10)	Extracts of the test article did not show a significantly greater biological reaction than the sites injected with the control article	Non-irritant
Hemocompatibility – Rabbit Blood Direct and Indirect Contact (ISO 10993-4)	The hemolysis index was 0.13% (direct contact) and 0.0% (indirect contact)	Non-hemolytic

Biocompatibility		
Test	Results	Conclusions
Hemocompatibility – Unactivated Partial Thromboplastin Time Assay Direct Contact (ISO 10993-4)	No statistically significant difference found between the Unactivated Partial Thromboplastin Time (UPTT) of the plasma exposed to the test article and that of the plasma exposed to either the negative control or the untreated control	No effect on coagulation
Hemocompatibility – Complement Activation Assay (ISO 10993-4)	C3a and SC5b-9 levels \leq negative and untreated controls	No effect on complement activation
Hemocompatibility – Thrombogenicity Study in Dogs (ISO 10993-4)	Minimal thrombosis observed with a Grade 0 in two out of two test sites and two out of two control sites	No significant thrombosis
Systemic Toxicity – Systemic Injection Test (ISO 10993-11)	Extracts of test article did not induce a significantly greater biological reaction than the control extracts when injected in Swiss Albino mice	No toxic effects
Systemic Toxicity - Rabbit Pyrogen Test (ISO 10993-11)	The temperature increases (maximum) was 0.03°C from baseline	Non-pyrogenic

Summary of Substantial Equivalence:

The data presented in this submission demonstrates the technological similarity and equivalency of the SOFIA Distal Access Catheter when compared with the predicate Chaperon Guiding Catheter (K082385) and the HEADWAY DUO Microcatheter (K120917) devices. The devices:

- Have an equivalent intended use,
- Use the same operating principle,
- Incorporate the same basic design,
- Are packaged and sterilized using the same methods.

In summary, the SOFIA Distal Access Catheter described in this submission is substantially equivalent to the predicate devices.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

November 22, 2013

Micro Vention, Inc.
c/o Ms. Naomi Gong
Sr. Regulatory Affairs Project Manager
1311 Valencia Avenue
Tustin, CA 92780

Re: K131482

Trade/Device Name: SOFIA Distal Access Catheter
Regulation Number: 21 CFR 870.1250
Regulation Name: Percutaneous Catheter
Regulatory Class: Class II
Product Code: DQY, DQO
Dated: October 22, 2013
Received: October 23, 2013

Dear Ms. Gong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA).

You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Ms. Naomi Gong

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Joyce M. Whang -S

for Victor Krauthamer, Ph.D.
Acting Director
Division of Neurological
and Physical Medicine Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K131482

Device Name: SOFIA Distal Access Catheter

Indications For Use:

The SOFIA Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of Center for Devices and Radiological Health (CDRH)

Joyce M. Whang -S

Page 1 of 1



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

November 22, 2013

Micro Vention, Inc.
c/o Ms. Naomi Gong
Sr. Regulatory Affairs Project Manager
1311 Valencia Avenue
Tustin, CA 92780

Re: K131482

Trade/Device Name: SOFIA Distal Access Catheter
Regulation Number: 21 CFR 870.1250
Regulation Name: Percutaneous Catheter
Regulatory Class: Class II
Product Code: DQY, DQO
Dated: October 22, 2013
Received: October 23, 2013

Dear Ms. Gong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA).

You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Ms. Naomi Gong

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Joyce M. Whang -S

for Victor Krauthamer, Ph.D.
Acting Director
Division of Neurological
and Physical Medicine Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Page 3 – Ms. Naomi Gong

Concurrence & Template History Page

[THIS PAGE IS INCLUDED IN IMAGE COPY ONLY]

Full Submission Number: K131482/S002

For Office of Compliance Contact Information:

http://insideportlets.fda.gov:9010/portal/page?_pageid=197,415881&_dad=portal&_schema=PORTAL&org=318

For Office of Surveillance and Biometrics Contact Information:

http://insideportlets.fda.gov:9010/portal/page?_pageid=197,415881&_dad=portal&_schema=PORTAL&org=423

Digital Signature Concurrence Table	
Reviewer Sign-Off	Samuel K. Shimp III 11/13/2013 @ 3:08PM
Branch Chief Sign-Off	Quynh Hoang November 21, 2013 DNPMD/NNDB
Division Sign-Off	Joyce M. Whang-S 2013.11.22 12:18:11 -05'00'

Template Name: K1(A) – SE after 1996

Template History:

Date of Update	By	Description of Update
7/27/09	Brandi Stuart	Added Updates to Boiler Table
8/7/09	Brandi Stuart	Updated HFZ Table
1/11/10	Diane Garcia	Liability/Warranty sentence added at bottom of 1 st page
10/4/11	M. McCabe Janicki	Removed IFU sheet and placed in Forms
9/25/12	Edwena Jones	Added digital signature format
12/12/12	M. McCabe Janicki	Added an extra line between letter signature block and the word "Enclosure". Also, added a missing digit in 4-digit extension on letterhead zip code: "002" should be "0002".
4/2/2013	M. McCabe Janicki	Edited sentence that starts "If you desire specific advice for your device on our labeling regulation (21 CFR Part 801)..." Replaced broken Compliance link with general link to DSMICA.
4/12/2013	Margaret McCabe Janicki	Fixed a typo: Paragraph 1, final sentence, "We remind you, however; that device labeling must be truthful..." Replaced incorrect semicolón with a comma.

Indications for Use

510(k) Number (if known): K131482

Device Name: SOFIA Distal Access Catheter

Indications For Use:

The SOFIA Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of Center for Devices and Radiological Health (CDRH)

Joyce M. Whang -S

Page 1 of 1

K131482

FDA CDRH DMC



MAY 23 2013

Received

U.S. Food and Drug Administration
Center for Devices and Radiologic Health
Document Mail Center (WO66-G609)
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

May 21, 2013

RE: Special 510(k) Notification:	SOFIA Distal Access Catheter
Predicate device:	Chaperon Guiding Catheter (K082385)
Classification:	II
Regulation Number:	21 CFR 870.1250
Product Code:	DQY
Classification Committee:	Cardiovascular

Dear Sir/Madam:

In accordance with Section 510(k) of the Federal Food, Drug and Cosmetic Act as amended by the Medical Device Amendment of 1976, MicroVention, Inc. hereby submits this Special Premarket Notification 510(k) for the SOFIA Distal Access Catheter. The SOFIA is a single lumen catheter designed to be introduced over a steerable guidewire to access the vasculature. The device has been designed, developed and tested according to the FDA special control guidance document: Short-Term and Long-Term Intravascular Catheter and the ISO 10555-1.

The SOFIA Distal Access Catheter utilizes the same fundamental scientific technology, basic design, operating principle and intended use as the predicate device, Chaperon Guiding Catheter (K082385). We believe the documentation included in this 510(k) submission supports our conclusion that it is substantially equivalent to the predicate devices.

An eCopy can be found on the included CD and it is an exact duplicate of the paper copy.

Statement of Confidentiality: MicroVention, Inc. considers the information in this submission to be confidential commercial information. We have not, to our knowledge, released this information through advertising or any other manner to anyone outside the employ of MicroVention, Inc. We ask that this notification and proprietary information herein be treated as confidential in accordance with the Freedom of Information Act.

Thank you in advance for your consideration of our application. If there are any questions, please contact me at (714) 247-8055 or (949) 282-3742

Naomi Gong
Sr. Regulatory Affairs Project Manager
Tel: (714) 247-8055 or (949) 282-3742
Fax: (714) 247-8014
naomi.gong@microvention.com

MicroVention, Inc. : 1311 Valencia Avenue : Tustin : CA 92780 : Main 714.247.8000 : microvention.com

35



U.S. Food and Drug Administration
Center for Devices and Radiologic Health
Document Mail Center **(WO66-G609)**
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

May 21, 2013

RE: **Special 510(k) Notification:** SOFIA Distal Access Catheter
Predicate device: Chaperon Guiding Catheter (K082385)
Classification: II
Regulation Number: 21 CFR 870.1250
Product Code: DQY
Classification Committee: Cardiovascular

Dear Sir/Madam:

In accordance with Section 510(k) of the Federal Food, Drug and Cosmetic Act as amended by the Medical Device Amendment of 1976, MicroVention, Inc. hereby submits this Special Premarket Notification 510(k) for the SOFIA Distal Access Catheter. The SOFIA is a single lumen catheter designed to be introduced over a steerable guidewire to access the vasculature. The device has been designed, developed and tested according to the FDA special control guidance document: Short-Term and Long-Term Intravascular Catheter and the ISO 10555-1.

The SOFIA Distal Access Catheter utilizes the same fundamental scientific technology, basic design, operating principle and intended use as the predicate device, Chaperon Guiding Catheter (K082385). We believe the documentation included in this 510(k) submission supports our conclusion that it is substantially equivalent to the predicate devices.

An eCopy can be found on the included CD and it is an exact duplicate of the paper copy.

Statement of Confidentiality: MicroVention, Inc. considers the information in this submission to be confidential commercial information. We have not, to our knowledge, released this information through advertising or any other manner to anyone outside the employ of MicroVention, Inc. We ask that this notification and proprietary information herein be treated as confidential in accordance with the Freedom of Information Act.

Thank you in advance for your consideration of our application. If there are any questions, please contact me at (714) 247-8055 or (949) 282-3742

Naomi Gong
Sr. Regulatory Affairs Project Manager
Tel: (714) 247-8055 or (949) 282-3742
Fax: (714) 247-8014
naomi.gong@microvention.com

Table of Contents

1.	FDA Forms	4
1.1.	Medical Device User Fee Cover Sheet	4
1.2.	CDRH Submission Coversheet FDA 3514	6
1.3.	Truthful and Accuracy Statement	12
1.4.	510(k) Summary	13
1.5.	Indication for Use.....	17
1.6.	Form FDA 3674	19
1.7.	Form FDA 3654	23
1.8.	Declaration of Conformity.....	42
1.9.	Design Control Activities Summary	43
2.	Executive Summary	46
3.	Device Name	47
4.	Address and Registration Number	47
5.	Device Class	47
6.	Predicate Device Information	48
7.	Labeling and Intended Use.....	48
8.	Device Overview	48
9.	Device Configurations and Dimensions.....	48
10.	Design Description	49
11.	Technological Characteristics Comparison	51
12.	Design Control and Risk Management Processes	51
13.	List of Voluntary Standards.....	52
14.	<i>In-Vitro</i> /Bench Verification and Shelf Life	52
14.1.	Simulated Use	55
14.2.	Dimensional & Physical Attributes.....	57
14.3.	Kink Resistance	57
14.4.	Tip Shapeability.....	59
14.5.	Radio-detectibility	59
14.6.	Hub Evaluation (ISO 594-2).....	59
14.7.	Catheter Stiffness	60
14.8.	Torque Strength Testing.....	60
14.9.	Catheter Flexural Fatigue.....	61
14.10.	Surface Contamination	61
14.11.	Force at Break	61
14.12.	Flow Rate	62

14.13. Static Burst Pressure..... 62

14.14. Freedom from Leakage – Liquid 62

14.15. Freedom from Leakage - Air 63

14.16. Dynamic Burst Pressure 63

14.17. Particulate Measurement Analysis 63

15. Biocompatibility 65

16. Packaging and Sterilization..... 66

17. Substantial Equivalence..... 67

18. ISO/EC Certification and Compliance..... 67

19. List of Attachments 68

1. FDA Forms

1.1. Medical Device User Fee Cover Sheet

DEPARTMENT OF HEALTH AND HUMAN SERVICES FOOD AND DRUG ADMINISTRATION MEDICAL DEVICE USER FEE COVER SHEET		PAYMENT IDENTIFICATION NUMBER: (b) (4) Write the Payment Identification number on
A completed cover sheet must accompany each original application or supplement subject to fees. If payment is sent by U.S. mail or courier, please include a copy of this completed form with payment. Payment and mailing instructions can be found at: http://www.fda.gov/oc/mdufma/coversheet.html		
1. COMPANY NAME AND ADDRESS (include name, street address, city state, country, and post office code) MICRO VENTION INC 1311 Valencia Avenue Tustin CA 92780 US 1.1 EMPLOYER IDENTIFICATION NUMBER (EIN) *****3774	2. CONTACT NAME Vin Cutarelli 2.1 E-MAIL ADDRESS vin.cutarelli@microvention.com 2.2 TELEPHONE NUMBER (include Area code) 714-247-8181 2.3 FACSIMILE (FAX) NUMBER (Include Area code) 714-247-8014	
3. TYPE OF PREMARKET APPLICATION (Select one of the following in each column; if you are unsure, please refer to the application descriptions at the following web site: http://www.fda.gov/oc/mdufma) <u>Select an application type:</u> <input checked="" type="checkbox"/> Premarket notification(510(k)); except for third party <input type="checkbox"/> 513(g) Request for Information <input type="checkbox"/> Biologics License Application (BLA) <input type="checkbox"/> Premarket Approval Application (PMA) <input type="checkbox"/> Modular PMA <input type="checkbox"/> Product Development Protocol (PDP) <input type="checkbox"/> Premarket Report (PMR) <input type="checkbox"/> Annual Fee for Periodic Reporting (APR) <input type="checkbox"/> 30-Day Notice		
3.1 Select a center <input checked="" type="checkbox"/> CDRH <input type="checkbox"/> CBER 3.2 Select one of the types below <input checked="" type="checkbox"/> Original Application <u>Supplement Types:</u> <input type="checkbox"/> Efficacy (BLA) <input type="checkbox"/> Panel Track (PMA, PMR, PDP) <input type="checkbox"/> Real-Time (PMA, PMR, PDP) <input type="checkbox"/> 180-day (PMA, PMR, PDP)		
4. ARE YOU A SMALL BUSINESS? (See the instructions for more information on determining this status) <input type="checkbox"/> YES, I meet the small business criteria and have submitted the required qualifying documents to FDA <input checked="" type="checkbox"/> NO, I am not a small business 4.1 If Yes, please enter your Small Business Decision Number:		
5. FDA WILL NOT ACCEPT YOUR SUBMISSION IF YOUR COMPANY HAS NOT PAID AN ESTABLISHMENT REGISTRATION FEE THAT IS DUE TO FDA. HAS YOUR COMPANY PAID ALL ESTABLISHMENT REGISTRATION FEES THAT ARE DUE TO FDA? <input checked="" type="checkbox"/> YES (All of our establishments have registered and paid the fee, or this is our first device, and we will register and pay the fee within 30 days of FDA's approval/clearance of this device.) <input type="checkbox"/> NO (If "NO," FDA will not accept your submission until you have paid all fees due to FDA. This submission will not be processed; see http://www.fda.gov/cdrh/mdufma for additional information)		
6. IS THIS PREMARKET APPLICATION COVERED BY ANY OF THE FOLLOWING USER FEE EXCEPTIONS? IF SO, CHECK THE APPLICABLE EXCEPTION. <input type="checkbox"/> This application is the first PMA submitted by a qualified small business, including any affiliates <input type="checkbox"/> This biologics application is submitted under section 351 of the Public Health Service Act for a product licensed for further manufacturing use only <input type="checkbox"/> The sole purpose of the application is to support conditions of use for a pediatric population <input type="checkbox"/> The application is submitted by a state or federal government entity for a device that is not to be distributed commercially		
7. IS THIS A SUPPLEMENT TO A PREMARKET APPLICATION FOR WHICH FEES WERE WAIVED DUE TO SOLE USE IN A PEDIATRIC POPULATION THAT NOW PROPOSES CONDITION OF USE FOR ANY ADULT POPULATION? (If so, the application is subject to the fee that applies for an original premarket approval application (PMA).) <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO		
PAPERWORK REDUCTION ACT STATEMENT Public reporting burden for this collection of information is estimated to average 18 minutes per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to the address below. Department of Health and Human Services, Food and Drug Administration, Office of Chief Information Officer, 1350 Piccard Drive, 4th Floor Rockville, MD 20850 [Please do NOT return this form to the above address, except as it pertains to comments on the burden estimate.]		
8. USER FEE PAYMENT AMOUNT SUBMITTED FOR THIS PREMARKET APPLICATION (b) (4)		12-Dec-2012

1.2. CDRH Submission Coversheet FDA 3514

CDRH PREMARKET REVIEW SUBMISSION COVER SHEET

Date of Submission 05/21/2013	User Fee Payment ID Number (b) (4)	FDA Submission Document Number (if known)
----------------------------------	--	---

SECTION A TYPE OF SUBMISSION

PMA <input type="checkbox"/> Original Submission <input type="checkbox"/> Premarket Report <input type="checkbox"/> Modular Submission <input type="checkbox"/> Amendment <input type="checkbox"/> Report <input type="checkbox"/> Report Amendment <input type="checkbox"/> Licensing Agreement	PMA & HDE Supplement <input type="checkbox"/> Regular (180 day) <input type="checkbox"/> Special <input type="checkbox"/> Panel Track (PMA Only) <input type="checkbox"/> 30-day Supplement <input type="checkbox"/> 30-day Notice <input type="checkbox"/> 135-day Supplement <input type="checkbox"/> Real-time Review <input type="checkbox"/> Amendment to PMA & HDE Supplement <input type="checkbox"/> Other	PDP <input type="checkbox"/> Original PDP <input type="checkbox"/> Notice of Completion <input type="checkbox"/> Amendment to PDP	510(k) <input checked="" type="checkbox"/> Original Submission: <input type="checkbox"/> Traditional <input checked="" type="checkbox"/> Special <input type="checkbox"/> Abbreviated (Complete section I, Page 5) <input type="checkbox"/> Additional Information <input type="checkbox"/> Third Party	Request for Feedback <input type="checkbox"/> Pre-Submission <input type="checkbox"/> Informational Meeting <input type="checkbox"/> Submission Issue Meeting <input type="checkbox"/> Day 100 Meeting <input type="checkbox"/> Agreement Meeting <input type="checkbox"/> Determination Meeting <input type="checkbox"/> Study Risk Determination <input type="checkbox"/> Other (specify):
IDE <input type="checkbox"/> Original Submission <input type="checkbox"/> Amendment <input type="checkbox"/> Supplement	Humanitarian Device Exemption (HDE) <input type="checkbox"/> Original Submission <input type="checkbox"/> Amendment <input type="checkbox"/> Supplement <input type="checkbox"/> Report <input type="checkbox"/> Report Amendment	Class II Exemption Petition <input type="checkbox"/> Original Submission <input type="checkbox"/> Additional Information	Evaluation of Automatic Class III Designation (De Novo) <input type="checkbox"/> Original Submission <input type="checkbox"/> Additional Information	Other Submission <input type="checkbox"/> 513(g) <input type="checkbox"/> Other (describe submission):

Have you used or cited Standards in your submission? Yes No (If Yes, please complete Section I, Page 5)

SECTION B SUBMITTER, APPLICANT OR SPONSOR

Company / Institution Name MicroVention, Inc.		Establishment Registration Number (if known) 2032493	
Division Name (if applicable)		Phone Number (including area code) 714-247-8055	
Street Address 1311 Valencia Avenue		FAX Number (including area code) 714-247-8014	
City Tustin	State / Province CA	ZIP/Postal Code 92780	Country USA
Contact Name Naomi Gong			
Contact Title Sr. Regulatory Affairs Project Manager		Contact E-mail Address naomi.gong@microvention.com	

SECTION C APPLICATION CORRESPONDENT (e.g., consultant, if different from above)

Company / Institution Name			
Division Name (if applicable)		Phone Number (including area code)	
Street Address		FAX Number (including area code)	
City	State / Province	ZIP Code	Country
Contact Name			
Contact Title		Contact E-mail Address	

SECTION D1

REASON FOR APPLICATION - PMA, PDP, OR HDE

<input type="checkbox"/> New Device <input type="checkbox"/> Withdrawal <input type="checkbox"/> Additional or Expanded Indications <input type="checkbox"/> Request for Extension <input type="checkbox"/> Post-approval Study Protocol <input type="checkbox"/> Request for Applicant Hold <input type="checkbox"/> Request for Removal of Applicant Hold <input type="checkbox"/> Request to Remove or Add Manufacturing Site	<input type="checkbox"/> Change in design, component, or specification: <input type="checkbox"/> Software / Hardware <input type="checkbox"/> Color Additive <input type="checkbox"/> Material <input type="checkbox"/> Specifications <input type="checkbox"/> Other (<i>specify below</i>)	<input type="checkbox"/> Location change: <input type="checkbox"/> Manufacturer <input type="checkbox"/> Sterilizer <input type="checkbox"/> Packager
<input type="checkbox"/> Process change: <input type="checkbox"/> Manufacturing <input type="checkbox"/> Packaging <input type="checkbox"/> Sterilization <input type="checkbox"/> Other (<i>specify below</i>)	<input type="checkbox"/> Labeling change: <input type="checkbox"/> Indications <input type="checkbox"/> Instructions <input type="checkbox"/> Performance Characteristics <input type="checkbox"/> Shelf Life <input type="checkbox"/> Trade Name <input type="checkbox"/> Other (<i>specify below</i>)	<input type="checkbox"/> Report Submission: <input type="checkbox"/> Annual or Periodic <input type="checkbox"/> Post-approval Study <input type="checkbox"/> Adverse Reaction <input type="checkbox"/> Device Defect <input type="checkbox"/> Amendment
<input type="checkbox"/> Response to FDA correspondence:		<input type="checkbox"/> Change in Ownership <input type="checkbox"/> Change in Correspondent <input type="checkbox"/> Change of Applicant Address

Other Reason (*specify*):

SECTION D2

REASON FOR APPLICATION - IDE

<input type="checkbox"/> New Device <input type="checkbox"/> New Indication <input type="checkbox"/> Addition of Institution <input type="checkbox"/> Expansion / Extension of Study <input type="checkbox"/> IRB Certification <input type="checkbox"/> Termination of Study <input type="checkbox"/> Withdrawal of Application <input type="checkbox"/> Unanticipated Adverse Effect <input type="checkbox"/> Notification of Emergency Use <input type="checkbox"/> Compassionate Use Request <input type="checkbox"/> Treatment IDE <input type="checkbox"/> Continued Access	<input type="checkbox"/> Change in: <input type="checkbox"/> Correspondent / Applicant <input type="checkbox"/> Design / Device <input type="checkbox"/> Informed Consent <input type="checkbox"/> Manufacturer <input type="checkbox"/> Manufacturing Process <input type="checkbox"/> Protocol - Feasibility <input type="checkbox"/> Protocol - Other <input type="checkbox"/> Sponsor	<input type="checkbox"/> Response to FDA Letter Concerning: <input type="checkbox"/> Conditional Approval <input type="checkbox"/> Deemed Approved <input type="checkbox"/> Deficient Final Report <input type="checkbox"/> Deficient Progress Report <input type="checkbox"/> Deficient Investigator Report <input type="checkbox"/> Disapproval <input type="checkbox"/> Request Extension of Time to Respond to FDA <input type="checkbox"/> Request Meeting <input type="checkbox"/> Request Hearing
<input type="checkbox"/> Report submission: <input type="checkbox"/> Current Investigator <input type="checkbox"/> Annual Progress Report <input type="checkbox"/> Site Waiver Report <input type="checkbox"/> Final		

Other Reason (*specify*):

SECTION D3

REASON FOR SUBMISSION - 510(k)

<input type="checkbox"/> New Device	<input type="checkbox"/> Additional or Expanded Indications	<input type="checkbox"/> Change in Technology
-------------------------------------	---	---

Other Reason (*specify*):

Catheter is similar to predicate device with same intended use (indications for use)

SECTION E ADDITIONAL INFORMATION ON 510(k) SUBMISSIONS

Product codes of devices to which substantial equivalence is claimed								Summary of, or statement concerning, safety and effectiveness information <input checked="" type="checkbox"/> 510 (k) summary attached <input type="checkbox"/> 510 (k) statement
1	DQY	2		3		4		
5		6		7		8		

Information on devices to which substantial equivalence is claimed (if known)

	510(k) Number	Trade or Proprietary or Model Name	Manufacturer
1	K082385	Chaperon Guiding Catheter	MicroVention, Inc. 1311 Valencia Avenue, Tustin, CA 92780
2			
3			
4			
5			
6			

SECTION F PRODUCT INFORMATION - APPLICATION TO ALL APPLICATIONS

Common or usual name or classification name
Percutaneous catheter

	Trade or Proprietary or Model Name for This Device	Model Number
1	SOFIA Distal Access Catheter	DA5125ST
2		
3		
4		
5		

FDA document numbers of all prior related submissions (regardless of outcome)

1	2	3	4	5	6
7	8	9	10	11	12

Data Included in Submission
 Laboratory Testing Animal Trials Human Trials

SECTION G PRODUCT CLASSIFICATION - APPLICATION TO ALL APPLICATIONS

Product Code DQY	C.F.R. Section (if applicable) 21 CFR 870.1250	Device Class <input type="checkbox"/> Class I <input checked="" type="checkbox"/> Class II <input type="checkbox"/> Class III <input type="checkbox"/> Unclassified
Classification Panel Cardiovascular Devices		

Indications (from labeling)
The SOFIA Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

SECTION H MANUFACTURING / PACKAGING / STERILIZATION SITES RELATING TO A SUBMISSION

<input checked="" type="checkbox"/> Original <input type="checkbox"/> Add <input type="checkbox"/> Delete		Facility Establishment Identifier (FEI) Number 2032493		<input checked="" type="checkbox"/> Manufacturer <input type="checkbox"/> Contract Sterilizer <input type="checkbox"/> Contract Manufacturer <input type="checkbox"/> Repackager / Relabeler	
Company / Institution Name MicroVention, Inc.			Establishment Registration Number 2032493		
Division Name (if applicable)			Phone Number (including area code) 714-247-8000		
Street Address 1311 Valencia Avenue			FAX Number (including area code) 714-247-8005		
City Tustin		State / Province CA	ZIP Code 92780	Country USA	
Contact Name Naomi Gong		Contact Title Sr. RA Project Manager		Contact E-mail Address naomi.gong@microvention.com	

<input checked="" type="checkbox"/> Original <input type="checkbox"/> Add <input type="checkbox"/> Delete		Facility Establishment Identifier (FEI) Number		<input type="checkbox"/> Manufacturer <input checked="" type="checkbox"/> Contract Sterilizer <input type="checkbox"/> Contract Manufacturer <input type="checkbox"/> Repackager / Relabeler	
Company / Institution Name Sterigenics			Establishment Registration Number 2011171		
Division Name (if applicable)			Phone Number (including area code) 951-340-0700		
Street Address 4900 South Griffith Avenue			FAX Number (including area code)		
City Los Angeles		State / Province CA	ZIP Code 90058	Country USA	
Contact Name Sharon Huges		Contact Title Representative		Contact E-mail Address losangelessales@sterigenics.com	

<input type="checkbox"/> Original <input type="checkbox"/> Add <input type="checkbox"/> Delete		Facility Establishment Identifier (FEI) Number		<input type="checkbox"/> Manufacturer <input type="checkbox"/> Contract Sterilizer <input type="checkbox"/> Contract Manufacturer <input type="checkbox"/> Repackager / Relabeler	
Company / Institution Name			Establishment Registration Number		
Division Name (if applicable)			Phone Number (including area code)		
Street Address			FAX Number (including area code)		
City		State / Province	ZIP Code	Country	
Contact Name		Contact Title		Contact E-mail Address	

SECTION I

UTILIZATION OF STANDARDS

Note: Complete this section if your application or submission cites standards or includes a "Declaration of Conformity to a Recognized Standard" statement.

	Standards No.	Standards Organization	Standards Title	Version	Date
1					
2					
3					
4					
5					
6					
7					

Please include any additional standards to be cited on a separate page.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF ADDRESS BELOW.

The burden time for this collection of information is estimated to average 0.5 hour per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
 Food and Drug Administration
 Office of Chief Information Officer
 Paperwork Reduction Act (PRA) Staff
 1350 Piccard Drive, Room 400
 Rockville, MD 20850

An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.

1.3. Truthful and Accuracy Statement

[As Required by 21 CFR 807.87(k)]

I certify that, in my capacity as Sr. Regulatory Affairs Project Manager of MicroVention, Inc., I believe to the best of my knowledge, that all data and information submitted in the premarket notification are truthful and accurate and that no material fact has been omitted.

(Signature)

Naomi Gong _____
(Typed Name)

(Date)

1.4. 510(k) Summary

510(k) Summary

Trade Name: SOFIA Distal Access Catheter
Generic Name: Percutaneous Catheter
Classification: Class II, 21 CFR 870.1250 (DQY)
Submitted By: MicroVention, Inc
 1311 Valencia Avenue
 Tustin, California U.S.A.
Contact: Naomi Gong
Date: 2013 May 21
Predicate Device: Chaperon Guiding Catheter (K082385)

Device Description:

The SOFIA™ Distal Access Catheter is a single-lumen, flexible catheter designed with coil and braid reinforcement. The distal segment is steam-shapeable and it has a hydrophilic coating for navigation through the vasculature. The radiopaque marker is located at the distal end of the catheter for visualization under fluoroscopy. An introducer sheath and shaping mandrel are also provided.

Indications For Use:

The SOFIA™ Distal Access Catheter is indicated for general intravascular use, including the neuro and peripheral vasculature. The SOFIA™ Distal Access Catheter can be used to facilitate introduction of diagnostic or therapeutic devices. The SOFIA™ Distal Access Catheter is not intended for use in coronary arteries.

Verification and Test Summary:

Bench Testing	Result
Surface and physical attributes	Pass
Distal tensile strength	Pass
Hub tensile strength	Pass
Hub test (ISO 594-2)	Pass
Leakage (liquid and air)	Pass
Static and dynamic burst pressure	Pass
Simulated use	Pass
Durability/lubricity of hydrophilic coating	Pass
Compatibility with devices	Pass
Flow rate	Pass
Kink resistance	Pass
Catheter stiffness	Pass
Radio-detectability	Pass
Catheter flexural fatigue	Pass

Torque strength	Pass
Particulate test	Pass
Tip shapeability	Pass
Biocompatibility	Result
Cytotoxicity (ISO 10993-5) - MEM elution assay - Agarose overlay	Pass
Sensitization/Irritation (ISO 10993-10) - Guinea pig maximization sensitization - Intracutaneous reactivity	Pass
Hemocompatibility (ISO 10993-4) - Hemolysis - Prothrombin time assay - Complement activation C3a and SC5b-9 - 4 hour thromboresistance in dogs	Pass
Systemic Toxicity (ISO 10993-11) - Systemic toxicity - Rabbit pyrogen test	Pass

Technological Comparison:

	Chaperon Guiding Catheter	SOFIA Distal Access Catheter
Intended Use	Intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.	Same
Material	Outer layer of polyester elastomer, stainless steel braid, inner liner of PTFE (polytetrafluoroethylene). Tungsten radiopaque marker, nylon hub, and polyester elastomer strain relief.	Outer layer of polyolefin elastomer, polyurethane elastomer, polyether block amide; inner layer of PTFE, stainless steel braid and coil. Pt/Ir radiopaque marker, nylon hub, polyurethane strain relief. Pebax introducer sheath and stainless steel shaping mandrel.
Catheter size	5F (Outer catheter)	Same
ID	1.5 mm	1.4 mm
OD	1.7 mm	Same
Effective Length	Outer: 95 cm Inner: 117 cm	125 cm
Coating	Yes	Same
Tip Configuration	Preshaped	Steam shapeable by user
Guidewire compatibility	0.035" or 0.038"	Same
Accessories	N/A	Introducer sheath and shaping mandrel

Summary of Substantial Equivalence:

The data presented in this submission demonstrates the technological similarity and equivalency of the SOFIA Distal Access Catheter when compared with the predicate device, Chaperon Guiding Catheter (K082385).

The devices,

- Have the same intended use,
- Use the same operating principle,
- Incorporate the same basic design,
- Are packaged and sterilized using same methods.

In summary, the SOFIA Distal Access Catheter described in this submission is substantially equivalent to the predicate device.

1.5. Indications for Use

Indications for Use

510(k) Number (if known): _____

Device Name: SOFIA™ Distal Access Catheter

Indications For Use:

The SOFIA Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

Prescription Use X AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

1.6. Form FDA 3674



DEPARTMENT OF HEALTH AND HUMAN SERVICES

Records processed under FOIA Request #2015-6024; Released by CDRH on 09-26-2016.
FOOD AND DRUG ADMINISTRATION

Certification of Compliance, under 42 U.S.C. § 282(j)(5)(B), with Requirements of ClinicalTrials.gov Data Bank (42 U.S.C. § 282(j))

Form Approved: OMB No. 0910-0616
Expiration Date: 06-30-2008
See OMB Statement on Reverse

(For submission with an application/submission, including amendments, supplements, and resubmissions, under §§ 505, 515, 520(m), or 510(k) of the Federal Food, Drug, and Cosmetic Act or § 351 of the Public Health Service Act.)

SPONSOR/APPLICANT/SUBMITTER INFORMATION

1. NAME OF SPONSOR/APPLICANT/SUBMITTER: Naomi Gong
2. DATE OF THE APPLICATION/SUBMISSION WHICH THIS CERTIFICATION ACCOMPANIES: May 21, 2013
3. ADDRESS (Number, Street, State, and Zip Code): MicroVention, Inc. 1311 Valencia Avenue Tustin, CA 92780
4. TELEPHONE AND FAX NUMBER (Include Area Code): (T) +1 (714) 247-8005 (F) +1 (714) 247-8014

PRODUCT INFORMATION

5. FOR DRUGS/BIOLOGICS: Include Any/All Available Established, Proprietary and/or Chemical/Biochemical/Blood/Cellular/Gene Therapy Product Name(s)
FOR DEVICES: Include Any/All Common or Usual Name(s), Classification, Trade or Proprietary or Model Name(s) and/or Model Number(s)
(Attach extra pages as necessary)
SOFIA Distal Access Catheter

APPLICATION/SUBMISSION INFORMATION

6. TYPE OF APPLICATION/SUBMISSION WHICH THIS CERTIFICATION ACCOMPANIES
[] IND [] NDA [] ANDA [] BLA [] PMA [] HDE [X] 510(k) [] PDP [] Other
7. INCLUDE IND/NDA/ANDA/BLA/PMA/HDE/510(k)/PDP/OTHER NUMBER (If number previously assigned)
8. SERIAL NUMBER ASSIGNED TO APPLICATION/SUBMISSION WHICH THIS CERTIFICATION ACCOMPANIES

CERTIFICATION STATEMENT/INFORMATION

9. CHECK ONLY ONE OF THE FOLLOWING BOXES
(See instructions for additional information and explanation)
[X] A. I certify that the requirements of 42 U.S.C. § 282(j), Section 402(j) of the Public Health Service Act, enacted by 121 Stat. 823, Public Law 110-85, do not apply because the application/submission which this certification accompanies does not reference any clinical trial.
[] B. I certify that the requirements of 42 U.S.C. § 282(j), Section 402(j) of the Public Health Service Act, enacted by 121 Stat. 823, Public Law 110-85, do not apply to any clinical trial referenced in the application/submission which this certification accompanies.
[] C. I certify that the requirements of 42 U.S.C. § 282(j), Section 402(j) of the Public Health Service Act, enacted by 121 Stat. 823, Public Law 110-85, apply to one or more of the clinical trials referenced in the application/submission which this certification accompanies and that those requirements have been met.
10. IF YOU CHECKED BOX C, IN # 9, PROVIDE THE NATIONAL CLINICAL TRIAL (NCT) NUMBER(S) FOR ANY "APPLICABLE CLINICAL TRIAL(S)," UNDER 42 U.S.C. § 282(j)(1)(A)(i), SECTION 402(j)(1)(A)(i) OF THE PUBLIC HEALTH SERVICE ACT, REFERENCED IN THE APPLICATION/SUBMISSION WHICH THIS CERTIFICATION ACCOMPANIES
(Attach extra pages as necessary)
NCT Number(s)

The undersigned declares, to the best of my knowledge, information, and belief, that the information furnished herein is true and correct. I understand that the failure to submit the certification required by 42 U.S.C. § 282(j)(5)(B), section 402(j)(5)(B) of the Public Health Service Act, and the knowing submission of a false certification under such section are prohibited acts under 21 U.S.C. § 331, section 301 of the Federal Food, Drug, and Cosmetic Act.

Warning: A willfully and knowingly false statement is a criminal offense, U.S. Code, title 18, section 1001.

11. SIGNATURE OF SPONSOR/APPLICANT/SUBMITTER OR AN AUTHORIZED REPRESENTATIVE (SIGN) _____	12. NAME AND TITLE OF THE PERSON WHO SIGNED IN #11 Naomi Gong Sr. Regulatory Affairs Project Manager
13. ADDRESS (Number, Street, State, and Zip Code) (of person identified in #11 & 12) 1311 Valencia Avenue Tustin, CA 92780	14. TELEPHONE AND FAX NUMBER (Include Area Code) (T) +1 (714) 247-8055 (F) +1 (714) 247-8014

15. DATE OF CERTIFICATION May 21, 2013

Paperwork Reduction Act Statement

Public Reporting Burden for this collection of information is estimated to average 15 minutes and 45 minutes (depending on the type of application/submission) per response, including time for reviewing instructions. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to the applicable address below.

Food and Drug Administration Center for Drug Evaluation and Research Central Document Room Form No. FDA 3674 5901-B Ammendale Road Beltsville, MD 20705-1266	Food and Drug Administration Center for Biologics Evaluation and Research 1401 Rockville Pike Rockville, MD 20852-1448	Food and Drug Administration Center for Devices and Radiological Health Program Operations Staff (HFZ-403) 9200 Corporate Blvd. Rockville, MD 20850
---	---	---

An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information, unless it displays a currently valid OMB control number.

Certification of Compliance, under 42 U.S.C. § 282(j)(5)(B), with Requirements of ClinicalTrials.gov Data Bank (42 U.S.C. § 282(j))
Form 3674 must accompany an application/submission, including amendments, supplements, and resubmissions, submitted under §§ 505, 515, 520(m), or 510(k) of the Federal Food, Drug, and Cosmetic Act or § 351 of the Public Health Service Act.

- 1. Name of Sponsor/Applicant/Submitter** - This is the name of the sponsor/applicant/submitter of the drug/biologic/device application/submission which the certification accompanies. The name must be identical to that listed on the application/submission.
- 2. Date** - This is the date of the application/submission which the certification accompanies.
- 3. & 4.** - Provide complete address, telephone number and fax number of the sponsor/applicant/submitter.
- 5. Product Information** - For Drugs/Biologics: Provide the established, proprietary name, and/or chemical/biochemical/blood product/cellular/gene therapy name(s) for the product covered by the application/ submission. Include all available names by which the product is known. For Devices: Provide the common or usual name, classification, trade or proprietary or model name(s), and/or model number(s). Include all available names/model numbers by which the product is known.
- 6. Type of Application/Submission** - Identify the type of application/submission which the certification accompanies by checking the appropriate box. If the name of the type of application/submission is not identified, check the box labeled "Other."
- 7. IND/NDA/ANDA/BLA/PMA/HDE/510(k)/PDP/Other Number** - If FDA has previously assigned a number associated with the application/ submission which this certification accompanies, list that number in this field. For example, if the application/submission accompanied by this certification is an IND protocol amendment and the IND number has already been issued by FDA, that number should be provided in this field.
- 8. Serial Number** - In some instances a sequential serial number is assigned to the application. If there is such a serial number, provide it in this field.
- 9. Certification** - This section contains three different check-off boxes.

Box A should be checked if the sponsor/applicant/submitter has concluded that the requirements of 42 U.S.C. § 282(j), section 402(j) of the Public Health Service Act, do not apply because no clinical trials are included, relied upon, or otherwise referred to, in the application/submission which the certification accompanies.

Box B should be checked if the sponsor/applicant/submitter has concluded that the requirements of 42 U.S.C. § 282(j), section 402(j) of the Public Health Service Act, do not apply at the time of submission to any clinical trials that are included, relied upon, or otherwise referred to, in the application/submission which the certification accompanies. This means that, at the time the application/submission is being made, the requirements of 42 U.S.C. § 282(j), section 402(j) of the Public Health Service Act, do not apply to any of the clinical trials included, relied upon, or otherwise referred to, in the application/submission which this certification accompanies.

Box C should be checked if the sponsor/applicant/submitter has concluded that the requirements of 42 U.S.C. § 282(j), section 402(j) of the Public Health Service Act, do apply at the time of submission to some or all of the clinical trials that are included, relied upon, or otherwise referred to, in the application/submission which the certification accompanies. This means that, at the time the application/submission is being made, the requirements of 42 U.S.C. § 282(j), section 402(j) of the Public Health Service Act, apply to one or more of the clinical trials included, relied upon, or otherwise referred to, in the application/submission which this certification accompanies.

10. National Clinical Trial (NCT) Numbers - If you have checked Box C in # 9 (Certification), provide the NCT Number obtained from www.ClinicalTrials.gov for each clinical trial that is an "applicable clinical trial" under 42 U.S.C. § 282(j)(1)(A)(i), section 402(j)(1)(A)(i) of the Public Health Service Act, and that is included, relied upon, or otherwise referred to, in the application/submission which the certification accompanies. Type only the number, as NCT will be added automatically before number. Include any and all NCT numbers assigned to the clinical trials included, relied upon, or otherwise referred to, in the application/submission which this certification accompanies. Multiple NCT numbers may be required for a particular certification, depending on the number of "applicable clinical trials" included, relied upon, or otherwise referred to, in the application/submission which the certification accompanies.

11. Signature of Sponsor/Applicant/Submitter or an Authorized Representative - The person signing the certification must sign in this field.

12. Name and Title of Person Who Signed in #11. - Include the name and title of the person who is signing the certification. If the person signing the certification is not the sponsor/applicant/submitter of the application/submission, he or she must be an authorized representative of the sponsor/applicant/submitter.

13. & 14. & 15. - Provide the full address, telephone and fax number of the person who is identified in number 11 and signs the certification in number 12. Provide the date the certification is signed. This date may be different from the date provided in #2.

1.7. Form FDA 3654

Department of Health and Human Services
 Food and Drug Administration
STANDARDS DATA REPORT FOR 510(k)s
 (To be filled in by applicant)

This report and the Summary Report Table are to be completed by the applicant when submitting a 510(k) that references a national or international standard. A separate report is required for each standard referenced in the 510(k).

TYPE OF 510(K) SUBMISSION

Traditional Special Abbreviated

STANDARD TITLE ¹

ISO 10555-1: 2004, Sterile single use intravascular catheters

Please answer the following questions

Yes No

Is this standard recognized by FDA ²?

FDA Recognition number ³ #6-161

Was a third party laboratory responsible for testing conformity of the device to this standard identified in the 510(k)?

Is a summary report ⁴ describing the extent of conformance of the standard used included in the 510(k)?
 If no, complete a summary report table.

Does the test data for this device demonstrate conformity to the requirements of this standard as it pertains to this device?

Does this standard include acceptance criteria?
 If no, include the results of testing in the 510(k).

Does this standard include more than one option or selection of tests?
 If yes, report options selected in the summary report table.

Were there any deviations or adaptations made in the use of the standard?.....
 If yes, were deviations in accordance with the FDA supplemental information sheet (SIS) ⁵?

Were deviations or adaptations made beyond what is specified in the FDA SIS?.....
 If yes, report these deviations or adaptations in the summary report table.

Were there any exclusions from the standard?
 If yes, report these exclusions in the summary report table.

Is there an FDA guidance ⁶ that is associated with this standard?.....
 If yes, was the guidance document followed in preparation of this 510k?

Title of guidance: Guidance on 510k submission for short and long-term intravascular catheter (1995)

¹ The formatting convention for the title is: [SDO] [numeric identifier] [title of standard] [date of publication]

² Authority [21 U.S.C. 360d], <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm>

³ <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁴ The summary report should include: any adaptations used to adapt to the device under review (for example, alternative test methods); choices made when options or a selection of methods are described; deviations from the standard; requirements not applicable to the device; and the name and

address of the test laboratory or certification body involved in conformance assessment to this standard. The summary report includes information on all standards utilized during the development of the device.

⁵ The supplemental information sheet (SIS) is additional information which is necessary before FDA recognizes the standard. Found at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁶ The online search for CDRH Guidance Documents can be found at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/default.htm>

EXTENT OF STANDARD CONFORMANCE SUMMARY REPORT TABLE		
STANDARD TITLE ISO 10555-1: 2004, Sterile single use intravascular catheters		
CONFORMANCE WITH STANDARD SECTIONS*		
SECTION NUMBER All	SECTION TITLE (as applicable)	CONFORMANCE? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
<p>* For completeness list all sections of the standard and indicate whether conformance is met. If a section is not applicable (N/A) an explanation is needed under "justification." Some standards include options, so similar to deviations, the option chosen needs to be described and adequately justified as appropriate for the subject device. Explanation of all deviations or description of options selected when following a standard is required under "type of deviation or option selected," "description" and "justification" on the report. More than one page may be necessary.</p> <p>† Types of deviations can include an exclusion of a section in the standard, a deviation brought out by the FDA supplemental information sheet (SIS), a deviation to adapt the standard to the device, or any adaptation of a section.</p>		
Paperwork Reduction Act Statement		
<p>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to:</p> <p style="margin-left: 40px;">Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer 1350 Piccard Drive, Room 400 Rockville, MD 20850</p> <p style="text-align: right; margin-right: 40px;"><i>An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.</i></p>		

Department of Health and Human Services
 Food and Drug Administration
STANDARDS DATA REPORT FOR 510(k)s
 (To be filled in by applicant)

This report and the Summary Report Table are to be completed by the applicant when submitting a 510(k) that references a national or international standard. A separate report is required for each standard referenced in the 510(k).

TYPE OF 510(K) SUBMISSION

Traditional Special Abbreviated

STANDARD TITLE ¹

ISO 14971:20012 Medical Device - Application of Risk Management to Medical Devices

Please answer the following questions

Yes No

Is this standard recognized by FDA ²?

FDA Recognition number ³ # 5-40

Was a third party laboratory responsible for testing conformity of the device to this standard identified in the 510(k)?

Is a summary report ⁴ describing the extent of conformance of the standard used included in the 510(k)?
 If no, complete a summary report table.

Does the test data for this device demonstrate conformity to the requirements of this standard as it pertains to this device?

Does this standard include acceptance criteria?
 If no, include the results of testing in the 510(k).

Does this standard include more than one option or selection of tests?
 If yes, report options selected in the summary report table.

Were there any deviations or adaptations made in the use of the standard?.....
 If yes, were deviations in accordance with the FDA supplemental information sheet (SIS) ⁵?

Were deviations or adaptations made beyond what is specified in the FDA SIS?.....
 If yes, report these deviations or adaptations in the summary report table.

Were there any exclusions from the standard?
 If yes, report these exclusions in the summary report table.

Is there an FDA guidance ⁶ that is associated with this standard?.....
 If yes, was the guidance document followed in preparation of this 510k?

Title of guidance: _____

¹ The formatting convention for the title is: [SDO] [numeric identifier] [title of standard] [date of publication]

² Authority [21 U.S.C. 360d], <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm>

³ <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁴ The summary report should include: any adaptations used to adapt to the device under review (for example, alternative test methods); choices made when options or a selection of methods are described; deviations from the standard; requirements not applicable to the device; and the name and

address of the test laboratory or certification body involved in conformance assessment to this standard. The summary report includes information on all standards utilized during the development of the device.

⁵ The supplemental information sheet (SIS) is additional information which is necessary before FDA recognizes the standard. Found at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁶ The online search for CDRH Guidance Documents can be found at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/default.htm>

EXTENT OF STANDARD CONFORMANCE SUMMARY REPORT TABLE		
STANDARD TITLE ISO 14971:2012, Medical Device - Application of Risk Management to Medical Devices		
CONFORMANCE WITH STANDARD SECTIONS*		
SECTION NUMBER All	SECTION TITLE (as applicable)	CONFORMANCE? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
<p>* For completeness list all sections of the standard and indicate whether conformance is met. If a section is not applicable (N/A) an explanation is needed under "justification." Some standards include options, so similar to deviations, the option chosen needs to be described and adequately justified as appropriate for the subject device. Explanation of all deviations or description of options selected when following a standard is required under "type of deviation or option selected," "description" and "justification" on the report. More than one page may be necessary.</p> <p>* Types of deviations can include an exclusion of a section in the standard, a deviation brought out by the FDA supplemental information sheet (SIS), a deviation to adapt the standard to the device, or any adaptation of a section.</p>		
Paperwork Reduction Act Statement		
<p>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to:</p> <p style="text-align: center;">Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer 1350 Piccard Drive, Room 400 Rockville, MD 20850</p> <p style="text-align: right;"><i>An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.</i></p>		

Department of Health and Human Services
 Food and Drug Administration
STANDARDS DATA REPORT FOR 510(k)s
 (To be filled in by applicant)

This report and the Summary Report Table are to be completed by the applicant when submitting a 510(k) that references a national or international standard. A separate report is required for each standard referenced in the 510(k).

TYPE OF 510(K) SUBMISSION

Traditional Special Abbreviated

STANDARD TITLE ¹

EN 556-1:2001, Sterilization of medical devices, Requirement for medical devices designated "Sterile"

Please answer the following questions

Yes No

Is this standard recognized by FDA ²?

FDA Recognition number ³ # N/A

Was a third party laboratory responsible for testing conformity of the device to this standard identified in the 510(k)?

Is a summary report ⁴ describing the extent of conformance of the standard used included in the 510(k)?
 If no, complete a summary report table.

Does the test data for this device demonstrate conformity to the requirements of this standard as it pertains to this device?

Does this standard include acceptance criteria?
 If no, include the results of testing in the 510(k).

Does this standard include more than one option or selection of tests?
 If yes, report options selected in the summary report table.

Were there any deviations or adaptations made in the use of the standard?.....
 If yes, were deviations in accordance with the FDA supplemental information sheet (SIS) ⁵?

Were deviations or adaptations made beyond what is specified in the FDA SIS?.....
 If yes, report these deviations or adaptations in the summary report table.

Were there any exclusions from the standard?
 If yes, report these exclusions in the summary report table.

Is there an FDA guidance ⁶ that is associated with this standard?.....
 If yes, was the guidance document followed in preparation of this 510k?

Title of guidance: _____

¹ The formatting convention for the title is: [SDO] [numeric identifier] [title of standard] [date of publication]

² Authority [21 U.S.C. 360d], <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm>

³ <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁴ The summary report should include: any adaptations used to adapt to the device under review (for example, alternative test methods); choices made when options or a selection of methods are described; deviations from the standard; requirements not applicable to the device; and the name and

address of the test laboratory or certification body involved in conformance assessment to this standard. The summary report includes information on all standards utilized during the development of the device.

⁵ The supplemental information sheet (SIS) is additional information which is necessary before FDA recognizes the standard. Found at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁶ The online search for CDRH Guidance Documents can be found at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/default.htm>

EXTENT OF STANDARD CONFORMANCE SUMMARY REPORT TABLE		
STANDARD TITLE EN 556-1:2001, Sterilization of medical devices, Requirement for medical devices designated "Sterile"		
CONFORMANCE WITH STANDARD SECTIONS*		
SECTION NUMBER All	SECTION TITLE (as applicable)	CONFORMANCE? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
<p>* For completeness list all sections of the standard and indicate whether conformance is met. If a section is not applicable (N/A) an explanation is needed under "justification." Some standards include options, so similar to deviations, the option chosen needs to be described and adequately justified as appropriate for the subject device. Explanation of all deviations or description of options selected when following a standard is required under "type of deviation or option selected," "description" and "justification" on the report. More than one page may be necessary.</p> <p>✦ Types of deviations can include an exclusion of a section in the standard, a deviation brought out by the FDA supplemental information sheet (SIS), a deviation to adapt the standard to the device, or any adaptation of a section.</p>		
Paperwork Reduction Act Statement		
<p>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to:</p> <p style="margin-left: 40px;">Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer 1350 Piccard Drive, Room 400 Rockville, MD 20850</p> <p style="text-align: right; margin-right: 40px;"><i>An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.</i></p>		

Department of Health and Human Services
 Food and Drug Administration
STANDARDS DATA REPORT FOR 510(k)s
 (To be filled in by applicant)

This report and the Summary Report Table are to be completed by the applicant when submitting a 510(k) that references a national or international standard. A separate report is required for each standard referenced in the 510(k).

TYPE OF 510(K) SUBMISSION

Traditional Special Abbreviated

STANDARD TITLE ¹

ISO 11135-1:2007, Sterilization of health care products - Ethylene oxide - Part 1

Please answer the following questions

Yes No

Is this standard recognized by FDA ²?

FDA Recognition number ³ #14-228

Was a third party laboratory responsible for testing conformity of the device to this standard identified in the 510(k)?

Is a summary report ⁴ describing the extent of conformance of the standard used included in the 510(k)?
 If no, complete a summary report table.

Does the test data for this device demonstrate conformity to the requirements of this standard as it pertains to this device?

Does this standard include acceptance criteria?
 If no, include the results of testing in the 510(k).

Does this standard include more than one option or selection of tests?
 If yes, report options selected in the summary report table.

Were there any deviations or adaptations made in the use of the standard?.....
 If yes, were deviations in accordance with the FDA supplemental information sheet (SIS) ⁵?

Were deviations or adaptations made beyond what is specified in the FDA SIS?.....
 If yes, report these deviations or adaptations in the summary report table.

Were there any exclusions from the standard?
 If yes, report these exclusions in the summary report table.

Is there an FDA guidance ⁶ that is associated with this standard?.....
 If yes, was the guidance document followed in preparation of this 510k?

Title of guidance: _____

¹ The formatting convention for the title is: [SDO] [numeric identifier] [title of standard] [date of publication]

² Authority [21 U.S.C. 360d], <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm>

³ <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁴ The summary report should include: any adaptations used to adapt to the device under review (for example, alternative test methods); choices made when options or a selection of methods are described; deviations from the standard; requirements not applicable to the device; and the name and

address of the test laboratory or certification body involved in conformance assessment to this standard. The summary report includes information on all standards utilized during the development of the device.

⁵ The supplemental information sheet (SIS) is additional information which is necessary before FDA recognizes the standard. Found at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁶ The online search for CDRH Guidance Documents can be found at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/default.htm>

EXTENT OF STANDARD CONFORMANCE SUMMARY REPORT TABLE		
STANDARD TITLE ISO 11135-1:2007, Sterilization of health care products - Ethylene oxide - Part 1		
CONFORMANCE WITH STANDARD SECTIONS*		
SECTION NUMBER All	SECTION TITLE (as applicable)	CONFORMANCE? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
<p>* For completeness list all sections of the standard and indicate whether conformance is met. If a section is not applicable (N/A) an explanation is needed under "justification." Some standards include options, so similar to deviations, the option chosen needs to be described and adequately justified as appropriate for the subject device. Explanation of all deviations or description of options selected when following a standard is required under "type of deviation or option selected," "description" and "justification" on the report. More than one page may be necessary.</p> <p>* Types of deviations can include an exclusion of a section in the standard, a deviation brought out by the FDA supplemental information sheet (SIS), a deviation to adapt the standard to the device, or any adaptation of a section.</p>		
Paperwork Reduction Act Statement		
<p>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to:</p> <p style="text-align: center;">Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer 1350 Piccard Drive, Room 400 Rockville, MD 20850</p> <p style="text-align: right;"><i>An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.</i></p>		

Department of Health and Human Services
 Food and Drug Administration
STANDARDS DATA REPORT FOR 510(k)s
 (To be filled in by applicant)

This report and the Summary Report Table are to be completed by the applicant when submitting a 510(k) that references a national or international standard. A separate report is required for each standard referenced in the 510(k).

TYPE OF 510(K) SUBMISSION

Traditional Special Abbreviated

STANDARD TITLE ¹

ISO 13485:2003/2009 Particular requirement for application of ISO 9001

Please answer the following questions

Yes No

Is this standard recognized by FDA ²?

FDA Recognition number ³ # N/A

Was a third party laboratory responsible for testing conformity of the device to this standard identified in the 510(k)?

Is a summary report ⁴ describing the extent of conformance of the standard used included in the 510(k)?
 If no, complete a summary report table.

Does the test data for this device demonstrate conformity to the requirements of this standard as it pertains to this device?

Does this standard include acceptance criteria?
 If no, include the results of testing in the 510(k).

Does this standard include more than one option or selection of tests?
 If yes, report options selected in the summary report table.

Were there any deviations or adaptations made in the use of the standard?.....
 If yes, were deviations in accordance with the FDA supplemental information sheet (SIS) ⁵?

Were deviations or adaptations made beyond what is specified in the FDA SIS?.....
 If yes, report these deviations or adaptations in the summary report table.

Were there any exclusions from the standard?
 If yes, report these exclusions in the summary report table.

Is there an FDA guidance ⁶ that is associated with this standard?.....
 If yes, was the guidance document followed in preparation of this 510k?

Title of guidance: _____

¹ The formatting convention for the title is: [SDO] [numeric identifier] [title of standard] [date of publication]

² Authority [21 U.S.C. 360d], <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm>

³ <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁴ The summary report should include: any adaptations used to adapt to the device under review (for example, alternative test methods); choices made when options or a selection of methods are described; deviations from the standard; requirements not applicable to the device; and the name and

address of the test laboratory or certification body involved in conformance assessment to this standard. The summary report includes information on all standards utilized during the development of the device.

⁵ The supplemental information sheet (SIS) is additional information which is necessary before FDA recognizes the standard. Found at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁶ The online search for CDRH Guidance Documents can be found at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/default.htm>

EXTENT OF STANDARD CONFORMANCE SUMMARY REPORT TABLE		
STANDARD TITLE ISO 13485:2003/2009, Particular requirement for application of ISO 9001		
CONFORMANCE WITH STANDARD SECTIONS*		
SECTION NUMBER All	SECTION TITLE (as applicable)	CONFORMANCE? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
<p>* For completeness list all sections of the standard and indicate whether conformance is met. If a section is not applicable (N/A) an explanation is needed under "justification." Some standards include options, so similar to deviations, the option chosen needs to be described and adequately justified as appropriate for the subject device. Explanation of all deviations or description of options selected when following a standard is required under "type of deviation or option selected," "description" and "justification" on the report. More than one page may be necessary.</p> <p>* Types of deviations can include an exclusion of a section in the standard, a deviation brought out by the FDA supplemental information sheet (SIS), a deviation to adapt the standard to the device, or any adaptation of a section.</p>		
Paperwork Reduction Act Statement		
<p>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to:</p> <p style="text-align: center;">Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer 1350 Piccard Drive, Room 400 Rockville, MD 20850</p> <p style="text-align: right;"><i>An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.</i></p>		

Department of Health and Human Services
 Food and Drug Administration
STANDARDS DATA REPORT FOR 510(k)s
 (To be filled in by applicant)

This report and the Summary Report Table are to be completed by the applicant when submitting a 510(k) that references a national or international standard. A separate report is required for each standard referenced in the 510(k).

TYPE OF 510(K) SUBMISSION

Traditional Special Abbreviated

STANDARD TITLE ¹

ISO 10993-1:2010, Biological Evaluation of Medical Devices

Please answer the following questions

Yes No

Is this standard recognized by FDA ²?

FDA Recognition number ³ # 2-98

Was a third party laboratory responsible for testing conformity of the device to this standard identified in the 510(k)?

Is a summary report ⁴ describing the extent of conformance of the standard used included in the 510(k)?
 If no, complete a summary report table.

Does the test data for this device demonstrate conformity to the requirements of this standard as it pertains to this device?

Does this standard include acceptance criteria?
 If no, include the results of testing in the 510(k).

Does this standard include more than one option or selection of tests?
 If yes, report options selected in the summary report table.

Were there any deviations or adaptations made in the use of the standard?.....
 If yes, were deviations in accordance with the FDA supplemental information sheet (SIS) ⁵?

Were deviations or adaptations made beyond what is specified in the FDA SIS?.....
 If yes, report these deviations or adaptations in the summary report table.

Were there any exclusions from the standard?
 If yes, report these exclusions in the summary report table.

Is there an FDA guidance ⁶ that is associated with this standard?.....
 If yes, was the guidance document followed in preparation of this 510k?

Title of guidance: FDA Blue Book Memorandum G95-1

¹ The formatting convention for the title is: [SDO] [numeric identifier] [title of standard] [date of publication]

² Authority [21 U.S.C. 360d], <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm>

³ <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁴ The summary report should include: any adaptations used to adapt to the device under review (for example, alternative test methods); choices made when options or a selection of methods are described; deviations from the standard; requirements not applicable to the device; and the name and

address of the test laboratory or certification body involved in conformance assessment to this standard. The summary report includes information on all standards utilized during the development of the device.

⁵ The supplemental information sheet (SIS) is additional information which is necessary before FDA recognizes the standard. Found at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁶ The online search for CDRH Guidance Documents can be found at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/default.htm>

EXTENT OF STANDARD CONFORMANCE SUMMARY REPORT TABLE		
STANDARD TITLE ISO 10993-1:2010, Biological Evaluation of Medical Devices		
CONFORMANCE WITH STANDARD SECTIONS*		
SECTION NUMBER All	SECTION TITLE (as applicable)	CONFORMANCE? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
<p>* For completeness list all sections of the standard and indicate whether conformance is met. If a section is not applicable (N/A) an explanation is needed under "justification." Some standards include options, so similar to deviations, the option chosen needs to be described and adequately justified as appropriate for the subject device. Explanation of all deviations or description of options selected when following a standard is required under "type of deviation or option selected," "description" and "justification" on the report. More than one page may be necessary.</p> <p>* Types of deviations can include an exclusion of a section in the standard, a deviation brought out by the FDA supplemental information sheet (SIS), a deviation to adapt the standard to the device, or any adaptation of a section.</p>		
Paperwork Reduction Act Statement		
<p>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to:</p> <p style="text-align: center;">Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer 1350 Piccard Drive, Room 400 Rockville, MD 20850</p> <p style="text-align: right;"><i>An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.</i></p>		

Department of Health and Human Services
 Food and Drug Administration
STANDARDS DATA REPORT FOR 510(k)s
 (To be filled in by applicant)

This report and the Summary Report Table are to be completed by the applicant when submitting a 510(k) that references a national or international standard. A separate report is required for each standard referenced in the 510(k).

TYPE OF 510(K) SUBMISSION

Traditional Special Abbreviated

STANDARD TITLE ¹

EN 980:2008, Graphical symbol used in labeling of Medical Devices

Please answer the following questions

Yes No

Is this standard recognized by FDA ²?

FDA Recognition number ³ # N/A

Was a third party laboratory responsible for testing conformity of the device to this standard identified in the 510(k)?

Is a summary report ⁴ describing the extent of conformance of the standard used included in the 510(k)?
 If no, complete a summary report table.

Does the test data for this device demonstrate conformity to the requirements of this standard as it pertains to this device?

Does this standard include acceptance criteria?
 If no, include the results of testing in the 510(k).

Does this standard include more than one option or selection of tests?
 If yes, report options selected in the summary report table.

Were there any deviations or adaptations made in the use of the standard?
 If yes, were deviations in accordance with the FDA supplemental information sheet (SIS) ⁵?

Were deviations or adaptations made beyond what is specified in the FDA SIS?
 If yes, report these deviations or adaptations in the summary report table.

Were there any exclusions from the standard?
 If yes, report these exclusions in the summary report table.

Is there an FDA guidance ⁶ that is associated with this standard?
 If yes, was the guidance document followed in preparation of this 510k?

Title of guidance: _____

¹ The formatting convention for the title is: [SDO] [numeric identifier] [title of standard] [date of publication]

² Authority [21 U.S.C. 360d], <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm>

³ <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁴ The summary report should include: any adaptations used to adapt to the device under review (for example, alternative test methods); choices made when options or a selection of methods are described; deviations from the standard; requirements not applicable to the device; and the name and

address of the test laboratory or certification body involved in conformance assessment to this standard. The summary report includes information on all standards utilized during the development of the device.

⁵ The supplemental information sheet (SIS) is additional information which is necessary before FDA recognizes the standard. Found at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁶ The online search for CDRH Guidance Documents can be found at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/default.htm>

EXTENT OF STANDARD CONFORMANCE SUMMARY REPORT TABLE		
STANDARD TITLE EN 980:2008, Graphical symbol used in labeling of Medical Devices		
CONFORMANCE WITH STANDARD SECTIONS*		
SECTION NUMBER All	SECTION TITLE (as applicable)	CONFORMANCE? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
<p>* For completeness list all sections of the standard and indicate whether conformance is met. If a section is not applicable (N/A) an explanation is needed under "justification." Some standards include options, so similar to deviations, the option chosen needs to be described and adequately justified as appropriate for the subject device. Explanation of all deviations or description of options selected when following a standard is required under "type of deviation or option selected," "description" and "justification" on the report. More than one page may be necessary.</p> <p>* Types of deviations can include an exclusion of a section in the standard, a deviation brought out by the FDA supplemental information sheet (SIS), a deviation to adapt the standard to the device, or any adaptation of a section.</p>		
Paperwork Reduction Act Statement		
<p>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to:</p> <p style="text-align: center;">Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer 1350 Piccard Drive, Room 400 Rockville, MD 20850</p> <p style="text-align: right;"><i>An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.</i></p>		

Department of Health and Human Services
 Food and Drug Administration
STANDARDS DATA REPORT FOR 510(k)s
 (To be filled in by applicant)

This report and the Summary Report Table are to be completed by the applicant when submitting a 510(k) that references a national or international standard. A separate report is required for each standard referenced in the 510(k).

TYPE OF 510(K) SUBMISSION

Traditional Special Abbreviated

STANDARD TITLE ¹

ISO 11607-1, -2:2006, Packaging for Terminally Sterilized Medical Devices

Please answer the following questions

Yes No

Is this standard recognized by FDA ²?

FDA Recognition number ³ # 14-193, 14-194

Was a third party laboratory responsible for testing conformity of the device to this standard identified in the 510(k)?

Is a summary report ⁴ describing the extent of conformance of the standard used included in the 510(k)?
 If no, complete a summary report table.

Does the test data for this device demonstrate conformity to the requirements of this standard as it pertains to this device?

Does this standard include acceptance criteria?
 If no, include the results of testing in the 510(k).

Does this standard include more than one option or selection of tests?
 If yes, report options selected in the summary report table.

Were there any deviations or adaptations made in the use of the standard?
 If yes, were deviations in accordance with the FDA supplemental information sheet (SIS) ⁵?

Were deviations or adaptations made beyond what is specified in the FDA SIS?
 If yes, report these deviations or adaptations in the summary report table.

Were there any exclusions from the standard?
 If yes, report these exclusions in the summary report table.

Is there an FDA guidance ⁶ that is associated with this standard?
 If yes, was the guidance document followed in preparation of this 510k?

Title of guidance: _____

¹ The formatting convention for the title is: [SDO] [numeric identifier] [title of standard] [date of publication]
² Authority [21 U.S.C. 360d], <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm>
³ <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>
⁴ The summary report should include: any adaptations used to adapt to the device under review (for example, alternative test methods); choices made when options or a selection of methods are described; deviations from the standard; requirements not applicable to the device; and the name and

address of the test laboratory or certification body involved in conformance assessment to this standard. The summary report includes information on all standards utilized during the development of the device.
⁵ The supplemental information sheet (SIS) is additional information which is necessary before FDA recognizes the standard. Found at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>
⁶ The online search for CDRH Guidance Documents can be found at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/default.htm>

EXTENT OF STANDARD CONFORMANCE SUMMARY REPORT TABLE		
STANDARD TITLE ISO 11607-1, -2:2006, Packaging for Terminally Sterilized Medical Devices		
CONFORMANCE WITH STANDARD SECTIONS*		
SECTION NUMBER All	SECTION TITLE (as applicable)	CONFORMANCE? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
<p>* For completeness list all sections of the standard and indicate whether conformance is met. If a section is not applicable (N/A) an explanation is needed under "justification." Some standards include options, so similar to deviations, the option chosen needs to be described and adequately justified as appropriate for the subject device. Explanation of all deviations or description of options selected when following a standard is required under "type of deviation or option selected," "description" and "justification" on the report. More than one page may be necessary.</p> <p>* Types of deviations can include an exclusion of a section in the standard, a deviation brought out by the FDA supplemental information sheet (SIS), a deviation to adapt the standard to the device, or any adaptation of a section.</p>		
Paperwork Reduction Act Statement		
<p>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to:</p> <p style="text-align: center;">Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer 1350 Piccard Drive, Room 400 Rockville, MD 20850</p> <p style="text-align: right;"><i>An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.</i></p>		

Department of Health and Human Services
 Food and Drug Administration
STANDARDS DATA REPORT FOR 510(k)s
 (To be filled in by applicant)

This report and the Summary Report Table are to be completed by the applicant when submitting a 510(k) that references a national or international standard. A separate report is required for each standard referenced in the 510(k).

TYPE OF 510(K) SUBMISSION

Traditional Special Abbreviated

STANDARD TITLE ¹

EN 1041: 2008, Terminology, Symbols, and Information Supplied with Devices

Please answer the following questions

Yes No

Is this standard recognized by FDA ²?

FDA Recognition number ³ # N/A

Was a third party laboratory responsible for testing conformity of the device to this standard identified in the 510(k)?

Is a summary report ⁴ describing the extent of conformance of the standard used included in the 510(k)?
 If no, complete a summary report table.

Does the test data for this device demonstrate conformity to the requirements of this standard as it pertains to this device?

Does this standard include acceptance criteria?
 If no, include the results of testing in the 510(k).

Does this standard include more than one option or selection of tests?
 If yes, report options selected in the summary report table.

Were there any deviations or adaptations made in the use of the standard?
 If yes, were deviations in accordance with the FDA supplemental information sheet (SIS) ⁵?

Were deviations or adaptations made beyond what is specified in the FDA SIS?
 If yes, report these deviations or adaptations in the summary report table.

Were there any exclusions from the standard?
 If yes, report these exclusions in the summary report table.

Is there an FDA guidance ⁶ that is associated with this standard?
 If yes, was the guidance document followed in preparation of this 510k?

Title of guidance: _____

¹ The formatting convention for the title is: [SDO] [numeric identifier] [title of standard] [date of publication]

² Authority [21 U.S.C. 360d], <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm>

³ <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁴ The summary report should include: any adaptations used to adapt to the device under review (for example, alternative test methods); choices made when options or a selection of methods are described; deviations from the standard; requirements not applicable to the device; and the name and

address of the test laboratory or certification body involved in conformance assessment to this standard. The summary report includes information on all standards utilized during the development of the device.

⁵ The supplemental information sheet (SIS) is additional information which is necessary before FDA recognizes the standard. Found at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁶ The online search for CDRH Guidance Documents can be found at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/default.htm>

EXTENT OF STANDARD CONFORMANCE SUMMARY REPORT TABLE		
STANDARD TITLE EN 1041: 2008, Terminology, Symbols, and Information Supplied with Devices		
CONFORMANCE WITH STANDARD SECTIONS*		
SECTION NUMBER All	SECTION TITLE (as applicable)	CONFORMANCE? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED *		
DESCRIPTION		
JUSTIFICATION		
<p>* For completeness list all sections of the standard and indicate whether conformance is met. If a section is not applicable (N/A) an explanation is needed under "justification." Some standards include options, so similar to deviations, the option chosen needs to be described and adequately justified as appropriate for the subject device. Explanation of all deviations or description of options selected when following a standard is required under "type of deviation or option selected," "description" and "justification" on the report. More than one page may be necessary.</p> <p>* Types of deviations can include an exclusion of a section in the standard, a deviation brought out by the FDA supplemental information sheet (SIS), a deviation to adapt the standard to the device, or any adaptation of a section.</p>		
Paperwork Reduction Act Statement		
<p>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to:</p> <p style="margin-left: 40px;">Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer 1350 Piccard Drive, Room 400 Rockville, MD 20850</p> <p style="margin-left: 40px;"><i>An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.</i></p>		


1.8. Declaration of Conformity

Declaration of Conformity With Design Controls

SOFIA Distal Access Catheter

Verification Activities:

To the best of my knowledge, the verification activities required by the risk analysis, for the above referenced device were performed by the designated individual(s) in accordance with the MicroVention Quality Procedure Design and Development Process requirements, and the results demonstrated that the predetermined acceptance criteria were met.



Joe Gulachenski
Director of Research and Development

5-16-2013
Date

=====

Manufacturing Facility:

The manufacturing facility, MicroVention Inc., is in conformance with the design control requirements as specified in 21 CFR 820.30, and the records are available for review.



Tom Sternweiler
Senior Director of Quality

5-16-13
Date

1.9. Design Control Activities Summary

SOFIA Catheter - Design Control Activities Summary

Device Modification	Risk	Verification Activity	Acceptance Criteria	Results of Verification
Predicate Device: Chaperon Guiding Catheter (K082385)				
<ol style="list-style-type: none"> 1. <u>Catheter length</u> Chaperon (predicate): 95-117 cm SOFIA: 125 cm 2. <u>Catheter ID</u> Chaperon: 1.5 mm SOFIA: 1.4 mm 3. <u>Radiopaque marker</u> Chaperon: Imbedded tungsten SOFIA: Pt/Ir marker band 4. <u>Hydrophilic coating</u> Both have, but SOFIA is the same as Headway DUO (K120917) 5. <u>Tip configuration</u> Chaperon: Preshaped SOFIA: Steam shapeable by user 6. <u>Additions to SOFIA</u> <ol style="list-style-type: none"> i. A stainless steel coil (304V) over the inner liner (length of minimum 15 cm to the distal tip) ii. The distal outer shaft consists of different durometers and different lengths of polyblend 1100, pellethane 80A and Plexar 3080 of natural color. The range of polyblend durometers is between 30A to 60A 	<p>All modifications were evaluated for risk and were determined to be tolerable.</p>	<p>Bench Testing:</p> <ol style="list-style-type: none"> 1. Surface contamination 2. Physical Attributes 3. Simulated use, insertion tool performance and equipment interface 4. Kink resistance 5. Tip shapeability 6. Radio detectability 7. Hub (ISO 594-2) 8. Durability/Lubricity of hydrophilic coating 9. Catheter stiffness testing 10. Torque strength test 11. Catheter flexural fatigue 12. Force at break (distal and hub) 13. Flow rate 14. Static burst 15. Leakage at 46 psi 16. Air Leakage 17. Dynamic burst 18. Catheter particle testing <p>Packaging and Pouch Testing:</p>	<p>Per TP12-280:</p> <ol style="list-style-type: none"> 1. Free from unacceptable conditions per TP12-280. 2. Physical attributes criteria per TP11-280 3. Simulated use and equipment interface 4. Equivalent to or better kink resistance than competitor products during simulated use testing (rating of 3 or better per TP12-280) 5. For reference only 6. The catheter must have acceptable results per TP12-280 7. Testing meets ISO 594-2 8. The catheter must have acceptable results per TP12-280 9. Document stiffness for reference use only. 10. The catheter must have acceptable results per TP12-280 11. The catheter must have acceptable results per TP12-280 12. ≥ 2.25 lbf for outer diameters $\geq 0.045''$ and $< 0.072''$ 13. Equivalent to or better flow rates than competitor products (for reference only) 14. Catheter shaft will not burst below 46 psi. 15. No liquid leaking from hub and catheter shaft at 46 psi for 30 second duration. 16. No air leaking into syringe for 15 seconds (ignore air bubbles for the first 5 seconds of the test. 17. Catheter will not burst below 300 psi 18. Less than 25 particles/ml (> 10 microns) and less than 3 particles/ml (> 25 microns) <p>Per TP12-288</p>	<p>All test data met acceptance criteria and results are documented in TR12-280.</p> <p>For Verification:</p> <ul style="list-style-type: none"> • All testing passed with 90% confidence level at 80% reliability <p>For Validation:</p> <ul style="list-style-type: none"> • All testing passed with 90% confidence level of 90% reliability <p>Met acceptance criteria (TR12-288)</p>

<p>iii. Introducer sheath and shaping mandrel included in packaged device</p> <p>iv. Packaging: Added a dispenser tube to hold the catheter in place on the packaging card</p>		<p>Biocompatibility testing</p>	<p>Per ISO 10993-1</p>	<p>Met acceptance criteria (Toxikon Reports in Appendix 7)</p>
--	--	--	------------------------	--

2. Executive Summary

The MicroVention SOFIA™ (Soft torqueable catheter **Optimized For Intracranial Access**) Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

The SOFIA Catheter is a single lumen catheter designed to be introduced over a steerable guidewire to access small, tortuous vasculature. The semi-rigid proximal section transitions to a flexible distal tip to facilitate advancement through vessels. A single radiopaque marker at the distal end facilitates fluoroscopic visualization. The outer surface of the catheter is coated with a hydrophilic coating to reduce friction during navigation in the vasculature. A luer fitting on the microcatheter hub is used for the attachment of accessories. The hub – strain relief provides kink resistance for the proximal end. A steam shaping mandrel and introducer sheath are also packaged with the catheter.

The SOFIA Catheter has the same intended use as the predicate, Chaperon Guiding Catheter (K082385) and is similar in terms of design and operating principle. The SOFIA Catheter has been designed to facilitate increased trackability and provide greater accessibility to the vasculature. The distal straight tip is designed to be steam shaped by the user at the time of use.

The *in vitro* testing covered the physical, mechanical, and functional performance of the SOFIA Catheter using the FDA Special Controls Draft Guidance Document: Short-Term and Long-Term Intravascular Catheters Dated March 16, 1995 and applicable sections of the ISO 10555-1 international standard for Sterile, Single-Use Intravascular Catheters. These tests validated the performance characterization of the SOFIA Catheter. The combined conclusion from these tests demonstrates that the *in vitro* behavior of device is well characterized within design specifications.

A full panel of biocompatibility tests was conducted according to ISO10993-1 and provides assurance that the device has a safe biocompatibility profile for use as an externally communicating, circulating blood, limited contact (≤ 24 hrs) device.

The SOFIA Catheter is intended to be sold sterile, for single use. The sterilization method (ethylene oxide) has been validated as well as the packaging configuration.

It is on this basis that it can be concluded the safety profile of SOFIA Catheter is well within acceptable safety limits to be used for its indications for use.

Note: During the product development, some of the documents reference Hybrid Access Catheter (HAC), Distal Access Catheter (DAC), or Project # RD11-012. These are all previous names used for the SOFIA Distal Access Catheter (final product name of the device).

3. Device Name

The device trade names and common/classification names are:

Device Trade Name	SOFIA Distal Access Catheter
Device Generic Name	Percutaneous Catheter
Classification Name	Percutaneous Catheter
CFR Classification	21 CFR 870.1250
Device Class	Class II
Classification Committee	Cardiovascular
Product Code	DQY

4. Address and Registration Number

The address and registration number of the manufacturer and sterilization sites for the device are:

Manufacturer	MicroVention, Inc. 1311 Valencia Avenue Tustin, California U.S.A
Establishment Registration No.	2032493
Contact	Naomi Gong Sr. Regulatory Affairs Project Manager 1311 Valencia Avenue Tustin, California U.S.A. Phone: (714) 247-8055 Fax: (714) 247-8014
Sterilization Site	Sterigenics 4900 South Griffith Ave. Los Angeles, California, 90058 FDA Registration No= 2011171

5. Device Class

The SOFIA Catheter is a percutaneous catheter and is classified as Class II, DQY. The product has been designed, developed and tested using the FDA Guidance 510(k) Submission for Short-Term and Long-Term Intravascular Catheters (1995).

6. Predicate Device Information

K082385, MicroVention, Inc., Chaperon Guiding Catheter

7. Labeling and Intended Use

Draft labels and Instructions For Use are provided in [Appendix 1](#).

Intended Use

The intended use is the same as the predicate device, Chaperon Guiding Catheter and is stated in the product labeling as follows:

The SOFIA™ Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

8. Device Overview

The SOFIA Catheter is a single lumen (0.055” or 1.4 mm inner diameter) catheter with a shapeable distal tip. The semi-rigid proximal section transitions to a flexible distal tip to facilitate advancement through the vasculature.

A single radiopaque marker (Pt/Ir) at the distal end facilitates fluoroscopic visualization. The outer surface of the catheter (distal 60 cm) is coated with a hydrophilic polymer coating to reduce friction during navigation in the vasculature.

A luer fitting on the microcatheter hub is used for the attachment of accessories. The hub/strain relief provides kink resistance for the proximal end. The tip configuration is provided straight, but is shapeable by the user. A steam shaping mandrel and introducer sheath accessories are packaged with the catheter.

The device is placed in a dispenser tube to keep the device in position during shipping and handling. The dispenser tube is placed on a packaging card and then placed into a tyvek pouch. The pouch is put into a chipboard carton box prior to sterilization. The device is provided sterile, for single use.


9. Device Configurations and Dimensions

The SOFIA Distal Access Catheter has an inner lumen of 0.055” and is recommended to be used with guidewires from 0.035” to 0.038”.

Device	Catalogue Number	Working Length	Outer Diameter/ Inner Diameter	Recommended Guidewire
SOFIA Distal Access Catheter	DA5125ST	125 mm	0.068”/0.055” (1.7/1.4 mm)	0.035” to 0.038”

10. Design Description

The SOFIA Catheter is a stainless steel braid plus coil reinforced, a single lumen, variable stiffness catheter with a shapeable distal segment. It consists of a lubricious inner liner and a radiopaque marker located at distal tip to facilitate fluoroscopic visualization. The outer distal section of catheter has a hydrophilic coating for lubricious insertion and delivery. The semi-rigid proximal section transitions to a flexible distal tip to facilitate advancement through vessels.



A	Diameter (outer), proximal	0.068 in/ 1.7 mm
B	Diameter (outer), distal	0.068 in/ 1.7 mm
C	Diameter (inner)	0.055 in
D	Working length	125 cm
E	Distal section	16 cm
F	Tip configuration	Straight Shapeable

The catheter body is constructed with a stainless steel coil (less 2 cm of catheter length) over the inner lumen liner comprised of polytetrafluoroethylene (PTFE) and polyolefin elastomer. To provide additional shaft support, a stainless steel wire braiding has been added over the stainless steel coil from the proximal end to distal end. A platinum/iridium alloy radiopaque marker band is located at the distal tip of the catheter. This is all covered by an outer layer of varying durometers of polyurethane and polyether block amide.

The outer layer consists of a range durometers and lengths of polyurethane (Polyblend and Pellethane) and polyether block amide (Pebax) - distal and proximal, respectively. The most proximal outer shaft section consists of a polyamide (Grilamid).

A nylon hub is attached to the proximal end of the catheter. A strain relief made from polyurethane is placed at the proximal end of the catheter and distal end of the hub.

The outer surface of the catheter (distal 60 cm) is coated with a hydrophilic coating to reduce friction during navigation in the vasculature.


A luer fitting on the catheter hub is used for the attachment of accessories. The hub – strain relief provides for the kink resistance from the proximal end.

A steam shaping (stainless steel) mandrel with an outer diameter of 0.0255” is provided with the catheter to be used by the physician for tip shaping. An introducer sheath (Pebax) is also included to facilitate the introduction of the catheter into the y-connector. These two components are the identical to the Headway Duo with the exception of dimensional differences to accommodate the ID/OD of the SOFIA Catheter.

The device is provided sterile and for single use. The catheter is placed in a dispenser tube (HDPE) and is placed on a packaging card (polyethylene) that is provided in a sterile barrier tyvek pouch and placed in a carton box.

The predicate Chaperon Guiding Catheter is a two-catheter system comprised of the outer catheter and the inner catheter. The Chaperon Guiding Catheter system can be used individually with 0.035” or 0.038” guidewire or together with the Inner Catheter to access the desired anatomy.

Chaperon Guiding Catheter



The SOFIA has the same basic design of the Outer catheter, but incorporates a hydrophilic coating like that of the Inner catheter.

Sample product drawing for the SOFIA Distal Access Catheter is provided in the [Appendix 2](#).

11. Technological Characteristics Comparison

The following table compares the technological characteristics of the SOFIA Catheter to the predicate device, Chaperon.

	Chaperon Guiding Catheter – K082385 (Outer catheter)	SOFIA Distal Access Catheter
Intended Use	Intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.	Same
Material	Outer layer of polyester elastomer, stainless steel braid, inner liner of PTFE (polytetrafluoroethylene). Tungsten radiopaque marker, nylon hub, and polyester elastomer strain relief.	Outer layer of polyolefin elastomer, polyurethane elastomer, polyether block amide, polyamide; inner layer of PTFE/polyolefin; stainless steel braid and coil. Pt/Ir radiopaque marker, nylon hub, polyurethane strain relief. Pebax introducer sheath and stainless steel shaping mandrel.
Catheter size	5F (Outer catheter)	Same
ID	1.5 mm (0.059")	1.4 mm (0.055")
OD	1.7 mm (0.068")	Same
Effective Length	Outer Catheter : 95 cm Inner Catheter: 117 cm	125 cm
Coating	Hydrophilic coating (inner catheter)	Hydrophilic coating (same as Headway Duo, K120917)
Tip Configuration	Preshaped	Steam shapeable by user
Guidewire compatibility	0.035" or 0.038"	Same
Accessories	N/A	Introducer sheath and shaping mandrel (same as Headway Duo)
Packaging	Packaging Card/Tyvek pouch/Carton	Same, with dispenser tube to hold catheter on packaging card
Method of supply	Sterile and single use	Same

12. Design Control and Risk Management Processes

The SOFIA Catheter is designed, developed and tested in accordance with the MicroVention Design and Development procedure in which the impact of modifications on device safety and performance is assessed in accordance with the ISO 14971-1 (Medical Device Risk Management) – Part 1, and with the MicroVention quality system procedure for Risk Management. Possible hazards and associated risk related to the device modification and

clinical usage of the device were identified, examined and found to be acceptable after the implementation of the countermeasures such as physician training program, labeling warnings, specify possible mitigation.

Copies of the relevant procedures and documents are included as shown below:

[Attachment 3](#) – QP 4.1, Design and Development Process procedure

[Attachment 4](#) – QP 4.8, Risk Management procedure

[Attachment 5](#) – RA060004, Risk Assessment and Control Analysis
- FA060010, dFMEA

13. List of Voluntary Standards

The SOFIA Catheter was designed, developed, and tested using the applicable requirements of the following standards:

Standard No.	Standard Name	Edition
FDA Guidance	Guidance on Premarket Notification 510(k) Submission for Short-term and Long-term Intravascular Catheter	1995
ISO 10555-1	Sterile, Single Use Intravascular Catheters	2004
Medical Device Directive	Council Directive 93/42/EEC	2003/2007
ISO/EN 14971	Medical Device – Application of Risk Management to medical devices	2012
ANSI/AAMI/ISO11135	Medical Devices - Validation and routine control of ethylene oxide sterilization, overkill method	2007
EN 556-1	Sterilization of medical devices, Requirements for medical devices designated “STERILE”	2001
ISO 13485	Particular requirement for application of ISO 9001	2003/2009
ISO 10993-1	Biological evaluation of medical devices	2010
EN 980	Graphical Symbol used in Labeling of Medical Devices	2008
ISO 11607 -1, -2	Packaging for Terminally Sterilized Medical Devices	2006
EN 1041	“Terminology, Symbols and Information Supplied with Devices.”	2008

14. In-Vitro/Bench Verification and Shelf Life

The SOFIA Catheter was tested and verified in the laboratory setting according to written protocol TP12-280, Design Verification/Validation and Accelerated Aging Study of HAC (aka SOFIA). All samples met the established design specification and shelf life testing for 6 months for parametric attributes as well as the determined confidence/reliability level for variable data.

The purpose of the study was to evaluate the effects of aging on the physical and functional properties of the balloon catheters assemblies. For these tests, packaged and sterilized catheters are placed in a temperature-controlled oven at 60°C per ASTM-F-1980 for 29 days

(equivalent to 1 year). Accelerated aging studies are based on the theory that a rise in temperature of 10° C will double the rate of a chemical reaction. The formula used for calculating aging study time period is:

$$\begin{aligned} \text{Shelf Life} &= t \times q^y \\ t &= \text{Accelerated Exposure Time (Days)} \\ q &= \text{Acceleration Factor (Coefficient of Aging)} \\ y &= \frac{\text{Elevated Storage Temp (°C)} - \text{Ambient Storage Temp (22 °C)}}{10} \end{aligned}$$

During the testing, it was found that the Vicat Softening Point for one of the materials (b)(4) Confidential and Proprietary Information – which impacted the leakage (liquid) test. Therefore, additional samples were aged at 45°C for 44 days to simulate 6 months of shelf life and the tests that were impacted were repeated. The tests samples passed all acceptance criteria at 6 months or 1 year, therefore the device will initially be labeled for 6 month shelf life per the testing regimen conducted. The shelf life will be extended as additional aging studies are successfully completed.

Test Sample Configuration

Product	Model
SOFIA	DA5125ST

Side by side testing with competitive devices, Chaperon, ev3 (Covidien) Navien, and Penumbra 5MAX catheters, has also been performed. Design verification test protocol and test report, TP/TR 12-280 are provided in [Appendix 6](#). The following table summarizes the verification/validation/shelf life testing as conducted:

SOFIA Testing Summary

Bench Testing	Acceptance Criteria	Result
Simulated use, insertion tool performance and equipment interface	Simulated use and equipment interface <ul style="list-style-type: none"> • Rating of 3 or higher in all test categories per TM11050 • Equivalent or better simulated use rating when compare to competitor products • Compatible with ≤ 0.035" guidewire • Compatible with 6F or larger guide catheter or guiding sheath • Compatible with all common RHVs and stopcocks • Compatible with ≤ 0.027" microcatheter (Headway) • Ease entering RHV using insertion tool 	Passed
Physical attributes <ul style="list-style-type: none"> • Catheter OD • Catheter ID • Catheter working length • Length of distal section 	<ul style="list-style-type: none"> • Catheter OD 0.0660" – 0.0685" • ≥ 0.055" • 125 ± 2cm and reaches the M1 vessel in the vessel model • ≥ 15cm 	Passed

Bench Testing	Acceptance Criteria	Result
<ul style="list-style-type: none"> Length of marker band to distal tip Total length of proximal hub and strain relief Strain relief color 	<ul style="list-style-type: none"> 0.025" \pm 0.005" Approximately 5cm Gray 	
Kink resistance	<ul style="list-style-type: none"> Equivalent to or better kink resistance than competitor products during simulated use testing (rating of 3 or better per TM11050) Equivalent to or better kink resistance than competitor products per TM218 (for reference only) 	Passed
Tip shapeability	For reference only	N/A
Radio detectability	Equivalent to or better radiopacity than competitor products	Passed
Gauging (ISO 594-2)	Verify gauging pin and hub align in limit planes.	Passed
Separation force (ISO 594-2)	Mating parts separation force is greater than 25N.	Passed
Unscrewing torque (ISO 594-2)	Catheter luer remains attached after applying an unscrewing torque not less than 0.02N m for a minimum of 10 seconds.	Passed
Stress cracking (ISO 594-2)	Verify there are no stress cracks on catheter hub	Passed
Ease of assembly (ISO 594-2)	<ul style="list-style-type: none"> No resistance observed between the catheter luer and the reference fitting. Components must fit together securely. 	Passed
Resistance to overriding (ISO 594-2)	Catheter luer does not override reference fitting threads	Passed
Durability/Lubricity of hydrophilic coating	Rating of 3 or higher in tested categories	Passed
Catheter stiffness testing	Testing for reference use only.	N/A
Torque strength test	50 rotations without breakage or equivalent to competitive products	Passed
Catheter flexural fatigue	Catheter shall have acceptable results from: <ul style="list-style-type: none"> Simulated Use Force at Break Flow rate Static burst pressure Freedom from leakage (liquid) 	Passed
Surface contamination	<ul style="list-style-type: none"> Free from uncured hydrophilic coating. No surface particulate > 0.02 mm² per tappi chart. 	Passed


Bench Testing	Acceptance Criteria	Result
	<ul style="list-style-type: none"> Free from surface defect/no sharp edges Embedded particulate acceptable if OD is in specification Distal tip smooth and round at lumen opening PTFE liner is not delaminated at distal tip 	
Force at break (distal and hub)	≥ 2.25 lbf for outer diameters ≥ 0.045" and < 0.072"	Passed
Flow rate	Equivalent to or better flow rates than competitor products (for reference only)	N/A
Static burst	Catheter shaft will not burst below 46 psi.	Passed
Leakage at 46 psi	No liquid leaking from hub and catheter shaft at 46 psi for 30 second duration.	Passed
Air leakage	No air leaking into syringe for 15 seconds (ignore air bubbles for the first 5 seconds of the test).	Passed
Dynamic burst	<ul style="list-style-type: none"> Equivalent or greater burst pressure than competitor product Microcatheter will not burst below 300 psi 	Passed
Catheter particle testing	Less than 25 particles greater than 10 microns and less than 3 particles greater than 25 micron	Passed

As summarized in the table, SOFIA Testing Summary, the following tests were conducted to establish the performance characteristics of the device:

14.1. Simulated Use

To verify that SOFIA Catheter meets the established performance specifications in a clinically simulated environment, MicroVention has tested total of 22 samples of SOFIA. Samples underwent simulated use testing that included delivery, trackability (with and without guidewire), guidewire lock-up, catheter ovalization, microcatheter stability during coil deployment, coil interaction with microcatheter lumen. All samples met the established acceptance criterion of ≥ 3 rating.

The test simulates a neurointerventional embolization procedure using 37°C fluid with catheters, guidewires, and vessel tortuosity.



The following table summarizes the results from total of 22 test samples tested using MicroVention SOFIA.

Simulated Use per TM 11050

	SOFIA (1 yr) N = 22	SOFIA (6 mos) N = 22	Navien N = 3	5MAX N = 3
	Minimum/Maximum rating obtained			
Preparation/Ease of Assembly	5/5	5/5	5/5	5/5
Introducer Sheath Interaction	5/5	5/5	1/1	4/4
Introducer Peel Away	5/5	5/5	1/1	2/2
Tracking Test with Guidewire/Microcatheter	4/5	5/5	1/1	3/3
Microcatheter/Guidewire Lockup	5/5	5/5	1/1	1/1
Lubricity and Durability of Hydrophilic Coating	4/5	4/5	1/1	1/1
Microcatheter/Guidewire Removal	5/5	5/5	1/1	1/1
Removal/aspiration of clot	5/5			
Mechanical clot retriever	4/5			
Stent delivery	5/5			
Particles	No particles	No particles		
Acceptance Criteria	Rating ≥ 3 (Scale of 1 -5)			
Results	Pass	Pass	N/A	N/A

As a result of the above testing, guidewires of 0.035” and 0.038” were determined to be compatible for use with the SOFIA.

14.2. Dimensional & Physical Attributes

The purpose of this inspection is to 1) verify that finished catheter meet dimensional specifications; and, 2) to ensure integrity of these units subsequently used for physical verification tests. A microscope and standard measuring scales are used for verification of defined dimensions. Results are summarized below:

Specification (N= 11)	Catheter OD			Catheter ID ≥ 0.055in. (1.40mm)	Overall Working Length 125 ± 2cm (49.2 ± 0.8in.)
	0.0660in. – 0.0685in. (1.68mm – 1.74mm)				
	Distal Tip OD	Distal OD	Proximal OD		
Min	0.0665 (1.69)	0.0675 (1.71)	0.0660 (1.68)	NA	123 (48.4)
Max	0.0675 (1.71)	0.0680 (1.73)	0.0665 (1.69)	NA	125 (48.4)
Average	0.0670 (1.70)	0.0677 (1.72)	0.0660 (1.68)	0.055 (1.40)	124.3 (48.9)
St. Dev.	0.0005 (0.013)	0.0003 (0.008)	0.0002 (0.005)	NA	0.8 (0.3)
Results	All units in test group met the criteria – PASS				

Specification (N=11)	Length of Distal Section	Length of distal tip to marker band	Total length of hub and strain relief
	≥ 15cm (5.9in.)	0.025in. ± 0.005in. (0.635 ± 0.127mm)	Approximately 5cm (1.97in.)
Min	15.5 (6.1)	0.022 (0.559)	NA
Max	17.5 (6.9)	0.029 (0.737)	NA
Average	16.1 (6.3)	0.025 (0.635)	5.5 (2.2)
St. Dev.	0.6 (0.2)	0.0021 (0.0533)	NA
Results	All units in test group met the criteria – PASS		

14.3. Kink Resistance

The purpose of the test is to assess kink resistance of SOFIA Catheter when tested side-by-side with the Navien and 5MAX. Results are then correlated to the data recorded from kink test fixture – smallest bend diameter at which the catheter begins to kink.

Kink Resistance (Minimum bend diameter at kink point)

Distance from distal tip	1cm (0.39in.)	4cm (1.57in.)	12cm (4.72in.)	25cm (9.84in.)
SOFIA (N = 11)	No units kinked using 0.030in. (0.076cm) pin gauge	No units kinked using 0.030in. (0.076cm) pin gauge	No units kinked using 0.030in. (0.076cm) pin gauge	No units kinked using 0.250in. (0.635cm) pin gauge
Chaperon (N=3)	No units kinked using 0.030in. (0.076cm) pin gauge	No units kinked using 0.030in. (0.076cm) pin gauge	No units kinked using 0.030in. (0.076cm) pin gauge	No units kinked using 0.170in. (0.432cm) pin gauge
Navien (N=3)	All units kinked using 0.080in. (0.203cm) pin gauge	All units kinked using 0.080in. (0.203cm) pin gauge	All units kinked using 0.080in. (0.203cm) pin gauge	All units kinked using 0.250in. (0.635cm) pin gauge
5 MAX	No units kinked using 0.080in. (0.203cm) pin gauge	No units kinked using 0.080in. (0.203cm) pin gauge	All units kinked using 0.080in. (0.203cm) pin gauge	All units kinked using 0.250in. (0.635cm) pin gauge
Specification	For Reference data only			

Kink Resistance (Simulated Use Testing)

	SOFIA (1 yr)	SOFIA (6 mos)	Navien	5 MAX
	Minimum/Maximum rating obtained			
N	22	22	3	3
Track test w/ guidewire	4/5	4/5	1/1	3/3
Guidewire lock up	5/5	5/5	1/1	1/1
Acceptance Criteria	Equivalent or better than competitive device			
Pass/Fail	Pass	Pass	N/A	N/A

14.4. Tip Shapeability

The objective of this test is to characterize shapeability distal-tip after steam-shaping. Results are listed below:

Tip Shapeability

	For reference only					
	SOFIA (N = 11)		Navien 058 (N = 3)		5 MAX (N = 3)	
	Steam shaping tip angle	Tip angle measured after steam shaping	Steam shaping tip angle	Tip angle measured after steam shaping	Steam shaping tip angle	Tip angle measured after steam shaping
Min	90.53	52.02	90.20	64.97	90.18	75.65
Max		67.36		83.66		86.32
Average		61.10		71.73		79.94
St. Dev.		4.34		10.36		5.64

14.5. Radio-detectibility

The purpose of this test is to compare the radiopacity under angiographic test comparing the SOFIA to the Chaperon and Navien 058. The results are summarized below:

	Radiopacity Test
Specification	Marker band radiopacity equivalent to or better than competitor products
Results	Distal marker band visible under fluoroscopy. Image intensity equal to that of predicate products.
Pass/Fail	Pass

14.6. Hub Evaluation (ISO 594-2)


The purpose of the test is to assess the hub performance of the microcatheter per ISO 594-2 requirements and methods.

Hub Testing (ISO 594-2)

N = 11	Acceptance criteria	Results
Hub gauging	Gauging pin and hub align at limit planes	Pass
Separation force	Mating parts separation force is great than 25 N	Pass
Unscrewing torque	Microcatheter luer remains attached under unscrewing torque of 0.02Nm for 10 seconds	Pass
Resistance to override	Microcatheter luer does not override reference fitting threads	Pass
Stress cracking	No stress cracks on the microcatheter hub	Pass

14.7. Catheter Stiffness

The purpose is to determine the stiffness, defined by applied load, of catheters in flexure at the bend point using an Instron Test Stand. This test provides a quantitative reference metric for evaluation of the catheter bending stiffness throughout its distal region which may influence its behavior while advancing to its target site.



Note: HAC is the same as SOFIA

14.8. Torque Strength Testing

The purpose of this test is to evaluate the torque strength of the catheter as defined by the number of turns to failure. The distal tip of the catheter is stabilized and the catheter is rotated in a single direction for up to 50 turns or until failure.

	SOFIA N = 11	Chaperon N = 3	Navien/5 MAX N= 1 ea
# of rotations	50	50	50
Catheter breakage	No breaks	No breaks	No breaks
Acceptance Criteria	50 rotations without catheter breakage or equivalent to competitive product catheters.		
Pass/fail	Pass		

14.9. Catheter Flexural Fatigue

The purpose of this test method is to evaluate the catheter tensile strength and pressure characteristics after undergoing repeated fatigue cycles. This method covers two types of fatigue which the microcatheter is exposed to; 1- Flexural fatigue from repeated bending of the catheter during simulated use testing and 2- Flexural fatigue from repeated hoop stress on the catheter after pressure testing and air aspiration testing. This method combines many catheter test methods which stresses the catheter as summarized below:

Catheter Flexural Fatigue

N = 22	Simulated Use	Force at Break	Flow Rate	Static Burst Pressure	Catheter Leakage (liquid)
SOFIA	Pass	Pass	Pass	Pass	Pass
Specification	Pass required specified acceptance criteria after completing fatigue testing				

14.10. Surface Contamination

The purpose of the inspection is 1) to ensure the catheter surface is free from uncured hydrophilic coating. The surface must be free of visible functional or cosmetic defect attributable to the manufacturing, packaging and/or sterilization processes; 2) to ensure the distal tip is smooth, rounded and tapered; 3) to ensure integrity of catheter subsequently used for physical verification tests. Test Results are provided below:

Surface Contamination

	N	Free from uncured coating	No surface particulate > .02 mm ²	Free from surface defect	Embedded particulate = OD spec.	Distal Tip	PTFE Liner
SOFIA	11	Pass	Pass	Pass	Pass	Pass	Pass

14.11. Force at Break

To determine the minimum force at break of the catheter shaft and hub-catheter connection is the objective of the test. Results are listed below:

Force at Break for OD ≥ .045" (Distal) to < .072" (Hub)

N = 22	Distal (lbf)	Hub (lbf)
Minimum	5.92 (26.33)	8.79 (39.10)
Maximum	7.66 (34.07)	14.02 (62.36)
Average	7.09 (31.54)	10.70 (47.60)
Std. Dev.	0.41 (1.82)	1.12 (4.98)
Acceptance Criteria	≥ 2.25 lbf (10.01 N)	
Result	Pass	Pass

14.12. Flow Rate

To measure the catheter injection flow rate of saline and/or contrast media as a function of the applied pressure is the purpose of this test. This test also evaluated and established the catheter compatibility with diagnostic agents (contrast media).

Flow Rate (ml/sec)

SOFIA N=11	Saline (100%)		50/50% contrast/saline		60/40% contrast/saline		76/24% contrast/saline	
	100 psi	300 psi	100 psi	300 psi	100 psi	300 psi	100 psi	300 psi
Minimum	11.0	20.0	9.6	17.0	8.8	16.0	7.0	15.0
Maximum	15.0	22.0	11.0	19.0	11.0	18.0	8.6	17.0
Average	11.9	20.7	10.1	17.9	9.5	17.0	7.7	16.4
Chaperon (Avg)	13.7	25.0	13.7	22.3	12.0	21.7	10.7	21.3
Navien (Avg)							8.5	17.7
5 MAX (Avg)							10.7	21.7
Specification	Reference data for IFU							

14.13. Static Burst Pressure

The purpose of the testing is to evaluate the fluid leakage and damage of pressurized catheter as outlined in the ISO 10555-1. The catheter is connected via a leak-proof connection to a syringe. A hydraulic pressure is applied to the hub assembly. The pressure is then measured after 2 seconds. Results are summarized below:

Static Burst Strength

	SOFIA (6 mos) N = 11	Chaperon N = 3	Navien N = 3	5 MAX N = 3
Minimum	122	1300	1107	129
Acceptance Criteria	≥ 46 psi			
Result	Pass	N/A	N/A	N/A

14.14. Freedom from Leakage – Liquid

The purpose of the test is to ensure the device does not leak under pressure of fluid per the ISO 10555-1 Annex C (Fluid). The formation of fluid droplets or presence of escaped fluid on the surface of the microcatheter shaft or any part of the hub connection due to intraluminal pressure is recorded under the test condition. Results are summarized below:

Freedom from Leakage – liquid low pressure - long duration (6 mos)

N = 11	
Acceptance criteria	No liquid leaking from device at 46 psi for 30 seconds
Result	Pass

14.15. Freedom from Leakage - Air

The purpose of the test is to ensure air does not leak into the device according to the ISO 10555-1 Annex D (Air). A 10cc syringe is filled with at least 10cc of boiled and cooled water is inserted and filled into the hub. Once a droplet is formed, the syringe volume is then adjusted according to the standard. The bubble formation after the first 5 seconds is then recorded. Results are summarized below:

Freedom from Leakage – Air (6 mos)

N = 11	
Acceptance criteria	No air leaking (at hub) into syringe for 15 seconds
Result	Pass

14.16. Dynamic Burst Pressure

The purpose of this test is to determine the dynamic burst pressure of catheter which is the hydrostatic pressure at which a free-flowing catheter fails or ruptures. Total of 22 units of SOFIA were evaluated.

Dynamic Burst Pressure (6 mos)

N = 22	
Acceptance criteria	Microcatheter will not burst below 300 psi (2068kPa)
Result	Pass

14.17. Particulate Measurement Analysis

As outlined in the USP <788>, analysis was performed to quantify particulate matter in injections of the SOFIA Catheter after advancement/retraction procedures.

The “particulate matter in injection” was performed by PMT (Particle Measurement Technology Co.), an independent laboratory per USP Section <788>. A total of 3 SOFIA were tested by automated light obscuration particle counter.

Samples were tested in the simulated intra-cranial silicon aneurysms tortuous flow model. Samples were prepared according to the IFU. The samples were subjected to worst-case anticipated clinical conditions by cycling (advancement and retraction) ten times. The

sample flushes collecting particulates generated during tracking were then tested according to the USP <788>.

As documented in the PMT Particle Analysis Report ([Appendix 6](#)), the result meets the requirements of the particle test if the statistical particle count does not exceed 25 particles equal to or greater than 10 microns in size, and/or 3 particles equal to or greater than 25 microns in size.

Particulate Analysis

	Particles/mL (Avg)		Particles/Device (Total)		Particles >70µm
	>10µm	> 25µm	>10µm	> 25µm	
SOFIA					
Sample #1	1.8	0.20	108	12	No particles > 70 microns detected
Sample #2	1.2	0.30	72	18	No particles > 70 microns detected
Sample #3	0.9	0.00	54	0	No particles > 70 microns detected
Chaperon (N=1)	2.8	0.40	168	24	No particles > 70 microns detected
Navien 058	1.6	0.20	96	12	No particles > 70 microns detected
Acceptance Criteria	< 25 particles/mL (for >10 microns) < 3 particles/mL (for > 25 microns)				
Results	All units in test group met the criteria – PASS.				

15. Biocompatibility

A full panel of biocompatibility studies was performed on representative samples of the SOFIA Retrieval Device. The biological safety has been verified in accordance with the ISO10993-1, Biological Evaluation of Medical Devices by an independent laboratory, Toxikon, located in Bedford, Massachusetts. The table below summarizes the tests conducted and the results provide assurance that the device [classified as externally communicating, circulating blood, limited contact (≤ 24 hrs)] has a safe biocompatibility profile. The Toxikon reports are provided in [Appendix 7](#).

Biocompatibility Summary for SOFIA Distal Access Catheter

Test Method	Standard	Test Material	Sample Preparation	Results	Conclusion
Cytotoxicity					
L929 MEM Elution Test - ISO	ISO 10993-5 No deviations	Distal Access Catheter	Test material extracted at 37°C for 24 hrs in MEM (6cm ² / ml)	Grade 0 to 1(no to slight reactivity)	Non-cytotoxic (b)(4) Confidential and Proprietary
Sensitization					
Kligman Maximization Test - ISO	ISO 10993-10 No deviations	Distal Access Catheter	Test material extracted with NaCl and cottonseed oil at 70°C for 24 hrs at ratio of 6cm ² / mL.	Grade I: Weak allergic potential	Not significant (b)(4) Confidential and Proprietary Information
Irritation					
Intracutaneous Injection Test - ISO	ISO 10993-10 No deviations	Distal Access Catheter	Test material extracted with NaCl and cottonseed oil at 70°C for 24 hrs at ratio of 6cm ² / mL.	Comparative between control and test article = 0	Non-irritant (b)(4) Confidential and Proprietary
Hemocompatibility					
Hemolysis - Rabbit Blood ASTM (Direct and Indirect)	ISO 10993-4 No deviations	Distal Access Catheter	-Direct: Test material was added to PBS at 6 cm ² /mL -Indirect: Test material extracted in PBS at 70°C for 24 hrs. -Tests incubated with rabbit blood for 3 hrs at 37°C.	Hemolysis index: 0.13% (direct) and 0.0% (indirect)	Non-hemolytic (b)(4) Confidential and Proprietary
Unactivated Partial Thromboplastin Time Assay (Direct)	ISO 10993-4 No deviations	Distal Access Catheter	Human plasma exposed to test material (6 cm ² / mL) incubated at 37 °C for 15 min.	Test article = 88.9-92.4 seconds (Avg) (Not statistically different from negative and untreated)	No effect on coagulation (b)(4) Confidential and Proprietary Information
Complement Activation C3a and SC5b-9	ISO 10993-4 No deviations	Distal Access Catheter	Test material exposed to human plasma at ratio of 6cm ² /ml at 37 °C for 90 min.	C3a and SC5b-9 levels \leq negative and untreated controls	No effect on complement activation (b)(4) Confidential and Proprietary Information
Dog Thrombogenicity	ISO 10993-4 No deviations	Distal Access Catheter	Test article was surgically inserted into jugular vein for a period of 4 hours.	Grade (test art.) = 0 Grade (control) = 0	Not significant. (b)(4) Confidential and Proprietary Information

Test Method	Standard	Test Material	Sample Preparation	Results	Conclusion
Systemic Toxicity					
Systemic toxicity-ISO	ISO 10993-11 No deviations	Distal Access Catheter	Test material extracted with NaCl and cottonseed oil at 70°C for 24 hrs at ratio of 6cm ² / ml. Extracts injected in albino mice and monitored at 24, 48, and 72 hrs.	No significantly greater biological reaction than control.	No reaction (b)(4) Confidential and Proprietary Information
Rabbit Pyrogen Test (material mediated) - ISO	ISO 10993-11 No deviations	Distal Access Catheter	Test material extracted with NaCl at 70°C for 24 hrs at ratio of 6cm ² / ml. Extracts injected in rabbit and monitored for body temperature.	Temperature (max) increase was 0.3°C from baseline.	Non-pyrogenic (b)(4) Confidential and Proprietary Information

16. Packaging and Sterilization

The packaging for the SOFIA Catheter is fundamentally unchanged from the Chaperon with the addition of a dispenser tube that holds the catheter in place on the packaging card. As a result of the change, we have validated the packaging configuration and the results are documented in TR 12-288 (see [Appendix 6](#)).

Packaging Configuration

Packaging	Existing Chaperon	SOFIA Catheter
Material	1. Packaging card: Polyethylene 2. Pouch: Tyvek 3. Carton Box: Bleached Sulfate	1. Dispenser tube: HDPE 2. Packaging card: polyethylene 3. Pouch: Tyvek 4. Carton Box: Bleached Sulfate
Package Configuration	Catheter is placed on packaging card which is inserted into the pouch. The pouch is placed inside a carton box.	Catheter is place in a dispenser tube, placed on a packaging card that is then inserted into the pouch. Placed in carton box.
Method of Supplying	Sterile and single use.	Same
Method of Sterilization	Ethylene oxide	Same

The SOFIA Distal Access Catheter is intended to be sold sterile, for single use only. The device is sterilized using 100% ethylene oxide (EO) gas in the same manner as our existing Headway microcatheters. Sterilization is performed by an outside contractor, (b)(4) Confidential and Proprietary Information

The validation and routine EO sterilization method of the SOFIA is in accordance with the requirement of the ANSI/AAMI/ISO11135-2007, Medical Devices - Validation and routine control of ethylene oxide sterilization, overkill method.

Sterilization Summary	
Sterility Validation Method.	ANSI/AAMI/ISO11135-2007, Medical Devices-Validation and routine control of ethylene oxide sterilization, overkill method.
EO Residuals	EO and ECH residuals are below the limits established by ISO 10993-7 Biological evaluation of medical devices – Part 7: Ethylene oxide sterilization residuals.
Sterilization Method	100% ethylene oxide gas
Sterility Assurance Level	(SAL) – 10^{-6}
Sterilization Location	(b)(4) Confidential and Proprietary Information

17. Substantial Equivalence

The data presented in this submission demonstrates the technological similarity and equivalency of the SOFIA Distal Access Catheter when compared with the predicate device: Chaperon Guiding Catheter (K082385).

The devices,

- have the same intended use,
- use the same operating principle,
- incorporate the same basic design,
- are packaged and sterilized using same material and processes.

In summary, the SOFIA Distal Access Catheter described in this submission is, in our opinion, substantially equivalent to the predicate devices.

18. ISO/EC Certification and Compliance

MicroVention develops and manufactures their products under its certified quality system (ISO13485:2003 +AC:2009, CMDCAS). All MicroVention products are developed and tested based upon design control procedures that include risk analysis, in vitro, in vivo and clinical studies (as appropriate). The MicroVention facility is US FDA registered as well as licensed by the California State Department of Health.

Copy of the MicroVention ISO 13485 Certificate is provided in the [Appendix 8](#).

19. List of Appendices

Appendix 1	Product Labels, Instructions For Use
Appendix 2	Product Drawing
Appendix 3	QP 4.1, Design and Development Quality Procedure
Appendix 4	QP 4.8, Risk Management Quality Procedure
Appendix 5	RA 060004, Risk Assessment and Control Analysis FA060010, dFMEA
Appendix 6	TP/TR 12-280, Design Verification and Validation, Protocol and Report (with 6 month accelerated aging) TR12-288, Packaging Validation
Appendix 7	Biocompatibility Reports (Toxikon)
Appendix 8	MicroVention ISO Certificate

Appendix 1

SOFIA™ Distal Access Catheter

Instructions for Use

Carefully read all instructions prior to use.

DEVICE DESCRIPTION

The SOFIA™ Distal Access Catheter is a non-tapered, single-lumen, flexible catheter equipped with the coil and the braid reinforcement. The distal segment is steam-shapeable to facilitate vessel selection and also has a hydrophilic coating for navigation through the vasculatures. The radiopaque marker is located at the distal end of the catheter for visualization under fluoroscopy.

CONTENTS

One Distal Access Catheter
One Introducer Sheath
One Shaping Mandrel

INDICATIONS FOR USE

The SOFIA™ Distal Access Catheter is indicated for general intravascular use, including the neuro and peripheral vasculature. The SOFIA™ Distal Access Catheter can be used to facilitate introduction of diagnostic or therapeutic devices. The SOFIA™ Distal Access Catheter is not intended for use in coronary arteries.

CONTRAINDICATIONS

There are no known contraindications.

CAUTION

Rx Only: Federal (USA) law restricts this device to sale by or on the order of a physician.

Do not use if pouch is opened or damaged.

This device is intended for single use only. Do not reuse, reprocess or resterilize. Reuse, reprocessing or resterilization may compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury, illness, or death. Reuse, reprocessing, or resterilization may also create a risk of contamination of the device and/or cause patient infection or cross-infection, including, but not limited to, the transmission of infectious disease(s) from one patient to another. Contamination of the device may lead to injury, illness or death of the patient.

After use, dispose in accordance with hospital, administrative and/or local government policy.

WARNINGS

The SOFIA™ Distal Access Catheter should only be used by physicians who have received appropriate training in interventional techniques.

The SOFIA™ Distal Access Catheter is provided sterile and non-pyrogenic. Do not use if the packaging is breached or damaged.

Inspect the SOFIA™ Distal Access Catheter prior to use. Do not use the device if any damages or irregularities are observed.

Appropriate anti-coagulation and anti-platelet therapy should be administered per standard medical practice.

The SOFIA™ Distal Access Catheter should be manipulated under fluoroscopic guidance. Do not advance or withdraw the device when excessive resistance is met until the cause of resistance is determined.

Do not use the SOFIA™ Distal Access Catheter with Ethiodol or Lipiodol contrast media or other such contrast media which includes the components of those agents.

Do not use organic solvents as the device may be damaged.

Do not exceed 2070 kPa (300 psi) maximum recommended infusion pressure. Excess pressure may damage the device or injure the patient. Carefully monitor placement of the distal tip when using a power injector to infuse.

Do not make more than 90 degree angle on the Shaping Mandrel. Steaming of the distal tip with more than 90 degree angle may result in damage to the device.

Do not repeat steaming of the same device more than once, which may result in damage to the device.

Torquing the SOFIA™ Distal Access Catheter excessively while kinked may damage the device resulting in separation of the device. Withdraw the entire device (the device, microcatheter, and guidewire) if the device is severely kinked.

The Introducer Sheath is not intended for use inside the patient body. Ensure that the Introducer Sheath is removed from the SOFIA™ Distal Access Catheter once the distal shaft of the SOFIA™ Distal Access Catheter is placed inside the patient body.

PRECAUTIONS

Exercise care in handling the SOFIA™ Distal Access Catheter to reduce the chance of accidental damage.

Verify compatibility of the SOFIA™ Distal Access Catheter when using other ancillary devices commonly used in intravascular procedures. The physician must be familiar with percutaneous, intravascular techniques and possible complications associated with the procedure.

Use caution when manipulating the SOFIA™ Distal Access Catheter in tortuous vasculature to avoid damage. Avoid advancing or withdrawal against resistance until the cause of resistance is determined.

Presence of calcifications, irregularities, or other devices may damage the SOFIA™ Distal Access Catheter and potentially affect its insertion or removal.

Maintain perfusion of heparinized saline for inner lumen of the SOFIA™ Distal Access Catheter to prevent thrombus formation.

The hydrophilic coating on the SOFIA™ Distal Access Catheter should be hydrated with heparinized saline before use. Keep the coating hydrated and do not allow the coating to dry.

POTENTIAL COMPLICATIONS

Potential complications include, but are not limited to: vessel or aneurysm perforation, vasospasm, hematoma at the site of entry, embolism, ischemia, intracerebral/intracranial hemorrhage, pseudo aneurysm, seizure, stroke, infection, vessel dissection, thrombus formation, and death.

COMPATIBILITY

Refer to product label for device dimensions. Use the information on labeling provided with other devices to determine device compatibility.

PREPARATION FOR USE

1. Carefully remove the SOFIA™ Distal Access Catheter and the Introducer Sheath from the package.

2. Inspect the SOFIA™ Distal Access Catheter for any damage.

WARNING: Do not use the device if any damages or irregularities are observed.

If steam shaping is desired, use the technique outlined in the step 3. Otherwise proceed to the step 4.

3. **STEAM SHAPING**

- a. Bend the Shaping Mandrel for desired shape.

WARNING: Do not make more than 90 degree angle on the Shaping Mandrel. Steaming of the distal tip with more than 90 degree angle may result in damage to the device.

- b. Carefully insert the Shaping Mandrel into the distal tip of the SOFIA™ Distal Access Catheter.
- c. Hold the distal segment together with the Shaping Mandrel and steam it for 30 seconds.
- d. Immediately place the shaped distal segment into heparinized saline to set the shape.
- e. Inspect the distal shaft for any damage.

WARNING: Do not use the device if any damages or irregularities are observed.

- f. Remove the Shaping Mandrel from the SOFIA™ Distal Access Catheter.

Do not use the device if any damages or irregularities are observed.

WARNING: Do not repeat steaming of the same device more than once, which may result in damage to the device.

4. Flush the lumen of the SOFIA™ Distal Access Catheter with heparinized saline. Attach a rotating hemostatic valve (RHV) to the proximal hub of the SOFIA™ Distal Access Catheter. Set up the line for perfusion of heparinized saline through the sidearm of the RHV.
5. Hydrate the hydrophilic coating on the SOFIA™ Distal Access Catheter with heparinized saline before use. Keep the coating hydrated and do not allow the coating to dry.

DELIVERY OF THE HYBRID DISTAL ACCESS CATHETER

6. Go to the step 7 or 8, depending on the situation described below and choose appropriate devices for navigation of the SOFIA™ Distal Access Catheter.

7. **Navigation through the vasculature, except for the intracranial vasculature**

- a. Prepare 0.035" or 0.038" Guidewire for navigation of the SOFIA™ Distal Access Catheter.
- b. Insert the guidewire into the SOFIA™ Distal Access Catheter and advance the Guidewire until the Guidewire and the SOFIA™ Distal Access Catheter are aligned at the distal end.
- c. Using the Introducer Sheath provided in the package, carefully insert the SOFIA™ Distal Access Catheter and the Guidewire through a hemostatic valve of the femoral sheath
- d. Remove the Introducer Sheath from the SOFIA™ Distal Access Catheter once the distal shaft of the SOFIA™ Distal Access Catheter is placed inside the patient body.

WARNING: Introducer Sheath is not intended for use inside the patient body.

- e. Under fluoroscopic guidance, advance or withdraw the SOFIA™ Distal Access Catheter over the guidewire until desired position is attained or before the intracranial position is achieved. Select vessels by slowly torquing the SOFIA™ Distal Access Catheter if necessary.

WARNING: Do not advance or withdraw the device when excessive resistance is met until the cause of resistance is determined.

WARNING: Torquing the SOFIA™ Distal Access Catheter excessively while kinked may damage the device resulting in separation of the device. Withdraw the entire device (the device, Microcatheter, and Guidewire) if the device is severely kinked.

WARNING: Do not exceed 2070 kPa (300 psi) maximum recommended infusion pressure.

Excess pressure may damage the device or injure the patient. Carefully monitor placement of the distal tip when using a power injector to infuse.

- f. Go to the step 8 for navigation through the intracranial vasculatures. Otherwise proceed to the step 9.

8. **Navigation through the intracranial vasculature**

- a. Prepare Microcatheter and compatible Guidewire for navigation of the SOFIA™ Distal Access Catheter.
- b. Slowly remove, if any, devices previously inserted in the SOFIA™ Distal Access Catheter. Insert the Microcatheter with the Guidewire into the SOFIA™ Distal Access Catheter.

- c. Under fluoroscopic guidance, advance or withdraw the SOFIA™ Distal Access Catheter over the Microcatheter and the Guidewire until desired position is attained. Select vessels by slowly torquing the SOFIA™ Distal Access Catheter if necessary.

WARNING: Do not advance or withdraw the device when excessive resistance is met until the cause of resistance is determined.

WARNING: Torquing the SOFIA™ Distal Distal Access Catheter excessively while kinked may damage the device resulting in separation of the device. Withdraw the entire device (the device, Microcatheter, and Guidewire) if the device is severely kinked.

WARNING: Do not exceed 2070 kPa (300 psi) maximum recommended infusion pressure. Excess pressure may damage the device or injure the patient. Carefully monitor placement of the distal tip when using a power injector to infuse.

- 9. Slowly remove the Guidewire or the Microcatheter if necessary. Make sure that continuous perfusion of heparinized saline is maintained through the sidearm of the RHV.

NOTE: The Microcatheter used to navigate the SOFIA™ Distal Access Catheter may be kept for the rest of procedure.

The physician has the discretion to modify described manipulations of the SOFIA™ Distal Access Catheter to accommodate the complexity and variation in procedures. Any technique modification must be consistent with previously described instructions, warnings, precautions and patient safety information.













STORAGE

Avoid exposure to water, sunlight, extreme temperatures and high humidity during storage. Store the SOFIA™ Distal Access Catheter under controlled room temperature. See the product label for the device shelf life. Do not use the device beyond the labeled shelf life.

MATERIALS

The SOFIA™ Distal Access Catheter does not contain latex or PVC materials.

SYMBOLS

	Lot Number		Do Not Reuse
	Catalog Number		Attention, Consult Accompanying Documents
	Contents		Use by Date
	Sterilized Using Ethylene Oxide		Date of Manufacture
	CE Mark		Manufacturer
	Authorized European Representative		Non-pyrogenic

WARRANTY

MicroVention, Inc. warrants that reasonable care has been used in the design and manufacture of this device. This warranty is in lieu of and excludes all other warranties not expressly set forth herein, whether expressed or implied by operation of law or otherwise, including, but not limited to, any implied warranties of merchantability or fitness. Handling, storage, cleaning and sterilization of the device as well as factors relating to the patient, diagnosis, treatment, surgical procedure and other matters beyond MicroVention’s control directly affect the device and the results obtained from its use. MicroVention’s obligation under this warranty is limited to the repair or replacement of this device and MicroVention shall not be liable for any incidental or consequential loss, damage or expense directly or indirectly arising from the use of this device. MicroVention neither assumes, nor authorizes any other person to assume for it, any other or additional liability or responsibility in connection with this device. MicroVention assumes no liability with respect to devices reused, reprocessed or resterilized and makes no warranties, expressed or implied, including, but not limited to, merchantability or fitness for intended use, with respect to such device.

Prices, specifications and model availability are subject to change without notice.



EC REP

Manufacturer:

MicroVention, Inc.
1311 Valencia Avenue
Tustin, CA 92780 USA
Tel: (714) 247-8000

www.microvention.com

CE 0297

Authorized European Representative:

MicroVention Europe
30 bis, rue du Vieil Abreuvoir
78100 Saint-Germain-en-Laye
France
Tel: +33 (0)1 39 21 77 46
Fax: +33 (0)1 39 21 16 01

PD110436 Rev. A
2013-2

© Copyright 2013 MicroVention, Inc. All rights reserved.

5F

SOFIA

REF "Catalog No."
Use By
YYYY-MM

125 cm / STR

"One Long Bar Code with Bc1 and BC2 information"


SOFIA

 Distal Access Catheter

Distal Access Catheter

Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter

Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter
Distal Access Catheter



REF Catalog Number
DA5125ST

LOT Lot Number
"Lot No."

<p>"Bar Code 1" REF: "Catalog No." "Bar Code 2" LOT NO: "Lot No."</p>	<p>SOFIA Distal Access Catheter 125 cm / STR</p>
<p>"Bar Code 1" REF: "Catalog No." "Bar Code 2" LOT NO: "Lot No."</p>	<p>SOFIA Distal Access Catheter 125 cm / STR</p>
<p>"Bar Code 1" REF: "Catalog No." "Bar Code 2" LOT NO: "Lot No."</p>	<p>SOFIA Distal Access Catheter 125 cm / STR</p>

STERILE EO Attention: Refer to Instructions For Use. ~~Do Not Reuse.~~ Rx-ONLY Non-pyrogenic

MicroVention, Inc.
1311 Valencia Avenue
Tustin, CA 92780 USA
PH: 714.247.8000
www.microvention.com

CONT Contents
1 Distal Access Catheter
1 Introducer Sheath
1 Shaping Mandrel

EC REP **MicroVention Europe**
30 bis, rue du Vieil Abreuvoir
78100 Saint-Germain-en-Laye
France

LOT "XXXXXXXXX"

Date of Manufacture YYYY-MM

Use By YYYY-MM


CE 0297
Made in U.S.A.

EDXXXX-XXX

LB11192A-DRAFT

AW11020 Rev. A 2013-04 (LB11190)

Appendix 2



This document contains Confidential and Proprietary Information of
 Microvention, Inc. All rights reserved. No part of this document may be reproduced
 without the prior written permission of Microvention, Inc. All other rights reserved.
 Also Verigo
 75, Columbia
 CA, 92058

MICROVENTION
TERUMO *innovating for life.*

**FINAL PACKAGING ASSY,
 HAC**

DWG. NO. PD060043 REV. X23
 SCALE: 1:1 DO NOT SCALE DRAWINGS SHEET 2 of 2

Appendix 3

Appendix 5

Appendix 6

Appendix 7

Appendix 8



CERTIFICATE



This is to certify that the company

MicroVention, Inc.

1311 Valencia Ave.
Tustin, CA 92780
United States of America

with the organizational units/sites as listed in the annex

has implemented and maintains a **Quality Management System**.

Scope:

Design, Development, Manufacturing and Distribution of Embolization Prostheses and Accessories, and Intravascular Access Devices and Accessories, Clot and Foreign Body Retrieval Devices

Through an audit, documented in a report, performed by DQS Medizinprodukte GmbH, it was verified that the management system fulfills the requirements of the following standard:

EN ISO 13485 : 2003 + AC : 2009

Certificate registration No.	411133 MP29
Certificate unique ID	170555867
Effective date	2012-10-12
Expiry date	2014-10-31
Frankfurt am Main	2012-10-12



DQS Medizinprodukte GmbH

Frank Graichen
Managing Director

Dr. Thomas Feldmann
Head of Certification Body



Annex to Certificate
Certificate registration No.: 411133 MP29
Certificate unique ID: 170555867
Effective date: 2012-10-12

MicroVention, Inc.

1311 Valencia Ave.
Tustin, CA 92780
United States of America

Location

Scope

MicroVention, Inc.
1311 Valencia Ave.
Tustin, CA 92780
United States of America

Design, Development, Manufacturing and Distribution of Embolization Prostheses and Accessories, and Intravascular Access Devices and Accessories, Clot and Foreign Body Retrieval Devices

MicroVention, Inc.
Production Site
75 Columbia
Aliso Viejo, CA 92656
United States of America

Manufacturing of Embolization Prostheses and Accessories, and Intravascular Access Devices and Accessories, Clot and Foreign Body Retrieval Devices

MicroVention Costa Rica, S.R.L.
Production Site
Zona Franca Coyol
Alajuela
Costa Rica

Manufacturing of Embolization Prostheses and Accessories, and Intravascular Access Devices and Accessories, Clot and Foreign Body Retrieval Devices

MicroVention, Inc.
Distribution Site
1800 E. Wilshire Ave.
Santa Ana 92705
United States of America

Distribution of Embolization Prostheses and Accessories, and Intravascular Access Devices and Accessories, Clot and Foreign Body Retrieval Devices





EC Design-Examination Certificate

DQS GmbH

Deutsche Gesellschaft zur Zertifizierung von Managementsystemen

hereby certifies to the manufacturer

MicroVention Inc.

1311 Valencia Ave.
Tustin, CA 92780
United States of America

that following product

Chaperon Guiding Catheter System

is conform to the

essential requirements of Annex I of the Council Directive 93/42/EEC of 14 June 1993 concerning medical devices

Examination Basis: Chaperon Dossier Summary_2008-Sep-14.pdf

The further examination basis is detailed in the following named report and the relating documents.

Examination reports:

Report Design Examination Chaperon Rev 2008-09-14.doc 07/2008

The results of the examination are detailed in the above mentioned report(s) and the therein named relating documents.

Validity:

This EC Design-Examination Certificate is based on section 4 of Annex II of the Council Directive 93/42/EEC of 14 June 1993 concerning medical devices and is only valid in connection with the valid DQS Certificate No.: 411133 MR2 according to Annex II of the Council Directive 93/42/EEC. Changes to the approved design must receive further approval from the notified body.

This certificate is valid until 2013-10-22

Certificate Registration No. 428948 MRA

Frankfurt am Main 2009-07-21



Ass. iur. M. Drechsel

MANAGING DIRECTORS

D-60433 Frankfurt am Main, August-Schanz-Straße 21



Dipl.-Bw. J. Böge

Number of DQS as notified body according to Council Directive 93/42EWG: 0297

Napperli, Jesse *

From: Napperli, Jesse *
Sent: Monday, November 25, 2013 4:24 PM
To: 'NAOMI.GONG@MICROVENTION.COM'
Cc: DCCLetters
Subject: K131482 SE LETTER
Attachments: K131482.pdf



COVER SHEET MEMORANDUM

Food and Drug Administration
Office of Device Evaluation &
Office of In Vitro Diagnostics and
Radiological Health

From: Reviewer Name Samuel K. Shimp III
Subject: 510(k) Number K131482
To: The Record

Please list CTS decision code: SE - Substantially Equivalent

- Refused to Accept (Note: this is considered the first review cycle. See screening checklist.)
- Hold (Additional Information or Telephone Hold)
- Final Decision (SE, SE with Limitations, NSE (select code below), Withdrawn, etc.)

Please complete the following for a final clearance decision (i.e, SE, SE with Limitations, etc.)	YES	NO
Indications for Use Page (<i>Attach IFU</i>)	X	
510(k) Summary or 510(k) Statement (<i>Attach Summary or Statement</i>)	X	
Truthful and Accurate Statement (<i>Must be present for a Final Decision</i>)	X	
Is the device Class III?		X
Does firm reference standards? (If yes, please attach <u>Form 3654</u> .)	X	
Is this a combination product?		X
Is this a reprocessed single use device? (See <u>Guidance for Industry and FDA Staff - MDUFMA - Validation Data in 510(k)s for Reprocessed Single-Use Medical Devices</u> .)		X
Is this device intended for pediatric use only?		X
Is this a prescription device? (If both prescription & OTC, check both boxes.)	X	
Is clinical data necessary to support the review of this 510(k)?		X
For United States based clinical studies only, did the application include a completed Form FDA 3674, Certification with Requirements of ClinicalTrials.gov Data Bank? (If study was conducted in the United States and Form FDA 3674 was not included or was incomplete, then applicant must be contacted to obtain completed form.)		X
Does this device include an Animal Tissue Source?		X
All Pediatric Patients age <= 21		X
Neonate/Newborn (Birth to 28 days)		X
Infant (29 days to < 2 years)		X
Child (2 years to <12 years)		X
Adolescent (12 years to <18 years)		X
Transitional Adolescent A (18 years to <21 years); Special considerations are being given to this group, different from adults age >= 21 (different device design or testing, different protocol procedures, etc.)		X
Transitional Adolescent B (18 years to <21 years); No special considerations compared to adults >= 21 years)		X

Nanotechnology		×
Is this device subject to the Tracking Regulation? (Medical Device Tracking Guidance)		×

Regulation Number: 870.1250
Class: II
Product Code: DQY
Additional Product Codes: DQO

Digital Signature Concurrence Table
(Not all signatures may be required)

Branch Chief Sign-Off	Quynh T. Hoang / S 2013.11.21 18:25:48 -05'00'
Division Sign-Off	Joyce M. Whang / S 2013.11.22 12:17:40 -05'00'

K131482-3001



FDA CDRH DMC

JUN 12 2013

Received

June 10, 2013

U.S. Food and Drug Administration
Center for Devices and Radiological Health
IDE Document Mail Center – WO66-G609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

RE: K131482 RTA1 Response
For SOFIA Catheter

Dear Dr. Shimp:

This is submitted in response to RTA notification (dated June 7, 2013) for the Special 510(k) for the SOFIA Catheter.

An eCopy can be found on the included CD and it is an exact duplicate of the paper copy. Please contact me if additional information is needed.

Statement of Confidentiality: MicroVention, Inc. considers the information in this submission to be confidential information. We ask that this proprietary information herein be treated as confidential in accordance with the Freedom of Information Act.

Sincerely,

Naomi Gong 6-11-2013

Naomi Gong
Regulatory Affairs Project Manager
MicroVention, Inc.
Ph: 714-247-8055
Fx: 714-247-8014
naomi.gong@microvention.com

37



K131482/81

June 10, 2013

U.S. Food and Drug Administration
Center for Devices and Radiological Health
IDE Document Mail Center – WO66-G609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

FDA CDRH DMC
JUN 11 2013
Received

RE: K131482 RTA1 Response
For SOFIA Catheter

Dear Dr. Shimp:

This is submitted in response to RTA notification (dated June 7, 2013) for the Special 510(k) for the SOFIA Catheter.


An eCopy can be found on the included CD and it is an exact duplicate of the paper copy. Please contact me if additional information is needed.

Statement of Confidentiality: MicroVention, Inc. considers the information in this submission to be confidential information. We ask that this proprietary information herein be treated as confidential in accordance with the Freedom of Information Act.

Sincerely,

Naomi Gong
Regulatory Affairs Project Manager
MicroVention, Inc.
Ph: 714-247-8055
Fx: 714-247-8014
naomi.gong@microvention.com

18



June 10, 2013

U.S. Food and Drug Administration
Center for Devices and Radiological Health
IDE Document Mail Center – WO66-G609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

RE: K131482 RTA1 Response
For SOFIA Catheter

Dear Dr. Shimp:

This is submitted in response to RTA notification (dated June 7, 2013) for the Special 510(k) for the SOFIA Catheter.

An eCopy can be found on the included CD and it is an exact duplicate of the paper copy. Please contact me if additional information is needed.

Statement of Confidentiality: MicroVention, Inc. considers the information in this submission to be confidential information. We ask that this proprietary information herein be treated as confidential in accordance with the Freedom of Information Act.

Sincerely,

Naomi Gong
Regulatory Affairs Project Manager
MicroVention, Inc.
Ph: 714-247-8055
Fx: 714-247-8014
naomi.gong@microvention.com

From the Elements of a Complete Submission (RTA Items), please see our responses to the identified items noted by the FDA reviewer:

Section 8a –

FDA Comments: The submission does not state whether or not there were prior submissions. Section F of form 3514 was left blank which is not adequate for declaring that there were no prior submissions. You need to provide a statement declaring whether or not there were prior submissions related to the subject device.

MicroVention Response: An updated Form-3514 is included with Section F completed is provided in [Appendix A](#).

Section 18a – Proposed Labeling, all changes in proposed labeling resulting from device modification(s) are highlighted or prominently identified.

FDA Comments: Only a clean copy of the labeling is provided. The labeling provided does not identify any changes from the previous version of the labeling. There is no narrative description or highlighted/redlined version of the labeling to prominently identify the proposed labeling changes.

MicroVention Response: A copy of the proposed IFU with highlighted (in yellow) changes resulting from device modification(s) is provided in [Appendix B](#).

This information should fulfill the elements outlined in the administrative review of a Special 510(k) and remove the current RTA hold status of the submission.

Appendix A

CDRH PREMARKET REVIEW SUBMISSION COVER SHEET

Date of Submission 05/21/2013	User Fee Payment ID Number (b) (4)	FDA Submission Document Number (if known)
----------------------------------	--	---

SECTION A TYPE OF SUBMISSION

PMA <input type="checkbox"/> Original Submission <input type="checkbox"/> Premarket Report <input type="checkbox"/> Modular Submission <input type="checkbox"/> Amendment <input type="checkbox"/> Report <input type="checkbox"/> Report Amendment <input type="checkbox"/> Licensing Agreement	PMA & HDE Supplement <input type="checkbox"/> Regular (180 day) <input type="checkbox"/> Special <input type="checkbox"/> Panel Track (PMA Only) <input type="checkbox"/> 30-day Supplement <input type="checkbox"/> 30-day Notice <input type="checkbox"/> 135-day Supplement <input type="checkbox"/> Real-time Review <input type="checkbox"/> Amendment to PMA & HDE Supplement <input type="checkbox"/> Other	PDP <input type="checkbox"/> Original PDP <input type="checkbox"/> Notice of Completion <input type="checkbox"/> Amendment to PDP	510(k) <input checked="" type="checkbox"/> Original Submission: <input type="checkbox"/> Traditional <input checked="" type="checkbox"/> Special <input type="checkbox"/> Abbreviated (Complete section I, Page 5) <input type="checkbox"/> Additional Information <input type="checkbox"/> Third Party	Request for Feedback <input type="checkbox"/> Pre-Submission <input type="checkbox"/> Informational Meeting <input type="checkbox"/> Submission Issue Meeting <input type="checkbox"/> Day 100 Meeting <input type="checkbox"/> Agreement Meeting <input type="checkbox"/> Determination Meeting <input type="checkbox"/> Study Risk Determination <input type="checkbox"/> Other (specify):
IDE <input type="checkbox"/> Original Submission <input type="checkbox"/> Amendment <input type="checkbox"/> Supplement	Humanitarian Device Exemption (HDE) <input type="checkbox"/> Original Submission <input type="checkbox"/> Amendment <input type="checkbox"/> Supplement <input type="checkbox"/> Report <input type="checkbox"/> Report Amendment	Class II Exemption Petition <input type="checkbox"/> Original Submission <input type="checkbox"/> Additional Information	Evaluation of Automatic Class III Designation (De Novo) <input type="checkbox"/> Original Submission <input type="checkbox"/> Additional Information	Other Submission <input type="checkbox"/> 513(g) <input type="checkbox"/> Other (describe submission):

Have you used or cited Standards in your submission? Yes No (If Yes, please complete Section I, Page 5)

SECTION B SUBMITTER, APPLICANT OR SPONSOR

Company / Institution Name MicroVention, Inc.		Establishment Registration Number (if known) 2032493	
Division Name (if applicable)		Phone Number (including area code) 714-247-8055	
Street Address 1311 Valencia Avenue		FAX Number (including area code) 714-247-8014	
City Tustin	State / Province CA	ZIP/Postal Code 92780	Country USA
Contact Name Naomi Gong			
Contact Title Sr. Regulatory Affairs Project Manager		Contact E-mail Address naomi.gong@microvention.com	

SECTION C APPLICATION CORRESPONDENT (e.g., consultant, if different from above)

Company / Institution Name			
Division Name (if applicable)		Phone Number (including area code)	
Street Address		FAX Number (including area code)	
City	State / Province	ZIP Code	Country
Contact Name			
Contact Title		Contact E-mail Address	

SECTION D1

REASON FOR APPLICATION - PMA, PDP, OR HDE

<input type="checkbox"/> New Device <input type="checkbox"/> Withdrawal <input type="checkbox"/> Additional or Expanded Indications <input type="checkbox"/> Request for Extension <input type="checkbox"/> Post-approval Study Protocol <input type="checkbox"/> Request for Applicant Hold <input type="checkbox"/> Request for Removal of Applicant Hold <input type="checkbox"/> Request to Remove or Add Manufacturing Site	<input type="checkbox"/> Change in design, component, or specification: <input type="checkbox"/> Software / Hardware <input type="checkbox"/> Color Additive <input type="checkbox"/> Material <input type="checkbox"/> Specifications <input type="checkbox"/> Other (<i>specify below</i>)	<input type="checkbox"/> Location change: <input type="checkbox"/> Manufacturer <input type="checkbox"/> Sterilizer <input type="checkbox"/> Packager
<input type="checkbox"/> Process change: <input type="checkbox"/> Manufacturing <input type="checkbox"/> Packaging <input type="checkbox"/> Sterilization <input type="checkbox"/> Other (<i>specify below</i>)	<input type="checkbox"/> Labeling change: <input type="checkbox"/> Indications <input type="checkbox"/> Instructions <input type="checkbox"/> Performance Characteristics <input type="checkbox"/> Shelf Life <input type="checkbox"/> Trade Name <input type="checkbox"/> Other (<i>specify below</i>)	<input type="checkbox"/> Report Submission: <input type="checkbox"/> Annual or Periodic <input type="checkbox"/> Post-approval Study <input type="checkbox"/> Adverse Reaction <input type="checkbox"/> Device Defect <input type="checkbox"/> Amendment
<input type="checkbox"/> Response to FDA correspondence:		<input type="checkbox"/> Change in Ownership <input type="checkbox"/> Change in Correspondent <input type="checkbox"/> Change of Applicant Address
<input type="checkbox"/> Other Reason (<i>specify</i>):		

SECTION D2

REASON FOR APPLICATION - IDE

<input type="checkbox"/> New Device <input type="checkbox"/> New Indication <input type="checkbox"/> Addition of Institution <input type="checkbox"/> Expansion / Extension of Study <input type="checkbox"/> IRB Certification <input type="checkbox"/> Termination of Study <input type="checkbox"/> Withdrawal of Application <input type="checkbox"/> Unanticipated Adverse Effect <input type="checkbox"/> Notification of Emergency Use <input type="checkbox"/> Compassionate Use Request <input type="checkbox"/> Treatment IDE <input type="checkbox"/> Continued Access	<input type="checkbox"/> Change in: <input type="checkbox"/> Correspondent / Applicant <input type="checkbox"/> Design / Device <input type="checkbox"/> Informed Consent <input type="checkbox"/> Manufacturer <input type="checkbox"/> Manufacturing Process <input type="checkbox"/> Protocol - Feasibility <input type="checkbox"/> Protocol - Other <input type="checkbox"/> Sponsor	<input type="checkbox"/> Response to FDA Letter Concerning: <input type="checkbox"/> Conditional Approval <input type="checkbox"/> Deemed Approved <input type="checkbox"/> Deficient Final Report <input type="checkbox"/> Deficient Progress Report <input type="checkbox"/> Deficient Investigator Report <input type="checkbox"/> Disapproval <input type="checkbox"/> Request Extension of Time to Respond to FDA <input type="checkbox"/> Request Meeting <input type="checkbox"/> Request Hearing
	<input type="checkbox"/> Report submission: <input type="checkbox"/> Current Investigator <input type="checkbox"/> Annual Progress Report <input type="checkbox"/> Site Waiver Report <input type="checkbox"/> Final	
<input type="checkbox"/> Other Reason (<i>specify</i>):		

SECTION D3

REASON FOR SUBMISSION - 510(k)

<input type="checkbox"/> New Device	<input type="checkbox"/> Additional or Expanded Indications	<input type="checkbox"/> Change in Technology
<input checked="" type="checkbox"/> Other Reason (<i>specify</i>): Catheter is similar to predicate device with same intended use (indications for use)		

SECTION E ADDITIONAL INFORMATION ON 510(k) SUBMISSIONS

Product codes of devices to which substantial equivalence is claimed								Summary of, or statement concerning, safety and effectiveness information <input checked="" type="checkbox"/> 510 (k) summary attached <input type="checkbox"/> 510 (k) statement
1	DQY	2		3		4		
5		6		7		8		

Information on devices to which substantial equivalence is claimed (if known)

	510(k) Number	Trade or Proprietary or Model Name	Manufacturer
1	K082385	Chaperon Guiding Catheter	MicroVention, Inc. 1311 Valencia Avenue, Tustin, CA 92780
2			
3			
4			
5			
6			

SECTION F PRODUCT INFORMATION - APPLICATION TO ALL APPLICATIONS

Common or usual name or classification name
Percutaneous catheter

	Trade or Proprietary or Model Name for This Device	Model Number
1	SOFIA Distal Access Catheter	DA5125ST
2		
3		
4		
5		

FDA document numbers of all prior related submissions (regardless of outcome)

1	None	2		3		4		5		6	
7		8		9		10		11		12	

Data Included in Submission
 Laboratory Testing Animal Trials Human Trials

SECTION G PRODUCT CLASSIFICATION - APPLICATION TO ALL APPLICATIONS

Product Code DQY	C.F.R. Section (if applicable) 21 CFR 870.1250	Device Class <input type="checkbox"/> Class I <input checked="" type="checkbox"/> Class II <input type="checkbox"/> Class III <input type="checkbox"/> Unclassified
Classification Panel Cardiovascular Devices		

Indications (from labeling)
The SOFIA Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

SECTION H MANUFACTURING / PACKAGING / STERILIZATION SITES RELATING TO A SUBMISSION

<input checked="" type="checkbox"/> Original <input type="checkbox"/> Add <input type="checkbox"/> Delete		Facility Establishment Identifier (FEI) Number 2032493		<input checked="" type="checkbox"/> Manufacturer <input type="checkbox"/> Contract Sterilizer <input type="checkbox"/> Contract Manufacturer <input type="checkbox"/> Repackager / Relabeler	
Company / Institution Name MicroVention, Inc.			Establishment Registration Number 2032493		
Division Name (if applicable)			Phone Number (including area code) 714-247-8000		
Street Address 1311 Valencia Avenue			FAX Number (including area code) 714-247-8005		
City Tustin		State / Province CA	ZIP Code 92780	Country USA	
Contact Name Naomi Gong		Contact Title Sr. RA Project Manager		Contact E-mail Address naomi.gong@microvention.com	

<input checked="" type="checkbox"/> Original <input type="checkbox"/> Add <input type="checkbox"/> Delete		Facility Establishment Identifier (FEI) Number		<input type="checkbox"/> Manufacturer <input checked="" type="checkbox"/> Contract Sterilizer <input type="checkbox"/> Contract Manufacturer <input type="checkbox"/> Repackager / Relabeler	
Company / Institution Name Sterigenics			Establishment Registration Number 2011171		
Division Name (if applicable)			Phone Number (including area code) 951-340-0700		
Street Address 4900 South Griffith Avenue			FAX Number (including area code)		
City Los Angeles		State / Province CA	ZIP Code 90058	Country USA	
Contact Name Sharon Huges		Contact Title Representative		Contact E-mail Address losangelessales@sterigenics.com	

<input type="checkbox"/> Original <input type="checkbox"/> Add <input type="checkbox"/> Delete		Facility Establishment Identifier (FEI) Number		<input type="checkbox"/> Manufacturer <input type="checkbox"/> Contract Sterilizer <input type="checkbox"/> Contract Manufacturer <input type="checkbox"/> Repackager / Relabeler	
Company / Institution Name			Establishment Registration Number		
Division Name (if applicable)			Phone Number (including area code)		
Street Address			FAX Number (including area code)		
City		State / Province	ZIP Code	Country	
Contact Name		Contact Title		Contact E-mail Address	

SECTION I

UTILIZATION OF STANDARDS

Note: Complete this section if your application or submission cites standards or includes a "Declaration of Conformity to a Recognized Standard" statement.

	Standards No.	Standards Organization	Standards Title	Version	Date
1					
2					
3					
4					
5					
6					
7					

Please include any additional standards to be cited on a separate page.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF ADDRESS BELOW.

The burden time for this collection of information is estimated to average 0.5 hour per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
 Food and Drug Administration
 Office of Chief Information Officer
 Paperwork Reduction Act (PRA) Staff
 1350 Piccard Drive, Room 400
 Rockville, MD 20850

An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.

Appendix B

SOFIA™ Distal Access Catheter

Instructions for Use

Carefully read all instructions prior to use.

DEVICE DESCRIPTION

The SOFIA™ Distal Access Catheter is a non-tapered, single-lumen, flexible catheter equipped with the coil and the braid reinforcement. The distal segment is steam-shapeable to facilitate vessel selection and also has a hydrophilic coating for navigation through the vasculatures. The radiopaque marker is located at the distal end of the catheter for visualization under fluoroscopy.

CONTENTS

One Distal Access Catheter
One Introducer Sheath
One Shaping Mandrel

INDICATIONS FOR USE

The SOFIA™ Distal Access Catheter is indicated for general intravascular use, including the neuro and peripheral vasculature. The SOFIA™ Distal Access Catheter can be used to facilitate introduction of diagnostic or therapeutic devices. The SOFIA™ Distal Access Catheter is not intended for use in coronary arteries.

CONTRAINDICATIONS

There are no known contraindications.

CAUTION

Rx Only: Federal (USA) law restricts this device to sale by or on the order of a physician.

Do not use if pouch is opened or damaged.

This device is intended for single use only. Do not reuse, reprocess or resterilize. Reuse, reprocessing or resterilization may compromise the structural integrity of the device and/or lead to device failure which, in turn, may result in patient injury, illness, or death. Reuse, reprocessing, or resterilization may also create a risk of contamination of the device and/or cause patient infection or cross-infection, including, but not limited to, the transmission of infectious disease(s) from one patient to another. Contamination of the device may lead to injury, illness or death of the patient.

After use, dispose in accordance with hospital, administrative and/or local government policy.

WARNINGS

The SOFIA™ Distal Access Catheter should only be used by physicians who have received appropriate training in interventional techniques.

The SOFIA™ Distal Access Catheter is provided sterile and non-pyrogenic. Do not use if the packaging is breached or damaged.

Inspect the SOFIA™ Distal Access Catheter prior to use. Do not use the device if any damages or irregularities are observed.

Appropriate anti-coagulation and anti-platelet therapy should be administered per standard medical practice.

The SOFIA™ Distal Access Catheter should be manipulated under fluoroscopic guidance. Do not advance or withdraw the device when excessive resistance is met until the cause of resistance is determined.

Do not use the SOFIA™ Distal Access Catheter with Ethiodol or Lipiodol contrast media or other such contrast media which includes the components of those agents.

Do not use organic solvents as the device may be damaged.

Do not exceed 2070 kPa (300 psi) maximum recommended infusion pressure. Excess pressure may damage the device or injure the patient. Carefully monitor placement of the distal tip when using a power injector to infuse.

Do not make more than 90 degree angle on the Shaping Mandrel. Steaming of the distal tip with more than 90 degree angle may result in damage to the device.

Do not repeat steaming of the same device more than once, which may result in damage to the device.

Torquing the SOFIA™ Distal Access Catheter excessively while kinked may damage the device resulting in separation of the device. Withdraw the entire device (the device, microcatheter, and guidewire) if the device is severely kinked.

The Introducer Sheath is not intended for use inside the patient body. Ensure that the Introducer Sheath is removed from the SOFIA™ Distal Access Catheter once the distal shaft of the SOFIA™ Distal Access Catheter is placed inside the patient body.

PRECAUTIONS

Exercise care in handling the SOFIA™ Distal Access Catheter to reduce the chance of accidental damage.

Verify compatibility of the SOFIA™ Distal Access Catheter when using other ancillary devices commonly used in intravascular procedures. The physician must be familiar with percutaneous, intravascular techniques and possible complications associated with the procedure.

Use caution when manipulating the SOFIA™ Distal Access Catheter in tortuous vasculature to avoid damage. Avoid advancing or withdrawal against resistance until the cause of resistance is determined.

Presence of calcifications, irregularities, or other devices may damage the SOFIA™ Distal Access Catheter and potentially affect its insertion or removal.

Maintain perfusion of heparinized saline for inner lumen of the SOFIA™ Distal Access Catheter to prevent thrombus formation.

The hydrophilic coating on the SOFIA™ Distal Access Catheter should be hydrated with heparinized saline before use. Keep the coating hydrated and do not allow the coating to dry.

POTENTIAL COMPLICATIONS

Potential complications include, but are not limited to: vessel or aneurysm perforation, vasospasm, hematoma at the site of entry, embolism, ischemia, intracerebral/intracranial hemorrhage, pseudo aneurysm, seizure, stroke, infection, vessel dissection, thrombus formation, and death.

COMPATIBILITY

Refer to product label for device dimensions. Use the information on labeling provided with other devices to determine device compatibility.

PREPARATION FOR USE

1. Carefully remove the SOFIA™ Distal Access Catheter and the Introducer Sheath from the package.

2. Inspect the SOFIA™ Distal Access Catheter for any damage.

WARNING: Do not use the device if any damages or irregularities are observed.

If steam shaping is desired, use the technique outlined in the step 3. Otherwise proceed to the step 4.

3. STEAM SHAPING

- a. Bend the Shaping Mandrel for desired shape.

WARNING: Do not make more than 90 degree angle on the Shaping Mandrel. Steaming of the distal tip with more than 90 degree angle may result in damage to the device.

- b. Carefully insert the Shaping Mandrel into the distal tip of the SOFIA™ Distal Access Catheter.
- c. Hold the distal segment together with the Shaping Mandrel and steam it for 30 seconds.
- d. Immediately place the shaped distal segment into heparinized saline to set the shape.
- e. Inspect the distal shaft for any damage.

WARNING: Do not use the device if any damages or irregularities are observed.

- f. Remove the Shaping Mandrel from the SOFIA™ Distal Access Catheter.

Do not use the device if any damages or irregularities are observed.

WARNING: Do not repeat steaming of the same device more than once, which may result in damage to the device.

4. Flush the lumen of the SOFIA™ Distal Access Catheter with heparinized saline. Attach a rotating hemostatic valve (RHV) to the proximal hub of the SOFIA™ Distal Access Catheter. Set up the line for perfusion of heparinized saline through the sidearm of the RHV.
5. Hydrate the hydrophilic coating on the SOFIA™ Distal Access Catheter with heparinized saline before use. Keep the coating hydrated and do not allow the coating to dry.

DELIVERY OF THE HYBRID DISTAL ACCESS CATHETER

6. Go to the step 7 or 8, depending on the situation described below and choose appropriate devices for navigation of the SOFIA™ Distal Access Catheter.

7. Navigation through the vasculature, except for the intracranial vasculature

- a. Prepare 0.035" or 0.038" Guidewire for navigation of the SOFIA™ Distal Access Catheter.
- b. Insert the guidewire into the SOFIA™ Distal Access Catheter and advance the Guidewire until the Guidewire and the SOFIA™ Distal Access Catheter are aligned at the distal end.
- c. **Using the Introducer Sheath provided in the package**, carefully insert the SOFIA™ Distal Access Catheter and the Guidewire through a hemostatic valve of the femoral sheath
- d. Remove the Introducer Sheath from the SOFIA™ Distal Access Catheter once the distal shaft of the SOFIA™ Distal Access Catheter is placed inside the patient body.
WARNING: **Introducer Sheath is not intended for use inside the patient body.**
- e. Under fluoroscopic guidance, advance or withdraw the SOFIA™ Distal Access Catheter over the guidewire until desired position is attained or before the intracranial position is achieved. Select vessels by slowly torquing the SOFIA™ Distal Access Catheter if necessary.
WARNING: Do not advance or withdraw the device when excessive resistance is met until the cause of resistance is determined.
WARNING: Torquing the SOFIA™ Distal Access Catheter excessively while kinked may damage the device resulting in separation of the device. Withdraw the entire device (the device, Microcatheter, and Guidewire) if the device is severely kinked.
WARNING: **Do not exceed 2070 kPa (300 psi) maximum recommended infusion pressure.** Excess pressure may damage the device or injure the patient. Carefully monitor placement of the distal tip when using a power injector to infuse.
- f. Go to the step 8 for navigation through the intracranial vasculatures. Otherwise proceed to the step 9.

8. Navigation through the intracranial vasculature

- a. Prepare Microcatheter and compatible Guidewire for navigation of the SOFIA™ Distal Access Catheter.
- b. Slowly remove, if any, devices previously inserted in the SOFIA™ Distal Access Catheter. Insert the Microcatheter with the Guidewire into the SOFIA™ Distal Access Catheter.

- c. Under fluoroscopic guidance, advance or withdraw the SOFIA™ Distal Access Catheter over the Microcatheter and the Guidewire until desired position is attained. Select vessels by slowly torquing the SOFIA™ Distal Access Catheter if necessary.

WARNING: Do not advance or withdraw the device when excessive resistance is met until the cause of resistance is determined.

WARNING: Torquing the SOFIA™ Distal Distal Access Catheter excessively while kinked may damage the device resulting in separation of the device. Withdraw the entire device (the device, Microcatheter, and Guidewire) if the device is severely kinked.

WARNING: Do not exceed 2070 kPa (300 psi) maximum recommended infusion pressure.

Excess pressure may damage the device or injure the patient. Carefully monitor placement of the distal tip when using a power injector to infuse.

- 9. Slowly remove the Guidewire or the Microcatheter if necessary. Make sure that continuous perfusion of heparinized saline is maintained through the sidearm of the RHV.

NOTE: The Microcatheter used to navigate the SOFIA™ Distal Access Catheter may be kept for the rest of procedure.

The physician has the discretion to modify described manipulations of the SOFIA™ Distal Access Catheter to accommodate the complexity and variation in procedures. Any technique modification must be consistent with previously described instructions, warnings, precautions and patient safety information.













STORAGE

Avoid exposure to water, sunlight, extreme temperatures and high humidity during storage. Store the SOFIA™ Distal Access Catheter under controlled room temperature. See the product label for the device shelf life. Do not use the device beyond the labeled shelf life.

MATERIALS

The SOFIA™ Distal Access Catheter does not contain latex or PVC materials.

SYMBOLS

	Lot Number		Do Not Reuse
	Catalog Number		Attention, Consult Accompanying Documents
	Contents		Use by Date
	Sterilized Using Ethylene Oxide		Date of Manufacture
	CE Mark		Manufacturer
	Authorized European Representative		Non-pyrogenic

WARRANTY

MicroVention, Inc. warrants that reasonable care has been used in the design and manufacture of this device. This warranty is in lieu of and excludes all other warranties not expressly set forth herein, whether expressed or implied by operation of law or otherwise, including, but not limited to, any implied warranties of merchantability or fitness. Handling, storage, cleaning and sterilization of the device as well as factors relating to the patient, diagnosis, treatment, surgical procedure and other matters beyond MicroVention's control directly affect the device and the results obtained from its use. MicroVention's obligation under this warranty is limited to the repair or replacement of this device and MicroVention shall not be liable for any incidental or consequential loss, damage or expense directly or indirectly arising from the use of this device. MicroVention neither assumes, nor authorizes any other person to assume for it, any other or additional liability or responsibility in connection with this device. MicroVention assumes no liability with respect to devices reused, reprocessed or resterilized and makes no warranties, expressed or implied, including, but not limited to, merchantability or fitness for intended use, with respect to such device.

Prices, specifications and model availability are subject to change without notice.



EC REP

Manufacturer:

MicroVention, Inc.
1311 Valencia Avenue
Tustin, CA 92780 USA
Tel: (714) 247-8000

www.microvention.com

CE 0297

Authorized European Representative:

MicroVention Europe
30 bis, rue du Vieil Abreuvoir
78100 Saint-Germain-en-Laye
France
Tel: +33 (0)1 39 21 77 46
Fax: +33 (0)1 39 21 16 01

PD110436 Rev. A
2013-2

© Copyright 2013 MicroVention, Inc. All rights reserved.



October 22, 2013

S 131482/S002

FDA CDRH DMC

OCT 23 2013

Received

Samuel Shimp, Ph.D., Biomedical Engineer
U.S. Food and Drug Administration
Center for Devices and Radiological Health
Document Mail Center – WO66-0609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

RE: eCopy Hold Letter K131482/S002 – SOFIA Distal Access Catheter

The enclosed replacement eCopy is being provided in response to the FDA letter dated October 22, 2013.

Please contact me if you have any questions. Thank you.

A handwritten signature in blue ink that reads "Naomi Gong".

Naomi Gong

35

K131482/S02



October 18, 2013

FDA/CDRH/DCC

OCT 21 2013

RECEIVED

Samuel Shimp, Ph.D., Biomedical Engineer
U.S. Food and Drug Administration
Center for Devices and Radiological Health
Document Mail Center – WO66-0609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

RE: Telephone Hold for K131482/S002 – SOFIA Distal Access Catheter

The following information is being provided in response to the FDA letter dated August 9, 2013.

(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



A paper copy and an eCopy are being submitted. The eCopy is an exact duplicate of the paper copy. Please contact me if you have any additional questions. Thank you.

Yours truly,



Naomi Gong



October 18, 2013

Samuel Shimp, Ph.D., Biomedical Engineer
U.S. Food and Drug Administration
Center for Devices and Radiological Health
Document Mail Center – WO66-0609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

RE: Telephone Hold for K131482/S002 – SOFIA Distal Access Catheter

The following information is being provided in response to the FDA letter dated August 9, 2013.

(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



A paper copy and an eCopy are being submitted. The eCopy is an exact duplicate of the paper copy. Please contact me if you have any additional questions. Thank you.

Yours truly,



Naomi Gong

DEPARTMENT OF HEALTH AND HUMAN SERVICES
 FOOD AND DRUG ADMINISTRATION

Form Approval
 OMB No. 0910-0120
 Expiration Date: December 31, 2013
 See PRA Statement on page 5.

CDRH PREMARKET REVIEW SUBMISSION COVER SHEET

Date of Submission October 18, 2013	User Fee Payment ID Number (b) (4)	FDA Submission Document Number (if known) K131482
--	--	--

SECTION A TYPE OF SUBMISSION				
PMA <input type="checkbox"/> Original Submission <input type="checkbox"/> Premarket Report <input type="checkbox"/> Modular Submission <input type="checkbox"/> Amendment <input type="checkbox"/> Report <input type="checkbox"/> Report Amendment <input type="checkbox"/> Licensing Agreement	PMA & HDE Supplement <input type="checkbox"/> Regular (180 day) <input type="checkbox"/> Special <input type="checkbox"/> Panel Track (PMA Only) <input type="checkbox"/> 30-day Supplement <input type="checkbox"/> 30-day Notice <input type="checkbox"/> 135-day Supplement <input type="checkbox"/> Real-time Review <input type="checkbox"/> Amendment to PMA & HDE Supplement <input type="checkbox"/> Other	PDP <input type="checkbox"/> Original PDP <input type="checkbox"/> Notice of Completion <input type="checkbox"/> Amendment to PDP	510(k) <input type="checkbox"/> Original Submission: <input type="checkbox"/> Traditional <input type="checkbox"/> Special <input type="checkbox"/> Abbreviated (Complete section I, Page 5) <input checked="" type="checkbox"/> Additional Information <input type="checkbox"/> Third Party	Request for Feedback <input type="checkbox"/> Pre-Submission <input type="checkbox"/> Informational Meeting <input type="checkbox"/> Submission Issue Meeting <input type="checkbox"/> Day 100 Meeting <input type="checkbox"/> Agreement Meeting <input type="checkbox"/> Determination Meeting <input type="checkbox"/> Study Risk Determination <input type="checkbox"/> Other (specify):
IDE <input type="checkbox"/> Original Submission <input type="checkbox"/> Amendment <input type="checkbox"/> Supplement	Humanitarian Device Exemption (HDE) <input type="checkbox"/> Original Submission <input type="checkbox"/> Amendment <input type="checkbox"/> Supplement <input type="checkbox"/> Report <input type="checkbox"/> Report Amendment	Class II Exemption Petition <input type="checkbox"/> Original Submission <input type="checkbox"/> Additional Information	Evaluation of Automatic Class III Designation (De Novo) <input type="checkbox"/> Original Submission <input type="checkbox"/> Additional Information	Other Submission <input type="checkbox"/> 513(g) <input type="checkbox"/> Other (describe submission):

Have you used or cited Standards in your submission? Yes No (If Yes, please complete Section I, Page 5)

SECTION B SUBMITTER, APPLICANT OR SPONSOR			
Company / Institution Name MicroVention, Inc.		Establishment Registration Number (if known) 2032493	
Division Name (if applicable)		Phone Number (including area code) 714-247-8055	
Street Address 1311 Valencia Avenue		FAX Number (including area code) 714-247-8014	
City Tustin	State / Province CA	ZIP/Postal Code 92780	Country U.S.
Contact Name Naomi Gong			
Contact Title Sr. Regulatory Affairs Project Manager		Contact E-mail Address naomi.gong@microvention.com	

SECTION C APPLICATION CORRESPONDENT (e.g., consultant, if different from above)			
Company / Institution Name			
Division Name (if applicable)		Phone Number (including area code)	
Street Address		FAX Number (including area code)	
City	State / Province	ZIP Code	Country
Contact Name			
Contact Title		Contact E-mail Address	

SECTION D1			REASON FOR APPLICATION - PMA, PDP, OR HDE		
<input type="checkbox"/> New Device <input type="checkbox"/> Withdrawal <input type="checkbox"/> Additional or Expanded Indications <input type="checkbox"/> Request for Extension <input type="checkbox"/> Post-approval Study Protocol <input type="checkbox"/> Request for Applicant Hold <input type="checkbox"/> Request for Removal of Applicant Hold <input type="checkbox"/> Request to Remove or Add Manufacturing Site	<input type="checkbox"/> Change in design, component, or specification: <input type="checkbox"/> Software / Hardware <input type="checkbox"/> Color Additive <input type="checkbox"/> Material <input type="checkbox"/> Specifications <input type="checkbox"/> Other (<i>specify below</i>)	<input type="checkbox"/> Location change: <input type="checkbox"/> Manufacturer <input type="checkbox"/> Sterilizer <input type="checkbox"/> Packager			
<input type="checkbox"/> Process change: <input type="checkbox"/> Manufacturing <input type="checkbox"/> Packaging <input type="checkbox"/> Sterilization <input type="checkbox"/> Other (<i>specify below</i>)	<input type="checkbox"/> Labeling change: <input type="checkbox"/> Indications <input type="checkbox"/> Instructions <input type="checkbox"/> Performance Characteristics <input type="checkbox"/> Shelf Life <input type="checkbox"/> Trade Name <input type="checkbox"/> Other (<i>specify below</i>)	<input type="checkbox"/> Report Submission: <input type="checkbox"/> Annual or Periodic <input type="checkbox"/> Post-approval Study <input type="checkbox"/> Adverse Reaction <input type="checkbox"/> Device Defect <input type="checkbox"/> Amendment			
<input type="checkbox"/> Response to FDA correspondence:		<input type="checkbox"/> Change in Ownership <input type="checkbox"/> Change in Correspondent <input type="checkbox"/> Change of Applicant Address			
<input type="checkbox"/> Other Reason (<i>specify</i>):					

SECTION D2			REASON FOR APPLICATION - IDE		
<input type="checkbox"/> New Device <input type="checkbox"/> New Indication <input type="checkbox"/> Addition of Institution <input type="checkbox"/> Expansion / Extension of Study <input type="checkbox"/> IRB Certification <input type="checkbox"/> Termination of Study <input type="checkbox"/> Withdrawal of Application <input type="checkbox"/> Unanticipated Adverse Effect <input type="checkbox"/> Notification of Emergency Use <input type="checkbox"/> Compassionate Use Request <input type="checkbox"/> Treatment IDE <input type="checkbox"/> Continued Access	<input type="checkbox"/> Change in: <input type="checkbox"/> Correspondent / Applicant <input type="checkbox"/> Design / Device <input type="checkbox"/> Informed Consent <input type="checkbox"/> Manufacturer <input type="checkbox"/> Manufacturing Process <input type="checkbox"/> Protocol - Feasibility <input type="checkbox"/> Protocol - Other <input type="checkbox"/> Sponsor	<input type="checkbox"/> Response to FDA Letter Concerning: <input type="checkbox"/> Conditional Approval <input type="checkbox"/> Deemed Approved <input type="checkbox"/> Deficient Final Report <input type="checkbox"/> Deficient Progress Report <input type="checkbox"/> Deficient Investigator Report <input type="checkbox"/> Disapproval <input type="checkbox"/> Request Extension of Time to Respond to FDA <input type="checkbox"/> Request Meeting <input type="checkbox"/> Request Hearing			
	<input type="checkbox"/> Report submission: <input type="checkbox"/> Current Investigator <input type="checkbox"/> Annual Progress Report <input type="checkbox"/> Site Waiver Report <input type="checkbox"/> Final				
<input type="checkbox"/> Other Reason (<i>specify</i>):					

SECTION D3			REASON FOR SUBMISSION - 510(k)		
<input type="checkbox"/> New Device	<input type="checkbox"/> Additional or Expanded Indications	<input type="checkbox"/> Change in Technology			
<input checked="" type="checkbox"/> Other Reason (<i>specify</i>): Response to FDA letter dated August 9, 2013					

SECTION E ADDITIONAL INFORMATION ON 510(K) SUBMISSIONS

Product codes of devices to which substantial equivalence is claimed				Summary of, or statement concerning, safety and effectiveness information <input checked="" type="checkbox"/> 510 (k) summary attached <input type="checkbox"/> 510 (k) statement
1 DQY	2	3	4	
5	6	7	8	

Information on devices to which substantial equivalence is claimed (if known)

	510(k) Number	Trade or Proprietary or Model Name	Manufacturer
1	K082385	Chaperon Guiding Catheter	MicroVention, Inc. 1311 Valencia Avenue, Tustin, CA 92780
2	K120917	HEADWAY DUO Microcatheter	MicroVention, Inc. 1311 Valencia Avenue, Tustin, CA 92780
3			
4			
5			
6			

SECTION F PRODUCT INFORMATION - APPLICATION TO ALL APPLICATIONS

Common or usual name or classification name
 Percutaneous catheter

	Trade or Proprietary or Model Name for This Device	Model Number
1	SOFIA Distal Access Catheter	DA5125ST
2		
3		
4		
5		

FDA document numbers of all prior related submissions (regardless of outcome)

1 K131482	2	3	4	5	6
7	8	9	10	11	12

Data Included in Submission
 Laboratory Testing
 Animal Trials
 Human Trials

SECTION G PRODUCT CLASSIFICATION - APPLICATION TO ALL APPLICATIONS

Product Code DQY	C.F.R. Section (if applicable) 21 CFR 870.1250	Device Class <input type="checkbox"/> Class I <input checked="" type="checkbox"/> Class II <input type="checkbox"/> Class III <input type="checkbox"/> Unclassified
Classification Panel Cardiovascular Devices		

Indications (from labeling)
 The SOFIA Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

Note: Submission of the information entered in Section H does not affect the need to submit device establishment registration.		FDA Document Number (if known) K131482	
SECTION H MANUFACTURING / PACKAGING / STERILIZATION SITES RELATING TO A SUBMISSION			
<input checked="" type="checkbox"/> Original <input type="checkbox"/> Add <input type="checkbox"/> Delete		Facility Establishment Identifier (FEI) Number 2032493	
<input checked="" type="checkbox"/> Manufacturer <input type="checkbox"/> Contract Manufacturer		<input type="checkbox"/> Contract Sterilizer <input type="checkbox"/> Repackager / Relabeler	
Company / Institution Name MicroVention, Inc.		Establishment Registration Number 2032493	
Division Name (if applicable)		Phone Number (including area code) 714-247-8000	
Street Address 1311 Valencia Avenue		FAX Number (including area code) 714-247-8005	
City Tustin		State / Province CA	ZIP Code 92780
Country U.S.			
Contact Name Naomi Gong	Contact Title Sr. Regulatory Affairs Project Manager	Contact E-mail Address naomi.gong@microvention.com	
<input checked="" type="checkbox"/> Original <input type="checkbox"/> Add <input type="checkbox"/> Delete		Facility Establishment Identifier (FEI) Number	
<input type="checkbox"/> Manufacturer <input type="checkbox"/> Contract Manufacturer		<input checked="" type="checkbox"/> Contract Sterilizer <input type="checkbox"/> Repackager / Relabeler	
Company / Institution Name Sterigenics		Establishment Registration Number 2011171	
Division Name (if applicable)		Phone Number (including area code) 951-340-0700	
Street Address 4900 South Griffith Avenue		FAX Number (including area code)	
City Los Angeles		State / Province CA	ZIP Code 90058
Country U.S.			
Contact Name Sharon Huges	Contact Title Representative	Contact E-mail Address losangelessales@sterigenics.com	
<input type="checkbox"/> Original <input type="checkbox"/> Add <input type="checkbox"/> Delete		Facility Establishment Identifier (FEI) Number	
<input type="checkbox"/> Manufacturer <input type="checkbox"/> Contract Manufacturer		<input type="checkbox"/> Contract Sterilizer <input type="checkbox"/> Repackager / Relabeler	
Company / Institution Name		Establishment Registration Number	
Division Name (if applicable)		Phone Number (including area code)	
Street Address		FAX Number (including area code)	
City		State / Province	ZIP Code
Country			
Contact Name	Contact Title	Contact E-mail Address	

SECTION I UTILIZATION OF STANDARDS

Note: Complete this section if your application or submission cites standards or includes a "Declaration of Conformity to a Recognized Standard" statement.

	Standards No.	Standards Organization	Standards Title	Version	Date
1					
2					
3					
4					
5					
6					
7					

Please include any additional standards to be cited on a separate page.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF ADDRESS BELOW.

The burden time for this collection of information is estimated to average 0.5 hour per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
 Food and Drug Administration
 Office of Chief Information Officer
 Paperwork Reduction Act (PRA) Staff
 1350 Piccard Drive, Room 400
 Rockville, MD 20850

An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.

Attachment 1

Bench Testing	Acceptance Criteria	Result
	<ul style="list-style-type: none"> • Free from surface defect/no sharp edges • Embedded particulate acceptable if OD is in specification • Distal tip smooth and round at lumen opening • PTFE liner is not delaminated at distal tip 	
Force at break (distal and hub)	≥ 2.25 lbf for outer diameters ≥ 0.045" and < 0.072"	Passed
Flow rate	Equivalent to or better flow rates than competitor products (for reference only)	N/A
Static burst	Catheter shaft will not burst below 46 psi.	Passed
Leakage at 46 psi	No liquid leaking from hub and catheter shaft at 46 psi for 30 second duration.	Passed
Air leakage	No air leaking into syringe for 15 seconds (ignore air bubbles for the first 5 seconds of the test.	Passed
Dynamic burst	<ul style="list-style-type: none"> • Equivalent or greater burst pressure than competitor product • Microcatheter will not burst below 300 psi 	Passed
Catheter particle testing	Less than 25 particles greater than 10 microns and less than 3 particles greater than 25 micron	Passed

As summarized in the table, SOFIA Testing Summary, the following tests were conducted to establish the performance characteristics of the device:

14.1. Simulated Use

To verify that SOFIA Catheter meets the established performance specifications in a clinically simulated environment, MicroVention has tested total of 22 samples of SOFIA. Samples underwent simulated use testing that included delivery, trackability (with and without guidewire), guidewire lock-up and catheter ovalization. All samples met the established acceptance criterion of ≥ 3 rating.

The test simulates a neurointerventional embolization procedure using 37°C fluid with catheters, guidewires, and vessel tortuosity.

Attachment 2

Attachment 3

Attachment 4

Attachment 5

Attachment 8

Attachment 9

Department of Health and Human Services
Food and Drug Administration
STANDARDS DATA REPORT FOR 510(k)s
(To be filled in by applicant)

This report and the Summary Report Table are to be completed by the applicant when submitting a 510(k) that references a national or international standard. A separate report is required for each standard referenced in the 510(k).

TYPE OF 510(K) SUBMISSION

Traditional Special Abbreviated

STANDARD TITLE ¹

ISO 594-2 (1998) Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment - Part 2

Please answer the following questions

Yes No

Is this standard recognized by FDA ²?

FDA Recognition number ³ #6-129

Was a third party laboratory responsible for testing conformity of the device to this standard identified in the 510(k)?

Is a summary report ⁴ describing the extent of conformance of the standard used included in the 510(k)?
If no, complete a summary report table.

Does the test data for this device demonstrate conformity to the requirements of this standard as it pertains to this device?

Does this standard include acceptance criteria?
If no, include the results of testing in the 510(k).

Does this standard include more than one option or selection of tests?
If yes, report options selected in the summary report table.

Were there any deviations or adaptations made in the use of the standard?.....
If yes, were deviations in accordance with the FDA supplemental information sheet (SIS) ⁵?

Were deviations or adaptations made beyond what is specified in the FDA SIS?.....
If yes, report these deviations or adaptations in the summary report table.

Were there any exclusions from the standard?
If yes, report these exclusions in the summary report table.

Is there an FDA guidance ⁶ that is associated with this standard?.....
If yes, was the guidance document followed in preparation of this 510k?

Title of guidance: _____

¹ The formatting convention for the title is: [SDO] [numeric identifier] [title of standard] [date of publication]

² Authority [21 U.S.C. 360d], <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Standards/default.htm>

³ <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁴ The summary report should include: any adaptations used to adapt to the device under review (for example, alternative test methods); choices made when options or a selection of methods are described; deviations from the standard; requirements not applicable to the device; and the name and

address of the test laboratory or certification body involved in conformance assessment to this standard. The summary report includes information on all standards utilized during the development of the device.


⁵ The supplemental information sheet (SIS) is additional information which is necessary before FDA recognizes the standard. Found at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

⁶ The online search for CDRH Guidance Documents can be found at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/default.htm>


EXTENT OF STANDARD CONFORMANCE SUMMARY REPORT TABLE		
STANDARD TITLE SO 594-2 (1998) Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment - Part 2		
CONFORMANCE WITH STANDARD SECTIONS*		
SECTION NUMBER All	SECTION TITLE (as applicable)	CONFORMANCE? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED ✧		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED ✧		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED ✧		
DESCRIPTION		
JUSTIFICATION		
SECTION NUMBER	SECTION TITLE	CONFORMANCE? <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
TYPE OF DEVIATION OR OPTION SELECTED ✧		
DESCRIPTION		
JUSTIFICATION		
<p>* For completeness list all sections of the standard and indicate whether conformance is met. If a section is not applicable (N/A) an explanation is needed under "justification." Some standards include options, so similar to deviations, the option chosen needs to be described and adequately justified as appropriate for the subject device. Explanation of all deviations or description of options selected when following a standard is required under "type of deviation or option selected," "description" and "justification" on the report. More than one page may be necessary.</p> <p>✧ Types of deviations can include an exclusion of a section in the standard, a deviation brought out by the FDA supplemental information sheet (SIS), a deviation to adapt the standard to the device, or any adaptation of a section.</p>		
Paperwork Reduction Act Statement		
<p>Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to:</p> <p style="text-align: center;">Department of Health and Human Services Food and Drug Administration Office of Chief Information Officer 1350 Piccard Drive, Room 400 Rockville, MD 20850</p> <p style="text-align: right;"><i>An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number.</i></p>		

Attachment 10

ATTACHMENT 2



(b)(4) Confidential and Proprietary Information -Drawing




ATTACHMENT 3


ATTACHMENT 4

ATTACHMENT 1

Attachment 13

(b)(4) Confidential and Proprietary Information -Drawing





This document contains Confidential and Proprietary Information of Microvention, Inc., and all design, manufacturing, reproduction, use and sale rights for such information are expressly reserved.


Microvention
TERUMO
FINAL PACKAGING ASSY
SOFIA

PROJ. NO. PD060043 REV. C
SCALE: XX DO NOT SCALE DRAWING SHEET 1 OF 2

Attachment 15


(b)(4) Confidential and Proprietary Information - Drawing





(b)(4) Confidential and Proprietary Information - Drawing





Attachment 16

510(k) Summary	
Trade Name:	SOFIA™ Distal Access Catheter
Generic Name:	Percutaneous Catheter
Classification:	Class II, 21 CFR 870.1250 (DQY)
Submitted By:	MicroVention, Inc 1311 Valencia Avenue Tustin, California U.S.A.
Contact:	Naomi Gong
Date:	2013 May 21
Predicate Device:	Chaperon Guiding Catheter (K082385) HEADWAY DUO Microcatheter (K120917)

Device Description:

The SOFIA Distal Access Catheter is a single-lumen, flexible catheter designed with coil and braid reinforcement. The distal segment is steam-shapeable and it has a hydrophilic coating for navigation through the vasculature. The radiopaque marker is located at the distal end of the catheter for visualization under fluoroscopy. An introducer sheath and shaping mandrel are also provided.

Indications For Use:

The SOFIA Distal Access Catheter is indicated for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate introduction of diagnostic or therapeutic devices. It is not intended for use in coronary arteries.

Technological Comparison:

	Chaperon Guiding Catheter	SOFIA Distal Access Catheter	HEADWAY DUO Microcatheter
Intended Use	Intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.	Intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.	Intended for general intravascular use, including the peripheral and coronary vasculature for the infusion of diagnostic agents, such as contrast media, and therapeutic agents, such as embolization materials. Intended for neurovascular use for the infusion of diagnostic agents, such as contrast media, and therapeutic agents that have been cleared or approved for use in the neurovasculature and are compatible with the inner diameter of the microcatheter.
Material Catheter Body	Outer layer of polyester elastomer; stainless steel braid; inner layer of PTFE (polytetrafluoroethylene).	Outer layer of polyurethane elastomer (Polyblend and Pellethane), polyether block amide (Pebax) and polyamide (Grilamid); inner layer of stainless steel braid/coil, PTFE and polyolefin elastomer	Outer layer of polyether block amide (Pebax) and polyamide (Grilamid); inner layer of stainless steel braid/coil and PTFE
Marker	Tungsten	Platinum/Iridium	Platinum/Iridium
Hub	Nylon	Nylon	Nylon
Strain Relief	Polyester elastomer	Polyurethane	Pebax
Introducer	Not applicable	Pebax	Pebax
Shaping Mandrel	Not applicable	Stainless steel	Stainless steel

	Chaperon Guiding Catheter	SOFIA Distal Access Catheter	HEADWAY DUO Microcatheter
Catheter size	5 F (outer catheter)	5 F	1.6 - 2.1 F
ID	0.059 inch (1.5 mm)	0.055 inch (1.4 mm)	0.0165 inch (0.42 mm)
OD	0.068 inch (1.7 mm)	0.068 inch (1.7 mm)	0.023 – 0.0275 inch (0.58 – 0.70 mm)
Effective Length	95 cm (outer) 117 cm (inner)	125 cm	157 and 168 cm
Coating	Hydrophilic coating (Terumo proprietary coating)	Hydrophilic coating (b) (4)	Hydrophilic coating (b) (4)
Tip Configuration	Preshaped	Steam shapeable by user	Steam shapeable by user
Guidewire Compatibility	0.035 inch or 0.038 inch	0.035 inch or 0.038 inch	0.014 inch or smaller
Accessories	N/A	Introducer sheath and shaping mandrel	Introducer sheath and shaping mandrel
Method of Supply	Sterile and single use	Sterile and single use	Sterile and single use
Sterilization Method	Ethylene Oxide	Ethylene Oxide	Ethylene Oxide
Packaging Configuration	Catheter placed on packaging card that is inserted into Tyvek® pouch. Pouch and IFU placed in bleached sulfate carton box.	Catheter placed into a HDPE dispenser tube. Dispenser tube, introducer and shaping mandrel placed on a polyethylene packaging card that is inserted into a Tyvek® pouch. Pouch and IFU placed in bleached sulfate carton box.	Catheter placed into a polyethylene dispenser hoop. Introducer and shaping mandrel placed on polyethylene packaging card. Dispenser hoop and packaging card inserted into Tyvek® pouch. Pouch and IFU placed in bleached sulfate carton box.

Verification and Test Summary:

Bench Testing		
Test	Results	Conclusions
Simulated Use	Test articles achieved a rating ≥ 3 for preparation/ease of assembly, introducer sheath interaction, introducer peel away, tracking with guidewire/microcatheter, microcatheter/guidewire lockup, lubricity and durability of hydrophilic coating, microcatheter/guidewire removal, removal/aspiration of clot, mechanical clot retriever and stent delivery with no particles	Device performs as intended under simulated use conditions
Equipment Interface	Test articles compatible with 0.035-inch and 0.038-inch guidewires, 6F or larger guide catheter/guiding sheath, common RHVs using insertion tool, stopcocks and ≤ 0.027 -inch microcatheters	Device compatible with recommended accessories commonly used in intravascular procedures
Dimensional and Physical Attributes	Test articles met the specified dimensional requirements for catheter OD, catheter ID, overall working length, length of distal section, length of distal tip to marker band and total length of hub/strain relief	Device met established dimensional and physical specifications
Kink Resistance	No kinks at 1 cm, 4 cm, 12 cm and 25 cm from distal tip when wrapped around 0.025-inch and 0.030-inch pin gauges No kinks noted during simulated use testing	Device resistant to kinking around small radii turns
Tip Shapeability	Tip angle of test article equivalent to competitive devices after steam shaping using mandrel with an angle of approximately 90°	Shapeability of distal tip after steam shaping equivalent to competitive devices
Radio Detectability	Distal marker band visible under fluoroscopy	Device radiopacity equivalent to or better than predicate and competitive devices

Bench Testing		
Test	Results	Conclusions
Gauging (ISO 594-2)	Gauging pin and hub align in limit planes	Device hub meets the requirements of ISO 594-2
Separation Force (ISO 594-2)	Mating parts separation force greater than 25 N	Device hub meets the requirements of ISO 594-2
Unscrewing Torque (ISO 594-2)	Test article luer remains attached after applying an unscrewing torque not less than 0.02 Nm for a minimum of 10 seconds	Device hub meets the requirements of ISO 594-2
Stress Cracking (ISO 594-2)	No stress cracks on test article hub	Device hub meets the requirements of ISO 594-2
Ease of Assembly (ISO 594-2)	Components fit together securely with no resistance observed between test article luer and reference fitting	Device hub meets the requirements of ISO 594-2
Resistance to Overriding (ISO 594-2)	Test article luer does not override reference fitting threads	Device hub meets the requirements of ISO 594-2
Durability/Lubricity of Hydrophilic Coating	Test article achieved a rating of ≥ 3 during simulated use testing for coating durability and lubricity.	Device tracks easily with no coating cracking or separation
Catheter Stiffness	Device stiffness equivalent to predicate and competitive devices	Device tracks in tortuous anatomy while advancing to target site
Torque Strength	No catheter breakage after 50 rotations	Device torque strength same as predicate device
Catheter Flexural Fatigue	No flexural fatigue following repeated bending during simulated use testing and repeated hoop stress following pressure and air aspiration testing	Device integrity suitable for intended clinical use
Surface Contamination	Test article free from surface contaminants from uncured coating surface particulates $> 0.02 \text{ mm}^2$, embedded particulates Distal tip smooth and tapered PTFE inner layer not delaminated	Device integrity suitable for intended clinical use
Force at Break (Distal and Hub)	Catheter force at break $\geq 2.25\text{lb}$ for distal section and hub/catheter junction	Tensile strength test results equivalent to predicate and competitive devices

Bench Testing		
Test	Results	Conclusions
Flow Rate	Flow rate at 100 psi and 300 psi with diagnostic agents (e.g., saline, contrast media) equivalent to or better than competitive devices	Device meets specified requirements for delivery of diagnostic agents
Static Burst Pressure	No damage of pressurized catheter at 46 psi	Device integrity suitable for intended clinical use and met requirements of ISO 10555-1
Fluid Leakage at > 46 psi	No liquid leakage from hub and catheter shaft at 46 psi for 30 seconds	Device integrity suitable for intended clinical use and met requirements of ISO 10555-1
Air Leakage	No air leakage at hub into syringe for 15 seconds	Device integrity suitable for intended clinical use and met requirements of ISO 10555-1
Dynamic Burst	Test articles did not burst at or below 300 psi	Device met labeled maximum infusion pressure of 300 psi
Particulate Test	Less than 25 particles greater than 10 microns per ml volume and less than 3 particles less than 25 microns per ml volume No particles greater than 70 microns	Device met specifications for maximum allowable particles

Biocompatibility		
Test	Results	Conclusions
Cytotoxicity – MEM Elution Assay (ISO 10993-5)	Cell culture treated with test article exhibited slight reactivity (Grade 1)	Non-cytotoxic
Sensitization/Irritation – Kligman Maximization Test (ISO 10993-10)	Extracts of the test article elicited no reaction at the challenge (0% sensitization) following the induction phase (Grade 1).	Weak allergic potential or sensitizing capacity
Sensitization/Irritation – Intracutaneous Injection Test (ISO 10993-10)	Extracts of the test article did not show a significantly greater biological reaction than the sites injected with the control article	Non-irritant
Hemocompatibility – Rabbit Blood Direct and Indirect Contact (ISO 10993-4)	The hemolysis index was 0.13% (direct contact) and 0.0% (indirect contact)	Non-hemolytic

Biocompatibility		
Test	Results	Conclusions
Hemocompatibility – Unactivated Partial Thromboplastin Time Assay Direct Contact (ISO 10993-4)	No statistically significant difference found between the Unactivated Partial Thromboplastin Time (UPTT) of the plasma exposed to the test article and that of the plasma exposed to either the negative control or the untreated control	No effect on coagulation
Hemocompatibility – Complement Activation Assay (ISO 10993-4)	C3a and SC5b-9 levels \leq negative and untreated controls	No effect on complement activation
Hemocompatibility – Thrombogenicity Study in Dogs (ISO 10993-4)	Minimal thrombosis observed with a Grade 0 in two out of two test sites and two out of two control sites	No significant thrombosis
Systemic Toxicity – Systemic Injection Test (ISO 10993-11)	Extracts of test article did not induce a significantly greater biological reaction than the control extracts when injected in Swiss Albino mice	No toxic effects
Systemic Toxicity - Rabbit Pyrogen Test (ISO 10993-11)	The temperature increases (maximum) was 0.03°C from baseline	Non-pyrogenic

Summary of Substantial Equivalence:

The data presented in this submission demonstrates the technological similarity and equivalency of the SOFIA Distal Access Catheter when compared with the predicate Chaperon Guiding Catheter (K082385) and the HEADWAY DUO Microcatheter (K120917) devices. The devices:

- Have an equivalent intended use,
- Use the same operating principle,
- Incorporate the same basic design,
- Are packaged and sterilized using the same methods.

In summary, the SOFIA Distal Access Catheter described in this submission is substantially equivalent to the predicate devices.

Attachment 17

6. Predicate Device Information

K082385, MicroVention, Inc., Chaperon Guiding Catheter

K120917, MicroVention, Inc., HEADWAY DUO Microcatheter

7. Labeling and Intended Use

Draft labels and Instructions For Use are provided in Appendix 1.

Intended Use

The intended use is the same as the predicate device, Chaperon Guiding Catheter and is stated in the product labeling as follows:

The SOFIA™ Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

8. Device Overview

The SOFIA Catheter is a single lumen (0.055” or 1.4 mm inner diameter) catheter with a shapeable distal tip. The semi-rigid proximal section transitions to a flexible distal tip to facilitate advancement through the vasculature.

A single radiopaque marker (Pt/Ir) at the distal end facilitates fluoroscopic visualization. The outer surface of the catheter (distal 60 cm) is coated with a hydrophilic polymer coating to reduce friction during navigation in the vasculature.

A luer fitting on the microcatheter hub is used for the attachment of accessories. The hub/strain relief provides kink resistance for the proximal end. The tip configuration is provided straight, but is shapeable by the user. A steam shaping mandrel and introducer sheath accessories are packaged with the catheter.

The device is placed in a dispenser tube to keep the device in position during shipping and handling. The dispenser tube is placed on a packaging card and then placed into a tyvek pouch. The pouch is put into a chipboard carton box prior to sterilization. The device is provided sterile, for single use.

9. Device Configurations and Dimensions

The SOFIA Distal Access Catheter has an inner lumen of 0.055” and is recommended to be used with guidewires from 0.035” to 0.038”.

Device	Catalogue Number	Working Length	Outer Diameter/ Inner Diameter	Recommended Guidewire
SOFIA Distal Access Catheter	DA5125ST	125 mm	0.068”/0.055” (1.7/1.4 mm)	0.035” to 0.038”

From: [Hoang, Quynh T.](#)
To: [510K Program](#)
Cc: [Shimp, Samuel](#)
Subject: (Due 6/7) K131482 Conversion fr Special to Traditional. Thanks.
Date: Wednesday, June 05, 2013 5:20:07 PM

Dear 510(k) staff:

For your concurrence is the conversion from a Special to a Traditional.

The conversion form is **file #1**, at

<https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Once you are done, we'll inform the sponsor of the conversion and log out the RTAA (for a Traditional).

Thanks.

Form for Converting a Special 510(k) to a Traditional or Abbreviated 510(k)

Note: Please send this to 510k Staff electronically. You do not need anyone to sign this in person.

Date: 6/28/2013

Reviewer: Samuel Shimp

510(k) Number: K131482

Device Name: Sofia Distal Access Catheter

Reason for Conversion (select one):

Change in Indications for Use (please list old and new indications below)

Change in Technology (select one):

We have not seen this change before in this device type

We have seen this change before in this device type, but we need to see the data (please provide a brief statement below regarding why summary data/risk analysis are insufficient)

Other (e.g. submission included unsolicited data or sponsor is modifying a device that is not their own – please specify below)

(b)(4) Confidential and Proprietary Information



Digital Signature Concurrence Table	
Reviewer Sign-Off	Samuel Shimp 6/28/2013
Branch Chief Sign-Off	Quynh Hoang, 7/2/13
Division Sign-Off (please obtain before calling or e-mailing POS)	
POS Sign-Off	

Date of Phone Conversation with Sponsor:

(The reviewer or Branch Chief must contact the sponsor to notify them of the conversion. At this time the reviewer or Branch Chief may request additional information that was not submitted in the special.)

MEMORANDUM

Date: July 25, 2013

From: Jeffrey Toy, Ph.D., Toxicologist

Subject: K131482 MicroVention SOFIA Distal Access Catheter
Biocompatibility Review

To: The Record

Through: Samuel Shimp, Ph.D.
Team Leader

RECOMMENDATION: ADDITIONAL INFORMATION

PURPOSE:

MicroVention request 510k clearance for their SOFIA Distal Access Catheter.

Excerpted text is italicized

INDICATION FOR USE

The SOFIA Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

(b)(4) Confidential and Proprietary Information and (b)(5)



K131482 Biocomp

(b)(4) Confidential and Proprietary Information and (b)(5)



K131482 Biocomp

(b)(4) Confidential and Proprietary Information and (b)(5)



K131482 Biocomp

(b)(4) Confidential and Proprietary Information and (b)(5)



K131482 Biocomp

(b)(4) Confidential and Proprietary Information and (b)(5)



K131482 Biocomp

(b)(4) Confidential and Proprietary Information and (b)(5)



K131482 Biocomp

(b)(4) Confidential and Proprietary Information and (b)(5)



RECOMMENDATION: ADDITIONAL INFORMATION

Reviewer Sign-Off:	
--------------------	--

Date: August 8, 2013
From: Myra Smith, DNPMD/NNDB
To: Samuel Shimp, Ph.D. DNPMD/NNDB
cc: The Record
Subject: K131482/S001 - Microbiology Review
Device Trade Name: SOFIA Distal Access Catheter
Sponsor: MicroVention Inc.

Classification

Generic Name: Percutaneous Catheter Classification:
Class II 21 CFR 870.1250 (DQY)

Recommendation – Additional information request from a microbiology standpoint

Purpose of Submission: would like to introduce the modified Catheter into interstate commerce.

Predicate Device

K082385 Chaperon Guiding Catheter, MicroVention, Inc

Device Description

The SOFIA™ Distal Access Catheter is a single-lumen, flexible catheter designed with coil and braid reinforcement. The distal segment is steam-shapeable and it has a hydrophilic coating for navigation through the vasculature. The radiopaque marker is located at the distal end of the catheter for visualization under fluoroscopy. An introducer sheath and shaping mandrel are also provided.

Indications for Use

The SOFIA™ Distal Access Catheter is indicated for general intravascular use, including the neuro and peripheral vasculature. The SOFIA™ Distal Access Catheter can be used to facilitate introduction of diagnostic or therapeutic devices. The SOFIA™ Distal Access Catheter is not intended for use in coronary arteries.

Reviewer Comments

The Indications for Use are the identical to that of the predicate device and are as follows:

Comparison to Predicate Device

	Chaperon Guiding Catheter – K082385 (Outer catheter)	SOFIA Distal Access Catheter
Intended Use	Intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.	Same
Material	Outer layer of polyester elastomer, stainless steel braid, inner liner of PTFE (polytetrafluoroethylene). Tungsten radiopaque marker, nylon hub, and polyester elastomer strain relief.	Outer layer of polyolefin elastomer, polyurethane elastomer, polyether block amide, polyamide; inner layer of PTFE/polyolefin; stainless steel braid and coil. Pt/Ir radiopaque marker, nylon hub, polyurethane strain relief. Pebax introducer sheath and stainless steel shaping mandrel.
Catheter size	5F (Outer catheter)	Same
ID	1.5 mm (0.059")	1.4 mm (0.055")
OD	1.7 mm (0.068")	Same
Effective Length	Outer Catheter : 95 cm Inner Catheter: 117 cm	125 cm
Coating	Hydrophilic coating (inner catheter)	Hydrophilic coating (same as Headway Duo, K120917)
Tip Configuration	Preshaped	Steam shapeable by user
Guidewire compatibility	0.035" or 0.038"	Same
Accessories	N/A	Introducer sheath and shaping mandrel (same as Headway Duo)
Packaging	Packaging Card/Tyvek pouch/Carton	Same, with dispenser tube to hold catheter on packaging card
Method of supply	Sterile and single use	Same

(b)(4) Confidential and Proprietary Information and (b)(5)

(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)





DEPARTMENT OF HEALTH AND HUMAN SERVICES

M E M O R A N D U M

Food and Drug Administration
Office of Device Evaluation
9200 Corporate Boulevard
Rockville, MD 20850

**Premarket Notification [510(k)] Review
Traditional/Abbreviated**

K131482/S001

Date: August 8, 2013
To: The Record
From: Samuel Shimp, Ph.D.

Office: ODE
Division: DNPMD/NNDB

510(k) Holder: MicroVention, Inc.
Device Name: SOFIA Distal Access Catheter
Contact: Naomi Gong
Phone: (714) 247-8055
Fax: (714) 247-8014
Email: Naomi.Gong@microvention.com

Reviewer Text	<i>Sponsor Text</i>	<i>Comments</i>
---------------	---------------------	-----------------

RECOMMENDATION: TELEPHONE HOLD (TH) and request additional information.

(b)(4) Confidential and Proprietary Information and (b)(5)

(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



	Chaperon Guiding Catheter – K082385 (Outer catheter)	SOFIA Distal Access Catheter
Intended Use	Intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.	Same
Material	Outer layer of polyester elastomer, stainless steel braid, inner liner of PTFE (polytetrafluoroethylene). Tungsten radiopaque marker, nylon hub, and polyester elastomer strain relief.	Outer layer of polyolefin elastomer, polyurethane elastomer, polyether block amide, polyamide; inner layer of PTFE/polyolefin; stainless steel braid and coil. Pt/Ir radiopaque marker, nylon hub, polyurethane strain relief. Pebax introducer sheath and stainless steel shaping mandrel.
Catheter size	5F (Outer catheter)	Same
ID	1.5 mm (0.059")	1.4 mm (0.055")
OD	1.7 mm (0.068")	Same
Effective Length	Outer Catheter : 95 cm Inner Catheter: 117 cm	125 cm
Coating	Hydrophilic coating (inner catheter)	Hydrophilic coating (same as Headway Duo, K120917)
Tip Configuration	Preshaped	Steam shapeable by user
Guidewire compatibility	0.035" or 0.038"	Same
Accessories	N/A	Introducer sheath and shaping mandrel (same as Headway Duo)
Packaging	Packaging Card/Tyvek pouch/Carton	Same, with dispenser tube to hold catheter on packaging card
Method of supply	Sterile and single use	Same

(b)(4) Confidential and Proprietary Information and (b)(5)

(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



(b)(4) Confidential and Proprietary Information and (b)(5)



Digital Signature Concurrence Table	
Reviewer Sign-Off	
Branch Chief Sign-Off (optional)	
Division Sign-Off (optional)	

MEMORANDUM

Date: November 13, 2013

From: Jeffrey Toy, Ph.D., Toxicologist

Subject: K131482/S001 MicroVention SOFIA Distal Access Catheter
Biocompatibility Review

To: The Record

Through: Samuel Shimp, Ph.D.
Team Leader

RECOMMENDATION: SUBSTANTIALLY EQUIVALENT

PURPOSE:

MicroVention request 510k clearance for their SOFIA Distal Access Catheter.

Excerpted text is italicized

INDICATION FOR USE

The SOFIA Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

(b)(4) Confidential and Proprietary Information and (b)(5)



K131482/S001 Biocomp

(b)(4) Confidential and Proprietary Information and (b)(5)



K131482/S001 Biocomp

(b)(4) Confidential and Proprietary Information and (b)(5)



K131482/S001 Biocomp

(b)(4) Confidential and Proprietary Information and (b)(5)



RECOMMENDATION: SUBSTANTIALLY EQUIVALENT

Reviewer Sign-Off:	
--------------------	--

Date: November 18, 2013
From: Myra Smith, DNPMD/NNDB
To: Samuel Shimp, Ph.D. DNPMD/NNDB
cc: The Record
Subject: K131482/S002 - Microbiology Review
Device Trade Name: SOFIA Distal Access Catheter
Sponsor: MicroVention Inc.

Classification

Generic Name: Percutaneous Catheter Classification:
Class II 21 CFR 870.1250 (DQY)

Recommendation –Substantial Equivalence (SE) from a microbiology standpoint

Purpose of Submission: would like to introduce the modified SOFIA Distal Access Catheter into interstate commerce.

(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



Conclusion/Recommendation: Substantial Equivalence (SE) from a microbiology standpoint. MicroVention has addressed outstanding microbiology deficiencies.



DEPARTMENT OF HEALTH AND HUMAN SERVICES

MEMORANDUM

Food and Drug Administration
Office of Device Evaluation
9200 Corporate Boulevard
Rockville, MD 20850

**Premarket Notification [510(k)] Review
Traditional/Abbreviated**

K131482/S002

Date: November 18, 2013
To: The Record
From: Samuel Shimp, Ph.D.

Office: ODE
Division: DNPMD/NNDB

510(k) Holder: MicroVention, Inc.
Device Name: SOFIA Distal Access Catheter
Contact: Naomi Gong
Phone: (714) 247-8055
Fax: (714) 247-8014
Email: Naomi.Gong@microvention.com

Reviewer Text	<i>Sponsor Text</i>	<i>Comments</i>
---------------	---------------------	-----------------

RECOMMENDATION: SUBSTANTIALLY EQUIVALENT (SE): All deficiencies have been resolved and the performance testing, sterilization process, and biocompatibility data all support the SE determination.

(b)(4) Confidential and Proprietary Information - (b)(5)

(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



(b)(4) Confidential and Proprietary Information - (b)(5)



XVII. Recommendation: Substantially Equivalent: Benchtop performance testing, sterilization, and biocompatibility are adequate.

Regulation Number: 21 CFR 870.1250

Regulation Name: Percutaneous Catheter

Regulatory Class: Class II

Product Code: DQY, DQO

Digital Signature Concurrence Table	
Reviewer Sign-Off	
Branch Chief Sign-Off (optional)	
Division Sign-Off (optional)	

From: [Hoang Quynh T.](#)
To: [Krauthamer, Victor](#); [Whang, Joyce M](#); [Lendor, Marisol](#)
Cc: [Shimp, Samuel](#)
Subject: Due today, K131482 Special RTA1. Thanks.
Date: Friday, June 07, 2013 3:28:18 PM

Dear Victor and Joyce:

During the call to inform the sponsor of the conversion from Special to Traditional because of they submitted test results and identified 2 new tests, the sponsor offered to withdraw both the data and tests. Since we have allowed other companies to send an email withdrawal, we agreed to this as well.

Pls concur to the RTA1 Special (**file #03**)

<https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

CTS log out

<http://webapps.cdrh.fda.gov/division-tracking/findTrackable.do?docNum=K131482>

Thanks.

From: [Hoang Quynh T.](#)
To: [Krauthamer, Victor](#); [Whang, Joyce M](#)
Cc: [Shimp, Samuel](#)
Subject: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.
Date: Tuesday, July 02, 2013 11:16:46 PM

Dear Victor and Joyce:

In this S1 round, we RTAA file. However, upon an in depth review, we determined that file needs to be converted.

Explanations for the conversion are in document **#05**, available in DocMan at <https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Pls reply to all so we know to forward form to 510(k) staff after your concurrence.

Thanks.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

November 22, 2013

Micro Vention, Inc.
c/o Ms. Naomi Gong
Sr. Regulatory Affairs Project Manager
1311 Valencia Avenue
Tustin, CA 92780

Re: K131482

Trade/Device Name: SOFIA Distal Access Catheter
Regulation Number: 21 CFR 870.1250
Regulation Name: Percutaneous Catheter
Regulatory Class: Class II
Product Code: DQY, DQO
Dated: October 22, 2013
Received: October 23, 2013

Dear Ms. Gong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Ms. Naomi Gong

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

for Victor Krauthamer, Ph.D.
Acting Director
Division of Neurological
and Physical Medicine Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Concurrence & Template History Page

[THIS PAGE IS INCLUDED IN IMAGE COPY ONLY]

Full Submission Number: K131482/S002

For Office of Compliance Contact Information:

http://insideportlets.fda.gov:9010/portal/page?_pageid=197,415881&_dad=portal&_schema=PORTAL&org=318

For Office of Surveillance and Biometrics Contact Information:

http://insideportlets.fda.gov:9010/portal/page?_pageid=197,415881&_dad=portal&_schema=PORTAL&org=423

Digital Signature Concurrence Table	
Reviewer Sign-Off	Samuel K. Shimp III 11/13/2013 @ 3:08PM
Branch Chief Sign-Off	Quynh Hoang November 21, 2013 DNPMD/NNDB
Division Sign-Off	

Template Name: K1(A) – SE after 1996

Template History:

Date of Update	By	Description of Update
7/27/09	Brandi Stuart	Added Updates to Boiler Table
8/7/09	Brandi Stuart	Updated HFZ Table
1/11/10	Diane Garcia	Liability/Warranty sentence added at bottom of 1 st page
10/4/11	M. McCabe Janicki	Removed IFU sheet and placed in Forms
9/25/12	Edwena Jones	Added digital signature format
12/12/12	M. McCabe Janicki	Added an extra line between letter signature block and the word "Enclosure". Also, added a missing digit in 4-digit extension on letterhead zip code: "002" should be "0002".
4/2/2013	M. McCabe Janicki	Edited sentence that starts "If you desire specific advice for your device on our labeling regulation (21 CFR Part 801)..." Replaced broken Compliance link with general link to DSMICA.
4/12/2013	Margaret McCabe Janicki	Fixed a typo: Paragraph 1, final sentence, "We remind you, however; that device labeling must be truthful..." Replaced incorrect semicolon with a comma.

510(k) Summary

Trade Name:	SOFIA™ Distal Access Catheter
Generic Name:	Percutaneous Catheter
Classification:	Class II, 21 CFR 870.1250 (DQY), 21 CFR 870.1200 (DQO)
Submitted By:	MicroVention, Inc 1311 Valencia Avenue Tustin, California U.S.A.
Contact:	Naomi Gong
Date:	2013 November 13
Predicate Device:	Chaperon Guiding Catheter (K082385) Headway Duo Microcatheter (K120917)

Device Description:

The SOFIA Distal Access Catheter is a single-lumen, flexible catheter designed with coil and braid reinforcement. The distal segment is steam-shapeable and it has a hydrophilic coating for navigation through the vasculature. The radiopaque marker is located at the distal end of the catheter for visualization under fluoroscopy. An introducer sheath and shaping mandrel are also provided.

Indications For Use:

The SOFIA Distal Access Catheter is indicated for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate introduction of diagnostic or therapeutic devices. It is not intended for use in coronary arteries.

Technological Comparison:

	Chaperon Guiding Catheter	SOFIA Distal Access Catheter	Headway Duo Microcatheter
Intended Use	Intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.	Intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.	Intended for general intravascular use, including the peripheral and coronary vasculature for the infusion of diagnostic agents, such as contrast media, and therapeutic agents, such as embolization materials. Intended for neurovascular use for the infusion of diagnostic agents, such as contrast media, and therapeutic agents that have been cleared or approved for use in the neurovasculature and are compatible with the inner diameter of the microcatheter.
Material Catheter Body	Outer layer of polyester elastomer; stainless steel braid; inner layer of PTFE (polytetrafluoroethylene).	Outer layer of polyurethane elastomer (Polyblend and Pellethane), polyether block amide (Pebax) and polyamide (Grilamid); inner layer of stainless steel braid/coil, PTFE and polyolefin elastomer	Outer layer of polyether block amide (Pebax) and polyamide (Grilamid); inner layer of stainless steel braid/coil and PTFE
Marker	Tungsten	Platinum/Iridium	Platinum/Iridium
Hub	Nylon	Nylon	Nylon
Strain Relief	Polyester elastomer	Polyurethane	Pebax
Introducer	Not applicable	Pebax	Pebax
Shaping Mandrel	Not applicable	Stainless steel	Stainless steel

	Chaperon Guiding Catheter	SOFIA Distal Access Catheter	Headway Duo Microcatheter
Catheter size	5 F (outer catheter)	5 F	1.6 - 2.1 F
ID	0.059 inch (1.5 mm)	0.055 inch (1.4 mm)	0.0165 inch (0.42 mm)
OD	0.068 inch (1.7 mm)	0.068 inch (1.7 mm)	0.023 – 0.0275 inch (0.58 – 0.70 mm)
Effective Length	95 cm (outer) 117 cm (inner)	125 cm	157 and 168 cm
Coating	Hydrophilic coating (Terumo proprietary coating)	Hydrophilic coating (Hydak [®] – same)	Hydrophilic coating (Hydak [®] – same)
Tip Configuration	Preshaped	Steam shapeable by user	Steam shapeable by user
Guidewire Compatibility	0.035 inch or 0.038 inch	0.035 inch or 0.038 inch	0.014 inch or smaller
Accessories	N/A	Introducer sheath and shaping mandrel	Introducer sheath and shaping mandrel
Method of Supply	Sterile and single use	Sterile and single use	Sterile and single use
Sterilization Method	Ethylene Oxide	Ethylene Oxide	Ethylene Oxide
Packaging Configuration	Catheter placed on packaging card that is inserted into Tyvek [®] pouch. Pouch and IFU placed in bleached sulfate carton box.	Catheter placed into a HDPE dispenser tube. Dispenser tube, introducer and shaping mandrel placed on a polyethylene packaging card that is inserted into a Tyvek [®] pouch. Pouch and IFU placed in bleached sulfate carton box.	Catheter placed into a polyethylene dispenser hoop. Introducer and shaping mandrel placed on polyethylene packaging card. Dispenser hoop and packaging card inserted into Tyvek [®] pouch. Pouch and IFU placed in bleached sulfate carton box.

Verification and Test Summary:

Bench Testing		
Test	Results	Conclusions
Simulated Use	Test articles achieved a rating ≥ 3 for preparation/ease of assembly, introducer sheath interaction, introducer peel away, tracking with guidewire/microcatheter, microcatheter/guidewire lockup, lubricity and durability of hydrophilic coating, microcatheter/guidewire removal, removal/aspiration of clot, mechanical clot retriever and stent delivery with no particles	Device performs as intended under simulated use conditions
Equipment Interface	Test articles compatible with 0.035-inch and 0.038-inch guidewires, 6F or larger guide catheter/guiding sheath, common RHVs using insertion tool, stopcocks and ≤ 0.027 -inch microcatheters	Device compatible with recommended accessories commonly used in intravascular procedures
Dimensional and Physical Attributes	Test articles met the specified dimensional requirements for catheter OD, catheter ID, overall working length, length of distal section, length of distal tip to marker band and total length of hub/strain relief	Device met established dimensional and physical specifications
Kink Resistance	No kinks at 1 cm, 4 cm, 12 cm and 25 cm from distal tip when wrapped around 0.025-inch and 0.030-inch pin gauges No kinks noted during simulated use testing	Device resistant to kinking around small radii turns
Tip Shapeability	Tip angle of test article equivalent to competitive devices after steam shaping using mandrel with an angle of approximately 90°	Shapeability of distal tip after steam shaping equivalent to competitive devices
Radio Detectability	Distal marker band visible under fluoroscopy	Device radiopacity equivalent to or better than predicate and competitive devices

Bench Testing		
Test	Results	Conclusions
Gauging (ISO 594-2)	Gauging pin and hub align in limit planes	Device hub meets the requirements of ISO 594-2
Separation Force (ISO 594-2)	Mating parts separation force greater than 25 N	Device hub meets the requirements of ISO 594-2
Unscrewing Torque (ISO 594-2)	Test article luer remains attached after applying an unscrewing torque not less than 0.02 Nm for a minimum of 10 seconds	Device hub meets the requirements of ISO 594-2
Stress Cracking (ISO 594-2)	No stress cracks on test article hub	Device hub meets the requirements of ISO 594-2
Ease of Assembly (ISO 594-2)	Components fit together securely with no resistance observed between test article luer and reference fitting	Device hub meets the requirements of ISO 594-2
Resistance to Overriding (ISO 594-2)	Test article luer does not override reference fitting threads	Device hub meets the requirements of ISO 594-2
Durability/Lubricity of Hydrophilic Coating	Test article achieved a rating of ≥ 3 during simulated use testing for coating durability and lubricity.	Device tracks easily with no coating cracking or separation
Catheter Stiffness	Device stiffness equivalent to predicate and competitive devices	Device tracks in tortuous anatomy while advancing to target site
Torque Strength	No catheter breakage after 50 rotations	Device torque strength same as predicate device
Catheter Flexural Fatigue	No flexural fatigue following repeated bending during simulated use testing and repeated hoop stress following pressure and air aspiration testing	Device integrity suitable for intended clinical use
Surface Contamination	Test article free from surface contaminants from uncured coating surface particulates $> 0.02 \text{ mm}^2$, embedded particulates Distal tip smooth and tapered PTFE inner layer not delaminated	Device integrity suitable for intended clinical use
Force at Break (Distal and Hub)	Catheter force at break $\geq 2.25 \text{ lbf}$ for distal section and hub/catheter junction	Tensile strength test results equivalent to predicate and competitive devices

Bench Testing		
Test	Results	Conclusions
Flow Rate	Flow rate at 100 psi and 300 psi with diagnostic agents (e.g., saline, contrast media) equivalent to or better than competitive devices	Device meets specified requirements for delivery of diagnostic agents
Static Burst Pressure	No damage of pressurized catheter at 46 psi	Device integrity suitable for intended clinical use and met requirements of ISO 10555-1
Fluid Leakage at > 46 psi	No liquid leakage from hub and catheter shaft at 46 psi for 30 seconds	Device integrity suitable for intended clinical use and met requirements of ISO 10555-1
Air Leakage	No air leakage at hub into syringe for 15 seconds	Device integrity suitable for intended clinical use and met requirements of ISO 10555-1
Dynamic Burst	Test articles did not burst at or below 300 psi	Device met labeled maximum infusion pressure of 300 psi
Particulate Test	Less than 25 particles greater than 10 microns per ml volume and less than 3 particles less than 25 microns per ml volume No particles greater than 70 microns	Device met specifications for maximum allowable particles

Biocompatibility		
Test	Results	Conclusions
Cytotoxicity – MEM Elution Assay (ISO 10993-5)	Cell culture treated with test article exhibited slight reactivity (Grade 1)	Non-cytotoxic
Sensitization/Irritation – Kligman Maximization Test (ISO 10993-10)	Extracts of the test article elicited no reaction at the challenge (0% sensitization) following the induction phase (Grade 1).	Weak allergic potential or sensitizing capacity
Sensitization/Irritation – Intracutaneous Injection Test (ISO 10993-10)	Extracts of the test article did not show a significantly greater biological reaction than the sites injected with the control article	Non-irritant
Hemocompatibility – Rabbit Blood Direct and Indirect Contact (ISO 10993-4)	The hemolysis index was 0.13% (direct contact) and 0.0% (indirect contact)	Non-hemolytic

Biocompatibility		
Test	Results	Conclusions
Hemocompatibility – Unactivated Partial Thromboplastin Time Assay Direct Contact (ISO 10993-4)	No statistically significant difference found between the Unactivated Partial Thromboplastin Time (UPTT) of the plasma exposed to the test article and that of the plasma exposed to either the negative control or the untreated control	No effect on coagulation
Hemocompatibility – Complement Activation Assay (ISO 10993-4)	C3a and SC5b-9 levels \leq negative and untreated controls	No effect on complement activation
Hemocompatibility – Thrombogenicity Study in Dogs (ISO 10993-4)	Minimal thrombosis observed with a Grade 0 in two out of two test sites and two out of two control sites	No significant thrombosis
Systemic Toxicity – Systemic Injection Test (ISO 10993-11)	Extracts of test article did not induce a significantly greater biological reaction than the control extracts when injected in Swiss Albino mice	No toxic effects
Systemic Toxicity - Rabbit Pyrogen Test (ISO 10993-11)	The temperature increases (maximum) was 0.03°C from baseline	Non-pyrogenic

Summary of Substantial Equivalence:

The data presented in this submission demonstrates the technological similarity and equivalency of the SOFIA Distal Access Catheter when compared with the predicate Chaperon Guiding Catheter (K082385) and the HEADWAY DUO Microcatheter (K120917) devices. The devices:

- Have an equivalent intended use,
- Use the same operating principle,
- Incorporate the same basic design,
- Are packaged and sterilized using the same methods.

In summary, the SOFIA Distal Access Catheter described in this submission is substantially equivalent to the predicate devices.

Indications for Use

510(k) Number (if known): K131482

Device Name: SOFIA Distal Access Catheter

Indications For Use:

The SOFIA Distal Access Catheter is intended for general intravascular use, including the neuro and peripheral vasculature. It can be used to facilitate the introduction of diagnostic and therapeutic agents. It is not intended for use in coronary arteries.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of Center for Devices and Radiological Health (CDRH)

From: [Hoang, Quynh T.](#)
To: [Tang, Xiaorui](#)
Cc: [Shimp, Samuel](#)
Subject: Pls forward to Samuel the text you used in your email. Thanks.
Date: Tuesday, June 04, 2013 2:02:52 PM

Dear Xiaorui:

I remember that in a couple of cases, you contacted the sponsor of a Special 510(k) to tell them that they shouldn't have included the test data in their Special submission and should send an email back to withdraw them from consideration. Please forward that email to Samuel so he has a sample for his case.

Thanks!

From: [Hoang, Quynh T.](#)
To: [Shimp, Samuel](#)
Subject: Pls inform the sponsor (w/ another FDA staff present, pls): K131482 Conversion fr Special to Traditional. Thanks.
Date: Thursday, June 06, 2013 12:00:06 PM

From: 510K Program
Sent: Thursday, June 06, 2013 7:06 AM
To: Hoang, Quynh T.; 510K Program
Cc: Shimp, Samuel; Krauthamer, Victor
Subject: RE: (Due 6/7) K131482 Conversion fr Special to Traditional. Thanks.

Good Day,

I concur with this conversion because the company provided raw data unsolicited. I have converted K131482 from Special to Traditional. The new (90th day) due date is August 21, 2013. I cannot e-sign the new conversion form, so please include this e-mail in the record as documentation of POS concurrence. Please let the company know about this conversion.

I have also put in a CTS ticket for you, CTSP-15062, to have the dates re-set to align w/Traditional timeframes.

Please let me know if you need anything further.

Margaret

From: Hoang, Quynh T.
Sent: Wednesday, June 05, 2013 5:20 PM
To: 510K Program
Cc: Shimp, Samuel
Subject: (Due 6/7) K131482 Conversion fr Special to Traditional. Thanks.

Dear 510(k) staff:

For your concurrence is the conversion from a Special to a Traditional.

The conversion form is **file #1**, at

<https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Once you are done, we'll inform the sponsor of the conversion and log out the RTAA (for a Traditional).

Thanks.

From: [Hoang, Quynh T.](#)
To: [Krauthamer, Victor](#); [Whang, Joyce M](#)
Cc: [Shimp, Samuel](#)
Subject: Pls reply to all, (due 6/7) K131482 Conversion fr Special to Traditional. Thanks.
Date: Tuesday, June 04, 2013 5:01:17 PM

Dear Victor and Joyce:

Pls reply to all, so we know to forward to 510(k) staff for their concurrence of the conversion from a Special to a Traditional.

The conversion form is **file #1**, at

<https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

[Your signature is not required for **file #2** as it's a RTAA. I will then log file out of CTS w/ an RTAA.]

Thanks.

From: [510K Program](#)
To: [Hoang, Quynh T.](#); [510K Program](#)
Cc: [Shimp, Samuel](#); [Krauthamer, Victor](#)
Subject: RE: (Due 6/7) K131482 Conversion fr Special to Traditional. Thanks.
Date: Thursday, June 06, 2013 7:05:34 AM

Good Day,

I concur with this conversion because the company provided raw data unsolicited. I have converted K131482 from Special to Traditional. The new (90th day) due date is August 21, 2013. I cannot e-sign the new conversion form, so please include this e-mail in the record as documentation of POS concurrence. Please let the company know about this conversion.

I have also put in a CTS ticket for you, CTSP-15062, to have the dates re-set to align w/Traditional timeframes.

Please let me know if you need anything further.

Margaret

From: Hoang, Quynh T.
Sent: Wednesday, June 05, 2013 5:20 PM
To: 510K Program
Cc: Shimp, Samuel
Subject: (Due 6/7) K131482 Conversion fr Special to Traditional. Thanks.

Dear 510(k) staff:

For your concurrence is the conversion from a Special to a Traditional.

The conversion form is **file #1**, at

<https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Once you are done, we'll inform the sponsor of the conversion and log out the RTAA (for a Traditional).

Thanks.

From: [Chao, Kuo](#)
To: [Shimp, Samuel](#)
Subject: RE: Another Tortuous Model - pls comment
Date: Tuesday, November 12, 2013 11:03:31 PM

It looks reasonable to me.

From: Shimp, Samuel
Sent: Tuesday, November 12, 2013 2:50 PM
To: Misra, Sanjay; Chao, Kuo; Froehler, Michael
Cc: Toy, Jeffrey; Zheng, Xiaolin
Subject: Another Tortuous Model - pls comment

(b) (5)



- Samuel

Samuel K. Shimp III, Ph.D.
Mechanical/Biomedical Engineer
FDA/CDRH/ODE/DNPMD/NNDB
WO66, Rm 1430

(301) 796-6610
Samuel.Shimp@fda.hhs.gov

From: [Toy, Jeffrey](#)
To: [Shimp, Samuel](#)
Subject: RE: Are these adequate tests?
Date: Thursday, June 27, 2013 3:33:59 PM

There are no other tests.

Yes, you can. In the future, I'm going to adjust the deficiency to outline the type of info I expect if a sponsor should wishes to omit genotox testing.

Jeffrey Toy, Ph.D.

From: Shimp, Samuel
Sent: Thursday, June 27, 2013 3:28 PM
To: Toy, Jeffrey
Subject: RE: Are these adequate tests?

Any other tests that we would require besides genotox? Can I regurgitate your genotox deficiency from Reverse Medical?

- Samuel

From: Toy, Jeffrey
Sent: Thursday, June 27, 2013 3:27 PM
To: Shimp, Samuel
Subject: RE: Are these adequate tests?

Samuel,

The sponsor has not addressed genotoxicity risk. If the sponsor wishes to omit genotoxicity testing the justification should be based on data and not hand waiving or a statement like the device material has a long history of use. Sponsors should know the chemicals that went into the manufacture of the device and demonstrate the residuals that leach out is consistent with what they put in (no unexplained new peaks on a chromatogram).

Jeffrey Toy, Ph.D.

From: Shimp, Samuel
Sent: Thursday, June 27, 2013 2:52 PM

To: Toy, Jeffrey

Subject: Are these adequate tests?

Hi Jeff:

For a catheter with a near-complete change of materials, would these biocompatibility tests be adequate? It does not look to me like they did genotox. Are there other tests that they would need to do?

- Samuel

From: [Toy, Jeffrey](#)
To: [Shimp, Samuel](#)
Subject: RE: Are these adequate tests?
Date: Thursday, June 27, 2013 3:27:14 PM

Samuel,

The sponsor has not addressed genotoxicity risk. If the sponsor wishes to omit genotoxicity testing the justification should be based on data and not hand waiving or a statement like the device material has a long history of use. Sponsors should know the chemicals that went into the manufacture of the device and demonstrate the residuals that leach out is consistent with what they put in (no unexplained new peaks on a chromatogram).

Jeffrey Toy, Ph.D.

From: Shimp, Samuel
Sent: Thursday, June 27, 2013 2:52 PM
To: Toy, Jeffrey
Subject: Are these adequate tests?

Hi Jeff:

For a catheter with a near-complete change of materials, would these biocompatibility tests be adequate? It does not look to me like they did genotox. Are there other tests that they would need to do?

- Samuel

From: [Ghosh, Molly](#)
To: [Shimp, Samuel](#)
Subject: RE: Cancelled Consult for K131482
Date: Friday, June 07, 2013 3:05:59 PM

Ok. Thx.

From: Shimp, Samuel
Sent: Friday, June 07, 2013 3:05 PM
To: Ghosh, Molly
Subject: Cancelled Consult for K131482

Hi Molly,

I had to cancel the consult for the file because the sponsor insisted that they wanted the file to remain as a special 510(k) and not be converted. Once I did the special 510(k) review we found that it needed to be RTA1. So, there is no need for the consult until they fix their submission and resubmit. I'll let you know if I need you, otherwise, Jeff might end up doing it at that time.

- Samuel

Samuel K. Shimp III, Ph.D.
Biomedical Engineer
FDA/CDRH/ODE/DNPMD/NNDB
WO66, Rm 2250

(301) 796-6610
Samuel.Shimp@fda.hhs.gov

From: [Whang, Joyce M](#)
To: [Lendor, Marisol](#)
Cc: [Shimp, Samuel](#); [Hoang, Quynh T.](#); [Krauthamer, Victor](#)
Subject: RE: Due today, K131482 Special RTA1. Thanks.
Date: Friday, June 07, 2013 3:48:27 PM

Concurred.

Marisol: Please return this document to DCC with the RTA1 coversheet. Thanks.

From: Hoang, Quynh T.
Sent: Friday, June 07, 2013 3:28 PM
To: Krauthamer, Victor; Whang, Joyce M; Lendor, Marisol
Cc: Shimp, Samuel
Subject: Due today, K131482 Special RTA1. Thanks.

Dear Victor and Joyce:

During the call to inform the sponsor of the conversion from Special to Traditional because of they submitted test results and identified 2 new tests, the sponsor offered to withdraw both the data and tests. Since we have allowed other companies to send an email withdrawal, we agreed to this as well.

Pls concur to the RTA1 Special (**file #03**)

<https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

CTS log out

<http://webapps.cdrh.fda.gov/division-tracking/findTrackable.do?docNum=K131482>

Thanks.

From: [Hoang, Quynh T.](#)
To: [Shimp, Samuel](#)
Subject: RE: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.
Date: Wednesday, July 03, 2013 9:42:12 AM

Dear Samuel:

If you haven't done so, pls forward to 510(k) staff for their OK before calling.

Pls alert me if you need my involvement; otherwise, I'll assume that it's a done deal.

Thanks.

From: Shimp, Samuel
Sent: Wednesday, July 03, 2013 9:39 AM
To: Hoang, Quynh T.
Subject: RE: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

I will give the firm a call today and let them know. I'll get Jeff or Lin to attend the call.

- Samuel

From: Whang, Joyce M
Sent: Wednesday, July 03, 2013 9:31 AM
To: Hoang, Quynh T.; Krauthamer, Victor
Cc: Shimp, Samuel
Subject: RE: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

I have concurred in DocMan.

From: Hoang, Quynh T.
Sent: Tuesday, July 02, 2013 11:17 PM
To: Krauthamer, Victor; Whang, Joyce M
Cc: Shimp, Samuel
Subject: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

Dear Victor and Joyce:

In this S1 round, we RTAA file. However, upon an in depth review, we determined that file needs to be converted.

Explanations for the conversion are in document **#05**, available in DocMan at <https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Pls reply to all so we know to forward form to 510(k) staff after your concurrence.

Thanks.

From: [Whang, Joyce M](#)
To: [Hoang, Quynh T.](#); [Krauthamer, Victor](#)
Cc: [Shimp, Samuel](#)
Subject: RE: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.
Date: Wednesday, July 03, 2013 9:31:12 AM

I have concurred in DocMan.

From: Hoang, Quynh T.
Sent: Tuesday, July 02, 2013 11:17 PM
To: Krauthamer, Victor; Whang, Joyce M
Cc: Shimp, Samuel
Subject: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

Dear Victor and Joyce:

In this S1 round, we RTAA file. However, upon an in depth review, we determined that file needs to be converted.

Explanations for the conversion are in document **#05**, available in DocMan at <https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Pls reply to all so we know to forward form to 510(k) staff after your concurrence.

Thanks.

From: [Lee, Patty](#)
To: [510K Program](#); [Shimp, Samuel](#)
Cc: [Hoang, Quynh T.](#)
Subject: RE: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.
Date: Wednesday, July 03, 2013 10:34:57 AM

Geeta,

converted

Patty Lee

OIM/DSM/CDRH IP Team
301-796-7783 (Office)
240-506-7039 (BB)

From: 510K Program
Sent: Wednesday, July 03, 2013 9:58 AM
To: Shimp, Samuel; 510K Program; Lee, Patty
Cc: Hoang, Quynh T.
Subject: RE: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.

Hi Patty,

510k staff concurs with the request to convert this special to traditional. This file is on S1 and I am unable to convert in APPS. Will you please complete this conversion when you get a chance?

Thank you,
Geeta

From: Shimp, Samuel
Sent: Wednesday, July 03, 2013 9:45 AM
To: 510K Program
Cc: Hoang, Quynh T.
Subject: FW: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.

Please follow the link below. Please call or email if you have any questions.

- Samuel

From: Whang, Joyce M
Sent: Wednesday, July 03, 2013 9:31 AM
To: Hoang, Quynh T.; Krauthamer, Victor
Cc: Shimp, Samuel
Subject: RE: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

I have concurred in DocMan.

From: Hoang, Quynh T.
Sent: Tuesday, July 02, 2013 11:17 PM
To: Krauthamer, Victor; Whang, Joyce M
Cc: Shimp, Samuel
Subject: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

Dear Victor and Joyce:

In this S1 round, we RTAA file. However, upon an in depth review, we determined that file needs to be converted.

Explanations for the conversion are in document **#05**, available in DocMan at <https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Pls reply to all so we know to forward form to 510(k) staff after your concurrence.

Thanks.

From: [510K Program](#)
To: [Lee, Patty](#); [510K Program](#); [Shimp, Samuel](#)
Cc: [Hoang, Quynh T.](#)
Subject: RE: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.
Date: Wednesday, July 03, 2013 10:36:35 AM

As always, many thanks for your quick action, Patty. Have a good 4th

From: Lee, Patty
Sent: Wednesday, July 03, 2013 10:35 AM
To: 510K Program; Shimp, Samuel
Cc: Hoang, Quynh T.
Subject: RE: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.

Geeta,

converted

Patty Lee

OIM/DSM/CDRH IP Team
301-796-7783 (Office)
240-506-7039 (BB)

From: 510K Program
Sent: Wednesday, July 03, 2013 9:58 AM
To: Shimp, Samuel; 510K Program; Lee, Patty
Cc: Hoang, Quynh T.
Subject: RE: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.

Hi Patty,

510k staff concurs with the request to convert this special to traditional. This file is on S1 and I am unable to convert in APPS. Will you please complete this conversion when you get a chance?

Thank you,
Geeta

From: Shimp, Samuel
Sent: Wednesday, July 03, 2013 9:45 AM
To: 510K Program
Cc: Hoang, Quynh T.
Subject: FW: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.

Please follow the link below. Please call or email if you have any questions.

- Samuel

From: Whang, Joyce M
Sent: Wednesday, July 03, 2013 9:31 AM
To: Hoang, Quynh T.; Krauthamer, Victor
Cc: Shimp, Samuel
Subject: RE: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

I have concurred in DocMan.

From: Hoang, Quynh T.
Sent: Tuesday, July 02, 2013 11:17 PM
To: Krauthamer, Victor; Whang, Joyce M
Cc: Shimp, Samuel
Subject: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

Dear Victor and Joyce:

In this S1 round, we RTAA file. However, upon an in depth review, we determined that file needs to be converted.

Explanations for the conversion are in document #05, available in DocMan at <https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Pls reply to all so we know to forward form to 510(k) staff after your concurrence.

Thanks.

From: [Lee, Patty](#)
To: [Shimp, Samuel](#)
Cc: [Jasper, Chad B *](#)
Subject: RE: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.
Date: Monday, July 08, 2013 9:57:54 AM

Samuel – I'm only handle APPS data. In APPS, K131482 has been converted to traditional last Friday (7/3). S001 was received on 6/12/2013 and it is due on 09/10/2013.

Chad – Please check CTS, see Samuel Shimp's email below, thanks.

Patty Lee

OIM/DSM/CDRH IP Team
301-796-7783 (Office)
240-506-7039 (BB)

From: Shimp, Samuel
Sent: Monday, July 08, 2013 9:46 AM
To: Lee, Patty
Subject: RE: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.

Hi Patty,

K131482/S001 still shows up as a special (i.e. the deadlines have not changed in CTS). Can you please look into this and see that the deadlines are updated to reflect that it is a traditional?

- Samuel

From: Lee, Patty
Sent: Wednesday, July 03, 2013 10:35 AM
To: 510K Program; Shimp, Samuel
Cc: Hoang, Quynh T.
Subject: RE: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.

Geeta,

converted

Patty Lee

OIM/DSM/CDRH IP Team
301-796-7783 (Office)
240-506-7039 (BB)

From: 510K Program
Sent: Wednesday, July 03, 2013 9:58 AM
To: Shimp, Samuel; 510K Program; Lee, Patty
Cc: Hoang, Quynh T.
Subject: RE: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.

Hi Patty,

510k staff concurs with the request to convert this special to traditional. This file is on S1 and I am unable to convert in APPS. Will you please complete this conversion when you get a chance?

Thank you,
Geeta

From: Shimp, Samuel
Sent: Wednesday, July 03, 2013 9:45 AM
To: 510K Program
Cc: Hoang, Quynh T.
Subject: FW: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.

Please follow the link below. Please call or email if you have any questions.

- Samuel

From: Whang, Joyce M

Sent: Wednesday, July 03, 2013 9:31 AM
To: Hoang, Quynh T.; Krauthamer, Victor
Cc: Shimp, Samuel
Subject: RE: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

I have concurred in DocMan.

From: Hoang, Quynh T.
Sent: Tuesday, July 02, 2013 11:17 PM
To: Krauthamer, Victor; Whang, Joyce M
Cc: Shimp, Samuel
Subject: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

Dear Victor and Joyce:

In this S1 round, we RTAA file. However, upon an in depth review, we determined that file needs to be converted.

Explanations for the conversion are in document **#05**, available in DocMan at
<https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Pls reply to all so we know to forward form to 510(k) staff after your concurrence.

Thanks.

(b) (5)



(b) (5)



From: [510K Program](#)
To: [Shimp, Samuel](#); [510K Program](#); [Lee, Patty](#)
Cc: [Hoang, Quynh T.](#)
Subject: RE: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.
Date: Wednesday, July 03, 2013 9:57:55 AM

Hi Patty,

510k staff concurs with the request to convert this special to traditional. This file is on S1 and I am unable to convert in APPS. Will you please complete this conversion when you get a chance?

Thank you,
Geeta

From: Shimp, Samuel
Sent: Wednesday, July 03, 2013 9:45 AM
To: 510K Program
Cc: Hoang, Quynh T.
Subject: FW: For your concurrence, K131482/S1 convert fr Special to Traditional. Thanks.

Please follow the link below. Please call or email if you have any questions.

- Samuel

From: Whang, Joyce M
Sent: Wednesday, July 03, 2013 9:31 AM
To: Hoang, Quynh T.; Krauthamer, Victor
Cc: Shimp, Samuel
Subject: RE: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

I have concurred in DocMan.

From: Hoang, Quynh T.
Sent: Tuesday, July 02, 2013 11:17 PM
To: Krauthamer, Victor; Whang, Joyce M
Cc: Shimp, Samuel
Subject: For division concurrence (pls reply to all), K131482/S1 convert fr Special to Traditional. Thanks.

Dear Victor and Joyce:

In this S1 round, we RTAA file. However, upon an in depth review, we determined that file needs to be converted.

Explanations for the conversion are in document #05 , available in DocMan at <https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Pls reply to all so we know to forward form to 510(k) staff after your concurrence.

Thanks.

From: [Hoang, Quynh T.](#)
To: [Shimp, Samuel](#)
Subject: RE: K131482 - Convert Special to Traditional: Unsolicited Data/Full Test Reports
Date: Tuesday, June 04, 2013 3:02:37 PM

Dear Samuel:

New tests / performance criteria mean that the existing design controls do not adequately cover the new device. This removes the new device from a Special. Sorry, no email will fix this one.

Pls convert. Thanks.

From: Shimp, Samuel
Sent: Tuesday, June 04, 2013 2:58 PM
To: Hoang, Quynh T.
Subject: RE: K131482 - Convert Special to Traditional: Unsolicited Data/Full Test Reports

Hi Quynh,

Even though the old device was a 2 catheter system, I talked to Mike Froehler and he felt that the device was the same, no new technology. Same basic usage principles as the predicate. However, there are two new tests being conducted on the new device and one of the repeated tests seems to have new criteria for the current sub. I would be comfortable reviewing it as a special as long as I could ask for clarification (interactively) on test protocol for one or two of the new tests.

- Samuel

From: Hoang, Quynh T.
Sent: Tuesday, June 04, 2013 1:52 PM
To: Shimp, Samuel
Subject: RE: K131482 - Convert Special to Traditional: Unsolicited Data/Full Test Reports

Dear Samuel:

Is the inclusion of data the only reason for the conversion? If so, we just make a call or send an email to ask the sponsor to respond in an email that they inadvertently submitted the data and request to withdraw those data and that the FDA would ignore them in our review.

Thanks.

From: Shimp, Samuel
Sent: Tuesday, June 04, 2013 8:31 AM
To: Hoang, Quynh T.
Subject: K131482 - Convert Special to Traditional: Unsolicited Data/Full Test Reports

Hi Quynh,

The Sofia Distal Access Catheter file (K131482) included full test reports for bench top and biocompatibility testing. The file seems to have considerable changes to materials and possibly design and I think the inclusion of the reports is warranted, however, it required review as a traditional and not a special 510(k). The conversion form is uploaded to DM in the [Reviewer Docs](#). I am now working on the traditional 510(k) RTA. (RTA is due Friday)

- Samuel

Samuel K. Shimp III, Ph.D.
Biomedical Engineer
FDA/CDRH/ODE/DNPMD/NNDB
WO66, Rm 2250

(301) 796-6610
Samuel.Shimp@fda.hhs.gov

From: [Hoang, Quynh T.](#)
To: [Shimp, Samuel](#)
Subject: RE: K131482 - Convert Special to Traditional: Unsolicited Data/Full Test Reports
Date: Tuesday, June 04, 2013 1:52:22 PM

Dear Samuel:

Is the inclusion of data the only reason for the conversion? If so, we just make a call or send an email to ask the sponsor to respond in an email that they inadvertently submitted the data and request to withdraw those data and that the FDA would ignore them in our review.

Thanks.

From: Shimp, Samuel
Sent: Tuesday, June 04, 2013 8:31 AM
To: Hoang, Quynh T.
Subject: K131482 - Convert Special to Traditional: Unsolicited Data/Full Test Reports

Hi Quynh,

The Sofia Distal Access Catheter file (K131482) included full test reports for bench top and biocompatibility testing. The file seems to have considerable changes to materials and possibly design and I think the inclusion of the reports is warranted, however, it required review as a traditional and not a special 510(k). The conversion form is uploaded to DM in the [Reviewer Docs](#). I am now working on the traditional 510(k) RTA. (RTA is due Friday)

- Samuel

Samuel K. Shimp III, Ph.D.
Biomedical Engineer
FDA/CDRH/ODE/DNPMD/NNDB
WO66, Rm 2250

(301) 796-6610
Samuel.Shimp@fda.hhs.gov

From: [Ghosh, Molly](#)
To: [Shimp, Samuel](#)
Cc: [Marjenin, Timothy](#)
Subject: RE: K131482 - Sofia Catheter
Date: Friday, June 07, 2013 8:42:13 AM

No problem! 7/1 works.

From: Shimp, Samuel
Sent: Friday, June 07, 2013 7:55 AM
To: Ghosh, Molly
Cc: Marjenin, Timothy
Subject: K131482 - Sofia Catheter

Hi Molly,
(cc: Tim "Consult Gatekeeper" Marjenin)

Since Jeff is away for a few weeks, would you mind taking a look at this catheter for biocompatibility? Just as an FYI, they submitted the file as a special but included full test reports and a couple of new tests that got it converted to a traditional 510(k).

Please let me know if you have any questions or concerns. I set the due date for the consult to 7/1. I'm not sure how busy you are so let me know if you need more time. Thanks!

- Samuel

Samuel K. Shimp III, Ph.D.
Biomedical Engineer
FDA/CDRH/ODE/DNPMD/NNDB
WO66, Rm 2250

(301) 796-6610
Samuel.Shimp@fda.hhs.gov

From: [Lee, Patty](#)
To: [McCabe-Janicki, Margaret](#); [Shimp, Samuel](#)
Cc: [Clayton, Jeff](#)
Subject: RE: K131482/s001 - Please convert back to Special
Date: Friday, June 21, 2013 1:19:58 PM

Margaret – This 510k has been converted to special

Patty Lee

OIM/DSM/CDRH IP Team
301-796-7783 (Office)
240-506-7039 (BB)

From: McCabe-Janicki, Margaret
Sent: Friday, June 21, 2013 12:12 PM
To: Shimp, Samuel; Lee, Patty
Cc: Clayton, Jeff
Subject: RE: K131482/s001 - Please convert back to Special

Hi Samuel – sure. If we convert solely because they provided the raw data, then they do have the option of withdrawing the data and giving us summary data & risk analysis instead. Then we have to convert it back to Special.

Hi Patty or Jeff – I can't do this conversion b/c the file has had a supplement. Can you please convert the subject 510(k) back to Special?

Thx!
Margaret

From: Shimp, Samuel
Sent: Friday, June 21, 2013 12:08 PM
To: McCabe-Janicki, Margaret
Subject: K131482/s001 - Special Converted to Traditional but now needs to go back to Special

Margaret:

This is an annoying file. They submitted as special but we determined it should be Traditional because they submitted new tests and complete data. When we called to notify of the conversion they insisted on retracting the data and tests and having it converted back to special. However as a special it did not make the RTA cut so it was RTA'ed. I had thought it was reconverted back to special last time around. Now it is back and will be accepted by RTA (RTAA) but I noticed the file is still listed as a traditional. Bottom line, can you please make the file "special"...for now ☺...

- Samuel

Samuel K. Shimp III, Ph.D.
Biomedical Engineer
FDA/CDRH/ODE/DNPMD/NNDB
WO66, Rm 1430

(301) 796-6610

Samuel.Shimp@fda.hhs.gov

From: [McCabe-Janicki, Margaret](#)
To: [Lee, Patty](#); [Shimp, Samuel](#)
Cc: [Clayton, Jeff](#)
Subject: RE: K131482/s001 - Please convert back to Special
Date: Friday, June 21, 2013 1:21:59 PM

Hi Patty – thank you.

Hi Samuel – I have put in a CTS ticket for you, CTSP-15241, to have the dates set back to Special timelines.

Margaret

From: Lee, Patty
Sent: Friday, June 21, 2013 1:20 PM
To: McCabe-Janicki, Margaret; Shimp, Samuel
Cc: Clayton, Jeff
Subject: RE: K131482/s001 - Please convert back to Special

Margaret – This 510k has been converted to special

Patty Lee

OIM/DSM/CDRH IP Team
301-796-7783 (Office)
240-506-7039 (BB)

From: McCabe-Janicki, Margaret
Sent: Friday, June 21, 2013 12:12 PM
To: Shimp, Samuel; Lee, Patty
Cc: Clayton, Jeff
Subject: RE: K131482/s001 - Please convert back to Special

Hi Samuel – sure. If we convert solely because they provided the raw data, then they do have the option of withdrawing the data and giving us summary data & risk analysis instead. Then we have to convert it back to Special.

Hi Patty or Jeff – I can't do this conversion b/c the file has had a supplement. Can you please convert the subject 510(k) back to Special?

Thx!
Margaret

From: Shimp, Samuel
Sent: Friday, June 21, 2013 12:08 PM
To: McCabe-Janicki, Margaret
Subject: K131482/s001 - Special Converted to Traditional but now needs to go back to Special

Margaret:

This is an annoying file. They submitted as special but we determined it should be Traditional because they submitted new tests and complete data. When we called to notify of the conversion

they insisted on retracting the data and tests and having it converted back to special. However as a special it did not make the RTA cut so it was RTA'ed. I had thought it was reconverted back to special last time around. Now it is back and will be accepted by RTA (RTAA) but I noticed the file is still listed as a traditional. Bottom line, can you please make the file "special"...for now 😊...

- Samuel

Samuel K. Shimp III, Ph.D.
Biomedical Engineer
FDA/CDRH/ODE/DNPMD/NNDB
WO66, Rm 1430

(301) 796-6610

Samuel.Shimp@fda.hhs.gov

From: [McCabe-Janicki, Margaret](#)
To: [Shimp, Samuel](#); [Lee, Patty](#)
Cc: [Clayton, Jeff](#)
Subject: RE: K131482/s001 - Please convert back to Special
Date: Friday, June 21, 2013 12:11:39 PM

Hi Samuel – sure. If we convert solely because they provided the raw data, then they do have the option of withdrawing the data and giving us summary data & risk analysis instead. Then we have to convert it back to Special.

Hi Patty or Jeff – I can't do this conversion b/c the file has had a supplement. Can you please convert the subject 510(k) back to Special?

Thx!
Margaret

From: Shimp, Samuel
Sent: Friday, June 21, 2013 12:08 PM
To: McCabe-Janicki, Margaret
Subject: K131482/s001 - Special Converted to Traditional but now needs to go back to Special

Margaret:

This is an annoying file. They submitted as special but we determined it should be Traditional because they submitted new tests and complete data. When we called to notify of the conversion they insisted on retracting the data and tests and having it converted back to special. However as a special it did not make the RTA cut so it was RTA'ed. I had thought it was reconverted back to special last time around. Now it is back and will be accepted by RTA (RTAA) but I noticed the file is still listed as a traditional. Bottom line, can you please make the file "special"...for now ☺...

- Samuel

Samuel K. Shimp III, Ph.D.
Biomedical Engineer
FDA/CDRH/ODE/DNPMD/NNDB
WO66, Rm 1430

(301) 796-6610
Samuel.Shimp@fda.hhs.gov

From: [Hoang, Quynh T.](#)
To: [Shimp, Samuel](#)
Subject: RE: K131482/S002 I received file but it is not yet assigned in CTS.
Date: Wednesday, October 30, 2013 4:36:03 PM

Thanks, Samuel. I must have forgotten to save on CTS after assigning file to you. It's now fixed.

From: Shimp, Samuel
Sent: Wednesday, October 30, 2013 3:28 PM
To: Hoang, Quynh T.
Subject: K131482/S002 I received file but it is not yet assigned in CTS.

Samuel K. Shimp III, Ph.D.
Mechanical/Biomedical Engineer
FDA/CDRH/ODE/DNPM/NNDB
WO66, Rm 1430

(301) 796-6610
Samuel.Shimp@fda.hhs.gov

From: [Getzoff, Natalie B](#)
To: [Shimp, Samuel](#)
Subject: RE: K131482-S001 MicroVention Sofia Catheter RTAA
Date: Friday, June 21, 2013 1:26:02 PM

So do I just concur?

Natalie Getzoff, MD, Medical Officer
Office of Combination Products ([detail](#))
Food and Drug Administration

tel: 301-796-6495
natalie.getzoff@fda.hhs.gov

From: Shimp, Samuel
Sent: Friday, June 21, 2013 12:04 PM
To: Getzoff, Natalie B
Subject: K131482-S001 MicroVention Sofia Catheter RTAA

They took care of the RTA deficiencies so we can accept this file now. The RTA Checklist is here:

<https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Note, the RTA is not due until 6/27 (next Thursday) so you may be able to ignore this and pass it on to Quynh.

- Samuel

Samuel K. Shimp III, Ph.D.
Biomedical Engineer
FDA/CDRH/ODE/DNPMD/NNDB
WO66, Rm 1430

(301) 796-6610
Samuel.Shimp@fda.hhs.gov

From: [Hoang, Quynh T.](#)
To: [Shimp, Samuel](#)
Subject: RE: Pls inform the sponsor (w/ another FDA staff present, pls): K131482 Conversion fr Special to Traditional. Thanks.
Date: Thursday, June 06, 2013 4:13:32 PM

Thanks. Samuel. If they are unavailable when you call tomorrow, pls email the conversion form. I'll then follow with the CTS log-out.

From: Shimp, Samuel
Sent: Thursday, June 06, 2013 3:55 PM
To: Hoang, Quynh T.
Subject: RE: Pls inform the sponsor (w/ another FDA staff present, pls): K131482 Conversion fr Special to Traditional. Thanks.

Lin and I tried to contact the sponsor but they did not answer so I emailed them to arrange a phone call for tomorrow.

- Samuel

From: Hoang, Quynh T.
Sent: Thursday, June 06, 2013 12:00 PM
To: Shimp, Samuel
Subject: Pls inform the sponsor (w/ another FDA staff present, pls): K131482 Conversion fr Special to Traditional. Thanks.

From: 510K Program
Sent: Thursday, June 06, 2013 7:06 AM
To: Hoang, Quynh T.; 510K Program
Cc: Shimp, Samuel; Krauthamer, Victor
Subject: RE: (Due 6/7) K131482 Conversion fr Special to Traditional. Thanks.

Good Day,

I concur with this conversion because the company provided raw data unsolicited. I have converted K131482 from Special to Traditional. The new (90th day) due date is August 21, 2013. I cannot e-sign the new conversion form, so please include this e-mail in the record as documentation of POS concurrence. Please let the company know about this conversion.

I have also put in a CTS ticket for you, CTSP-15062, to have the dates re-set to align w/Traditional timeframes.

Please let me know if you need anything further.

Margaret

From: Hoang, Quynh T.
Sent: Wednesday, June 05, 2013 5:20 PM
To: 510K Program
Cc: Shimp, Samuel
Subject: (Due 6/7) K131482 Conversion fr Special to Traditional. Thanks.

Dear 510(k) staff:

For your concurrence is the conversion from a Special to a Traditional.

The conversion form is **file #1**, at <https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

Once you are done, we'll inform the sponsor of the conversion and log out the RTAA (for a Traditional).

Thanks.

From: [Krauthamer, Victor](#)
To: [Hoang, Quynh T.](#)
Cc: [Shimp, Samuel](#); [Whang, Joyce M](#)
Subject: RE: Pls reply to all, (due 6/7) K131482 Conversion fr Special to Traditional. Thanks.
Date: Tuesday, June 04, 2013 5:36:39 PM

Quynh,

I signed both forms. Please do the CTS log out.

Victor

From: Hoang, Quynh T.
Sent: Tuesday, June 04, 2013 5:01 PM
To: Krauthamer, Victor; Whang, Joyce M
Cc: Shimp, Samuel
Subject: Pls reply to all, (due 6/7) K131482 Conversion fr Special to Traditional. Thanks.

Dear Victor and Joyce:

Pls reply to all, so we know to forward to 510(k) staff for their concurrence of the conversion from a Special to a Traditional.

The conversion form is **file #1**, at

<https://docs.fda.gov/share/page/site/submissions/documentlibrary#filter=path%7C%2FReviews%2F510%2528k%2529s%2F2013%2F1401%2520-%25201600%2FK131482%2FReviewer%2520Documents%7C&page=1>

[Your signature is not required for **file #2** as it's a RTAA. I will then log file out of CTS w/ an RTAA.]

Thanks.

From: [Hoang, Quynh T.](#)
To: [Shimp, Samuel](#)
Subject: RE: To RTA or not to RTA? (K131482)
Date: Tuesday, June 04, 2013 11:45:16 AM

Dear Samuel:

Your reasoning is sound. I don't believe Form 3514 is a part of the RTA process. The other item #9 is not RTA1 by itself.

Thanks.

From: Shimp, Samuel
Sent: Tuesday, June 04, 2013 9:15 AM
To: Hoang, Quynh T.
Subject: To RTA or not to RTA? (K131482)

K131482 needs to be converted from special to traditional (see previous email). However, it looks like it is acceptable (RTAA) for a traditional 510(k) except for 2 things:

- The firm cites standards and used form 3654 to identify the extent to which each standard was conformed to BUT they didn't fill out the section I of the CDRH Premarket Review Submission Cover Sheet (form 3514) in which they are supposed to give the list of standards they conform to. The way I read the deficiency, this is not technically an RTA offense (Traditional RTA checklist #8).
- The firm did not state whether or not there were prior submissions. (RTA checklist #9). My understanding is that this alone is not supposed to warrant an RTA but can be cited if there are other issues. Since this is a modification to an existing catheter I don't really think that there were prior submissions.

How do you suggest I proceed?

(Checklist attached for your reference.)

- Samuel

Samuel K. Shimp III, Ph.D.
Biomedical Engineer
FDA/CDRH/ODE/DNPMD/NNDB
WO66, Rm 2250

(301) 796-6610

Samuel.Shimp@fda.hhs.gov



DEPARTMENT OF HEALTH AND HUMAN SERVICES **MEMORANDUM**

Food and Drug Administration
Office of Device Evaluation

K131482/S001

Date: August 9, 2013
To: MicroVention
% Naomi Gong
Phone: (714) 247-8055
Fax: (714) 247-8014
Email: Naomi.Gong@microvention.com

From: Samuel Shimp, Ph.D., Biomedical Engineer

Subject: Telephone hold for K131482/S001 for the SOFIA Distal Access Catheter

Dear Ms. Gong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above. We cannot determine if the device is substantially equivalent to a legally marketed predicate device based solely on the information you provided; this document serves to notify you that we have placed the file on hold. To complete the review of your submission, we require that you respond to the following deficiencies:

(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



The deficiencies identified above represent the issues that we believe need to be resolved before our review of your 510(k) submission can be successfully completed. In developing the deficiencies, we carefully considered the statutory criteria as defined in Section 513(i) of the Federal Food, Drug, and Cosmetic Act (Act) for determining substantial equivalence of your device.

You may not market this device until you have provided adequate information described above and required by 21 CFR 807.87(l), and you have received a letter from FDA allowing you to do so. If you market the device without conforming to these requirements, you will be in violation of the Act. You may, however, distribute this device for investigational purposes to obtain clinical data if needed to establish substantial equivalence. Clinical investigations of this device

must be conducted in accordance with the investigational device exemption (IDE) regulations (21 CFR 812).

If the information, or a request for an extension of time, is not received within 180 days, we will consider your premarket notification to be withdrawn and your submission will be deleted from our system. If you submit the requested information after 180 days it will be considered and processed as a new 510(k)(21 CFR 807.87(l)); therefore, all information previously submitted must be resubmitted so that your new 510(k) is complete. For guidance on 510(k) actions, please see our guidance document entitled, "Guidance for Industry and FDA Staff: Interactive Review for Medical Device Submissions: 510(k)s, Original PMAs, PMA Supplements, Original BLAs, and BLA Supplements" at

<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089402.htm>.

The purpose of this document is to assist agency staff and the device industry in understanding how various FDA and industry actions that may be taken on 510(k)s should affect the review clock for purposes of meeting the Medical Device User Fee and Modernization Act. You may review this document at

<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Overview/MedicalDeviceUserFeeandModernizationActMDUFMA/default.htm>.

The requested information should reference your above 510(k) number and should be submitted in duplicate to:

U.S. Food and Drug Administration
Center for Devices and Radiological Health
Document Mail Center – WO66-0609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

If you have any questions concerning the contents of the letter, please contact me by email at Samuel.Shimp@fda.hhs.gov or by phone at (301) 796-6610. If you need information or assistance concerning the IDE regulations, please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or at (301) 796-7100, or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.



DEPARTMENT OF HEALTH AND HUMAN SERVICES **MEMORANDUM**

Food and Drug Administration
Office of Device Evaluation

K131482/S001

Date: August 9, 2013
To: MicroVention
% Naomi Gong
Phone: (714) 247-8055
Fax: (714) 247-8014
Email: Naomi.Gong@microvention.com

From: Samuel Shimp, Ph.D., Biomedical Engineer

Subject: Telephone hold for K131482/S001 for the SOFIA Distal Access Catheter

Dear Ms. Gong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above. We cannot determine if the device is substantially equivalent to a legally marketed predicate device based solely on the information you provided; this document serves to notify you that we have placed the file on hold. To complete the review of your submission, we require that you respond to the following deficiencies:

(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



(b)(4) Confidential and Proprietary Information



The deficiencies identified above represent the issues that we believe need to be resolved before our review of your 510(k) submission can be successfully completed. In developing the deficiencies, we carefully considered the statutory criteria as defined in Section 513(i) of the Federal Food, Drug, and Cosmetic Act (Act) for determining substantial equivalence of your device.

You may not market this device until you have provided adequate information described above and required by 21 CFR 807.87(l), and you have received a letter from FDA allowing you to do so. If you market the device without conforming to these requirements, you will be in violation of the Act. You may, however, distribute this device for investigational purposes to obtain clinical data if needed to establish substantial equivalence. Clinical investigations of this device

must be conducted in accordance with the investigational device exemption (IDE) regulations (21 CFR 812).

If the information, or a request for an extension of time, is not received within 180 days, we will consider your premarket notification to be withdrawn and your submission will be deleted from our system. If you submit the requested information after 180 days it will be considered and processed as a new 510(k)(21 CFR 807.87(l)); therefore, all information previously submitted must be resubmitted so that your new 510(k) is complete. For guidance on 510(k) actions, please see our guidance document entitled, "Guidance for Industry and FDA Staff: Interactive Review for Medical Device Submissions: 510(k)s, Original PMAs, PMA Supplements, Original BLAs, and BLA Supplements" at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089402.htm>.

The purpose of this document is to assist agency staff and the device industry in understanding how various FDA and industry actions that may be taken on 510(k)s should affect the review clock for purposes of meeting the Medical Device User Fee and Modernization Act. You may review this document at <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Overview/MedicalDeviceUserFeeandModernizationActMDUFMA/default.htm>.

The requested information should reference your above 510(k) number and should be submitted in duplicate to:

U.S. Food and Drug Administration
Center for Devices and Radiological Health
Document Mail Center – WO66-0609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

If you have any questions concerning the contents of the letter, please contact me by email at Samuel.Shimp@fda.hhs.gov or by phone at (301) 796-6610. If you need information or assistance concerning the IDE regulations, please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or at (301) 796-7100, or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.



Contains Nonbinding Recommendations

Print Form

Acceptance Checklist for Special 510(k)s

(Should be completed within 15 days of DCC receipt)

The following information is not intended to serve as a comprehensive review.

510(k) #: K131482

Date Received by DCC: May 23, 2013

Lead Reviewer: Samuel Shimp, Ph.D.

Branch: NNDB

Division: DNPMD

Center/Office: CDRH/ODE

Note: If an element is left blank on the checklist, it does not mean the checklist is incomplete. It means the reviewer did not assess the element during RTA and the element will be assessed during the substantive review.

Special 510(k) Criteria

The submission should not be reviewed as a Special 510(k) if "No" is selected for any of the 4 criteria below. Complete the Refuse to Accept Checklist for a Traditional 510(k) if submission is converted.

	Yes	No
1) 510(k) is submitted to modify a legally marketed device (predicate) AND the Special 510(k) submission is submitted by the holder of the 510(k) for the predicate device.	X	
Comments?		
2. Indications for Use of the proposed device are unchanged from the legally marketed device (predicate).	X	
Comments?		
3) Fundamental scientific technology of the proposed device is unchanged from the legally marketed device (predicate).	X	
Comments?		
4) The submission includes only summary-level information (i.e., NO test reports with performance data). Note that if performance data are provided and are conducted under design validation (21 CFR 820.30(g)), for example, to demonstrate continued conformance with a special control or recognized standard, then a Special 510(k) may be appropriate.		X
Comments?	The submission was not limited to only summary-level information. The submission included the full test reports for the Design Verification and Validation testing (appendix 6) and the full test reports for the Biocompatibility testing (appendix 7). Furthermore two new tests are used to validate the design and one of the existing tests has new testing criteria/measurement methods.	

Does the submission meet all 4 criteria above?

- Yes, submission meets criteria for a Special 510(k). Continue with the remainder of this checklist below.
- No, submission does not meet criteria for a Special 510(k). Discontinue this RTA checklist; convert to a Traditional using the form below and apply the Traditional checklist.

Form for Converting a Special 510(k) to a Traditional or Abbreviated 510(k)

Records processed under FOIA Request # 2019-0024, Released by CDRH on 09-26-2016

Note: Please send this to 510k Staff electronically. You do not need anyone to sign this in person.

Conversion Instructions

510(k) #: K131482

Date Received by DCC: May 23, 2013

Lead Reviewer: Samuel Shimp, Ph.D.

Device Name: Sofia Distal Access Catheter

- Change in Indications for Use (please list old and new indications below)
- Change in Technology (select one):
 - We have not seen this change before in this device type.
 - We have seen this change before in this device type, but we need to see the data (please provide a brief statement below regarding why summary data/risk analysis are insufficient)
- Other (e.g. submission included unsolicited data or sponsor is modifying a device that is not their own - please specify below)

The submission included unsolicited data including the full reports for design verification and validation as well as full reports for biocompatibility. Furthermore two new tests are used to validate the design and one of the existing tests has new testing criteria/ measurement methods, which indicate that the predicate's design controls do not adequately cover the new device.

Digital Signature Concurrence Table	
Reviewer Sign-Off	Samuel K. Shimp III 2013.06.04 15:08:37 -04'00'
Branch Chief Sign-Off	Quynh T. Hoang -S 2013.06.04 16:54:26 -04'00'
Division Sign-Off	Victor Krauthamer -S 2013.06.04 17:33:36 -04'00'



Contains Nonbinding Recommendations

Print Form

Acceptance Checklist for Traditional 510(k)s

(Should be completed within 15 days of DCC receipt)

The following information is not intended to serve as a comprehensive review.

510(k) #: K131482

Date Received by DCC: May 23, 2013

Lead Reviewer: Samuel Shimp, Ph.D.

Branch: NNDB

Division: DNPMD

Center/Office: CDRH/ODE

Note: If an element is left blank on the checklist, it does not mean the checklist is incomplete. It means the reviewer did not assess the element during RTA and the element will be assessed during the substantive review.

Preliminary Questions		
Answers in the shaded blocks indicate consultations with Center advisor is needed	Yes	No
<p>1) Is the product a device (per section 201(h) of the FD&C Act) or a combination product (per 21 CFR 3.2(e)) with a device constituent part subject to review in a 510(k)?</p> <p>If it appears not to be a device (per section 201(h) of the FD&C Act) or such a combination product, or you are unsure, consult with the CDRH Jurisdictional Officer or the CBER Office Jurisdiction Liaison to determine the appropriate action, and inform division management. <i>Provide a summary of the Jurisdictional Officer's/Liaison's determination.</i> If the product does not appear to be a device or such a combination product, mark "No."</p>	X	
Comments?		
<p>2. Is the application with the appropriate Center?</p> <p>If the product is a device or a combination product with a device constituent part, is it subject to review by the Center in which the submission was received? If you believe the application is not with the appropriate Center or you are unsure, consult with the CDRH Jurisdictional Officer or CBER Office Jurisdiction Liaison to determine the appropriate action and inform your division management. <i>Provide a summary of the Jurisdictional Officer's/Liaison's determination.</i> If application should not be reviewed by your Center mark "No."</p>	X	
Comments?		
<p>3) If a Request for Designation was submitted for the device or combination product with a device constituent part and assigned to your center, identify the RFD # and confirm the following:</p> <p>a) Is the device or combination product the same (e.g., design, formulation) as that presented in the RFD submission?</p> <p>b) Are the indications for use for the device or combination product identified in the 510(k) the same as those identified in the RFD submission ?</p> <p>If you believe the product or the indications presented in the 510(k) have changed from the RFD, or you are unsure, consult with the CDRH Jurisdictional Officer or appropriate CBER Jurisdiction Liaison to determine the appropriate action and inform your division management. Provide summary of Jurisdictional Officer's/Liaison's determination. If the answer to either question is no, mark "No." If there was no RFD, skip this question.</p>		
Comments?		
<p>4) Is this device type eligible for a 510(k) submission?</p> <p>If a 510(k) does not appear to be appropriate (e.g., Class III type and PMA required, or Class I or II type and 510(k)-exempt), you should consult with the CDRH 510(k) Program Director or appropriate CBER staff during the acceptance review. If 510(k) is not the appropriate regulatory submission, mark "No."</p>	X	
Comments?		

<p>5) Is there a pending PMA for the same device with the same indications for use? If yes, consult division management and the CDRH 510(k) Program Director or appropriate CBER staff to determine the appropriate action.</p>		×
<p>Comments?</p>		
<p>6) If clinical studies have been submitted, is the submitter the subject of an Application Integrity Policy (AIP)? If yes, consult with the CDRH Office of Compliance/Division of Bioresearch Monitoring (OC/DBM - BIMO) or CBER Office of Compliance and Biologics Quality/Division of Inspections and Surveillance/Bioresearch Monitoring Branch (OCBQ/DIS/BMB) to determine the appropriate action. Check on web at http://www.fda.gov/ICECI/EnforcementActions/ApplicationIntegrityPolicy/ucm134453.htm</p>		×
<p>Comments?</p>		

If the answer to 1 or 2 appears to be "No," then stop review of the 510(k) and issue the "Original Jurisdictional Product" letter.
 If the answer to 3a or 3b appears to be "No," then stop the review and contact the CDRH Jurisdictional Officer or CBER Office of Jurisdiction Liaison.
 If the answer to 4 is "No," the lead reviewer should consult division management and other Center resources to determine the appropriate action.
 If the answer to 5 is "Yes," then stop review of the 510(k), contact the CDRH 510(k) Staff and PMA Staff, or appropriate CBER staff.
 If the answer to 6 is "Yes," then contact CDRH/OC/DBM-BIMO or CBER/OCBQ/DIS/BMB, provide a summary of the discussion with the BIMO Staff, and indicate BIMO's recommendation/action.

Failure to include these items alone generally should not result in an RTA designation.

	Yes	No
1) Submission contains a Table of Contents	X	
2) Each section is labeled (e.g., headings or tabs designating Device Description section, Labeling section, etc.)	X	
3) All pages of the submission are numbered.	X	
4) Type of 510(k) is identified (i.e., traditional, abbreviated, or special)	X	
Comments?		

Elements of a Complete Submission (RTA Items)
(21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision. - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.	Yes	No	N/A	Comment
---	------------	-----------	------------	----------------

A. Administrative

1) All content used to support the submission is written in English (including translations of test reports, literature articles, etc.)	X			
2) Submission identifies the following (such as in CDRH Premarket Review Submission Cover Sheet (Form 3514) or 510(k) cover letter):	X			
a) Device trade name or proprietary name	X			
b) Device common name	X			
c) Device class and panel or Classification regulation or Statement that device has not been classified with rationale for that conclusion	X			
3) Submission contains Indications for Use Statement with Rx and/or OTC designation (see also 21 CFR 801.109).	X			
4) Submission contains 510(k) Summary or 510(k) Statement	X			
a) Summary contains all elements per 21 CFR 807.92 (See also 510(k) Summary Checklist)	X			
b) Statement contains all elements per 21 CFR 807.93			X	
5) Submission contains Truthful and Accuracy Statement per 21 CFR 807.87(k) See recommended format .	X			
6) Submission contains Class III Summary and Certification. See recommended content .			X	
7) Submission contains clinical data			X	
8) If submission references use of a national or international standard as part of demonstration of substantial equivalence, submission contains Standards Data Report for 510(k)s (Form 3654) or includes detailed information about how and the extent to which the standard has been followed.	X			X

Comments?

Section I of form 3514 (Utilization of Standards) was not completed and no standards were identified. However, several standards were referenced in the submission and form 3654 was provided for several standards.

This is provided as a comment and not a deficiency that warrants a refusal to accept.

9) The submission identifies prior submissions for the same device for which FDA provided feedback related to the data or information needed to support substantial equivalence (e.g., submission numbers for Pre-Submission, IDE, prior not substantially equivalent (NSE) determination, prior 510(k) that was deleted or withdrawn) or states that there were no prior submissions for the subject device.		X		X
a) If there were prior submissions, the submitter has identified where in the current submission any issues related to a determination of substantial equivalence outlined in prior communications are addressed. For additional information regarding the Pre-Submission process, please refer to the Draft Guidance " Medical Devices: The Pre-Submission Program and Meetings with FDA Staff ." Once finalized, this guidance will represent the Agency's current thinking on this topic.		X		

Elements of a Complete Submission (RTA Items) (21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision. - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.	Yes	No	N/A	Comment
---	------------	-----------	------------	----------------

Comments? The submission does not state whether or not there were prior submissions. Section F of form 3514 was left blank which is not adequate for declaring that there were no prior submissions. The firm should provide a statement declaring whether or not there were prior submissions related to the subject device. However, the reviewer is familiar with these devices from this firm and given the nature of the submission being submitted as a special 510(k), it is not believed this device had a prior submission.

This deficiency is the only deficiency found during the RTA review. Since this is the only deficiency I recommend accepting the submission (RTAA).

B. Device Description

10)				
a) If there are requirements regarding the device description, such as special controls, in a device-specific regulation that are applicable to the device, the submission includes device description information to establish that the submitter has followed the device-specific requirement.			X	
b) If there is a device-specific guidance, other than a special controls guidance document, applicable to the device, the submission includes device description information to establish that the submitter has addressed the recommendations or otherwise has met the applicable statutory or regulatory criteria through an alternative approach.	X			
11) Descriptive information is present and consistent within the submission (e.g., the device description section is consistent with the device description in the labeling), including:				
a) A description of the principle of operation and mechanism of action for achieving the intended effect.	X			
b) A description of proposed conditions of use, such as surgical technique for implants; anatomical location of use; user interface; how the device interacts with other devices; and/or how the device interacts with the patient.				
c) A list and description of each device for which clearance is requested.	X			
12) Submission contains representative engineering drawing(s), schematics, illustrations and/or figures of the device that are clear, legible, labeled, and include dimensions.	X			
13) If device is intended to be marketed with multiple components, accessories, and/or as part of a system				
a) Submission includes a list of all components and accessories to be marketed with the subject device.	X			
b) Submission includes a description (as detailed in item 11(a) and (b) and 12 above) of each component or accessory.	X			
c) A 510(k) number is provided for each component or accessory that received a prior 510(k) clearance.			X	

C. Substantial Equivalence Discussion

14) Submitter has identified a predicate device.	X			
--	---	--	--	--

Elements of a Complete Submission (RTA Items)

(21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision. - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.	Yes	No	N/A	Comment
a) Predicate's 510(k) number, trade name, and model number (if applicable) provided. For predicates that are preamendments devices, information is provided to document preamendments status. <i>Information regarding documenting preamendment status is available online.</i>	X			
b) The identified predicate(s) is consistent throughout the submission (i.e., the predicate(s) identified in the Substantial Equivalence section is the same as that listed in the 510(k) Summary (if applicable) and that used in comparative performance testing.	X			
15) Submission includes a comparison of the following for the predicate(s) and subject device				
a) Indications for Use	X			
b) Technology, including features, materials, and principles of operation	X			
16) Submission includes an analysis of why any differences between the subject device and predicate(s) do not render the device NSE (e.g., does not constitute a new intended use; and any differences in technological characteristics are accompanied by information that demonstrates the device is as safe and effective as the predicate and do not raise different questions of safety and effectiveness than the predicate), affect safety or effectiveness, or raise different questions of safety and effectiveness (see section 513(i)(1)(A) of the FD&C Act and 21 CFR 807.87(f))	X			
D. Proposed Labeling (see also 21 CFR part 801)				
<i>If in vitro diagnostic (IVD) device, criteria 17, 18, & 19 may be omitted.</i>				
17) Submission includes proposed package labels and labeling (e.g., instructions for use, package insert, operator's manual) that include a description of the device, its intended use, and the directions for use.	X			
a) Indications for use are stated in labeling and are identical to Indications for Use form and 510(k) Summary (if 510(k) Summary provided).	X			
b) Submission includes directions for use that - include statements of all conditions, purposes or uses for which the device is intended (e.g., hazards, warnings, precautions, contraindications) AND - includes directions for layperson (see 21 CFR 801.5) OR submission states that device qualifies for exemption per 21 CFR 801 Subpart D	X			
18) If indicated for prescription use, labeling includes the prescription use statement (see 21 CFR 801.109(b)(1)) or "Rx only" symbol [See also Alternative to Certain Prescription Device Labeling Requirements]	X			
19) General labeling provisions				
a) Labeling includes name and place of business of the manufacturer, packer, or distributor (21 CFR 801.1).	X			
b) Labeling includes device common or usual name. (21 CFR 801.61)	X			
20)				
a) If there are requirements regarding labeling, such as special controls, in a device-specific regulation that are applicable to the device, the submission includes labeling to establish that the submission has followed the device-specific requirements.			X	

Elements of a Complete Submission (RTA Items) (21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision. - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.	Yes	No	N/A	Comment
b) If there is a device-specific guidance, other than a special controls guidance document, applicable to the device, the submission includes labeling to establish that the submitter has addressed the recommendations or otherwise has met the applicable statutory or regulatory criteria through an alternative approach.			X	
c) If there is a special controls document applicable to the device, the submission includes labeling to establish that the submitter has complied with the particular mitigation measures set forth in the special controls document or uses alternative mitigation measures but provides a rationale to demonstrate that those alternative measures identified by the firm will provide at least an equivalent assurance of safety and effectiveness.			X	
21) If the device is an in vitro diagnostic device, provided labeling includes all applicable information required per 21 CFR 809.10 .			X	

E. Sterilization

If IVD device and sterilization is not applicable, select "N/A" and criteria below will be omitted from checklist.

Submission states that the device and/or accessories are: (one of the below must be checked)

provided sterile

provided non-sterile but sterilized by the end user

non-sterile when used

Information regarding the sterility status of the device is not provided.

This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.

22) Assessment of the need for sterilization information

a) Identification of device, and/or accessories, and/or components that are provided sterile.

X

b) Identification of device, and/or accessories, and/or components that are end user sterilized.

X

c) Identification of device, and/or accessories, and/or components that are reusable and cleaning /disinfection instructions are provided.

X

23) If the device, and/or accessory, and/or a component is provided sterile:

a) Sterilization method is stated for each component (including parameters such as dry time for steam sterilization, radiation dose, etc.).

X

b) A description of method to validate the sterilization parameters (e.g., half-cycle method and full citation of FDA-recognized standard, including date) is provided for each proposed sterilization method. *Note, the sterilization validation report is not required.*

X

c) For devices sterilized using chemical sterilants such as ethylene oxide (EO) and hydrogen peroxide, submission states maximum levels of sterilant residuals remaining on the device and sterilant residual limits.

X

Elements of a Complete Submission (RTA Items)
(21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision. - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.	Yes	No	N/A	Comment
d) Submission includes description of packaging and packaging contents (e.g., if multiple devices are included within the same package, Tyvek packaging, etc.)	X			
e) Sterility Assurance Level (SAL) is stated.	X			
24) If the device, and/or accessory, and/or a component is end user sterilized:			X	
25)				
a) If there are requirements regarding sterility, such as special controls, in a device-specific regulation that are applicable to the device, the submission includes sterility information to establish that the submitter has followed the device-specific requirement.			X	
b) If there is a device-specific guidance, other than a special controls guidance document, applicable to the device, the submission includes sterility information to establish that the submitter has addressed the recommendations or otherwise has met the applicable statutory or regulatory criteria through an alternative approach.			X	
c) If there is a special controls document applicable to the device, the submission includes sterility information to establish that the submitter has complied with the particular mitigation measures set forth in the special controls document or uses alternative mitigation measures but provides a rationale to demonstrate that those alternative measures identified by the firm will provide at least an equivalent assurance of safety and effectiveness.			X	

F. Shelf Life

26) Proposed shelf life/expiration date stated	X			
27) For sterile device, submission includes summary of methods used to establish that device will remain sterile through the proposed shelf life or a rationale for why testing to establish shelf life is not applicable.	X			
28) Submission includes summary of methods used to establish that device performance is not adversely affected by aging or includes a rationale for why the storage conditions are not expected to affect device safety or effectiveness.	X			

G. Biocompatibility

If IVD device, select "N/A" and the below criteria will be omitted from checklist.				
Submission states that there: (one of the below must be checked)				
X are direct or indirect (e.g., through fluid infusion) patient-contacting components.				
are no direct or indirect (e.g., through fluid infusion) patient-contacting components.				
Information regarding the patient contact status of the device is not provided.				
This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.				
29) Submission includes list of patient-contacting device components and associated materials of construction, including identification of color additives, if present	X			

Elements of a Complete Submission (RTA Items) (21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision. - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.	Yes	No	N/A	Comment
30) Submission identifies contact classification (e.g., surface-contacting, less than 24 hour duration, etc.)	X			
31) Biocompatibility assessment of patient-contacting components Submission includes: Test protocol (including identification and description of test article), methods, pass/fail criteria, and results provided for each completed test, OR a statement that biocompatibility testing is not needed with a rationale (e.g., materials and manufacturing/processing are identical to the predicate).	X			
H. Software				
Submission states that the device: (one of the below must be checked)				
does contain software/firmware.				
X	does not contain software/firmware.			
Information regarding whether the device contains software is not provided.				
This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.				
I. EMC and Electrical Safety				
Submission states that the device: (one of the below must be checked)				
does require EMC and Electrical Safety evaluation.				
X	does not require EMC and Electrical Safety evaluation.			
Information regarding whether the device requires EMC and Electrical Safety evaluation is not provided.				
This information will determine whether and what type of additional information may be necessary for a substantial equivalence determination.				
J. Performance Data - General				
If IVD device, select "N/A" and the below criteria will be omitted from checklist. Performance data criteria relating to IVD devices will be addressed in Section K.				
36) Full test report is provided for each completed test. A full test report includes: objective of the test, description of the test methods and procedures, study endpoint(s), pre-defined pass/fail criteria, results summary, conclusions, and an explanation of how the data generated from the test supports a finding of substantial equivalence.	X			
37)				
a) If there are requirements regarding performance data, such as special controls, in a device-specific regulation that are applicable to the device, the submission includes performance data to establish that the submitter has followed the device-specific requirement.	X			

Elements of a Complete Submission (RTA Items) (21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision. - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.	Yes	No	N/A	Comment
b) If there is a device-specific guidance, other than a special controls guidance document, applicable to the device, the submission includes performance data to establish that the submitter has addressed the recommendations or otherwise has met the applicable statutory or regulatory criteria through an alternative approach.			X	
c) If there is a special controls document applicable to the device, the submission includes performance data to establish that the submitter has complied with the particular mitigation measures set forth in the special controls document or uses alternative mitigation measures but provides a rationale to demonstrate that those alternative measures identified by the firm will provide at least an equivalent assurance of safety and effectiveness.			X	
38) If literature is referenced in the submission, submission includes:			X	
39) For each completed nonclinical (i.e., animal) study conducted			X	

K. Performance Characteristics - In Vitro Diagnostic Devices Only

(Also see [21 CFR 809.10\(b\)\(12\)](#))

Submission states that the device: (one of the below must be checked)

is an in vitro diagnostic device.

X is not an in vitro diagnostic device.

Decision: Accept Refuse to Accept
Records processed under FOIA Request #2015-6024; Released by CDRH on 09-26-2016.

If Accept, notify applicant.

If Refuse to Accept, notify applicant in writing and include a copy of this checklist.

Digital Signature Concurrence Table

Reviewer Sign-Off	Samuel K. Shimp III 2013.06.04 12:30:43 -04'00'
Branch Chief Sign-Off (digital signature optional)*	Quynh T. Hoang -S 2013.06.04 16:47:55 -04'00'
Division Sign-Off (digital signature optional)*	Victor Krauthamer -S 2013.06.04 17:31:22 -04'00'

* Branch and Division review of checklist and concurrence with decision required.
Branch and Division digital signature optional.



Contains Nonbinding Recommendations

Print Form

Acceptance Checklist for Special 510(k)s

(Should be completed within 15 days of DCC receipt)

The following information is not intended to serve as a comprehensive review.

510(k) #: K131482

Date Received by DCC: May 23, 2013

Lead Reviewer: Samuel Shimp, Ph.D.

Branch: NNDB

Division: DNPMD

Center/Office: CDRH/ODE

Note: If an element is left blank on the checklist, it does not mean the checklist is incomplete. It means the reviewer did not assess the element during RTA and the element will be assessed during the substantive review.

Special 510(k) Criteria

The submission should not be reviewed as a Special 510(k) if "No" is selected for any of the 4 criteria below. Complete the Refuse to Accept Checklist for a Traditional 510(k) if submission is converted.

	Yes	No
1) 510(k) is submitted to modify a legally marketed device (predicate) AND the Special 510(k) submission is submitted by the holder of the 510(k) for the predicate device.	X	
Comments?		
2. Indications for Use of the proposed device are unchanged from the legally marketed device (predicate).	X	
Comments?		
3) Fundamental scientific technology of the proposed device is unchanged from the legally marketed device (predicate).	X	
Comments?		
4) The submission includes only summary-level information (i.e., NO test reports with performance data). Note that if performance data are provided and are conducted under design validation (21 CFR 820.30(g)), for example, to demonstrate continued conformance with a special control or recognized standard, then a Special 510(k) may be appropriate.	X	
Comments?		

Does the submission meet all 4 criteria above?

- Yes, submission meets criteria for a Special 510(k). Continue with the remainder of this checklist below.
- No, submission does not meet criteria for a Special 510(k). Discontinue this RTA checklist; convert to a Traditional using the form below and apply the Traditional checklist.

Organizational Elements

Failure to include these items alone generally should not result in an RTA designation.

	Yes	No
1) Submission contains a Table of Contents	X	
2) Each section is labeled (e.g., headings or tabs designating Device Description section, Labeling section, etc.)	X	
3) All pages of the submission are numbered.	X	
4) Type of 510(k) is identified (i.e., traditional, abbreviated, or special)	X	
Comments?		

Elements of a Complete Submission (RTA Items) (21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision.
 - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.

Yes

No

N/A

Comment

A. Administrative

1) All content used to support the submission is written in English (including translations of test reports, literature articles, etc.)	X			
2) Submission identifies the following (such as in CDRH Premarket Review Submission Cover Sheet (Form 3514) or 510(k) cover letter):	X			
a) Device trade name or proprietary name	X			
b) Device common name	X			
c) Device class and panel or Classification regulation or Statement that device has not been classified with rationale for that conclusion	X			
3) Submission contains Indications for Use Statement with Rx and/or OTC designation (see also 21 CFR 801.109).	X			
4) Submission contains 510(k) Summary or 510(k) Statement	X			
a) Summary contains all elements per 21 CFR 807.92 (See also 510(k) Summary Checklist)	X			
b) Statement contains all elements per 21 CFR 807.93			X	
5) Submission contains Truthful and Accuracy Statement per 21 CFR 807.87(k) See recommended format .	X			
6) Submission contains Class III Summary and Certification. See recommended content .			X	
7) If submission references use of a national or international standard as part of demonstration of substantial equivalence, submission contains Standards Data Report for 510(k)s (Form 3654) or includes detailed information about how and the extent to which the standard has been followed.	X			X

Comments? Section I of form 3514 (Utilization of Standards) was not completed and no standards were identified. However, several standards were referenced in the submission and form 3654 was provided for several standards.
 This is provided as a comment and not a deficiency that warrants a refusal to accept.

8) The submission identifies prior submissions for the same device for which FDA provided feedback related to the data or information needed to support substantial equivalence (e.g., submission numbers for Pre-Submission, IDE, prior not substantially equivalent (NSE) determination, prior 510(k) that was deleted or withdrawn) or states that there were no prior submissions for the subject device.		X		X
a) If there were prior submissions, the submitter has identified where in the current submission any issues related to a determination of substantial equivalence outlined in prior communications are addressed. For additional information regarding the Pre-Submission process, please refer to the Draft Guidance " Medical Devices: The Pre-Submission Program and Meetings with FDA Staff ." Once finalized, this guidance will represent the Agency's current thinking on this topic.		X		

Comments? The submission does not state whether or not there were prior submissions. Section F of form 3514 was left blank which is not adequate for declaring that there were no prior submissions. You need to provide a statement declaring whether or not there were prior submissions related to the subject device.
 Questions? Contact FDA/CDRH/OCE/DID at CDRH-FOISTATUS@fda.hhs.gov or call 301-796-8118.

Elements of a Complete Submission (RTA Items) (21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision. - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.	Yes	No	N/A	Comment
---	------------	-----------	------------	----------------

B. Device Description

9)				
a) If there are requirements regarding the device description, such as special controls, in a device-specific regulation that are applicable to the device, the submission includes device description information to establish that the submitter has followed the device-specific requirement.			X	
b) If there is a device-specific guidance, other than a special controls guidance document, applicable to the device, the submission includes device description information to establish that the submitter has addressed the recommendations or otherwise has met the applicable statutory or regulatory criteria through an alternative approach.	X			
10) Descriptive information is present and consistent within the submission (e.g., the device description section is consistent with the device description in the labeling), including:				
a) A description of the principle of operation and mechanism of action for achieving the intended effect.	X			
b) A description of proposed conditions of use, such as surgical technique for implants; anatomical location of use; user interface; how the device interacts with other devices; and/or how the device interacts with the patient.	X			
c) A list and description of each device for which clearance is requested.	X			
11) A description of all device modification(s) including rationale for each modification.	X			
12) Submission contains representative engineering drawing(s), schematics, illustrations and/or figures of the device that are clear, legible, labeled, and include dimensions.	X			
13) If device is intended to be marketed with multiple components, accessories, and/or as part of a system				
a) Submission includes a list of all components and accessories to be marketed with the subject device.	X			
b) Submission includes a description (as detailed in item 10(a) and (b) and 12 above) of each component or accessory.	X			
c) A 510(k) number is provided for each component or accessory that received a prior 510(k) clearance.			X	

C. Substantial Equivalence Discussion

14) Submitter has identified a predicate device.	X			
a) Predicate's 510(k) number, trade name, and model number (if applicable) provided. For predicates that are preamendments devices, information is provided to document preamendments status. <i>Information regarding documenting preamendment status is available online.</i>	X			
b) The identified predicate(s) is consistent throughout the submission (i.e., the predicate(s) identified in the Substantial Equivalence section is the same as that listed in the 510(k) Summary (if applicable) and that used in comparative performance testing.	X			
15) Submission includes a comparison of the following for the predicate(s) and subject device: a) Intended use b) Indications for use c) Contraindications d) Warnings e) Precautions f) Side effects g) Adverse reactions h) Other information relevant to the safety and effectiveness of the device				

Elements of a Complete Submission (RTA Items)
(21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision.
 - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.

	Yes	No	N/A	Comment
a) Indications for Use	X			
b) Technology, including features, materials, and principles of operation	X			
16) Submission includes an analysis of why any differences between the subject device and predicate(s) do not render the device NSE (e.g., does not constitute a new intended use; and any differences in technological characteristics are accompanied by information that demonstrates the device is as safe and effective as the predicate and do not raise different questions of safety and effectiveness than the predicate), affect safety or effectiveness, or raise different questions of safety and effectiveness (see section 513(i)(1)(A) of the FD&C Act and 21 CFR 807.87(f))	X			

D. Design Control Activities

17) Design Control Activities Summary includes all of the following:				X
a) Identification of Risk Analysis methods(s) used to assess the impact of the modification on the device and its components AND the results of the analysis.	X			
b) Based on the Risk Analysis, an identification of the verification and/or validation activities required, including methods or tests used and acceptance criteria.	X			
c) Declaration of conformity with design controls, including: <i>All 3 must be present to answer "Yes."</i> i. Statement that all verification and validation activities were performed by designated individuals and results demonstrate that predetermined acceptance criteria were met. ii. Statement that manufacturing facility is in conformance with design control procedure requirements as specified in 21 CFR 820.30 . iii. Statement is signed by the individual responsible for these activities.	X			

Comments? You provide a Design Control Activities Summary in which you identify the verification and validation activities required based on the Risk Analysis. We recommend that you provide with the summary a justification for why the design controls in the predicate device are sufficient to verify and validate the changes implemented in the subject 510(k) device. For example, you specify that the hydrophilic coating in the new device is similar to that used in the Headway DUO (K120917) but is different from the cited predicate K082385. Please be aware that if the design control activities for the predicate are not sufficient to verify and validate the changes implemented in the subject device, a traditional 510(k) submission will be required.

We also recommend that for each design control activity you perform for the subject device you identify the corresponding design control activity performed for the predicate device and whether or not the performance criteria for that has been changed. Note that changes in performance criteria may also warrant a conversion to traditional 510(k).

These comments are provided for your convenience but are not deficiencies that warrant a refusal to accept.

E. Proposed Labeling (see also [21 CFR part 801](#))

18) Submission includes proposed package labels and labeling (e.g., instructions for use, package insert, operator's manual) that include a description of the device, its intended use, and the directions for use. Questions? Contact FDA/CDRH/OCE/DID at CDRH-FOISTATUS@fda.hhs.gov or call 301-796-8118.	X			X
--	---	--	--	---

Elements of a Complete Submission (RTA Items) (21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision. - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.	Yes	No	N/A	Comment
a) All changes in proposed labeling resulting from device modification(s) are highlighted or prominently identified.		X		
Comments? Only a clean copy of the labeling is provided. The labeling provided does not identify any changes from the previous version of the labeling. There is no narrative description or highlighted/redlined version of the labeling to prominently identify the proposed labeling changes.				
19) Statement that the intended use of the modified device, as described in the labeling, has not changed as a result of the modification(s).	X			

Decision: Accept Refuse to Accept
Records processed under FOIA Request #2015-6024; Released by CDRH on 09-26-2016.

If Accept, notify applicant.

If Refuse to Accept, notify applicant in writing and include a copy of this checklist.

Digital Signature Concurrence Table

Reviewer Sign-Off	Samuel K. Shimp III 2013.06.07 14:43:56 -04'00'
Branch Chief Sign-Off (digital signature optional)*	Quynh T. Hoang -S 2013.06.07 15:15:06 -04'00'
Division Sign-Off (digital signature optional)*	

* Branch and Division review of checklist and concurrence with with decision required.
Branch and Division digital signature optional.



Contains Nonbinding Recommendations

Print Form

Acceptance Checklist for Special 510(k)s

(Should be completed within 15 days of DCC receipt)

The following information is not intended to serve as a comprehensive review.

510(k) #: K131482

Date Received by DCC: Jun 12, 2013

Lead Reviewer: Samuel Shimp, Ph.D.

Branch: NNDB

Division: DNPMD

Center/Office: CDRH/ODE

Note: If an element is left blank on the checklist, it does not mean the checklist is incomplete. It means the reviewer did not assess the element during RTA and the element will be assessed during the substantive review.

Special 510(k) Criteria

The submission should not be reviewed as a Special 510(k) if "No" is selected for any of the 4 criteria below. Complete the Refuse to Accept Checklist for a Traditional 510(k) if submission is converted.

	Yes	No
1) 510(k) is submitted to modify a legally marketed device (predicate) AND the Special 510(k) submission is submitted by the holder of the 510(k) for the predicate device.	X	
Comments?		
2. Indications for Use of the proposed device are unchanged from the legally marketed device (predicate).	X	
Comments?		
3) Fundamental scientific technology of the proposed device is unchanged from the legally marketed device (predicate).	X	
Comments?		
4) The submission includes only summary-level information (i.e., NO test reports with performance data). Note that if performance data are provided and are conducted under design validation (21 CFR 820.30(g)), for example, to demonstrate continued conformance with a special control or recognized standard, then a Special 510(k) may be appropriate.	X	
Comments?		

Does the submission meet all 4 criteria above?

- Yes, submission meets criteria for a Special 510(k). Continue with the remainder of this checklist below.
- No, submission does not meet criteria for a Special 510(k). Discontinue this RTA checklist; convert to a Traditional using the form below and apply the Traditional checklist.

Organizational Elements

Records processed under FOIA Request #2016-0002, Released by CDRH on 09-26-2016.

Failure to include these items alone generally should not result in an RTA designation.

	Yes	No
1) Submission contains a Table of Contents	X	
2) Each section is labeled (e.g., headings or tabs designating Device Description section, Labeling section, etc.)	X	
3) All pages of the submission are numbered.	X	
4) Type of 510(k) is identified (i.e., traditional, abbreviated, or special)	X	
Comments?		

Elements of a Complete Submission (RTA Items) (21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision.
 - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.

Yes

No

N/A

Comment

A. Administrative

1) All content used to support the submission is written in English (including translations of test reports, literature articles, etc.)	X			
2) Submission identifies the following (such as in CDRH Premarket Review Submission Cover Sheet (Form 3514) or 510(k) cover letter):	X			
a) Device trade name or proprietary name	X			
b) Device common name	X			
c) Device class and panel or Classification regulation or Statement that device has not been classified with rationale for that conclusion	X			
3) Submission contains Indications for Use Statement with Rx and/or OTC designation (see also 21 CFR 801.109).	X			
4) Submission contains 510(k) Summary or 510(k) Statement	X			
a) Summary contains all elements per 21 CFR 807.92 (See also 510(k) Summary Checklist)	X			
b) Statement contains all elements per 21 CFR 807.93			X	
5) Submission contains Truthful and Accuracy Statement per 21 CFR 807.87(k) See recommended format .	X			
6) Submission contains Class III Summary and Certification. See recommended content .			X	
7) If submission references use of a national or international standard as part of demonstration of substantial equivalence, submission contains Standards Data Report for 510(k)s (Form 3654) or includes detailed information about how and the extent to which the standard has been followed.	X			
8) The submission identifies prior submissions for the same device for which FDA provided feedback related to the data or information needed to support substantial equivalence (e.g., submission numbers for Pre-Submission, IDE, prior not substantially equivalent (NSE) determination, prior 510(k) that was deleted or withdrawn) or states that there were no prior submissions for the subject device.	X			
a) If there were prior submissions, the submitter has identified where in the current submission any issues related to a determination of substantial equivalence outlined in prior communications are addressed. For additional information regarding the Pre-Submission process, please refer to the Draft Guidance " Medical Devices: The Pre-Submission Program and Meetings with FDA Staff ." Once finalized, this guidance will represent the Agency's current thinking on this topic.	X			

B. Device Description

9)				
a) If there are requirements regarding the device description, such as special controls, in a device-specific regulation that are applicable to the device, the submission includes device description information to establish that the submitter has followed the device-specific requirement.			X	

Questions? Contact FDA/CDRH/OCE/DID at CDRH-FOISTATUS@fda.hhs.gov or call 301-796-8118.

Elements of a Complete Submission (RTA Items) (21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision. - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.	Yes	No	N/A	Comment
b) If there is a device-specific guidance, other than a special controls guidance document, applicable to the device, the submission includes device description information to establish that the submitter has addressed the recommendations or otherwise has met the applicable statutory or regulatory criteria through an alternative approach.	X			
10) Descriptive information is present and consistent within the submission (e.g., the device description section is consistent with the device description in the labeling), including:				
a) A description of the principle of operation and mechanism of action for achieving the intended effect.	X			
b) A description of proposed conditions of use, such as surgical technique for implants; anatomical location of use; user interface; how the device interacts with other devices; and/or how the device interacts with the patient.	X			
c) A list and description of each device for which clearance is requested.	X			
11) A description of all device modification(s) including rationale for each modification.	X			
12) Submission contains representative engineering drawing(s), schematics, illustrations and/or figures of the device that are clear, legible, labeled, and include dimensions.	X			
13) If device is intended to be marketed with multiple components, accessories, and/or as part of a system				
a) Submission includes a list of all components and accessories to be marketed with the subject device.	X			
b) Submission includes a description (as detailed in item 10(a) and (b) and 12 above) of each component or accessory.	X			
c) A 510(k) number is provided for each component or accessory that received a prior 510(k) clearance.			X	
C. Substantial Equivalence Discussion				
14) Submitter has identified a predicate device.	X			
a) Predicate's 510(k) number, trade name, and model number (if applicable) provided. For predicates that are preamendments devices, information is provided to document preamendments status. <i>Information regarding documenting preamendment status is available online.</i>	X			
b) The identified predicate(s) is consistent throughout the submission (i.e., the predicate(s) identified in the Substantial Equivalence section is the same as that listed in the 510(k) Summary (if applicable) and that used in comparative performance testing.	X			
15) Submission includes a comparison of the following for the predicate(s) and subject device				
a) Indications for Use	X			
b) Technology, including features, materials, and principles of operation	X			

Elements of a Complete Submission (RTA Items) (21 CFR 807.87 unless otherwise indicated)

Submission should be designated RTA if not addressed.

Check "Yes" if item is present, "N/A" if it is not needed and "No" if it is not included but needed.

- Any "No" answer will result in a "Refuse to Accept" decision. - Each element on the checklist should be addressed within the submission. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (Yes). An assessment of the rationale will be considered during the review of the submission.	Yes	No	N/A	Comment
---	-----	----	-----	---------

16) Submission includes an analysis of why any differences between the subject device and predicate(s) do not render the device NSE (e.g., does not constitute a new intended use; and any differences in technological characteristics are accompanied by information that demonstrates the device is as safe and effective as the predicate and do not raise different questions of safety and effectiveness than the predicate), affect safety or effectiveness, or raise different questions of safety and effectiveness (see section 513(i)(1)(A) of the FD&C Act and 21 CFR 807.87(f))	X			
--	---	--	--	--

D. Design Control Activities

17) Design Control Activities Summary includes all of the following:				
a) Identification of Risk Analysis methods(s) used to assess the impact of the modification on the device and its components AND the results of the analysis.	X			
b) Based on the Risk Analysis, an identification of the verification and/or validation activities required, including methods or tests used and acceptance criteria.	X			
c) Declaration of conformity with design controls, including: <i>All 3 must be present to answer "Yes."</i> i. Statement that all verification and validation activities were performed by designated individuals and results demonstrate that predetermined acceptance criteria were met. ii. Statement that manufacturing facility is in conformance with design control procedure requirements as specified in 21 CFR 820.30 . iii. Statement is signed by the individual responsible for these activities.	X			

E. Proposed Labeling (see also [21 CFR part 801](#))

18) Submission includes proposed package labels and labeling (e.g., instructions for use, package insert, operator's manual) that include a description of the device, its intended use, and the directions for use.	X			
a) All changes in proposed labeling resulting from device modification(s) are highlighted or prominently identified.	X			
19) Statement that the intended use of the modified device, as described in the labeling, has not changed as a result of the modification(s).	X			

Decision: Accept Refuse to Accept
Records processed under FOIA Request #2015-6024; Released by CDRH on 09-26-2016.

If Accept, notify applicant.

If Refuse to Accept, notify applicant in writing and include a copy of this checklist.

Digital Signature Concurrence Table

Reviewer Sign-Off	Samuel K. Shimp III 2013.06.21 11:56:11 -04'00'
Branch Chief Sign-Off (digital signature optional)*	Quynh T. Hoang -S 2013.06.27 12:32:30 -04'00'
Division Sign-Off (digital signature optional)*	

* Branch and Division review of checklist and concurrence with with decision required.
Branch and Division digital signature optional.



From: Reviewer Name Samuel K. Shimp III
Subject: 510(k) Number K131482
To: The Record

Please list CTS decision code: SE - Substantially Equivalent

- Refused to Accept (Note: this is considered the first review cycle. See [screening checklist](#).)
- Hold (Additional Information or Telephone Hold)
- Final Decision (SE, SE with Limitations, NSE (select code below), Withdrawn, etc.)

Please complete the following for a final clearance decision (i.e, SE, SE with Limitations, etc.)	YES	NO
Indications for Use Page (<i>Attach IFU</i>)	X	
510(k) Summary or 510(k) Statement (<i>Attach Summary or Statement</i>)	X	
Truthful and Accurate Statement (<i>Must be present for a Final Decision</i>)	X	
Is the device Class III?		X
Does firm reference standards? (If yes, please attach Form 3654 .)	X	
Is this a combination product?		X
Is this a reprocessed single use device? (See Guidance for Industry and FDA Staff - MDUFMA - Validation Data in 510(k)s for Reprocessed Single-Use Medical Devices .)		X
Is this device intended for pediatric use only?		X
Is this a prescription device? (If both prescription & OTC, check both boxes.)	X	
Is clinical data necessary to support the review of this 510(k)?		X
For United States based clinical studies only, did the application include a completed Form FDA 3674, Certification with Requirements of ClinicalTrials.gov Data Bank? (If study was conducted in the United States and Form FDA 3674 was not included or was incomplete, then applicant must be contacted to obtain completed form.)		X
Does this device include an Animal Tissue Source?		X
All Pediatric Patients age <= 21		X
Neonate/Newborn (Birth to 28 days)		X
Infant (29 days to < 2 years)		X
Child (2 years to <12 years)		X
Adolescent (12 years to <18 years)		X
Transitional Adolescent A (18 years to <21 years); Special considerations are being given to this group, different from adults age >= 21 (different device design or testing, different protocol procedures, etc.)		X
Transitional Adolescent B (18 years to <21 years); No special considerations compared to adults >= 21 years)		X

Nanotechnology	Records processed under FOIA Request #2015-6024; Released by CDRH on 09-26-2016.		×
Is this device subject to the Tracking Regulation? (Medical Device Tracking Guidance)			×

Regulation Number: 870.1250
Class: II
Product Code: DQY
Additional Product Codes: DQO

Digital Signature Concurrence Table
 (Not all signatures may be required)

Branch Chief Sign-Off	Quynh T. Hoang -S 2013.11.21 18:25:48 -05'00'
Division Sign-Off	Joyce M. Whang -S 2013.11.22 12:17:40 -05'00'