
Helios Gene Gun System

Instruction Manual

**Catalog Numbers
165-2431 and 165-2432**



Warranty and Regulatory Notices

Warranty Statement

This warranty may vary outside of the continental United States. Contact your local Bio-Rad office for the exact terms of your warranty.

Bio-Rad Laboratories warrants to the customer that the Helios Gene Gun System (catalog number 165-2431 and 165-2432) will be free from defects in material and workmanship, and will meet all performance specifications for the period of one year from the date of shipment. This warranty covers all parts and labor.

In the event that the instrument must be returned to the factory for repair under warranty, the instrument must be packed for return in the original packaging.

Bio-Rad shall not be liable for any incidental, special or consequential loss, damage, or expense directly or indirectly arising from the use of the Helios Gene Gun System. Bio-Rad makes no warranty whatsoever in regard to products or parts furnished by third parties, such being subject to the warranty of their respective manufacturers. Service under this warranty shall be requested by contacting your nearest Bio-Rad office.

The following items are considered customer-installable consumables: tubing, desiccant pellets, and microcarriers. The battery, razor blade, O-rings, barrel liner, cartridge holders, and syringe are replacement parts (see Section 9.2). These consumables and replacement parts are not covered by this warranty and are warranted only to be free from defects in workmanship.

This warranty does not extend to any instruments or parts that have been subject to misuse, neglect, or accident, or that have been modified by anyone other than Bio-Rad, or that have been used in violation of Bio-Rad instructions.

The foregoing obligations are in lieu of all other obligations and liabilities including negligence and all warranties, of merchantability, fitness for a particular purpose or otherwise, expressed or implied in fact or by law, and state Bio-Rad's entire and exclusive liability and buyer's exclusive remedy for any claims or damages in connection with the furnishing of goods or parts, their design, suitability for use, installation or operation. Bio-Rad will in no event be liable for any special, incidental or consequential damages whatsoever, and Bio-Rad's liability under no circumstances will exceed the contract price for the goods for which liability is claimed. Bio-Rad is not responsible for any injury caused by the use of this instrument for purposes other than those for which it is intended.

Regulatory Notices

Important: This Bio-Rad instrument is designed and certified to meet EN55011, EN50082-1, and EN61010 requirements, which are internationally accepted electromagnetic compliance and electrical safety standards. Certified products are safe to use when operated in accordance with the instruction manual. This instrument should not be modified or altered in any way. Alteration of this instrument will result in the following:

Void the manufacturer's warranty.

Void the regulatory certifications.

Create a potential safety hazard.

Note: This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to Part 15 of the FCC rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at his own expense.

Patent License and Usage

Particle bombardment technology is covered by several patents which are held by E. I. duPont de Nemours & Co. and Auragen, Inc. Particle bombardment may be used for research purposes for gene delivery. Use of particle bombardment for commercial purposes requires a commercial license from the appropriate patent holder. The Helios Gene Gun is designed for research purposes only and is not intended for human or veterinary use. Licensed only for research use.

Table of Contents

	Page
Section 1 General Safety Information.....	1
1.1 Helios Gene Gun Safety.....	1
1.2 Pressurized Helium and Nitrogen Safety.....	1
1.3 Power Safety.....	1
1.4 Ear and Eye Protection.....	2
Section 2 Introduction to Particle Delivery	2
2.1 Particle Delivery Technology	2
2.2 The Helios Gene Gun.....	3
2.3 Operating Principle of the Helios Gene Gun System	3
2.4 Requirements for System Operation.....	4
Section 3 Product Description	7
3.1 Packing List.....	7
3.2 Identification of System Components and Controls.....	8
Section 4 Setting up the Helios Gene Gun System	12
4.1 Inserting the Battery into the Helios Gene Gun	12
4.2 Connecting the Helios Gene Gun to a Helium Source.....	13
4.3 Connecting the Tubing Prep Unit to a Nitrogen Source	14
Section 5 Operation of the Helios Gene Gun System.....	17
5.1 Quick Guide to Operation	17
5.2 Preparation of System Components Prior to Bombardment.....	18
5.3 Particle delivery using the Helios Gene Gun.....	23
5.4 Removing Used Cartridges, Depressurization, and Shut Down	28
Section 6 Preparation of Mammalian Cell Targets.....	29
6.1 <i>In vitro</i> Delivery to Adherent Cells	29
6.2 <i>In vitro</i> Delivery to Suspension Cultures.....	30
6.3 <i>in vivo</i> Delivery to Epidermis	31
Section 7 Optimization of Gene Gun Parameters	32
7.1 Overview	32
7.2 Parameters for <i>in vitro</i> Delivery.....	34
7.3 Parameters for <i>in vivo</i> Delivery	35
Section 8 Troubleshooting	35
8.1 DNA/Microcarrier Preparation	35
8.2 Cartridge Preparation	35
8.3 Helios Gene Gun Operation.....	36
8.4 <i>In Vitro</i> and <i>in Vivo</i> Targeting	36
Section 9 Product Information	37
9.1 Helios Gene Gun System	37
9.2 Spare Parts.....	38
9.3 Specifications	38
Section 10 Appendices	39
10.1 Precipitation of RNA onto Microcarriers	39
10.2 Replacing the O-rings and Inner Sleeve on the Helios Gene Gun.....	40
10.3 Replacing the O-ring on the Tubing Prep Unit.....	41
10.4 Replacing Tubing Cutter Razor Blade and Unit Disassembly.....	42
10.5 Cleaning and Sterilizing the Helios Gene Gun.....	43
10.6 Testing Cartridges for Microcarrier Penetration and Density	43
10.7 Quantitation of DNA in Cartridges.....	45
10.8 References	46
10.9 Quick Guide	47

Section 1

General Safety Information



Caution: In particle bombardment DNA-coated microparticles are accelerated to velocities in excess of 1,000 ft/sec in order to penetrate the cell membrane and through multiple layers of cells in tissues and organs. In the Helios Gene Gun, this accelerating force is supplied by a high pressure helium pulse. Numerous safety features have been designed into this instrument to protect both the user and bystanders. The parts used in manufacturing the Helios Gene Gun have been chosen because they are designed to work at the pressures required for operation and have a wide safety margin. General safety principles are indicated below. Specific safety recommendations are indicated in the appropriate sections throughout the manual.

1.1 Helios Gene Gun Safety



Caution: While the Helios Gene Gun has a trigger button which is time-activated by a safety interlock switch, accidental or unintentional discharge is still possible. Do not point the gun at people. The Helios Gene Gun is for research use only.

Refer to Section 4.2 for connecting the Helios Gene Gun to a helium source, to Section 5.3 for use of the Gene Gun, and to Section 5.4 for depressurization and shut down of the Gene Gun.

1.2 Pressurized Helium and Nitrogen Safety



Caution: Although helium and nitrogen are neither toxic nor flammable, all gases under pressure are potentially dangerous if used improperly. Always be sure pressurized tanks are properly secured. This may be accomplished by placing the tank in a floor stand or by using a wall-mounted or bench-mounted strap. Please follow the instructions provided with the helium cylinder from the supplier and those that are applicable for your institution (see your Site Safety Officer). Bio-Rad has supplied tubing, fittings, a control valve, and a pressure regulator capable of safely handling the high pressure helium gas used in the Helios Gene Gun. These components have been carefully selected and are the only parts to be used with the Helios Gene Gun System.

Refer to Section 2.4 for a description of the helium and nitrogen gases required for the Helios Gene Gun System.

1.3 Power Safety



Figure 1 shows the serial number certification label which is found underneath the molded case of the Helios Gene Gun. This label provides the manufacturing data about the instrument. This instrument is operational using a standard 9 volt battery. Change the battery only after detaching the Gene Gun from the helium hose.

Refer to Section 2.4 for a description of the battery required for the Gene Gun and to Section 4.1 for information on replacing the battery in the Gene Gun.




Fig. 1. Location of the instrument serial number label on the Helios Gene Gun.

1.4 Ear and Eye Protection



Caution: Expansion of gas from high pressure to low pressure produces a sound wave, the intensity of which is a function of the gas pressure. The intensity of the sound generated by discharging the Helios Gene Gun is ~108 decibels (db) at 400 psi; sustained noise levels of 85 db or brief noise levels of 110 db may lead to permanent hearing damage. Hearing protection should be worn by all those in the immediate vicinity when discharging the Helios Gene Gun. Earmuffs or ear plugs provide equivalent protection against hearing damage.

Refer to Section 2.4 for suggestions on ear protection. Eye protection should always be worn when working with high pressure gases.

Section 2 Introduction to Particle Delivery

2.1 Particle Delivery Technology

Particle bombardment is a physical method of cell transformation in which high density, sub-cellular sized particles are accelerated to high velocity to carry DNA into cells. The technique was first described as a method of gene transfer into plants (Klein *et al.*, 1987, 1988; McCabe *et al.*, 1988) and subsequently shown to be applicable to mammalian experimental systems (Zelenin *et al.*, 1989; Yang *et al.*, 1990; Williams *et al.*, 1991). Because it does not depend on specific ligand-receptors and/or the biochemical features of structural components typically present on cell surfaces, particle-mediated gene transfer can be readily applied to a variety of biological systems. Consequently, this procedure can be used to transform such diverse targets as bacteria (Shark *et al.*, 1991; Smith *et al.*, 1992), fungi (Armaleo *et al.*, 1990), and intracellular organelles (Johnston *et al.*, 1988; Boynton *et al.*, 1988). Since it is a physical method of gene delivery, particle bombardment also overcomes physical barriers to effective gene transfer, such as the stratum corneum of the epidermis and the cell wall of plants. Particle bombardment is a convenient method for transforming intact cells in culture since minimal pre- or post-bombardment manipulation is necessary. In addition, this technique is much easier and faster to perform than the tedious task of microinjection. Both transient and stable expression are possible with particle bombardment. In addition to DNA, RNA may also be transferred to cells by particle bombardment (Qiu, *et al.*, 1996). Table 1 lists some of the advantages of using particle bombardment for *in vitro* and *in vivo* transformation.

Table 1. Advantages of particle bombardment for *in vitro* and *in vivo* gene transfer.

-
- Easy to use, rapid, versatile gene delivery system
 - Independent of target cell type
 - Useful for both transient and stable expression
 - Requires only small amounts of DNA
 - No carrier DNA is needed
 - Requires only small numbers of cells
 - May obtain high levels of co-transformation
 - Large DNA fragments may be transferred
 - Direct intracellular delivery to many cells in the target area
 - Applicable to both *in vitro* and *in vivo* transformation
 - No extraneous genes or proteins are delivered
-

2.2 The Helios Gene Gun

The Helios Gene Gun is the second instrument in Bio-Rad's particle delivery product line. In contrast to the PDS-1000/He instrument where the overall size of the target to be transformed is limited by the size of the chamber and the target tissue is subjected to a vacuum during bombardment, the Helios Gene Gun requires no vacuum and any target accessible to the barrel can be transformed. Consequently, the Helios Gene Gun may be used in a much wider variety of gene transfer applications and provides a tool for both *in vitro* and *in vivo* transformations in the research lab. Essentially any type of cells which can be made accessible to its nozzle may be transformed.

Gene gun models have also been developed by Auragen, Inc., a Bio-Rad collaborator. Cell penetration, gene expression and other measures of performance vary with the model of gene gun used. Users must be careful to select operating parameters optimized for their particular model. The Accell® model used by Auragen Inc. may include modifications to be included in the future Helios models. The current Helios Gene Gun has been designed to serve a wide range of research uses.

In vertebrates, the epidermal cells of the skin are the most obvious target (Yang *et al.*, 1990; Williams *et al.*, 1991). *In vivo* experimental systems have targeted the skin for vaccination studies (Tang *et al.*, 1992; Fynan *et al.*, 1993; Eisenbraun *et al.*, 1993), and wound healing (Andree, *et al.*, 1994), and cytokine gene therapy studies in mouse tumor models (Sun *et al.*, 1995; Keller *et al.*, 1996; Rakhmilevich *et al.*, 1996). In addition to skin, muscle and internal organs, including liver, pancreas, spleen, kidney, etc., when appropriately exposed surgically, can also be targeted *in vivo* (Yang *et al.*, 1990; William *et al.*, 1991; Cheng *et al.*, 1993). Using the Accell Gene Gun, both primary and established cultures of mammalian cells have been transfected *in vitro* and *ex vivo* (Albertini *et al.*, 1996; Mahvi *et al.*, 1996; Rakhmilevich *et al.*, 1996). Additionally, transgenic expression of β -galactosidase, luciferase, IL-12, granulocyte macrophage-colony stimulating factor and a nuclear papillomavirus protein has been demonstrated following *in vivo* transformation (Sundaram *et al.*, 1996; Keller *et al.*, 1996; Rakhmilevich *et al.*, 1996). Meristematic tissues and leaves are obvious target cells for *in vivo* transformation of plants.

2.3 Operating Principle of the Helios Gene Gun System

The Helios Gene Gun System consists of all of the components needed to prepare DNA-coated microcarriers, coat the DNA-microcarrier suspension onto the inner surface of the Gold-Coat™ tubing, cut the tubing into cartridges which are used in the Helios Gene Gun, and finally propel the microcarriers and their associated DNA into cells.

Prior to transfection, the plasmid DNA must be attached to the gold particles. This is accomplished by precipitation of the DNA from solution in the presence of gold microcarriers and the polycation spermidine by the addition of CaCl_2 . The particles are then washed extensively with ethanol to remove the water and resuspended in ethanol. Using the Tubing Prep Station, the DNA/microcarrier solution is coated onto the inner wall of Gold-Coat tubing and dried. The tubing is then cut into 0.5" length cartridges using the Tubing Cutter. These cartridges, when inserted into the cartridge holder of the Helios Gene Gun are the source of the DNA which enters the target cells by the helium discharge.

The Helios Gene Gun employs a high velocity stream of helium to accelerate gold particles coated with plasmids or RNA to velocities sufficient to penetrate and transform cells, both *in vitro* and *in vivo* (Figure 2). The discharge is initiated by pressing the trigger buttons which activates the main valve, causing helium to travel down the bore of the particle delivery device. When the helium enters one of the bores of the cylinder containing the cartridge, the gold particles on the inside of the tubing are pulled from the surface, become entrained in the helium stream, and begin to pick up speed. Immediately past the acceleration channel, the barrel begins to open as a cone. The slope of the cone causes the gas to be pulled outward, a process known as the Coanda effect (Reba, 1966), expanding the high pressure jet into a less destructive low velocity pulse, while the gold particles maintain a high velocity. The expansion also helps spread the microcarriers from their original 1/16" diameter to an area approximately 1/2" in diameter at the target site.




Fig. 2. How the samples are delivered. Helium gas is pulsed through the cartridge loaded with DNA-coated microcarriers. This pulse sweeps the microcarriers from the inside wall of the cartridge. As the microcarriers enter the barrel liner they pick up speed in the acceleration channel then spread out as they travel down the barrel; the increased cross-sectional area of the barrel from the acceleration chamber to the spacer also moderates the helium shock wave so it is less intense when it reaches the target cells. The O-rings on each side of the cartridge holder direct the flow of helium through the cartridge and the acceleration channel. The spacer maintains optimal target distance and permits venting of the helium gas away from the target.

2.4 Requirements for System Operation

Selecting a Site for Operation

The Helios Gene Gun is a portable particle bombardment device. The range of its use is limited by its requirement for a supply of pressurized helium and the 6 foot length of pressurized helium hose. When using the Gene Gun, only a small area is needed for setting down the gun during an experiment, for loading the cartridges into the cartridge holders and exchanging cartridge holders during experiments. In addition, a clean and dry area is needed for working with the tissue samples.

Preparation of the gold/DNA tubes used in the Gene Gun requires an area approximately 1 m² for the Tubing Prep Station, for manipulating the tubing, precipitating the DNA onto the gold, and processing the tubing into cartridges. Additionally, the Tubing Prep Station requires an electrical outlet and a tank of pressurized nitrogen for evaporating the ethanol from the DNA-coated gold particles from the inner surface of the tubing.

User Supplied Components

Helium Supply

Only helium gas is to be used with the Helios Gene Gun. The low atomic weight of helium results in maximum gas expansion when the high pressure helium is released through the valve opening and enters the cartridge at atmospheric pressure. Thus, sufficient acceleration of the DNA-coated microcarriers is generated for penetration of the target cell membrane.

Compressed helium of grade 4.5 (99.995%) or higher should be used; lower grades may contain contaminating material which can obstruct gas flow within the Helios Gene Gun as well as contaminate the biological sample. A helium tank pressurized to 2,600 psi [approximately 5 ft (1.7 m) high, 291 cu ft standard in the United States] is recommended, although a smaller tank [~2.5 ft (~0.8 m) high] may be used. Follow all safety instructions provided by the helium supplier for helium tank installation.

The helium pressure regulator (supplied) has a CGA 580 female fitting (standard in the United States) for attachment to the user-supplied helium tank. An adaptor to this fitting may be required outside of the United States. Contact your local Bio-Rad office for information on the helium pressure regulator adaptor requirements in your location. The regulator supplied with the Helios Gene Gun is the only one that should be used with this instrument because of its three safety features: (1) a self-venting valve that permits decreasing the pressure in the Helios Gene Gun System in the event of battery failure or when it is necessary to reduce the pressure during an experiment; (2) an over-pressure relief valve that prevents the helium pressure in the Helios Gene Gun System from being adjusted above 700 psi $\pm$ 10%; and (3) a check valve that shuts off pressure if the helium hose is disconnected while the system is still pressurized (**Note:** a check valve is also present at the female connector of the helium hose where it connects to the Gene Gun to shut off pressure to the gun if it is disconnected while the system is still pressurized.) Refer to Section 4.2 for proper use of the helium regulator and to Section 5.4 for a description on shutting down the Gene Gun System.

A user supplied 10" or 12" (~25 cm) adjustable wrench or a 1 1/8" open end wrench is required for attachment of the helium regulator to the helium tank.

Nitrogen Supply

Compressed nitrogen of grade 4.8 (99.998%) or higher is required for cartridge preparation using the Tubing Prep Station. Nitrogen is used for this purpose because it is relatively inexpensive and provides a water-free atmosphere for evaporating the ethanol from the DNA/gold sample inside the tubing. As with the helium tank, the nitrogen tank should be properly secured on a floor stand or with a strap for safety.

A nitrogen regulator must be attached to the tank. A single stage regulator with an output gauge that registers a maximum of 30 psi is recommended since an output pressure of no more than 1–2 psi is needed to produce the 0.4 liters per minute (LPM) flow rate necessary for using the Tubing Prep Station. A regulator especially designed for this use, including a self-venting valve, an over-pressure relief valve, and a hose barb for attaching the nitrogen hose is available from Bio-Rad (catalog number 165-2425). Other regulators which are adjustable to give a low pressure output may also be used. Manufacturers of regulators include Victor and Matheson; examples of regulators which may be used include Victor Model No. SR250A-580 and Matheson Model No. 3537-580. Large scientific supply houses (e.g., VWR, Fisher, CMS, etc.) are also good sources for regulators. The nitrogen line provided for use with the Cartridge Prep Unit is 3/16" diameter Tygon tubing.

Battery

One battery is provided with the Helios Gene Gun System. Under normal use, it should provide approximately 1,000 discharges. For maximum life, only alkaline batteries should be used.

Laboratory Equipment

The following materials should be available before beginning any work with the Helios system.

- Ultrasonic cleaner (*e.g.*, Fisher FS3, Branson 1210)
- Vortex mixer
- Analytical balance
- Microfuge
- Peristaltic pump capable of pumping 5–8 ml/min (*e.g.* Bio-Rad Econo Pump, catalog number 731-8140)
- 1.5 ml microfuge tubes
- 20 µl, 200 µl, and 1,000 µl micropipettors and tips
- 5 ml and 10 ml pipettes and pipettors
- Lab timer
- Ear protection (*e.g.*, VWR catalog number 56610-728 (earmuffs) or catalog number 56610-680 (ear plugs))
- 1 1/8" open end or 10" or 12" (~25 cm) adjustable wrench
- Helium tank (grade 4.5 or higher)
- Nitrogen tank (grade 4.8 or higher)
- Nitrogen regulator (*e.g.*, Bio-Rad, catalog number 165-2425)
- Scissors
- Marking pen

Laboratory reagents

The following chemicals will be needed for coating plasmid onto the gold and for preparing the tubing:

- Gold microcarriers
- Polyvinylpyrrolidone, MW = 360,000
- 100% ethanol (*e.g.*, Spectrum Chemical, catalog number ET-107; it is extremely important that this be free of water; an unopened bottle should be used daily)
- Spermidine (*e.g.*, Sigma, catalog numbers S-0266 or S-4139)
- Calcium chloride (CaCl₂)
- Plasmid DNA (for most applications, this should be at a concentration of ~1 µg/µl)
- Plasmid DNA of high purity suitable for the Helios Gene Gun can be obtained through use of any of Bio-Rad's Quantum Prep[®] Plasmid Prep kits.

Catalog Number	Description
732-6100	Quantum Prep Plasmid Miniprep Kit, 100 preps
732-6120	Quantum Prep Plasmid Miniprep Kit, 20 preps
732-6130	Quantum Prep Plasmid Maxiprep Kit, 10 preps
732-6150	Quantum Prep HT/96 ClearVac Plasmid Miniprep Kit, 2 x 96 preps

Section 3

Product Description

3.1 Packing List

The Helios Gene Gun System (see Figures 3 and 4) is shipped with the following components. If items are missing or damaged, contact your local Bio-Rad office.

Helios Gene Gun Kit

- Instruction manual
- Warranty card (please complete and return)
- Helios Gene Gun
- 5 cartridge holders
- 5 barrel O-rings
- 5 barrel liners (four plus one installed in Gene Gun)
- 9 volt battery
- Cartridge extractor tool

Helium hose assembly

Helium regulator

Tubing Cutter and 10 razor blades

Tubing Prep Station (see Figure 4)

- Tubing Prep Unit (base, tubing support cylinder and power cord)
- Nitrogen hose [12 ft, (~4m), Nalgene tubing 8000-0030, 3/16" ID, 5/16 " OD]
- 3/16" barb-to-male Luer-Lok fitting
- 10 cc syringe sleeve
- 5 O-rings, Tubing Prep Station
- 2 1/8" barb-to-male Luer fittings
- 5/64" Allen wrench
- Syringe Kit
 - 5 10 cc syringes
 - 5 1/8" barb to female Luer fittings
 - 1 syringe adaptor tubing [silicone, 5ft, (~2.6 m), 0.104" ID, 0.192" OD]

Optimization Kit

- Gold-Coat Tubing [50 ft, (~26 m)]
- 1.6 μ gold microcarriers, 0.25 g
- 1.0 μ gold microcarriers, 0.25 g
- 0.6 μ gold microcarriers, 0.25 g
- Polyvinylpyrrolidone, 360,000 MW (0.5 g)
- 5 desiccant pellets (store tightly sealed)
- 5 cartridge collection/storage vials

Note: If any of the system components (Helios Gene Gun, Tubing Prep Station, Tubing Cutter, Helium Regulator, or Helium Hose) are dropped, check them for proper operation before use.




Fig. 3. Major components used for sample delivery with the Helios Gene Gun.




Fig. 4. Components of the Tubing Prep Station.

3.2 Identification of System Components and Controls

Helios Gene Gun

The locations front and back refer to the barrel end and LED display end of the Helios Gene Gun, respectively. The locations left and right refer to the left and right sides of the Gene Gun from the viewpoint of the user holding the device. Top and bottom refer to the side of the gun that the cartridge holder is on and the side of the gun that the helium hose connects, respectively (see Figure 5).




Fig. 5. Components and controls on the Helios Gene Gun.

Gene Gun Controls	Description
Cylinder Lock	Controls movement of the barrel pin. The cylinder lock is spring-loaded; its natural position is in the backward (locked) position so the barrel pin is inserted in the hole in the cartridge holder; this keeps the cartridge holder in its proper position for firing. Moving the cylinder lock forward disengages the barrel pin from the cartridge holder to permit removing the cartridge holder from the gun. Moving the cylinder lock forward and to the right latches the cylinder lock to permit removal of the cartridge holder; however, to prevent damage to the O-rings, the cartridge holder should only be removed after compressing the cylinder advance lever (see below).
Safety Interlock Switch	Switch that must be held down to permit the trigger button to be operational. Once this switch is depressed, the trigger button is functional for 30 sec; the LED ARMED display flashes quickly during this time. If the trigger button is not pressed within the allotted time, this safety interlock switch must be released and pressed again to re-activate the trigger button.
Trigger Button	Controls the flow of helium gas through the Gene Gun. This switch activates the solenoid, momentarily (for ~40 msec) opening the main valve, and permitting helium to enter the cartridge and barrel. The trigger button is only active for 30 sec after the safety interlock switch is depressed.
Cylinder Advance Lever	A multi-functional lever which is spring-activated by pulling the lever backwards. When inserting or removing a cartridge holder, pull back and hold in the cylinder advance lever; this moves the barrel liner forward to provide additional room for maneuvering the cartridge holder. After discharging the microcarriers from one cartridge, pull back and release the cylinder advance lever; this ratchets the cartridge holder, bringing the next cartridge into firing position—the number visible on the very top of the cartridge rim indicates the active sample position.

Push Bar	A metal bar that ratchets the cartridge holder from one position to the next when the Cylinder Advance Lever is pressed. Move this bar to the left (outward) prior to inserting a cartridge holder to provide additional room for maneuvering the cartridge holder.
LED Display	An 11 light display. The display is normally off; inserting a cartridge holder in the Gene Gun and advancing to position 1 activates the display. The left-most 7 LEDs act to indicate charging and ready status of the gun. After each firing of the gun and at reset, the CHARGING LEDs turn-on in bargraph fashion, left-to-right, throughout the 5 second charging time. Once the Gene Gun is fully charged, the CHARGED LED will flash, indicating the safety interlock switch can be pressed. Upon pressing the safety interlock, the ARMED LED's sequentially flash. When the trigger is pressed during the 30 sec armed period, the gun fires and the FIRED LED turns-on for 1 sec. The change bargraph then operates as described. The last light indicates battery status : good battery (steady green light) or low battery (flashing red light). The Gene Gun can be fired when the green light is illuminated. If the battery is low, neither the safety interlock switch or the trigger button is active and an alarm will beep three times every 15 sec.

Tubing Prep Station (see Figure 6)

Tubing Prep Unit Controls	Description
3 Position Switch	Located on the motor housing and controls rotation of the tubing support cylinder. At position (I), the tubing support cylinder turns continuously at 30 revolutions per minute (rpm). At position (II), the tubing support cylinder rotates only while the switch is depressed. At position (O), no rotation occurs and the unit is off.
Flow Meter	Registers the rate of nitrogen flow in liters per min (LPM) into the tubing support cylinder. The valve on the flowmeter is used to control the rate of nitrogen flow.
Tubing Support Cylinder	A 28" aluminum cylinder with an opening on the right side leading to a channel which holds the Gold-Coat tubing. The left side of the channel has a replaceable O-ring into which the tubing must be inserted. The tubing support cylinder can be removed by pushing it to the right/left to compress the spring which holds it in position.

Tubing Cutter (see Figure 7)

An instrument for rapid preparation of cartridges from Gold-Coat tubing. It cuts the tubing into the exact length and shape required by the Gene Gun.

Tubing Cutter Part	Description
Arm	A spring-loaded piece that holds a razor blade (used for cutting the tubing). The razor blade is held in place by the locking knob on the lock block.
Base	The support for the arm. It positions a storage vial under the tubing channels so that the cut tubing pieces fall into the vial.




Fig. 6. Components and controls on the Tubing Prep Station, fully assembled.




Fig. 7. The Tubing Cutter.

Cartridge Extractor Tool (see Figure 8)

A 12-prong tool for removal of tubes from the cartridge holder. One prong is longer than the others so it can be easily inserted into one of the bores of the cartridge holder to orient the remaining 11 prongs.




Fig. 8. Cartridge Holder and Cartridge Extractor Tool.

Section 4

Setting up the Helios Gene Gun System

4.1 Inserting the Battery into the Helios Gene Gun

The electrical system of the Helios Gene Gun is powered by a 9 volt battery. Under normal use, this should provide sufficient energy for 1,000 shots. The battery compartment is in the base of the handle near the attachment fitting for the helium hose (Figure 9).

To insert the battery, first remove the battery cover by sliding it toward the front of the Gene Gun. Slide the battery into the opening as shown in Figure 9. The battery must be oriented with the positive terminal toward the front of the Gene Gun indicated by the "+" symbol inside the battery compartment. (**Note:** If the battery is inserted correctly, a tone will be audible for ~5 sec and the battery status LED will light; if the battery is inserted backward, the electrical system will not operate and the Helios Gene Gun will be non-operational.) The battery is held in place by the battery cover.




Fig. 9. Battery compartment. The battery compartment is located at the base of the handle of the Gene Gun next to the connection for the helium hose and is protected by a battery cover that slides forward. The battery is inserted with the positive terminal (the smaller of the two terminals) facing forward.

4.2 Connecting the Helios Gene Gun to a Helium Source

Refer to Section 1, Safety Information, and Section 3.2, Identification of System Components and Controls, prior to system installation.

Helium Pressure Regulator Installation

Components needed

Pressure regulator for helium cylinder (with pressure relief valve, check valve, and female Swagelok® Quick-Connect fitting), provided with unit (Figure 10).

Helium cylinder of grade 4.5 or higher (minimum 99.995% pure); maximum pressure of 2,600 psi, user supplied.

1 1/8" open-end wrench or a 10" or 12" adjustable wrench, user supplied.

Note: The regulator is intended for use only with helium gas with a maximum pressure of 2,600 psi. The outlet on pressurized helium cylinders used in the United States is compatible with the fitting supplied on the pressure regulator with the Helios Gene Gun System (CGA 580, female fitting). Outside the US, contact your local Bio-Rad office for information regarding the proper cylinder/regulator fitting in your area.

Procedure

1. Secure the cylinder in a floor stand or to a wall or lab bench with a strap so it will not tip or fall during use.
2. Inspect the cylinder valve for dirt, dust, oil, grease or damaged threads. Remove dust and dirt with a clean cloth. Do not attach the regulator if you determine that the valve port is damaged or cannot be cleaned. Inform your gas supplier of this condition and request a replacement cylinder.
3. Clear the valve port of any foreign matter by standing to the side of the cylinder and quickly opening and closing the cylinder valve.
4. Attach the regulator to the cylinder valve and tighten securely with a 1 1/8" open-end wrench or a 10" or 12" adjustable wrench.

Attaching the Helios Gene Gun to the Helium Regulator

Components needed

- Helium regulator attached to a helium cylinder
- Helium hose assembly
- Helios Gene Gun

Procedure

1. Insert the stem of the Swagelok Quick-Connect fitting on the helium hose into the opening in the body of the Swagelok Quick-Connect fitting on the helium regulator and push until it clicks. The helium hose will be locked into the helium regulator (see Figure 10). **Note:** If the helium regulator has been pressurized, the stem and body will not lock. Turning the regulator valve counter-clockwise will depressurize the system.
2. In a similar manner, insert the stem of the Swagelok Quick-Connect fitting on the Helios Gene Gun into the opening in the body of the Swagelok Quick-Connect fitting on the helium hose until it clicks. The Gene Gun will be locked into the helium hose.




Fig. 10. Connecting the helium hose to the helium regulator.

4.3 Setting up the Tubing Prep Station

Refer to Section 3.2, Identification of System Components and Controls, prior to system installation. See Figure 11 for a diagram of the assembled Tubing Prep Station.

The Tubing Prep Station is shipped disassembled. The following sections describe assembly of the Tubing Prep Station, attachment of the syringes and tubing, installation of the nitrogen pressure regulator, and connection of the Tubing Prep Station to the nitrogen regulator.

A peristaltic pump is recommended for removal of the ethanol from the Gold-Coat tubing after the microcarriers have been loaded (see Section 5.1). If a peristaltic pump is not available, this may be done manually using a syringe. Assembly of the syringe and tubing is described in step 5 in the following section.

Assembly of the Tubing Prep Station and Syringes

Components needed

Tubing Prep Station, base
Tubing Prep Station, tubing support cylinder
Tubing Prep Station, power cord
O-rings, for tubing support cylinder
Syringe adapter tubing (silicone, 5 ft, 0.104" ID, 0.192" OD)
10 cc syringes (2)
10 cc syringe sleeve
1/8" barb-to-male Luer fittings (2)
1/8" barb-to-female Luer fittings (2)
Scissors, user supplied




Fig. 11. Components and controls on the Tubing Prep Station, fully assembled.

Procedure

1. Place an O-ring in the stainless steel end of the tubing support cylinder next to the spur gear.
2. Insert the end of the tubing support cylinder with the spring into the hole in the support bar on the base of the Tubing Prep Station. Push the tubing support cylinder to compress the spring, then slide the end of the tubing support cylinder containing the O-ring into the support on the motor housing. (**Note:** Carry the Tubing Station by the base, not by the tubing support cylinder.)
3. Cut a 4–5" piece of syringe adapter tubing; attach each end to the barb of the 1/8" barb-to-male Luer fittings. Attach one of the male Luer-Lok fittings to the female Luer on the top of the flowmeter. Attach the other male Luer-Lok fitting to the female Luer on the side of the motor housing on the Tubing Prep Station.

4. Cut a 12–13" piece of syringe adapter tubing; attach one end of the tubing to the barb of a 1/8" barb to female Luer fitting; attach the female Luer to a 10 cc syringe. This syringe and tubing will be used to load the DNA/microcarrier suspension into the Gold-Coat tubing (Section 5.1).
5. (Optional—if peristaltic pump not used) Cut a 16-18" piece of syringe adapter tubing; attach one end of the tubing to the barb of a 1/8" barb-to-female Luer fitting; attach the female Luer to a 10 cc syringe. Slide the syringe into the 10 cc syringe sleeve; fasten the sleeve (with the open side facing up) to the clamp on the base of the Tubing Prep Station using the plastic screw; the screw should be tightened sufficiently to hold the sleeve and syringe stationary but allow free movement of the plunger. This syringe and tubing will be used to remove the ethanol from the DNA/microcarrier suspension in the Gold-Coat tubing if a peristaltic pump is not available.
6. Attach the power cord to the three pronged receptacle on the back of the base of the Tubing Prep Station and plug into an appropriate electrical outlet.

Nitrogen Pressure Regulator Installation

Components needed

Pressure regulator for nitrogen cylinder, user supplied (see Section 2.4). A regulator is available from Bio-Rad (catalog number 165-2425) which is ready for use in the Helios Gene Gun System.

Nitrogen tank of grade 4.8 or higher (99.998% pure); maximum pressure of 2,600 psi, user supplied.

1 1/8" open-end wrench or a 10" or 12" (~25 cm) adjustable wrench, user supplied.

Procedure

1. Secure the cylinder in a floor stand or to a wall or lab bench with a strap so it will not tip or fall during use.
2. Inspect the cylinder valve for dirt, dust, oil, grease or damaged threads. Remove dust and dirt with a clean cloth. Do not attach the regulator if you determine that the valve port is damaged or cannot be cleaned. Inform your gas supplier of this condition and request a replacement cylinder.
3. Clear the valve port of any foreign matter by standing to the side of the cylinder and quickly opening and closing the cylinder valve.
4. If necessary, connect a male hose barb to the nitrogen regulator.
5. Attach the regulator to the cylinder valve and tighten securely with a 1 1/8" open-end wrench or a 10" or 12" (~25 cm) adjustable wrench.

Connecting the Tubing Prep Station to the Nitrogen Regulator

Components needed

Nitrogen regulator attached to a nitrogen cylinder

Nitrogen hose [12 ft (~4 m), of 3/16" ID Nalgene tubing]

3/16" barb-to-male Luer fitting

Tubing Prep Station

Procedure

1. Determine the length of nitrogen hose needed to connect the Tubing Prep Station to the nitrogen regulator and cut it if necessary.
2. Connect the 3/16" barb-to-male Luer fitting to one end of the nitrogen hose.
3. Push the other end of the nitrogen hose onto the male hose barb on the nitrogen regulator. Do not use hose clamps to secure the nitrogen hose on the Tubing Prep Station.

4. The nitrogen regulator should be turned on and adjusted to the correct pressure prior to connecting the nitrogen line to the Tubing Prep Station. Close the nitrogen pressure regulator by turning the regulator adjustment screw counterclockwise until the adjusting spring pressure is released and the screw moves without resistance. Release nitrogen into the pressure regulator by carefully and slowly opening the cylinder valve on the nitrogen tank. The cylinder pressure in the tank is indicated on the high pressure gauge (the gauge closest to the cylinder). Hold the nitrogen line in your hand and slowly turn the regulator adjustment screw clockwise until nitrogen can just be heard flowing from the nitrogen line. Clamp the nitrogen line with your fingers; the pressure on the output (low pressure) gauge should register no more than 1–2 psi. If the pressure is too high, turn the regulator screw counterclockwise to reduce the pressure.
5. Connect the male Luer-Lok fitting on the the nitrogen line to the female Luer fitting on the side of the Tubing Prep Station. The flow of nitrogen into the Tubing Prep Station can be adjusted with the valve on the flowmeter.

Section 5

Operation of the Helios Gene Gun System

5.1 Quick Guide to Operation

Before the Bombardment

1. Coat microcarriers with DNA, load into tubes, and prepare cartridges prior to day of experiment.
2. Check helium supply (50 psi in excess of desired delivery pressure).
3. Clean and/or sterilize the Gene Gun, tube holders, and barrel liners as appropriate.
4. Connect the Gene Gun to a helium source.
5. Activate the Gene Gun: turn on the flow of helium to the desired pressure and with an empty cartridge holder in place, make 2–3 “pre-shots” by engaging the safety interlock and firing the trigger.

Firing the Device

1. Load cartridges into the cartridge holder and place in Gene Gun.
2. Prepare/position target cells for bombardment.
3. Bombard sample: engage safety interlock and press the firing trigger

After the Bombardment

1. Remove cartridge holder from Gene Gun.
2. Remove cartridges from cartridge holder.
3. Turn off the helium pressure to the system.
4. Turn the regulator value counterclockwise to de-pressurize the system.
5. Disconnect the helium hose and Gene Gun.

5.2 Preparation of System Components Prior to Bombardment

Calculating the Amounts of Gold and Plasmid Required

Prior to precipitating DNA onto the gold particles and loading them into the Gold-Coat tubing, it is necessary to calculate the amount of DNA and gold required for each transformation. Points to consider in making these calculations are presented below. The amount of DNA loaded per mg of microcarriers is referred to as the DNA Loading Ratio (DLR). Typical DLRs range between 1 and 5 μg DNA/mg gold. Adding more DNA tends to cause agglomeration of the gold particles, probably as a result of DNA binding to more than one particle. The amount of microcarriers delivered per target is referred to as the Microcarrier Loading Quantity (MLQ). Typical MLQs range from 0.25 to 0.5 mg/cartridge for *in vivo* delivery to epidermal cells, but may be slightly lower for *in vitro* delivery to mammalian cells. Refer to Table 2 for representative starting amounts of microcarriers and plasmid to use for different MLQs and DLRs. Refer to Section 7 for suggestions on parameter optimization and starting conditions for using the Helios Gene Gun to deliver DNA to mammalian cells

Procedure 1: Determining the Microcarrier Loading Quantity (MLQ)

1. For most systems, delivering 0.5 mg of gold per target is a good starting point.
2. A 1 ml suspension will fill an 8.5" length of tubing; one cartridge is 0.5" long. Each 30" length of tubing can be filled with approximately 25" (3.0 ml) of DNA/gold suspension. (There will be a void space at each end.)
3. For delivering 0.5 mg of microcarriers per target (MLQ=0.5), resuspend the DNA/microcarrier sample at 8.5 mg of gold/ml ethanol. A 25" length of tubing will require 25 mg of gold resuspended in a volume of 3 ml of ethanol.

Procedure 2: Determining the DNA Loading Ratio (DLR)

1. For many applications, delivery of 1 μg of plasmid per target is a good starting point.
2. At a MLQ of 0.5 mg/cartridge, a DLR of 2 μg DNA/mg gold results in loading 1 μg of DNA/cartridge and in delivery of 1 μg of DNA per target. Preparation of two lengths of Gold-Coat tubing requires 100 μg of DNA and 50 mg of gold. The concentration of DNA should be approximately 1 $\mu\text{g}/\mu\text{l}$ and the volume of DNA should not exceed the volume of spermidine in Section 5.2, Precipitation of DNA onto Microcarriers, Step 3. If the DNA is too dilute, concentrate it by ethanol precipitation. If a high DLR is desired, increase the volume of spermidine and CaCl_2 so that equal volumes of each component are added (spermidine, DNA, and CaCl_2) up to a total volume of 1,200 μl .
3. For a detailed description on determining which MLQs and DLRs will work best for several mammalian targets, refer to Section 7.

Table 2. Microcarriers and DNA Required for Various Microcarrier Loading Quantities (MLQ) and DNA Loading Ratios (DLR)¹

Calculated Particle Delivery Conditions			Materials Required for Selected MLQ's and DLR's			
MLQ	DLR		Gold (mg)	DNA (µg)	Final volume (ml) ²	Tubing (total in) ³
(mg/shot)	(µg/mg gold)	(µg/shot)				
0.5	2	1	50	100	6.0	50
0.125	8	1	12.5	100	6.0	50
0.25	4	1	25	100	6.0	50
0.75	1.33	1	75	100	6.0	50
1.0	1	1	100	100	6.0	50
0.5	0.002	0.001	50	0.1	6.0	50
0.5	0.02	0.01	50	1	6.0	50
0.5	0.2	0.1	50	10	6.0	50
0.5	10	5	20	200	2.4	20

1 For most applications with mammalian cells, in initial experiments, use an MLQ of 0.5 and a DLR of 2.

2 Based on loading 1 ml of the DNA-coated microcarriers suspended in ethanol in 8.5 inches (22 cm) of Gold-Coat tubing.

3 Various lengths of tubing may be prepared. Adjust amounts of gold, DNA volume of ethanol in proportion to the change from the length of tubing listed above for each desired MLQ and DLR. Approximately 25 inches of tubing can be prepared in the Tubing Prep Station at one time; 50 inches of tubing will usually yield 80–90 cartridges.

Precipitation of DNA onto Microcarriers

It is important to use an unopened bottle of 100% ethanol each day this step is performed. Opened bottles of ethanol absorb water and the presence of water in the tubing while drying will lead to streaking, clumping, and uneven coating of the microcarriers over the inner surface of the Gold-Coat tubing, resulting in poor or unusable cartridges. All ethanol solutions should be opened only briefly when in use and kept tightly capped when not in use.

Polyvinylpyrrolidone (PVP) serves as an adhesive during the cartridge preparation process. At higher discharge pressures, preparing cartridges with PVP can increase the total number of particles delivered. The optimum amount of PVP to be used must be determined empirically. Typical PVP concentrations range from 0.01 to 0.1 mg/ml. For recommendations on the amount of PVP to use in initial experiments, refer to Section 7.

Materials

Provided

Gold microcarriers

Polyvinylpyrrolidone (PVP), 360,000 MW

To be supplied by user

Fresh 100% ethanol

15 ml disposable polypropylene centrifuge tubes

1.5 ml microfuge tubes

0.05 M spermidine

1 M CaCl₂

200 µl and 500 µl pipettors and tips

5 ml, 10 ml pipettes and pipette-aid

Purified plasmid DNA resuspended in distilled water or 10 mM Tris (pH 8.0), 1 mM EDTA

Ultrasonic cleaner (*e.g.*, Fisher FS3, Branson 1210)

Analytical balance capable of weighing microgram quantities

Microfuge

Procedure

Time considerations: preparation of the DNA/gold suspension requires approximately 30 min. Several samples may be prepared simultaneously without a significant increase in time.

1. Prepare a stock solution of 20 mg/ml PVP in ethanol in a small screw-cap container. Dilute this solution with ethanol to prepare PVP solutions at the desired concentration (generally 0.01–0.1 mg/ml); prepare 3.5 ml of the dilute solution for each 30" length of Gold-Coat tubing, (25" to be coated) in the Tubing Prep Station. Keep these solutions tightly capped when not in use. Prepare solution daily.
2. In a 1.5 ml microfuge tube, weigh out gold microcarriers. (Refer to Procedure 1 for a detailed description on determining MLQ. Refer to Table 2 for suggestions on the relative amounts of gold and microcarriers required and on the length of tubing produced.)
3. To the measured gold, add 100 μ l of 0.05 M spermidine. (However, if the volume of plasmid to be added in step 5 is greater than 100 μ l, refer to the discussion above for Procedure 2: Determining the DNA Loading Rate, and add the appropriate volume of spermidine.
4. Vortex the gold and spermidine mixture for a few seconds, then sonicate for 3–5 seconds using an ultrasonic cleaner to break up gold clumps.
5. To the gold and spermidine mixture, add the required volume of plasmid to achieve the desired DLR. (Refer to Procedure 2 for a detailed description on determining DLR. Refer to Table 2 for suggestions on the relative amounts of gold and microcarriers required and on the length of tubing produced.) For co-transfection of multiple plasmids, add each of the plasmids at this step. DNA does not associate with the microcarriers prior to addition of CaCl_2 .
6. Mix DNA, spermidine and gold by vortexing ~5 sec.
7. While vortexing the mixture at moderate rate on a variable speed vortexer, add 100 μ l of 1 M CaCl_2 dropwise to the mixture. The volume added should equal that of the spermidine in Step 3.
8. Allow the mixture to precipitate at room temperature for 10 min.
9. Most of the gold will now be in the pellet, but some may be on the sides of the tube. The supernatant should be relatively clear. Spin the microcarrier solution in a microfuge ~15 sec to pellet the gold. Remove the supernatant and discard.
10. Resuspend the pellet in the remaining supernatant by vortexing briefly. Wash the pellet three times with 1 ml of fresh 100% ethanol each time; spin ~5 sec in a microfuge between each wash. Discard the supernatants.
11. After the final ethanol wash, resuspend the pellet in 200 μ l of the ethanol solution containing the appropriate concentration of PVP prepared in step 1. Transfer this suspension to a 15 ml disposable polypropylene centrifuge tube with a screw cap. Rinse the microfuge tube once with 200 μ l with the same ethanol/PVP solution and add to the centrifuge tube. Add the necessary volume of the ethanol/PVP solution to the centrifuge tube to bring the DNA/microcarrier solution to the desired MLQ.
12. The suspension is now ready for tube preparation. Alternatively, the DNA/microcarrier suspensions can be stored for up to 2 months at -20°C . Prior to freezing, tighten the cap securely and put Parafilm® around the cap of the tube. After storage at -20°C , allow the particle suspension to come to room temperature prior to breaking the Parafilm seal.

Loading the DNA/Microcarrier Suspension into Gold-Coat Tubing Using the Tubing Prep Station

Materials

Supplied

Tubing Prep Station (see Section 4.3)

Gold-Coat tubing

To be Supplied by User

Microcarrier/DNA suspension(s) from Section 5.1, Precipitation of DNA onto Microcarriers, at room temperature

Ultrasonic cleaner

Vortexer

100% ethanol

Peristaltic pump

Minute timer

Nitrogen tank (see Section 2.4)

Nitrogen regulator (see Section 2.4)

Scissors

Procedure

Time considerations: Since only one piece of Gold-Coat tubing can be coated at a time, this procedure may take 15–30 min for the first sample and 15–20 min for additional samples.

1. Set up the Tubing Prep Station and connect to a nitrogen tank as described in Section 4.3.
2. Prior to using the Tubing Prep Station, a peristaltic pump should be set up and calibrated to be used later for removing ethanol from the tubing at the rate of 0.5–1.0"/sec. [**Note:** 1 ml of liquid occupies 8.5" (21.5 cm) of tubing, removing liquid at 0.5–1.0"/sec is the same as 0.06–0.12 ml/sec or 3.6–7.2 ml/min.] Using a 10 ml graduated cylinder, calibrate a peristaltic pump at 5.5–6.0 ml/min. The tubing at the end of the peristaltic pump which will be connected to the Gold-Coat tubing should have an inside diameter of 1/8".
3. If a peristaltic pump is not available, a syringe fitted with a 16–18" piece of silicone adaptor tubing and inserted into the syringe sleeve and clamped onto the base of the Tubing Prep Station may be used to remove the liquid from the Gold-Coat tubing (see Section 4.3). Liquid should be removed from the Tefzel tubing at 0.5–1.0"/sec. Practice removing liquid at this rate from the Tefzel tubing by cutting a ~30" piece of tubing, loading it with ~3.0 ml of ethanol (~24" (~60 cm) of tubing), inserting it into the Tubing Prep Station, and removing the liquid at 0.5–1.0"/sec. Marking the tubing at several points and using a timer which measures time in seconds should make removal more accurate. It should take 25–45 sec to draw the liquid from the entire length of tubing. (**Note:** use ethanol rather than water for practicing this step because, if water leaks into the Tubing Prep Station, it may contaminate subsequent pieces of Tefzel tubing leading to poorly coated tubes [see Section 8.2]).
4. Prior to preparing cartridges, ensure that the Gold-Coat tubing is completely dry by purging with nitrogen. Insert an uncut piece of tubing into the opening on the right side of the Tubing Prep Station. The edge of the hole is beveled to permit easier insertion. Push the tubing into the hole and into the tubing support cylinder. At the opposite end of the tubing support cylinder is an O-ring. There will be slight resistance as the tubing is pushed into the O-ring; insert the tubing another 1/2" (1 cm).
5. Using the knob on the flowmeter, turn on the nitrogen and adjust the flow to 0.3–0.4 LPM. Allow nitrogen to flow through the Gold-Coat tubing for at least 15 min immediately prior to using it in the following steps.

6. Remove the Gold-Coat tubing from the Tubing Prep Station. Turn off the flow of nitrogen to the Tubing Prep Station using the knob on the flowmeter.
7. From the dried Gold-Coat tubing cut a 29–30" (~75 cm) length of tubing for each 3 ml sample of Microcarrier/DNA suspension. (**Note:** Cutting the tubing with a scissors may distort the shape of the end; the tubing may be easier to insert into the tubing support cylinder if the end is subsequently cut in the tubing cutter.) Insert one end of the Gold-Coat tubing into the end of the adaptor tubing fitted to the 10 cc syringe (see Section 4.3, step 5).
8. Vortex the microcarrier suspension and, if necessary, sonicate briefly to achieve an even suspension of gold. Invert the tube several times to resuspend the gold; immediately remove the cap and quickly draw the gold suspension into the Gold-Coat tubing approximately 22–24", approximately 58 cm, (6–8" (~17 cm) from the end). **AVOID DRAWING BUBBLES INTO THE GOLD-COAT TUBING:** Do not vortex the microcarrier suspension while drawing it into the tubing. Do not try to remove all of the liquid from the container with the suspension. Remove the tubing from the suspension and continue drawing the suspension into the tubing another 2–3" (~6 cm) to leave some space at each end.
9. Immediately bring the Gold-Coat tubing to a horizontal position and slide the loaded tube, with syringe attached, into the tubing support cylinder in the Tubing Prep Station until the tubing passes through the O-ring.
10. Allow the microcarriers to settle for 3–5 min. Detach the Gold-Coat tubing from the adaptor tubing and attach to the tubing on the peristaltic pump or on the 10 cc syringe. (Be careful not to rotate the Gold-Coat tubing.) Remove ethanol at the rate of 0.5–1.0"/sec (this should require 30–45 sec).
11. Detach the peristaltic pump or syringe from the Gold-Coat tubing. Immediately turn the Gold-Coat tubing 180° while in the groove and allow the gold to begin coating the inside surface of the tubing for 3–4 sec.
12. Turn the switch on the Tubing Prep Station to ON (I) to start rotating the Tubing Prep Station. (**Warning:** Keep objects away from the gears of the Tubing Prep Station while in operation.)
13. Allow the gold to smear in the tube for 20–30 sec, then open the valve on the flowmeter to allow 0.35–0.4 LPM of nitrogen to dry the Gold-Coat tubing, while it continues to rotate.
14. Continue drying the Gold-Coat tubing while turning for 3–5 min.
15. Turn the motor on the Tubing Prep Station to OFF (O). Turn off the nitrogen by closing the valve on the flowmeter. Remove the tubing from the tubing support cylinder.

Preparing 0.5" Cartridges Using the Tubing Cutter

This procedure requires only a few minutes and should be performed as soon after loading the DNA/microparticle suspension into the tubing as possible. It is important to store the coated tubes in a desiccated environment. Tubes stored at 4 °C are stable for at least 8 months.

Materials

Supplied

- Tubing Cutter
- Razor blade
- Cartridge storage vial
- Desiccant pellets

To be Supplied by User

- Gold-Coat tubing coated with microcarriers from Section 5.2
- Scissors
- Marking pen

Procedure

1. Examine the coated the Gold-Coat tubing to verify that the microcarriers are evenly distributed over the length of the tubing. Ideally, the gold should be spread uniformly over the entire inside surface of the tubing; however, while drying, it may polarize to one side of the tubing. As long as there are no clumps or bare sections, the tubing can be used for cartridges.
2. Using scissors, cut off and discard sparsely and unevenly coated tubing from one of the ends. With a marking pen mark any sections of tubing to be discarded. This is usually limited to the outer 1–2" where the gold has settled, but may also include internal sections of tubing which are un unevenly coated with gold.
3. Use the Tubing Cutter to cut the remaining tubing into 0.5" pieces as follows:
 - a. Place a cartridge storage vial containing a desiccant pellet inside the base of the Tubing Cutter.
 - b. Insert the cut ends of one or two pieces of tubing into the tubing channels on the front face of the Tubing Cutter (Figure 7); be sure to push them in until they make contact with the rear plate.
 - c. Push down sharply on the handle; the tubes will drop into the storage vial.
 - d. Repeat the process of inserting the uncut tubing and cutting the cartridges until the entire length of usable tubing is cut.
4. Cap the vial tightly, label, wrap with Parafilm, and store at 4 °C.

5.3 Particle Delivery Using the Helios Gene Gun

This section describes the procedure for preparing the Helios Gene Gun for firing, discharging the device, loading cartridges into the cartridge holder, and delivering DNA to target cells. (**Warning:** Use the Helios Gene Gun in a well ventilated area.)

Materials

Supplied

- Helios Gene Gun
- Barrel liner (see Section 10.5 on sterilization)
- Cartridge holder
- Helium hose
- Helium regulator (see Section 4.2)

To be Supplied by User

- Helium tank (see Section 2.4)
- Ear protection
- Cartridges with coated microcarriers (from Section 5.2)

Activating the Helios Gene Gun for Gene Delivery

When first setting up the system and prior to discharging the Gene Gun into a target, it is important to pressurize the helium hose and the internal reservoirs of the gun with the correct helium pressure. This is accomplished by discharging the device prior to delivering loaded tubes to the target cells. It is important to have a cartridge holder in place during these “pre-shots” to keep the O-rings from being blown out (see Section 10.2).

1. Insert an empty cartridge holder into the Helios Gene Gun as follows:
 - a. Move the cylinder lock on the Gene Gun so it is latched in the forward position and the barrel pin does not protrude behind the barrel (Figure 12).
 - b. Unlatch the push bar by pulling it outward (Figure 12).




Fig. 12. Positioning the cylinder lock and the push bar in preparation for loading a cartridge holder into the Gene Gun. The cylinder lock has been pushed forward and into the slot on the right to move the locking pin away from the opening occupied by the cartridge holder. Likewise, the push bar has been pulled to the left to increase the space for inserting the cartridge holder.




Fig. 13. Inserting a cartridge holder into the Gene Gun. Pulling the cylinder advance lever in moves the barrel forward to increase the space for inserting the cartridge holder behind the barrel liner. The cartridge holder is inserted with the labeled position number 12 (visible at the top) in line with the center of the Gene Gun (see also Fig 12).

- c. Pull back and hold the cylinder advance lever to retract the inner barrel sleeve into the gun barrel (Figure 13).
- d. Place the empty cartridge holder into the Gene Gun with the position 12 label facing up and the numbered side of the cartridge holder facing the exit nozzle of the barrel (Figure 12).
- e. When the cartridge holder is in its correct position, the knob on the backside of the cartridge holder will slip into the notch on the barrel plate and the cartridge holder will be flush with the barrel plate.
- f. Release the cylinder advance lever; the O-ring on the inner barrel sleeve should hold the cartridge holder in position.
- g. Unlatch the cylinder lock; it should snap into position and the barrel pin should be inserted into the center hole in the cartridge holder.
- h. Push the push bar in; it should snap back into position and engage the cartridge holder in one of the deep crevices (see Figure 14).




Fig. 14. Correct assembly of the cartridge holder in the Gene Gun in preparation for discharge. Releasing the cylinder advance lever moves the barrel liner backward, bringing the O-ring on the back of the liner in contact with the cartridge holder. Releasing the cylinder lock inserts the locking pin into the center hole in the cartridge holder.

- i. Push in and release the cylinder advance lever to ratchet the cartridge holder one position, bringing the first cartridge into firing position. The number 12 should be seen at the top point of the cartridge holder. The Gene Gun is now ready for pressurizing with helium. **Note:** do not wiggle the cartridge holder after the cartridge is in firing position. The Helios Gene Gun has a self-centering mechanism which places the cartridge in the proper position for firing.
2. Set up the Helios Gene Gun and connect to a helium source as described in Section 4.2. Refer to Section 3.2 for identification of controls on the Gene Gun.




Fig. 15. LED display of the Helios Gene Gun. Using a pressurized system, once the cartridge holder of the Gene Gun is correctly inserted and engaged in position 1, the five CHARGING lights will be sequentially illuminated. Once the unit is fully charged (~5 sec), the CHARGED light will flash. If the safety interlock is then pressed, the ARMED lights will alternately flash. If the safety interlock continues to be pressed (in this ARMED state) and the trigger button is pressed, the helium will discharge and the FIRED light will then be illuminated for one second. The process of charging and arming the gun occurs automatically after it is fired. The green battery light in the BATTERY STATUS window indicates that the battery is good.

3. Start the flow of helium as follows: (**Note:** If a continuous gas leak is observed after opening either the cylinder valve or the regulator adjustment screw, close the appropriate valve and check the tightness of the fitting. Contact Bio-Rad if a system leak persists.)
 - a. Close the helium pressure regulator by turning the regulator adjustment screw counterclockwise until the adjusting spring pressure is released and the screw moves without resistance.
 - b. Release helium into the pressure regulator by slowly opening the cylinder valve on the helium tank; open the valve fully. The helium pressure in the tank is indicated on the high pressure gauge (the one closest to the cylinder). Verify that the pressure in the tank is at least as high as the desired discharge pressure.
 - c. Open the helium pressure regulator using the regulator adjustment screw and set the discharge pressure to the desired setting (100–600 psi). The pressure to the system is indicated by the low pressure gauge on the regulator. Do not set the discharge pressure above 600 psi; the helium regulator has an over-pressure relief valve that will vent the system at 700 ± 50 psi. Additionally, the regulator is self-venting; turning the the regulator adjustment screw counterclockwise will reduce pressure to the system.
4. Put on hearing and eye protection, point the device away from any by-standers, and depress the trigger 2–3 times to discharge the device as described below. If this step is performed without a cylinder in place, the O-rings may be blown out.
 - a. For right handed users (see Figure 16):
 - Engage the safety interlock switch by pushing in with the thumb.
 - Push the firing trigger with the index finger to discharge the cartridge.
 - b. For left handed users:
 - Engage the safety interlock switch by pushing it in with the lower part of the index finger.
 - Push the firing trigger with the index finger to discharge the cartridge.

Notes: (1) The firing trigger is functional only while the safety interlock switch is pushed in. The safety interlock switch activates the trigger button for 30 sec; if the Gene Gun is not fired within the 30 sec period, the safety interlock switch must be released, then pressed again to reactivate it. (2) After firing the gun, there is a 5 sec delay before the gun may be fired again. (3) Advance the cartridge holder by squeezing the cylinder advance lever between each shot.



Fig. 16. Position of hand over the safety interlock switch in preparation for firing.

5. Removing the cartridge holder from the Helios Gene Gun is accomplished as follows:
 - a. Move the cylinder lock on the Gene Gun forward and to the right so it is latched and the barrel pin does not protrude behind the barrel (Figure 12).
 - b. Unlatch the push bar by pulling it outward.
 - c. Pull back and hold the cylinder advance lever to retract the inner barrel sleeve into the gun barrel (Figure 13).
 - d. Pull the cartridge holder up and out of the Gene Gun. The gun will turn-off after 3 min.

Loading Cartridges into the Helios Gene Gun

Each cartridge holder has slots for 12 cartridges. Each slot is numbered along the edge of the cartridge holder, as seen when loading. The numbers correspond to the firing order when the cartridge holder is properly positioned in the Gene Gun. The cartridge holder should be loaded with cartridges beginning with position 1, then clockwise through position 12 (see Figure 17). Note, however, that the cartridge that is in firing position is at the bottom of the cartridge holder when it is inserted in the Gene Gun. Therefore, the cartridge in firing position will not be visible to the user but will be indicated by the number at the top of cartridge holder and visible along the cartridge rim.

1. Place the cartridge holder on a flat surface with the numbered edge facing up.
2. Starting with position 1, load up to 12 cartridges into the cartridge holder (Figure 17). We recommend loading only one type of DNA per cylinder to avoid possible confusion.
3. Invert the cartridge holder and push the cartridges against a flat surface so that they are flush with the numbered side of the cylinder.
4. Insert the loaded cartridge holder into the Helios Gene Gun as described above. When the LED on the back of the Gene Gun indicates that the first cartridge is in firing position (Figure 15) the device is ready to deliver DNA.




Fig. 17. Loading cartridges into the cartridge holder. The numbers located on the outer rim indicates sample number when delivered by the Gene Gun.

DNA delivery to target tissue (see Section 6 for suggestions on preparing mammalian target cells)

1. Touch the target area with the spacer so that the spacer is flush and the Gene Gun is perpendicular to the target surface. Activate the safety interlock switch and press the trigger button to deliver the DNA/microcarriers to the target.
2. Ratchet to the next cartridge by pulling in and releasing the cylinder advance lever. After approximately 5 sec, the Gene Gun is ready to deliver the next cartridge.
3. After all tubes have been discharged, remove the cartridge holder as described below.

5.4 Removing Used Cartridges, Depressurization, and Shut Down

Materials

Supplied

Cartridge extractor tool

Removing Cartridges from the Cartridge Holder

1. Hold the cartridge holder containing the cartridges to be extracted with the back side (the side with the knob in the center) facing up.
2. Insert the long prong of the Cartridge Extractor Tool into one of the bores of the cartridge holder; turn the Cartridge Extractor Tool until all of the prongs mesh with the bores in the cartridge holder.
3. Push the prongs of the tool into the cartridge holder until the cartridges are ejected.

Depressurizing and Shutting Down the Helios Gene Gun

1. After discharging the last cartridge, turn off the helium pressure to the system by closing the valve on the helium tank.

2. Turn the regulator valve counterclockwise until both the high and low pressure gauges on the helium regulator register 0 psi. Several increase/decrease adjustments on the regulator may be necessary. Listen for venting to ensure complete depressurization. The system is now depressurized and can be safely disassembled.
3. Disconnect the helium hose from the regulator by pulling the sleeve on the Swagelok Quick-Connect coupling toward the helium hose and pulling the fittings apart.
4. Disconnect the helium hose from the Gene Gun by pulling the sleeve on the Swagelok Quick-Connect coupling toward the Gene Gun and pulling the fitting apart. (**Warning:** For safety, do not leave the Helios Gene Gun unattended while attached to the helium regulator.)

Section 6

Preparation of Mammalian Cell Targets

Time Considerations: Setting up the device requires no more than 5 min, while the delivery process requires approximately 30–60 sec per cell target. *If necessary, sterilize parts as described in Section 10.5 before starting.*

6.1 *In vitro* Delivery to Adherent Cells

Method 1

Day 1

1. Trypsinize cells from flasks. Resuspend the cells in tissue culture media at a density so that inoculating 2 ml into a 35 mm tissue culture plate will produce a monolayer 60–80% confluent after 24 hr.
2. Inoculate 2 ml of cells into 35 mm tissue culture plates.
3. Incubate overnight under the appropriate conditions.

Day 2

1. Prepare the Helios Gene Gun for operation as described in Section 5.3.
2. Immediately prior to DNA delivery, aspirate the media from the dish.
3. Hold the dish perpendicular to the spacer and touch the end of the plastic spacer (if the spacer is sterile), or as close as possible to the target, and discharge the Gene Gun (Figure 18).
4. Add 1.5–2 ml of media to the plate and return to the incubator.

Method 2

1. Trypsinize cells from flasks. Resuspend the cells in tissue culture media buffered with 25 mM HEPES (pH 7.3) and count viable cells.
2. Dilute cells to 2×10^6 cells/ml.
3. Using a sterile pipet tip, spread 100 μ l of the cell suspension in a 1.5 cm diameter circle in the center of a 35 mm dish, being careful not to break the meniscus. It is useful to draw a 1.5 cm diameter circle on an index card to use as a template under the culture dish.
4. Allow the plates to incubate undisturbed for 30 min to permit the cells to attach.
5. Gently add 1 ml of media to each plate and place in the incubator. Incubate the cells for at least 4 hr.
6. Prepare the Helios Gene Gun for operation as described in Section 5.3.

7. Immediately prior to DNA delivery, aspirate the media from the dish.
8. Hold the dish perpendicular to the spacer and touch the end of the plastic spacer (if the spacer is sterile), or as close as possible to the target, and discharge the Gene Gun (Figure 18).
9. Add 1.5–2 ml of media to the plate and return to the incubator.

6.2 *In vitro* Delivery to Suspension Cultures

Method 1

1. Prepare the Helios Gene Gun for operation as described in Section 5.3.
2. Transfer the cells to a centrifuge tube and pellet in a table top centrifuge for 5 min at 250 x g.
3. Resuspend the cells in tissue culture media containing 25 mM HEPES (pH 7.3) at a concentration of 5×10^7 cells/ml.
4. Inoculate 20 μ l (1×10^6 cells) of the cell suspension into a 1.5 cm circle in the center of a 35 mm dish. It is useful to draw a 1.5 cm diameter circle on an index card for use as a template under the culture dish.
5. Hold the dish perpendicular to the spacer and touch the end of the plastic spacer (if the spacer is sterile), or as close as possible to the target, and discharge the Gene Gun (Figure 18).
6. Add 1.5–2 ml of media to the plate and return to the incubator.



Fig. 18. Correct placement of the Gene Gun when transfecting cells *in vitro* (photo courtesy of Auragen, Inc.).

Method 2

1. Prepare 35 mm tissue culture plates with Cell-Tak (Collaborative Research, Cambridge, MA) using manufacturer's instructions.
2. Transfer cells to a centrifuge tube and pellet in a table top centrifuge for 5 min at 250 x g.
3. Resuspend cells in tissue culture media without serum at a concentration such that inoculating 2 ml of cells into the 35 mm tissue culture plates treated with Cell-Tak will produce a monolayer 60–80% confluent.
4. Inoculate 2 ml of cells into 35 mm tissue culture plates.

5. Incubate 30 min to 4 hr under the appropriate conditions.
6. Prepare the Helios Gene Gun for operation as described in Section 5.3.
7. Immediately prior to DNA delivery, aspirate the media from the dish.
8. Hold the dish perpendicular to the spacer and touch the end of the plastic spacer (if the spacer is sterile), or as close as possible to the target, and discharge the Gene Gun (Figure 18).
9. Add 1.5–2 ml of media to the plate and return to the incubator.

6.3 *In vivo* Delivery to Epidermis

Animal Preparation

1. Anesthetize the animal if necessary for safe handling.
2. Clip fur as closely as possible over the desired target area using Oster clippers with a #40 surgical blade and brush or vacuum fur off.
3. (Optional) After clipping, a commercial depilatory such as Nair can be used to completely remove the animal's fur. This treatment removes the stratum corneum from the skin, completely exposing the epidermis. Carefully rinse the skin with warm water following depilatory treatment.
4. If the target site is wet or dirty, clean and dry with 70% ethanol.

Helios Gene Gun Operation

1. Prepare the Helios Gene Gun for operation as described in Section 5.3.
2. Hold the spacer directly against the target site (Figure 19) and discharge the device. (Discomfort following Helios Gene Gun bombardment of the skin is minimal. No apparent macroscopic disruption of the skin, external bleeding or hematoma should be observed. Some animals may show transient erythema or inflammation at the treatment site.)



Fig. 19. *In vivo* bombardment of murine epidermis (photo courtesy of Auragen, Inc.).

Section 7

Optimization of Gene Gun Parameters

7.1 Overview

The flexibility of the particle delivery system allows fine-tuning of experimental parameters; however, it is necessary for each laboratory to determine the optimal parameters for their particular instrument and *in vivo* or cell culture system. Any quantitative assay may be used to determine the optimum combination of critical parameters for the particular biological system under investigation. Important parameters to evaluate include the helium pressure, the PVP concentration, the Microcarrier Loading Quantities (MLQ), and the DNA loading ratio (DLR). It should be noted that absolute transgene expression levels are only a part of the processes leading to immune or other biological responses, thus, each researcher must identify those parameters that result in the appropriate level, location, and duration of transgene expression following particle-mediated delivery. The representative results shown below reveal a commonly observed and useful trait of the Helios Gene Gun: a broad, bell shaped distribution of effective delivery parameters.

Scientists are advised to optimize bombardment parameters for their particular gene gun and biological system. The experimental approach which appears to be most effective in determining the optimum bombardment conditions is the following.

- Expt 1:** Optimize helium pressure: Coat plasmid onto gold particles at 2 µg plasmid/mg gold. Prepare tubes with 0.05 mg/ml PVP and 500 µg gold/shot (1 µg plasmid/shot). Bombard the target tissue at 50, 100, 200 and 400 psi helium to determine the optimum helium pressure.
- Expt 2:** Optimize PVP concentration: Use gold particles coated with 2 µg plasmid/mg gold as in Experiment 1. Prepare tubes with 0, 0.05, and 0.1 mg/ml PVP. Bombard the target tissue using the optimum helium pressure determined in Experiment 1.
- Expt 3:** Optimize MLQ: Coat plasmid onto gold particles at 2 µg, 4 µg, and 8 µg plasmid/mg gold; prepare tubes with each DNA/gold sample at 1 µg plasmid/shot (this is 500 µg gold/shot, 250 µg gold/shot, and 125 µg gold/shot, respectively), and containing the optimum amount of PVP determined in Experiment 2. Bombard the target tissue at the optimum helium pressure determined in Experiment 1 or 2.
- Expt 4:** Optimize DLR: The lowest amount of DNA that has been found to give a detectable level of gene expression is 1 ng of plasmid, but this is dependent on the vector, the target, and the assay system. For most experimental systems, bombardment with 1 µg of plasmid produces near maximal expression. Increasing the amount of plasmid per bombardment does not result in a proportional increase in gene expression. However, bombardment with higher amounts of DNA may be important when using several expression vectors. To verify the optimum amount of plasmid per bombardment precipitate different amounts of plasmid onto gold particles—depending on how extensive a study desired, this may be as simple as a single two-fold dilution to multiple dilutions using between 1 ng and 5 µg of plasmid per bombardment; the other bombardment parameters should be as determined in the previous experiments.

For example, Figure 20 shows a test of discharge pressure on transient gene expression, in mouse skin transfected *in vivo* as described in Section 6.3. Optimum pressure was determined to occur at 200 psi. This pressure resulted in adequate penetration of the particles without excessive tissue damage, and deposition of the majority of the particles in the epithelial layer rather than the relatively acellular underlying dermal tissue.




Fig. 20. Luciferase expression in mouse skin transfected *in vivo* using the Helios Gene Gun. Skin homogenates were assayed 24 hours post-transfection.

As illustrated in Figure 21, the optimum MLQ is determined by adjusting the concentration of the DNA/microcarrier suspension loaded into the Gold-Coat tubing. For suspension cells, higher MLQs (0.5–1.0 mg gold/cartridge) are often needed due to the high cell concentrations in the aqueous cell smear. Adherent cell monolayers and intact tissues may require reduced MLQs (0.06–0.25 mg gold/cartridge) to minimize tissue damage while maximizing transfection.




Fig. 21. Effect of the amount of accelerated gold particles on *in vitro* expression levels. Suspension cultures of CHO cells were bombarded as described in Section 6.2 using the Accell Gene Gun. After 18 hours, the CHO cells that had reattached to the tissue culture plate were lysed and assayed for luciferase expression (Thompson *et al.* 1993).

Another variable parameter, DNA loading rate (DLR), is determined by varying the concentration of DNA precipitated on to the gold particles. Table 3 shows DNA dosage results for transfection of CHO cells using the method described in Section 6.2. Secretion of murine granulocyte macrophage colony stimulating factor (mGM-CSF) was greater at higher DNA dosages. Similar results have been found for several *in vivo* and *in vitro* systems, with maximal expression observed for 1–5 µg DNA/mg gold particles. At dosages above ~5 µg/mg, the DNA and gold can form a single clump which is unsuitable for cartridge preparation.

Table 3. Effect of DNA concentration on mGM-CSF expression in CHO cells^a

DNA loading rate ^b	Expression ^c (ng/10 ⁶ cells)
2.5	515
0.25	239
0.025	25
0.0025	0.9

a CHO cells were transfected as suspension cultures using the Accell Gene Gun.

b Irrelevant plasmid DNA was added to each preparation to bring the total amount of DNA precipitated constant.

c mGM-CSF in the media was assayed 24 hours post-bombardment by ELISA as described by Mahvi *et al.* (1996).

7.2 Parameters for *in vitro* Delivery

The following conditions have been found to be effective for transformation of mammalian cell lines. However, each laboratory should determine the optimum parameters for their particular application. (See previous section.)

Table 4. Suggested starting conditions for *in vitro* transformation of tissue culture cells using the Helios Gene Gun.

Parameter	Conditions
Helium pressure:	50–200 psi
PVP:	0–0.015 mg/ml
MLQ:	0.125–0.5 mg/tube
Microcarriers:	1.0 or 1.6 μ gold
DLR:	0.5–2.5 μ g DNA/mg gold (0.06–1.25 μ g DNA/cartridge)

7.3 Parameters for *in vivo* Delivery

The following are suggested starting conditions for optimizing DNA delivery to the skin of various species. Each laboratory should determine the optimum parameters for their particular application (see Section 7.1 and Table 5). Bombardment conditions are those used in references cited in Section 10.8 using the Accell Gene Gun. Use this information as a starting point when optimizing bombardment conditions with the Helios Gene Gun. Lower helium pressures should be tested early in the optimization process.

Table 5. Examples of parameters for delivery into skin of various species

Species	Site	MLQ(mg/target)	DLR(μ g/mg)	PVP(mg/ml)	psi
Dog ⁹	Dorsum	0.5	1–2	0.1	300
Monkey ^{7, 8, 9}	Abdomen	0.25	2	0–0.05	300–400
Mouse ^{17, 19}	Abdomen*	0.25–0.5	1–2.5	0–0.1	300–500
Pig ^{7, 8, 10}	Inner thigh (clipping not necessary)	0.25–0.5	0.5–2.5	0.05	500
Rabbit ²⁷	Back	0.5	1–2	None	350–400

* Do not use a depilatory agent to remove the fur (Section 6.3, step 3)

Section 8 Troubleshooting

8.1 DNA/Microcarrier Preparation

Problem	Microcarriers agglomerate after coating with DNA.
Possible solutions	Lower DNA Loading Rate.

8.2 Cartridge Preparation

Problem	Gold does not spread evenly in the Gold-Coat tubing (rings, clumps, uncoated sections, streaks).
Possible solutions	Eliminate any potential sources of water in the tube and in the final DNA/Microcarrier preparation.

1. Rewash remaining microcarrier prep 2–3 times using a fresh bottle of 100% ethanol.
2. Flush Gold-Coat tubing with nitrogen gas for 15 min prior to loading with DNA/gold microcarrier solution.
3. If using PVP, make a new stock of PVP in ethanol using a fresh bottle of alcohol.
4. Resuspend gold/DNA pellet prior to first ethanol wash.
5. After drawing off the ethanol, quickly turn the Gold-Coat tubing with your fingers prior to rotating.

8.3 Helios Gene Gun Operation

Problem	LED display not lit.
Possible solutions	Battery missing or discharged; replace battery.
Problem	Device does not fire (helium burst does not occur).
Possible solutions	1. Check that the helium tank and regulator knobs are opened and that there is sufficient helium in the tank. 2. Check that the helium hose is properly connected. 3. Check that the battery is good and that the LED display indicates that a cartridge is in firing position.
Problem	The first cartridge fires well but gold is not delivered in subsequent cartridges.
Possible solutions	Check that the O-rings are in place and are not worn (see Section 10.2). Check that the cartridge holder is properly positioned in the Gene Gun (see Section 5.2).
Problem	The device fires but gold is not delivered.
Possible solutions	1. Check that the O-rings are in place and are not worn. 2. Check that the barrel liner is tight (see Section 10.2). 3. Check that the cartridge holder is properly positioned in the Gene Gun (see Section 5.2). 4. Helium pressure is too low; increase the discharge pressure. 5. Cartridge preparation is bad; make new cartridges. 6. Too much PVP in tubes; increase helium pressure or reduce PVP concentration.
Problem	A discharge of helium occurs from the gun after first turning on the helium regulator and pressurizing the system.
Possible solutions	Open the helium regulator quickly one or two turns and pressurize the system to > 100 psi. The pressure can then be re-adjusted to another setting. If continuous helium discharge is observed after this, contact Bio-Rad.

8.4 *In Vitro* and *in Vivo* Targeting

Problem	Gold does not penetrate skin or organ.
Possible solutions	1. Check that the helium pressure has not dropped below the desired delivery pressure. 2. Increase the PVP concentration in the cartridge preparations. 3. Increase the delivery pressure. 4. Check that the gold has been discharged from the cartridge after firing. 5. Use larger gold particles.
Problem	Poor or no expression.

Possible solutions	1. Make sure DNA is resuspended in TE or distilled water, not saline. 2. Check that DNA has been precipitated onto the microcarriers (see Section 10.7).
Problem	Attached cells are lifting off the center of the plate after delivery of gold particles, or there is a dead zone in the center of the target.
Possible solutions	1. Lower the helium pressure and/or lower the microcarrier loading rate.
Problem	Contamination of the cell culture.
Possible solution	Wipe the end of the barrel with ethanol or dilute bleach.

Section 9

Product Information

9.1 Helios Gene Gun System

The two catalog numbers for the two Helios Gene Gun Systems (100/120V and 220/240V) and component descriptions are as follows:

Catalog Number	Product Description
165-2431	Helios Gene Gun System, 100/120 V includes
165-2411	Helios Gene Gun Kit , that includes Gene Gun Unit, 1
165-2416	O-rings, Helios Gene Gun, 5
165-2417	Barrel Liner, 5
165-2426	Cartridge Holder, white, 5
165-2435	Cartridge Extractor Tool, 1
165-2436	Battery, 9 V, 1
165-2412	Helium Hose Assembly , with Swagelock Quick-Connect fittings, 1
165-2413	Helium Regulator , CGA 580 female fitting, with pressure relief valve. Maximum pressure 2,600 psi., 1
165-2418	Tubing Prep Station , includes Tubing Prep Station, 100/120 V, Tubing Support Cylinder, 1 Power Cord, 1, O-rings, Tubing Prep Unit, 5 (165-2419), Nitrogen Regulator Hose, Nalgene, 12 ft, 3/16" Barb to Male Luer-Lock Fitting, 2, Nitrogen Flow Meter Fitting, 1/8" Barb-to-Male Luer-Lock Fitting, 2, Allen wrench, 5/64", 1, Syringe Holder, 10 cc, 1
165-2421	Syringe Kit , 1, includes: Syringe Adaptor Tubing, Silicone, 5 ft, 0.104" ID x 0.192" OD, 1, Syringe 10 cc, 5, Syringe Adaptor Fitting, 1/8" Barb-to-Female Luer-Lock Fitting, 5
165-2422	Tubing Cutter , Helios Gene Gun, that includes Tubing Cutter Unit and Razor blades, 10 (165-2423)
165-2424	Helios Gene Gun Optimization Kit , that includes
165-2262	0.6 micron gold, 0.25 g
165-2263	1.0 micron gold, 0.25 g
165-2264	1.6 micron gold, 0.25 g
165-2440	Cartridge Kit, Helios Gene Gun, 1, that contains PVP

(polyvinylpyrrolidone, MW 360,000), 0.5 g, Cartridge Collection/Storage Vials, 5, Dessicant Pellets, 5, Gold-Coat Tubing, 50 ft.

165-2432 **Helios Gene Gun System**, 220/240 V, as above for 165-2431, except 165-2420 Tubing Prep Station, 220/240 V, 1

9.2 Spare Parts

Part Number	Product Description
9204987	Door, Battery Compartment, Gene Gun
9204988	Cover, Molded Nose
9204995	External Barrel Handle
9101503	Screw , 6-32 x 1/4, FH
9100453	Thumbscrew , Nylon, 1/4 x 20 x 1"
9100620	Thumbscrew , Nylon, 1/4 x 20 x 1/2"
8004478	Assembly , C-shaft (Tube Support Assembly)
9009712	Fuse , 5 x 20 mm, 0 .125A
9007296	Fuse , 5 x 20 mm, 0.25A
9400359	Sealant , Pipe
9109223	Fitting , Elbow, Female Luer
9301185	Kit , Helium regulator, soft goods (O-rings), version A
165-2436	9 V Battery
165-2423	Razor Blades , 10
165-2416	O-rings , Gene Gun
165-2417	Barrel Liner
165-2426	Cartridge Holder

9.3 Gene Gun Specifications

Physical	
Dimensions	Approximately 8" x 10"
Weight	3.15 lbs
Construction	Super Epoxy or Polycarbonate
Cylinder	Acetal
Barrel Liner	Ryton
O-rings	Viton
Electrical	
Maximum Current	10 mA peak
Voltage Input	9 VDC alkaline battery, replaceable
Battery Life	1,000 discharges in continuous use
Functional	
Input Pressure	600 psi maximum Helium
Regulator Relief Pressure	700 psi ± 35 psi at regulator assembly
Regulator Adjustment	800 psi limit maximum
Discharges	12 per cylinder revolution
Indexing	Manual

Environmental

Operating	50 °F (10 °C) to 90 °F (32 °C) temp. 30–80 % humidity
Storage	32 °F (0 °C) to 140 °F (60 °C) temp. 10–90 % humidity

Tubing Prep Station Specifications

Physical

Dimensions	Approximately 33.5" x 4"
Weight	11.2 lbs
Construction	Aluminum and Acrylic

Electrical

Maximum Current	62 mA/125 mA peak
Voltage Input	220–240 Vac/100–120 Vac
Input Frequency	50/60 Hz

Functional

Input Pressure	30 psi N ₂
Relief Pressure	30 psi ± 1.5 psi at regulator assembly
Regulator Adjustment	TDB psi limit maximum
Speed	30 RPM nominal
Tubing Fill	manual

Environmental

Operating	50 °F (10 °C) to 90 °F (32 °C) temp 30–80 % humidity
Storage	32 °F (0 °C) to 140 °F (60 °C) temp 10–90 % humidity

Section 10

Appendices

10.1 Precipitation of RNA onto microcarriers

Primary RNA transcripts and mRNA may be effectively delivered *in vivo* and *in vitro* using the Helios Gene Gun System (Qiu *et al.*, 1996). The procedure is similar to that for DNA, but the precipitation step is performed with ammonium acetate and 2-propanol.

Materials

In addition to those identified in Section 5.2:

Purified RNA preparation (*e.g.* capped, polyadenylated, or oligo dT-selected mRNA).

5 M ammonium acetate

2-propanol

Experiments involving RNA require careful technique to prevent RNA degradation. In the studies published to date, it was not necessary to pre-treat the gold particles to prevent RNase contamination.

Procedure (sufficient for one length of Gold-Coat tubing):

1. In a 1.5 ml microfuge tube, weigh out 25 mg gold particles.

2. To measured gold, add aqueous solution of mRNA preparation.

The ratios of RNA to gold used are similar to those used for DNA (1–15 µg RNA/mg particles).

3. Add 1/10 volume of 5 M ammonium acetate and 2 volumes of 2-propanol, mix by vortexing. Place the microfuge tube at -20 °C for 1 hour.
4. Proceed with Section 5.2, steps 9–12.

10.2 Replacing the O-rings and Barrel Liner on the Helios Gene Gun




Fig. 22. Location of barrel liner and O-rings in the Gene Gun. The barrel (cover) is removed for visual clarity but is not required for barrel liner removal/insertion.

The Helios Gene Gun has three user-serviceable parts: the barrel liner and two O-rings (Figure 22). The barrel liner screws into the barrel of the Gene Gun. It is a molded plastic piece with the correct shape to permit optimum spread and penetration of the microcarriers at the target site. It may be autoclaved between experiments or replaced.

To remove the barrel liner:

Hold the exposed part of the barrel liner (the spacer) and turn counterclockwise 4–5 turns (see Figure 5).

To insert the barrel liner:

1. Remove the cartridge holder from the Gene Gun.
2. Place the barrel liner into the Gene Gun and turn clockwise until the threads just bind. Do not overtighten.

The O-rings function to keep the helium from expanding around the tube in the firing position in the cartridge holder. One O-ring is inserted at the end of the barrel liner that makes contact with the cartridge holder; the other O-ring surrounds the exit port of the main valve. Both O-rings fit in grooves with center supports.

To replace the O-ring on the end of the barrel liner:

1. Remove the barrel liner.
2. Remove the old O-ring from the barrel liner.
3. Place a new O-ring in the groove and press it in gently. Do not use grease or adhesive.

To replace the O-ring around the exit port of the main valve:

1. Using a pin or syringe needle, poke the O-ring from the edge to pop it out. Be careful not to scratch the surface of the valve.
2. Place a new O-ring in the groove and press it in gently. Do not use grease or adhesive.

10.3 Replacing the O-ring on the Tubing Prep Station

The Tubing Prep Station has two user-serviceable parts: O-rings at the end of the tubing support cylinder nearest the motor. If the O-rings are worn, nitrogen may leak around them and the amount of nitrogen indicated on the flowmeter will not be passing through the tubing. Under normal use, the lifetime of the O-rings should be longer than 1 year. The O-rings may be changed as follows (see Figure 24):

1. Remove the tubing support cylinder by pushing it to the right to contract the spring on the support bar, then raise the left side up and away from the base of the unit.
2. The outer O-ring should be located in the slot at the end of the tubing support cylinder, although it may remain in the support on the motor housing. It may be removed with a fingernail or a blunt object, such as a toothpick. If worn, replace the O-ring with a new one.
3. The inner O-ring is at the end of the tubing support cylinder next to the stem. To replace it, remove the stainless steel stem at the end of the tubing support cylinder by loosening the hex nut on the lower surface of the tubing support cylinder one-eighth turn using a 5/64" Allen wrench. The stem and spur gear can be pulled out of the tubing support cylinder. Remove the O-ring from the slot in the stem with a fingernail or blunt object and replace if worn. Insert the stem into the end of the tubing support cylinder; press gently on the end to hold the O-ring in place and tighten the hex nut. Replace the tubing support cylinder in the base of the Tubing Prep Station.




Fig. 24. Replacing the O-rings on the Tubing Prep Station. The O-rings are located in the tubing support cylinder on the side next to the motor housing. One O-ring (A) fits into a slot on the outside edge of the stem at the end of the tubing support cylinder. The other O-ring (B) is on the inside surface of the stem and can be removed only after loosening the hex nut and removing the stem from the tubing support cylinder.

10.4 Replacing the razor blade on the Tubing Cutter and disassembly of the unit

The cutting edge of the Tubing Cutter is a standard, single edge razor blade. Because cutting is done only on one end of the blade at once, it can be rotated after the first end becomes dull. The first sign of a dull blade will be incomplete cutting of two adjacent cartridges. Front refers to the side of the Tubing Cutter with the tubing channels. Right, left, and back refer to the sides of the Tubing Cutter when the user faces the front of the unit (see Figure 23).

1. Loosen the red locking knob on the back of the Tubing Cutter by turning it approximately one full turn until the lock block on the rear of the cutting arm is free to move.
2. Move the lock block to the right to free the razor blade retaining screw on the right side.
3. Lift up on the right side of the razor blade until it can be lifted free of the Tubing Cutter.
4. Insert a new blade (or rotate the old blade 180°) by inserting the notch on the left side of the blade into the retaining screw on the left side of the cutting arm.
5. Hold the razor blade so that the notch on the blade is next to retaining screw on the right side of the cutting arm. Using the red locking knob, move the retaining screw into the notch in the razor blade and turn the locking knob clockwise to lock the razor blade in place.




Fig. 23. The Tubing Cutter.

The Tubing Cutter may be disassembled for cleaning. For safety, remove the razor blade prior to disassembly.

1. Place the Tubing Cutter in front of you on a solid surface so that you are facing the front of the unit.
2. Hold the base between the thumb and index finger of your right hand. At the same time, push the left side of the Tubing Cutter arm to the right with the index finger of your left hand. This will disengage the arm from the pivot pin on the left side of the base and from the retaining pin on the right side of the base — the tension on the spring will push the arm up.

3. Raise the arm up and away from the base.
4. The Tubing Cutter may be cleaned with soap and water and /or with 70% or 100% ethanol. Do not autoclave the Tubing Cutter.
5. Assembly is the reverse of disassembly. With the spring in place, lay the cutting arm, correctly oriented, on top of the base. With the index and middle fingers of your left hand, press down on the left side of the arm to compress the spring about 3/16". Simultaneously, with the index and middle fingers of your right hand, push the arm to the left to engage the cutting arm under the stainless steel pivot pin on the left side of the base and under the stainless steel retaining pin on the right side of the base.

10.5 Cleaning and Sterilizing the Helios Gene Gun

In situations where sterility is a concern, such as in vitro targeting, the spacer may be brought as close to the target surface as possible without actually touching the surface. Alternatively, by wiping down the end of the barrel liner with 70% alcohol or dilute bleach followed by a rinse with distilled water, the spacer may touch the target surface.

It is possible to sterilize certain parts of the instrument by autoclaving. The following parts may be autoclaved:

- Barrel liner
- Cartridge holder
- Cartridge extractor tool

The following items should never be autoclaved or immersed in liquid:

- Gene Gun
- Tubing Prep Station
- Helium Regulator

These parts, along with the Tubing Cutter, may be cleaned using a cloth and a mild soap solution. Do not autoclave the Tubing Cutter.

Although Gold-Coat tubing may be autoclaved, this is not recommended since it may result in water condensation inside the tubing which will lead to poor quality cartridges.

10.6 Testing Cartridges for Microcarrier Penetration and Density (Optional)

Either of these procedures may be used to roughly determine the quality of the cartridge preparation. These procedures give only qualitative results and are therefore best when they can be compared against cartridges that have been shown to produce positive results in a biological assay. The results are very dependent on the size of the microcarriers and on the amount of PVP used in loading the microcarriers into the Gold-Coat tubing.



Fig. 25. Representative bombardment showing penetration of 1.6 μ gold particles into Parafilm at (left to right) 100, 200, 300, 400, 500, and 600 psi using the Helios Gene Gun.

Discharge into Parafilm

Materials

Parafilm laboratory film (American Can Company)
Glass plate

Procedure

1. Cut a piece of Parafilm from the roll; a 1" length is required for each cartridge to be assayed.
2. Smooth and press the Parafilm laboratory film (waxy surface down) onto the glass plate. Fold the edges around the sides then peel away the protective paper.
3. Connect the helium regulator, helium hose, and Helios Gene Gun as described in Section 4.2.
4. Pressurize the system as described in Section 5.3.
5. Discharge a cartridge at the test psi into a strip of Parafilm laboratory film. At this point, one can roughly determine the quality of the microcarrier preps by examining the pattern of gold particles. When examining the resulting pattern, it should appear circular with a dark center (Figure 25). The discharged cartridges should be completely empty.

Spot darkness	Proportional to particle penetration
Spot width	Proportional to particle spread
Spot shape	Uniformity of shot

Discharge into 3% Water Agar

Materials

Slides, coverslips and mounting solution
Microscope with 10X eyepiece micrometer
3% water agar in 60 mm petri dishes

Procedure

1. Discharge a cartridge at the test psi into 3% water agar. At this point, one can roughly determine the quality of the microcarrier preps by examining the pattern of gold particles. When examining the resulting pattern, it should appear circular with a dark center. The discharged cartridges should be completely empty.

Additionally, one can also examine microcarrier depth penetration into the water agar microscopically. Cut a thin slice approximately 0.5 cm long through the center of the agar target. Mount the slice onto a slide with mounting solution and a cover slip and analyze for microcarrier depth and concentration.

2. A band of microcarriers will be visible in the agar with increasing particle sizes at increased depths. The researcher will be able to determine areas of high, medium, and low microcarrier density in each slice (Figure 26).
3. To determine the depth of penetration, align the 0 of an eyepiece micrometer at the upper surface of the agar and measure the distance to the arbitrary floor of the microcarrier distribution. This value is the microcarrier depth.

The distribution pattern of the microcarriers is bell-curve shaped with respect to depth and area. The highest microcarrier density is concentrated at the center of the shot and at the surface of the target site. For this reason, in order to obtain a truly representative microcarrier distribution cross-section, be careful to examine the center and full length of your target site.




Fig. 26. Representative cross-section showing penetration of gold particles into water agar after discharge with a helium pulse.

10.7 Quantitation of DNA in Cartridges

This procedure can be used for estimating the amount of DNA in cartridges when 0.5 μg or more of DNA has been loaded per tube.

Materials

10 mM Tris, pH 8.0, 1 mM EDTA	1,000 μl micropipettor and tips
Ultrasonic cleaner	UV spectrophotometer
Vortexer	Quartz cuvettes
Microfuge	

Procedure

1. Place five 0.5" cartridges in a microfuge tube and add 500 μl of TE.
2. Sonicate and/or vortex to dislodge gold and solubilize the DNA. Spin for 5 min in a microfuge to pellet gold.
3. On a spectrophotometer determine the A_{260} by reading against a TE blank. If 100% of the DNA were precipitated onto the microcarriers, the A_{260} readings should be as follows:

Plasmid/cartridge	A_{260}
0.5 μg	0.1
1.0	0.2
2.0	0.4

4. DNA sample may also be analyzed by restriction enzyme digestion and analysis by agarose gel electrophoresis.

10.8 References

1. Albertini, M. R., Emler, C. A., Schell, K., Tans, K. J., King, D. M. and Sheeby, M. J., *Cancer Gene Ther.*, **3**, In press (1996).
2. Andree, C., Swain, W. F., Page, C. P., Macklin, M. D., Slama, J., Hatis, D. and Eriksson, E., *Proc. Natl. Acad. Sci. USA*, **91**, 12188-12192 (1994).
3. Armaleo, D., Ye, G. N., Shark, K. B., Sanford, J. C. and Johnston, S. A., *Curr. Genet.* **17**, 97-103 (1990).
4. Boynton, J. E., Gillham, N. W., Harris, E. H., Hosler, P. J., Johnson, A. M., Jones, A. R., Randolph-Anderson, B. L., Robertson, D., Klein, T. M., Shark, K. B. and Sanford, J. C., *Science* **240**, 1534-1538 (1988).
5. Cheng, L., Ziegelhoffer, P. R. and Yang, N. S., *Proc. Natl. Acad. Sci. USA*, **90**, 4455-4459 (1993).
6. Eisenbraun, M. D., Fuller, D. H. and Haynes, J. R., *DNA Cell Biol.*, **12**, 791-797 (1993).
7. Fuller, J. T., Fuller, D. H., McCabe, D., Haynes, J. R., and Widera, G., *Ann. N. Y. Acad. Sci.*, **772**, 282-284 (1995).
8. Fuller, J. T., Fuller, D. H., McCabe, D., Haynes, J. R., and Widera, G., *Vaccines*, **96**, 87-90 (1996).
9. Fynan, E. F., Webster, R. G., Fuller, D. H., Haynes, J. R., Santaro, J. C. and Robinson, H. L., *Proc. Natl. Acad. Sci. USA*, **90**, 11478-11482 (1993).
10. Haynes, J. R., McCabe, D. E., Swain, W. F., Widera, G., Fuller J. T., *J. Biotechnol.*, **44**, 37-42 (1996).
11. Johnston, S. A., Anziano, P. Q., Shark, K., Sanford, J. C. and Butow, R. A., *Science* **240**, 1538-1541 (1988).
12. Keller, E. T., Burkholder, J. K., Shi, F., Pugh, T. D., McCabe, D., Malter, J. S., MacEwen, E. G., Yang, N. S. and Ershler, W. B., *Cancer Gene*, **3**, In press (1996).
13. Klein, T. M., Fromm, M., Weissinger, A., Tomes, D., Schaaf, S., Sletten, M., and Sanford, J. C., *Proc. Natl. Acad. Sci. USA*, **85**, 4305-4309 (1988).
14. Klein, T. M., Wolf, E. D., Wu, R., Sanford, J. C., *Nature* **327**, 70-73 (1987).
15. Mahvi, D. M., Burkholder, J. K., Turner, J., Culp, J., Malter, J. S., Sondel, P. M., and Yang, N. S., *Human Gene Ther.*, In press (1996).
16. McCabe, D. E., Swain, W. F., Martinell, B. J. and Christou, P., *Bio/Technology*, **6**, 923-926 (1988).
17. Pertmer, T. M., Eisenbraun, M. D., McCabe, D., Prayaga, S. K., Fuller, D. H., and Haynes, J. R., *Vaccine*, **13**, 1427-1430 (1995).
18. Qiu, P., Ziegelhoffer, P., Sun, J. and Yang, N. S., *Gene Therapy*, **3**, 262-268 (1996).
19. Rakhmilevish, A. L., Turner, J., Ford, J. M., McCabe, D., Sun, W. H., Sondel, P. M., Grote, K. and Yang, N. S., *Proc. Natl. Acad. Sci. USA*, **93**, In press (1996).
20. Reba, I., *Scientific American*, **214**, 84-92 (1966).
21. Shark, K. B., Harpending, P. R., Rasmussen, J. L. and Sanford, J. C., *Appl. Environ. Microbiol.*, **57**, 480-485 (1991).
22. Smith, F. D., Harpending, P. R., and Sanford, J. C., *J. Gen. Microbiol.*, **138**, 239-248 (1992).
23. Sun, W. H., Burkholder, J. K., Sun, J., Culp, J., Turner, Joel, Lu, X. G., Pugh, T. D., Ershler, W. B. and Yang, N. S., *Proc. Natl. Acad. Sci. USA*, **92**, 2889-2893 (1995).
24. Sundaran, P., Xiao, W. and Brandsma, J. L., *Nuc. Acids Res.*, **24**, 1375-1377 (1996).
25. Tang, D. C., DeVit, M., and Johnston, S. A., *Nature* **356**, 152-154 (1992).
26. Williams, R. S., Johnston, S. A., Riedy, M., DeVit, M. J., McElligott, S. G. and Sanford, J. C., *Proc. Natl. Acad. Sci. USA*, **88**, 2726-2730 (1991).
27. Xiao, W. and Brandsma, J.L., *Nuc. Acids Res.*, **24**, 2620-2622 (1996).
28. Yang, N. S., Burkholder, J., Roberts, B., Martinell, B. and McCabe, D., *Proc. Natl. Acad. Sci. USA*, **87**, 9568-9572 (1990).
29. Zelenin, A. V., Titomirov A. V. and Kolesnikov, V. A., *FEBS Lett.*, **244**, 65-67 (1989).

The instruments used in these experiments vary. Use the above information as a starting point when optimizing bombardment conditions with the Helios Gene Gun. Lower helium pressures should be tested early in the optimization process.

10.9 Quick Guide to Operation

Before the Bombardment

1. Coat microcarriers with DNA, load into tubes, and prepare cartridges prior to day of experiment.
2. Check helium supply (50 psi in excess of desired rupture pressure).
3. Clean and/or sterilize the Gene Gun, tube holders, and barrel liner as appropriate.
4. Connect the Gene Gun to a helium source.
5. Activate the Gene Gun: turn on the flow of helium to the desired pressure and with an empty cartridge holder in place, make 2–3 “pre-shots” by engaging the safety interlock and firing the trigger.

Firing the Device

1. Load cartridges into the cartridge holder and place in Gene Gun.
2. Prepare/position target cells for bombardment.
3. Bombard sample: engage safety interlock and press the firing trigger

After the Bombardment

1. Remove cartridge holder from Gene Gun.
2. Remove cartridges from cartridge holder.
3. Turn off the helium pressure to the system.
4. Turn the regulator value counterclockwise to de-pressurize the system.
5. Disconnect the helium hose and Gene Gun.



Bio-Rad
Laboratories

*Life Science
Group*

Website www.bio-rad.com Bio-Rad Laboratories Main Office 2000 Alfred Nobel Drive, Hercules, CA 94547, Ph. (510) 741-1000, Fx. (510)741-5800
Also in: Australia Ph. 02 9914 2800, Fx. 02 9914 2889 Austria Ph. (01) 877 89 01, Fx. (01) 876 56 29 Belgium Ph. 09-385 55 11, Fx. 09-385 65 54
Canada Ph. (905) 712-2771, Fx. (905) 712-2990 China Ph. 86-10-62051850/51, Fx. 86-10-62051876 Denmark Ph. 45 39 17 99 47, Fx. 45 39 27 16 98
Finland Ph. 358 (0)9 804 2200, Fx. 358 (0)9 804 1100 France Ph. 01 43 90 46 90, Fx. 01 46 71 24 67 Germany Ph. 089 318 84-0, Fx. 089 318 84-100
Hong Kong Ph. 852-2789-3300, Fx. 852-2789-1257 India Ph. (91-11) 461-0103, Fx. (91-11) 461-0765 Israel Ph. 03 951 4127, Fx. 03 951 4129
Italy Ph. 39-02-216091, Fx.39-02-21609-399 Japan Ph. 03-5811-6270, Fx. 03-5811-6272 Korea Ph. 82-2-3473-4460, Fx. 82-2-3472-7003
Latin America Ph. 305-894-5950, Fx. 305-894-5960 Mexico Ph. 514-2210, Fx. 514-2209 The Netherlands Ph. 0318-540666, Fx. 0318-542216
New Zealand Ph. 64-9-4152280, Fx. 64-9-4152284 Norway Ph. 22-74-18-70, Fx. 22-74-18-71 Russia Ph. 7 095 979 98 00, Fx. 7 095 979 98 56
Singapore Ph. 65-2729877, Fx. 65-2734835 Spain Ph. 34-91-661-7085, Fx. 34-91-661-9698 Sweden Ph. 46 (0)8-55 51 27 00, Fx. 46 (0)8-55 51 27 80
Switzerland Ph. 01-809 55 55, Fx. 01-809 55 00 United Kingdom Ph. 0800-181134, Fx. 01442-259118