

Quick-Start Protocol

SARS-CoV-2 N1+N2 Assay Kits

The SARS-CoV-2 N1+N2 Assay Kits (cat. nos. 222015 and 222017) are intended for molecular biology application for epidemiological research using RT-PCR.

Further information

- Safety Data Sheets: www.qiagen.com/safety
- Technical assistance: [support.qiagen.com](mailto:support@qiagen.com)

Each tube contains primer probes for detection of SARS-CoV-2 based on the CDC design (<https://www.cdc.gov/coronavirus/2019-ncov/lab/rt-pcr-panel-primer-probes.html>), targeting the genes N1 and N2 of the viral genome. The two probes are coupled with FAM™ as the reporter dye and use ZEN™ quenchers (see the table in the next page).

Each tube from the SARS-CoV-2 N1+N2 Assay Kits contains a mixture of four primers and two probes purified by HPLC at a 20x concentration. Therefore, only 1 µl of the assay is needed per 20 µl reaction. Each tube contains 600 µl for 600 reactions.

	Concentration in 20x PPM [μ M]	Final concentration per reaction [μ M]	Sequence
2019-nCoV_N1-F	16	0.8	GAC CCC AAA ATC AGC GAA AT
2019-nCoV_N1-R	16	0.8	TCT GGT TAC TGC CAG TTG AAT CTG
2019-nCoV_N2-F	16	0.8	TTA CAA ACA TTG GCC GCA AA
2019-nCoV_N2-R	16	0.8	GCG CGA CAT TCC GAA GAA
2019-nCoV_N1-P	5	0.25	FAM-ACC CCG CAT /ZEN/ TAC GTT TGG TGG ACC-3IABkFQ
2019-nCoV_N2-P	5	0.25	FAM-ACA ATT TGC /ZEN/ CCC CAG CGC TTC AG-3IABkFQ

Note: Always follow the reaction setup and cycling condition by the suppliers of your RT-PCR chemistries and PCR cyclers to ensure best performance.

Document Revision History

Date	Changes
10/2020	Initial release

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual.

Trademarks: QIAGEN®, Sample to Insight® (QIAGEN Group); ZENTM (Integrated DNA Technologies, Inc.); FAM™ (Thermo Fisher Scientific or its subsidiaries). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.

1122887 10/2020 HB-2829-001 © 2020 QIAGEN, all rights reserved.