

## PRODUCT CERTIFICATION AND DECLARATION OF CONFORMITY

## FT-IR Spectrometer, Model Spectrum Two N

This PerkinElmer product conforms to the regulations stipulated in the CE Mark requirements for the EMC Directive (2014/30/EU), the Low Voltage Directive (2014/35/EU), and the RoHS 2 Directive (2011/65/EU as amended by (EU) 2015/863), and the corresponding UK Statutory Instruments (SI) for the UKCA Mark, the Electromagnetic Compatibility Regulations 2016 (SI 2016 No. 1091), the Electrical Equipment (Safety) Regulations 2016 (SI 2106 No. 1101), and the Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 (SI 2012 No. 3032):

EN 55011:2009 + A1:2010, Group 1, Class A, EMC -- RF Characteristics of ISM Equipment

EN 61326-1:2013, EMC -- Requirements for Electrical Equipment for Laboratory Use

EN 61000-4-2:2009, EMC -- Electrostatic Discharge Requirements

EN 61000-4-3:2006 + A1:2008 + A2:2010, EMC -- Radiated Electromagnetic Field Requirements

EN 61000-4-4:2004 + A1:2010, EMC -- Electrical Fast Transient/Burst Requirements

EN 61000-4-5:2006, EMC -- Surge Immunity Requirements

EN 61000-4-6:2009, EMC -- Conducted Disturbances (induced by RF fields) Requirements

EN 61000-4-11:2004, EMC -- Voltage Dips, Short Interruptions, Voltage Variations Requirements

EN 61000-3-2:2014 -- Harmonic Current Emissions

EN 61000-3-3:2013, EMC -- Voltage Fluctuations and Flicker

EN 61010-1:2010, Safety Requirements for Electrical Equipment for Laboratory Use

EN 61010-2-010:2014, Safety Requirements for Laboratory Equipment for the Heating of Materials

EN 61010-2-081:2015, Particular Requirements for Automatic & Semi-automatic Laboratory Equipment for Analysis & Other Purpose

EN 60825-1:2014, Safety of laser products - Part 1: Equipment classification and requirements

EN 63000:2018, Technical Documentation for the assessment of electrical and electronic products with respect to the RoHS

CAN/CSA C22.2 No. 61010-1-12, Safety Requirements for Electrical Equipment for Laboratory Use

UL 61010-1, 3rd edition, Safety Requirements for Electrical Equipment for Laboratory Use

CAN/CSA C22.2 No. 61010-2-010:15, Safety Requirements for Laboratory Equipment for the Heating of Materials

UL 61010-2-010, 3<sup>rd</sup> edition, Safety Requirements for Laboratory Equipment for the Heating of Materials

CAN/CSA C22.2 No. 61010-2-081:15, Particular Requirements for Automatic & Semi-automatic Laboratory Equipment for Analysis & Other Purposes

UL 61010-2-081, 2<sup>nd</sup> edition, Particular Requirements for Automatic and Semi-automatic Laboratory Equipment for Analysis and Other Purposes ICES-003, Class A, Radiated and Conducted Emissions

FCC Part 15, Class A, Radiated and Conducted Emissions

FDA CDRH 21 CFR 1000-1050 Laser Products

AS/NZS CISPR 11:2011

Korean Radio Waves Act, Article 58-2, Clause 3

This product also conforms to the Waste Electrical and Electronic Equipment Directive (WEEE, 2012/19/EU), and the Regulation on Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH, EC 1907/2006), and the corresponding UK Statutory Instruments (SI), the WEEE Regulations 2013 (SI 2013 No. 3113), and the REACH Enforcement Regulations 2008 (SI 2008 No. 2852).

This declaration of conformity is issued under the sole responsibility of PerkinElmer.

Signed for and on behalf of:

Alan Mears

Compliance Engineer

28 June 2021

2021

09932088C.doc PerkinElmer, 710 Bridgeport Avenue, Shelton, CT 06484 USA

[in the EU: DANI ANALITICA S.R.L., Part of PerkinElmer Group, Viale Brianza, 87 20093 Cologno Monzese MI - ITALY] [in the UK: PerkinElmer, Chalfont Road, Buckinghamshire, Seer Green, HP9 2FX, United Kingdom]

An ISO 9001 Company