



SYSTEM IMPLANT MANUAL

Inspire IV Implantable Pulse Generator Model 3028
Stimulation Lead Model 4063
Respiratory Sensing Lead Model 4340

Rx Only

Table of Contents

Explanation of Symbols on Product or Package Labeling v

Indications for Use 1

MRI Conditions for Use 1

Therapy Overview 1

Overview of the Manual 2

Sales Package Contents 2

Implanted Component Descriptions 3

Generator 3

Leads 4

Contraindications 5

Adverse Effects 5

Warnings and Precautions 6

Warnings 6

Precautions 7

Storage and Handling 9

Generator 9

Leads 10

Physician Training 11

System Implant 11

Implantable Components 11

Procedure Overview 12

Patient Preparation 12

Surgical Materials 12

Precautions for Handling Components 12

Stimulation Lead Implant 14

Securing the Stimulation Lead 16

Forming the Generator Pocket 18

Tunneling the Lead 18

Respiratory Sensing Lead Implant 20

Connecting the Leads and Generator 22

Implanting the generator 26

System Test 27

Completing the Implant Procedure 27

Postoperative Follow-up 28

Physician Instructions to Patient 29

Patient Registration 29

Therapy Activation 29

Therapy Titration 29

Surgical Revision and Explant 30

- Lead Repositioning 30
- Generator Replacement 31
- System or Generator Explant 32
- Explant Disposition 32

Clinical Summary 33

- Stimulation Therapy for Apnea Reduction (STAR) Clinical Trial 33
- Patients Studied 33
- Study Design and Methods 34
- Study Results 35

IPG Specifications 47

- Factory Settings 47
- Configurable Settings 48
- Battery Longevity 49
- Physical Description 49

Inspire Medical Systems Limited Warranty 50

Explanation of Symbols on Product or Package Labeling

Refer to the appropriate product for symbols that apply.



Open here



Do not reuse



Do not re-sterilize



Sterilized using ethylene-oxide gas



Use by



Serial number



Temperature limitation

SENSE

Lead that inserts into SENSE (sensing) port of IPG

STIM

Lead that inserts into STIM (stimulation) port of IPG



Caution, consult accompanying documents



Consult electronic instructions for use



Date of manufacture



Manufacturer



Reference number



The Inspire therapy system is Magnetic resonance (MR) conditional

The following is a trademark of Inspire Medical Systems, Inc.: Inspire®

This product and/or the use of this product in a method may be covered by one or more patents or patent applications, available at www.inspiresleep.com/patents.

Indications for Use

Inspire Upper Airway Stimulation (UAS) is used to treat a subset of patients with moderate to severe obstructive sleep apnea (OSA) (apnea-hypopnea index [AHI] of greater than or equal to 15 and less than or equal to 65). Inspire UAS is used in adult patients 22 years of age and older who have been confirmed to fail or cannot tolerate positive airway pressure (PAP) treatments (such as continuous positive airway pressure [CPAP] or bi-level positive airway pressure [BPAP] machines) and who do not have a complete concentric collapse at the soft palate level.

PAP failure is defined as an inability to eliminate OSA (AHI of greater than 15 despite PAP usage), and PAP intolerance is defined as:

- (1) Inability to use PAP (greater than 5 nights per week of usage; usage defined as greater than 4 hours of use per night), or
- (2) Unwillingness to use PAP (for example, a patient returns the PAP system after attempting to use it).

Inspire UAS is also indicated for OSA patients ages 18 to 21 years with moderate to severe OSA ($15 \leq \text{AHI} \leq 65$) who:

- Do not have complete concentric collapse at the soft palate level,
- Are contraindicated for, or not effectively treated by, adenotonsillectomy,
- Have been confirmed to fail, or cannot tolerate, PAP therapy despite attempts to improve compliance, or
- Have followed standard of care in considering all other alternative/adjunct therapies.

MRI Conditions for Use

The Inspire IV MR Conditional system consists of the Inspire IV model 3028 implantable pulse generator (IPG), the Inspire stimulation lead model 4063 and the Inspire respiratory sensing lead model 4323 or 4340. If certain criteria are met and the warnings and precautions provided by Inspire are followed, patients with an MR Conditional system are able to undergo an MRI scan. For details, refer to the “MRI Guidelines for Inspire UAS Therapy” manual at manuals.inspiresleep.com.

Therapy Overview

The implanted components of the Inspire therapy system consist of the Inspire IPG model 3028, the stimulation lead model 4063, and the respiratory sensing lead model 4340 (Figure 1).




Figure 1. Inspire system implanted components

When therapy is on, the Inspire system detects the patient's respiratory effort and maintains airway patency with mild stimulation of the hypoglossal nerve.

Therapy settings are stored in the generator and configured by the physician using an external programmer.

The patient uses their Inspire sleep remote to turn therapy on before they go to sleep and to turn therapy off when they wake up. The sleep remote also provides the ability to pause therapy and adjust stimulation amplitude within physician-defined limits.

Overview of the Manual

This manual provides physicians with implant procedure and follow-up care information for the Inspire system. The manual includes instructions for handling, storing, and implanting the leads and the generator. Critical therapy information is provided for you to discuss with your patient, as well as instructions for follow-up care. The generator and leads cannot be resterilized. Information on explanting the generator and leads is included. This manual also explains how to register your patient's medical devices.

Sales Package Contents

The leads and generator are provided in separate sterile packages.

Inspire Implantable Pulse Generator (Model 3028)

- Sterile Package Contents
 - One generator
 - One torque limiting hex wrench
- Product Literature (patient manual, patient registration form, patient ID card, and electronic labeling insert)

NOTE: The system implant manual is provided at the following website:
manuals.inspiresleep.com

Implanted Component Descriptions

The implanted components of the Inspire system consist of a generator, a respiratory sensing lead, and a stimulation lead. All implanted Inspire system components are intended for single-use only.

Generator

The generator (Figure 2) contains the battery and electronics that deliver Inspire therapy and store the therapy settings.



Figure 2. Generator

The generator has two 3.2 mm low-profile connector ports (Figure 3), which are compatible with the connectors on the stimulation lead and the respiratory sensing lead. After inserting the lead connectors into the generator connector ports, the lead connectors are secured using the set screws next to the connector ports.




Figure 3. Generator connector ports

Leads

The respiratory sensing lead (Figure 4) detects respiratory effort. The lead has a pressure-sensitive membrane that converts the mechanical energy of respiration into an electrical signal.




Figure 4. Respiratory sensing lead

The stimulation lead (Figure 5) delivers stimulation to the hypoglossal nerve. The lead has a flexible, self-sizing stimulation cuff. The stimulating electrodes are on the inner surface of the cuff.




Figure 5. Stimulation lead

Contraindications

Contraindications for the use of Inspire UAS therapy include the following:

- Central + mixed apneas > 25% of the total apnea–hypopnea index (AHI)
- Any anatomical finding that would compromise the performance of upper airway stimulation, such as the presence of complete concentric collapse of the soft palate
- Any condition or procedure that has compromised neurological control of the upper airway
- Patients unable to or do not have the necessary assistance to operate the sleep remote
- Patients who are pregnant or plan to become pregnant. UAS therapy has not been evaluated for safety or efficacy during pregnancy.
- Patients with an implantable device that may be susceptible to unintended interaction with the Inspire system. Consult the device manufacturer to assess the possibility of interaction.
- Patients who require magnetic resonance imaging (MRI) other than what is specified in the MR Conditional labeling

Adverse Effects

Possible adverse effects include, but are not limited to, the following patient-related conditions:

- Damage to blood vessels in the vicinity of implant
- Excessive bleeding
- Nerve trauma or damage
- Allergic and/or rejection response to the implanted materials
- Infection
- Local irritation, seroma, hematoma, erosion, or swelling
- Persistent pain, numbness, or inflammation at the implant site
- Discomfort from the stimulation
- Tongue movement restrictions, irritation resulting from tongue abrasions on preexisting sharp or broken teeth
- Tongue soreness or weakness
- Problems with swallowing or speaking
- Tongue paresis and atrophy
- Undesirable change in stimulation over time, possibly related to tissue changes around the electrode(s), shifts in electrode position, loose electrical connections, or lead fractures
- Fibrosis to the extent that it makes it difficult to remove the system without damaging surrounding structures
- Dry mouth
- Other acute symptoms (i.e., headaches, coughing, choking, dysphasia, and speech-related events)
- Insomnia
- Pneumothorax

Warnings and Precautions

Warnings

- **Training** — Physicians must be trained in the proper use and surgical procedure before implantation or operation of the device.
- **Pediatrics** — The safety of upper airway stimulation has not been evaluated in clinical studies for patients less than 18 years of age. There may be increased risk of nerve injury and stimulation-related adverse events in this population, particularly in younger children (e.g., less than 12 years of age).
- **Components** — The use of components not provided by Inspire Medical Systems may result in damaged components, improper operation, or increased risks to the patient.
- **Diathermy** — Do not use shortwave diathermy, microwave diathermy or therapeutic ultrasound diathermy (all now referred to as diathermy) on patients implanted with the Inspire system. Energy from diathermy can be transferred through the implanted system and can cause tissue damage at the location of the implanted electrodes, resulting in severe injury or death.

Diathermy can also damage the implanted system components, resulting in loss of therapy and requiring additional surgery for system explantation and replacement. Advise your patient to inform all their healthcare professionals that they should not be exposed to diathermy treatment.

Injury to the patient or damage to the device can occur during diathermy treatment when:

- The generator is turned on or off
- Diathermy is used anywhere on the body—not just at the location of the implanted Inspire system
- Diathermy delivers heat or no heat
- Any component of the Inspire system (leads or generator) remains in the body
- **Magnetic Resonance Imaging (MRI)** — An MRI is a type of medical imaging that uses magnetic fields to create an internal view of the body. If certain criteria are met and the warnings and precautions provided by Inspire are followed, patients with an MR Conditional device are able to undergo an MRI scan. For details, refer to the “MRI Guidelines for Inspire UAS Therapy” manual at manuals.inspiresleep.com.

Do not bring the sleep remote into the MRI environment. Bringing the remote into the MRI scanner room could cause damage to the remote and make it unable to function.

The patient is only eligible for certain MRI scans. If the precautions provided by Inspire are not followed, exposure to MRI can damage your stimulator or leads, cause serious injury, or result in unintended stimulation.

- **Sleep remote use** — When operating their Inspire sleep remote, patients should use special care near flammable or explosive atmospheres. The consequences of using the battery-powered sleep remote near flammable or explosive atmospheres are unknown.
- **Body Mass Index (BMI)** — BMI greater than 32 was not studied as part of the pivotal trial. Based on data from the feasibility study, it may be associated with decreased likelihood of response to treatment. Use of Inspire UAS in higher BMI patients is not recommended due to unknown effectiveness and safety.

Precautions

General

- **Expiration date** — Do not use any Inspire system product after its expiration date.
- **Component handling** — Precautions related to component handling during the implant procedure are located on page 12.
- **Storage temperature ranges**
 - Do not expose the generator to temperatures above 58°C (136°F) or below -35°C (-31°F).
 - Do not expose the respiratory sensing lead to temperatures above 70°C (158°F) or below -35°C (-31°F).
 - Do not expose the stimulation lead to temperatures above 55°C (131°F) or below -10°C (14°F).

Electromagnetic compatibility and medical procedures

For information on MRI, reference “Warnings” on page 6 and the “MRI Guidelines for Inspire UAS Therapy” manual at manuals.inspiresleep.com

For information on diathermy, see “Warnings” on page 6.

The generator is designed to ensure immunity from most common sources of electromagnetic disturbance. In most cases, turning off the electromagnetic disturbance source, or moving away from the electromagnetic disturbance source, will return the generator to normal operation. Extremely strong sources of electromagnetic disturbance could interfere with normal generator operation, causing the generator to reset and requiring the generator to be reconfigured. To reduce the possibility of electromagnetic interference (EMI), patients are recommended to use therapy only while asleep.

Medical environment

Electrocautery, irradiation, lithotripsy, RF-ablation, X-ray, and fluoroscopy are typical electromagnetic disturbance sources in hospital and clinical environments. Medical treatments that use ultrasonics, defibrillation, or radiation can adversely affect the Inspire system.

- **Electrocautery** — Electrocautery tools used near or in contact with the stimulator or leads can cause tissue damage, uncomfortable stimulation, or damage to the generator. Bipolar electrocautery should be used if alternatives are not available. Unipolar electrocautery can be transmitted along the lead body and could cause nerve damage. If electrocautery must be used in the vicinity of the generator, therapy should be turned off.
- **Radiation therapy** — The generator should not be directly irradiated by therapeutic levels of ionizing radiation (such as produced by cobalt machines or linear accelerators used for cancer treatment) because of the risk of permanent damage to the generator circuitry. If such therapy is required in the vicinity of the generator, shield the device and confirm its function after treatment.
- **RF-ablation** — RF-ablation should not be used directly over the implant sites.
- **X-ray and fluoroscopy** — Exposure to diagnostic X-ray or fluoroscopic radiation should not affect the generator or leads.
- **Therapeutic ultrasound** — Exposure to high ultrasonic frequencies may result in damage to the generator or leads. It is not recommended to use high-output ultrasonic

devices, such as an electrohydraulic lithotripter or bone growth stimulator on patients with an implanted generator.

- **Ultrasonic scanning** — While there is no danger to the patient, ultrasonic scanning equipment could cause mechanical damage to the generator or leads if used directly over the implant sites.
- **Defibrillation/Cardioversion** —
 - Utilize biphasic waveforms and minimize the energy delivered.
 - Position paddles as far as possible from implanted Inspire components.
 - Following defibrillation, confirm normal system operations.

Home or work environment

Based on laboratory tests of the generator, the device should not be affected by the normal operation of electrical equipment, household appliances, electric machine shop tools, microwave ovens, internal combustion engines, low-powered radio, or microwave frequency transmitters. All such equipment should be kept in good repair and properly grounded to avoid the possibility of electrical shock or interference with the proper operation of the generator.

Inspire therapy is intended for use during sleep only and should be turned off otherwise.

- **Equipment operation** — Patients should not operate potentially dangerous equipment, such as power tools, while therapy is on.
- **Theft detectors** — Theft detectors have been known to cause inadvertent and potentially uncomfortable stimulation in implanted stimulation systems. Patients should use care to avoid prolonged exposure to theft detectors and be aware in the presence of such systems.
- **High-powered electric fields** — Consult Inspire Medical Systems when the patient will be in an area where contact with current carrying conductors is possible or near high-powered electromagnetic fields radiated by arc welding units, induction furnaces, induction stoves, resistance welders, radio or microwave frequency transmitters, etc.
- **Mobile and cellular phones** — Maintain a separation of at least 15 cm (6 in) between a phone and the generator.

Pediatric use

For pediatric patients aged 18-21, OSA may possibly resolve without intervention. Therefore, a decision to surgically implant a device in this population should be discussed with the treating sleep physician. Even though the device may be explanted, when absolutely necessary, at times the cuff around the hypoglossal nerve may be left implanted permanently. While there is no definite way to identify who will remit in this population, please consider factors which may influence persistence OSA (e.g. comorbidities, ethnicity, obesity, and male gender.) Please consider following alternative/adjunct therapies prior to Inspire UAS therapy:

- Adenotonsillectomy in appropriate patients
- Intensive behavioral therapy and/or desensitization therapy to improve PAP adherence
- Medical pharmaceutical interventions
- Other less invasive interventions, e.g. lifestyle changes, positional therapy, oral appliances, and nasal devices.

In younger patients undergoing this surgical implantation, please take into consideration the need for a lifetime of serial surgical reimplantation for battery replacements, which may potentially be associated with future surgical complications.

Storage and Handling

Recommendations for storage and handling of the generator and leads are provided in this section. Inspire Medical Systems sterilizes the generator and leads with ethylene oxide (EtO) prior to shipment.

Information about precautions for handling components is located on page 12.

Generator

Inspect the generator sterile package prior to opening. If the generator package is damaged, the generator may be damaged as well. Return a damaged package to Inspire Medical Systems; see the back cover of this manual for addresses.

The generator box includes a sterilization indicator. This indicator is green after the device has been sterilized. **Do not use the generator if the indicator is red.**

Handling and Storage: Acceptable

Store and transport generator within the following environmental temperature limits: -35°C (-31°F) to +58°C (+136°F). A full or partial electrical reset condition may occur at temperatures below -18°C (0°F).

Unacceptable

Do not implant the generator if it has been dropped on a hard surface from a height of 30 cm (12 in) or greater.

Resterilization

The generator cannot be resterilized.

- Generators cannot be resterilized. If the sterile package seal is broken, or if the packages are otherwise damaged, do not use.
 - Return the package to your local Inspire Medical Systems representative, see back cover for address.
-

Leads

If the lead sterile package seal is broken or the package is otherwise damaged, return the package to Inspire Medical Systems. Leads cannot be resterilized.

Handling and Storage: Acceptable

Store and transport the respiratory sensing lead within the following environmental temperature limits: -35°C (-31°F) to +70°C (+158°F).

Store and transport the stimulation lead within the following environmental temperature limits: -10°C (14°F) to +55°C (+131°F).

Only use sterile-gloved hands to handle the lead; rinse sterile surgical gloves in sterile water before handling the lead.

Protect leads from materials that shed lint and dust.

Exercise care and appropriate instrument selection when handling the stimulation lead cuff with a surgical instrument.

Unacceptable

Do not implant a lead that was dropped.

Avoid excessive traction or sharp instruments.

Avoid severe bending, kinking, stretching, or handling with surgical instruments.

Do not immerse a lead in mineral oil or silicone oil.

Resterilization

Leads cannot be resterilized.

- If the sterile package seal is broken, or if the packages are otherwise damaged, do not use.
 - Return the package to Inspire Medical Systems; see the back cover of this manual for addresses.
-

Physician Training

Prior to implanting an Inspire system, surgeons will receive classroom instruction on Inspire implant techniques as well as cadaver training. Sleep physicians and sleep technicians will receive classroom instruction on how to titrate the device including hands-on operation of the programmer.

System Implant

This section describes a general implant procedure for the Inspire system.

Implantable Components

The Inspire system includes the following implantable components:

- Inspire implantable pulse generator (Model 3028)
- Inspire respiratory sensing lead (Model 4340)
- Inspire stimulation lead (Model 4063)

The generator has two lead connector ports (Figure 6). The connector port for the respiratory sensing lead is marked **SENSE**. The connector port for the stimulation lead is marked **STIM**.




Figure 6. Generator and connector ports

Procedure Overview

The implant procedure begins with preoperative planning. It is recommended that the stimulation lead be the first Inspire component to be implanted. Secondly, a subcutaneous pocket is created for the generator. The connector end of the leads will be tunneled to this pocket. After the stimulation lead is implanted, the respiratory sensing lead is implanted. After tunneling the connector end of the leads to the generator pocket, the leads are connected to the generator and the generator is secured in the subcutaneous pocket.

Patient Preparation

- Ensure the tongue is visible during the surgical procedure in order to observe the response to intraoperative test stimulation.
- The recommended body side for system implantation is the right side.
- Elevate the patient's right thorax and loosely tuck the patient's right arm to allow suitable approach for sensor implantation. If other active implanted devices are present, plan incision and tunneling locations to keep the Inspire system at least 15 cm (6 inches) away from the other devices.
- The patient's head and neck should be positioned to provide optimal access to the hypoglossal nerve.
- Iodine-impregnated adhesive drape over entire surgical area is recommended.
- Use only short-acting paralytic agent to preserve tongue response.
- A nerve monitoring system is recommended to locate the hypoglossal nerve and confirm nerve recruitment.
- The patient should be given antibiotics preoperatively and may be given antibiotics postoperatively, at the surgeon's discretion.

Surgical Materials

An Inspire system implant requires typical surgical equipment used during neck surgeries. The following is a list of additional materials typically used during the system implant procedure:

- Sterile sleeve, bag or equivalent (to bring the telemetry cable into the sterile field)
- Finer right angled forceps or hemostat (for cuff electrode placement)
- Narrow malleable ribbon retractor to form intercostal passage for sensing lead
- A nerve monitoring and stimulation system (to locate the hypoglossal nerve and confirm nerve recruitment)

Precautions for Handling Components

- The implanted components of this system should be carefully handled to avoid damage by excessive traction or sharp instruments. Any component showing signs of damage should not be used.

**Caution:**

- No instrument of any type should touch the silver sensor past the blue silicone boot. Touching the silver titanium components may result in damage to the sensor or damage to the electrical insulation.
- When handling the sensor, grip the device on the blue silicone boot, blue insulation tubing, or the suture anchors.
- Handle the sensor gently. Excessive force can cause deformation or damage to the insulation.
- Do not handle the lead body on the clear tubing with instruments, instead grip the lead at the anchors. When suturing the sensor lead in place do not tie directly on the lead body and do not damage the insulation.




Figure 7. Sensor membrane

- Generator drop — If the generator is dropped more than 30 cm (12 in) onto a hard surface, it should not be used.
- Setscrew cautions — Counterclockwise rotation of a set screw beyond one or two revolutions while retracting it from the connector port may disengage the set screw from the connector block. Do not use any hex wrench other than the one packaged with the generator.
- Set Screw Seals — Use care when inserting the hex wrench to avoid damage to the seals. Insert the hex wrench perpendicular to the seals.
- Leads should be handled with great care at all times. Any severe bending, kinking, stretching, or handling with surgical instruments may cause permanent damage to the lead body or the cuff. Do not implant a lead that has been dropped.
- Leads attract small particles, such as lint and dust; to minimize contamination, protect the lead from materials shedding these substances. Handle the lead with sterile surgical gloves that have been rinsed in sterile water.
- Do not immerse leads in mineral oil or silicone oil.

Stimulation Lead Implant

The stimulation lead is designed with a cuff that is placed around the hypoglossal nerve after the nerve is exposed.

The following is an overview of the recommended process for implanting the stimulation lead:

- Expose the hypoglossal nerve (see “Exposing the hypoglossal nerve” below).
- Place the cuff around the nerve and irrigate the cuff and nerve with sterile saline.
- Test the electrode placement using the generator or an external nerve stimulator.
- Secure the stimulation lead anchor to the digastric muscle or tendon with permanent sutures.
- Form the generator pocket and tunnel the lead connector to the pocket.

Exposing the hypoglossal nerve

1. Make a 4–6 cm (1.6–2.5 in) incision midline between the hyoid and mandible.
2. Retract the submandibular gland posterosuperior and the mylohyoid anteriorly.
3. Ligate and divide segment of vena comitans to expose distal hypoglossal neuroanatomy in a segment of 1.5 - 2.0 cm.
4. Once the nerve is identified, it may be stimulated at a low setting (typically 0.2 - 0.5 mA) using an external nerve stimulator to confirm nerve function. Do not over stimulate the nerve with the external device.



Cautions:

- Do not apply tension to the nerve and supporting tissue while exposing the nerve and placing the cuff.
 - Preserve the small nutrient blood vessels along the nerve fibers.
 - Maintain hemostasis. Fluid residuals increase the chances of hematoma formation and infection.
-

Placing the stimulation lead

To place the stimulation lead cuff, the cuff's short inner and long outer flaps (Figure 8) are wrapped around the hypoglossal nerve.




Figure 8. Stimulation lead cuff flaps

Refer to Figure 9 while completing cuff placement steps 1 through 4.

1. Using a right-angled forceps positioned under the nerve, grasp the long outer flap.



Caution: Do not force the cuff into position. Be sure that a sufficient opening has been cleared. Forcing the cuff into position may result in nerve damage.

2. Carefully bring the outer flap underneath and around the nerve and then unfurl so it lies flat.
3. Ensure the short flap covers the nerve, then release the outer flap to close around the inner flap.



Cautions:

- Be sure that the cuff flaps are properly placed.
- Do not suture the cuff around the nerve. The cuff is designed to expand and contract with the nerve. Suturing the cuff in place may result in nerve damage.

4. Ensure all target nerve branches are enclosed within both flaps, and without undue tension or filmy adhesions interfering with flap closure and settling.
5. Irrigate between the nerve and cuff (e.g. use 18 - 20 gauge Angiocloth) with sterile saline to facilitate adequate electrical contact between the electrodes and the nerve.




Figure 9. *Placing the cuff around the hypoglossal nerve*

Securing the Stimulation Lead

The stimulation lead is anchored to the tissues surrounding the hypoglossal nerve. The suggested method for anchoring the lead body is to anchor it to the digastric tendon using permanent sutures (Figure 10).

1. Position the cuff and stimulation lead:
 - Maintain the cuff and stimulation lead body parallel to the nerve to avoid placing torque or tension on the nerve.
 - It is recommended that the spine of the cuff is positioned inferior to the nerve.
2. Secure the stimulation lead with adequate strain relief by creating a lead loop in between the stimulation lead cuff and the anchoring site (e.g. digastric tendon or muscle).
3. Using both anchor recesses, tie permanent sutures to the anchor, then secure the anchor to the digastric tendon or muscle using the sutures.
4. It is recommended that the physician not close the neck incision until all system components are implanted and tested. Consider gently packing the neck incision with 4x4 gauze soaked in a saline/antibiotic solution. Remove such packing with care prior to closing the incision so as not to dislodge or disrupt the cuff placement.



Cautions:

- Make sure that the anchor points are located in tissue that moves with the hypoglossal nerve.
 - Do not loop the lead such that the lead body crosses and touches itself. Crossing the lead bodies can result in fibrosis at the intersection point and reduce the strain relief provided by the lead body.
 - Place sutures only around the lead anchor.
 - Surgical instruments should not be used to handle the lead body directly. The lead and lead insulation is easily damaged. Care should be used when handling the lead. Surgical instruments may be used for handling the lead anchor.
-




Figure 10. Anchoring the stimulation lead

Forming the Generator Pocket

When selecting the location for the generator pocket, consider patient lifestyle factors, such as the use of firearms, carrying backpacks, and other work or recreation-related activities. The following instructions reflect the typical generator pocket location.

1. Make a 4-5 cm (1.6-2.0 in) incision mid-line 4–5 cm (1.6–2.0 in) below the right clavicle, taking precautions to ensure that the patient's typical arm movements with activities of daily living will not cause the generator to ride up onto the clavicle.
2. Make a subcutaneous pocket of sufficient size to contain the generator and any excess lead wrap, which can typically be expected. The pocket should be created no deeper than 2.5cm (1.0 inch) below the skin to allow for reliable communication between the generator and external devices.
3. Place two anchoring sutures, permanent braided 2.0 silk or equivalent, in the medial-most pectoralis major fascia and 1.0 —1.5 cm lateral of the first, forming a V-shaped sling for securing the generator's rear anchor point.

Tunneling the Lead

These instructions apply to both the stimulation lead and the respiratory sensing lead. Use the tunneling tool to pass the lead connector from the point of lead implantation to the subcutaneous pocket, avoiding sharp angle bends of the lead body.

1. Locate the sterile tunneling tool (Figure 11) provided with the stimulation lead packaging.
 - Prior to assembly, the rod may be bent into a bow shape to aide tunneling. Generally, it is better to make multiple gentle bends than a single sharp bend.
 - The tool is assembled by threading the tip and the collet assembly to the stainless steel rod. Attach the tip first and the collet only after the tunnel is established.




Figure 11. Tunneling tool components

2. Simulate the final positioning by identifying where the lead connector will exit the eventual tunnel.
3. Subcutaneously advance the tunneling tool from the lead incision to the generator until the tip is exposed in the generator pocket. Complete the tunnel before attaching the collet.

**Cautions:**

- Follow the tunneling path established in step 2. Deep tunneling is not desirable. Pass the lead superficially to avoid damage to deep structures.
- To avoid damage to the lead or body tissue, do not use excessive force or surgical instruments when using the tunneling tool.
- Tunneling the lead under the clavicle bone is not recommended. A lead tunneled under the clavicle bone creates an increased risk of damage to veins and/or arteries.
- To avoid damage to the collet, do not attach it to the tunneling tool until the tunnel is established from the lead implant site to the generator pocket.

4. Insert the lead connector into the tunneling tool collet as follows:
 - a. Slide the collet sleeve down toward the tunneling tool tip to allow the lead connector to be inserted into the collet.
 - b. Insert the pin of the lead connector into the collet of the tunneling tool (Figure 12 A)
 - c. Slide the sleeve over the collet to lock the connector pin in place (Figure 12 B).
 - d. It is not necessary to exert excessive force to secure the sleeve over the collet.




Figure 12. Inserting lead connector into tunneling tool collet

5. Gently pull the lead out through the exit site in the generator pocket.



Caution: Be sure the lead is routed so as to avoid sharp bends or kinks in the lead body.

6. Remove the lead from the tunneling tool by sliding back the sleeve from the collet.



Caution: Leave a small amount of excess lead length at both sides of the subcutaneous tunnel so that normal body motions do not stretch the leads' body. The patient may be able to feel this stretching and it may cause damage to the lead.

Respiratory Sensing Lead Implant



Cautions:

- No instrument of any type should touch the silver sensor past the blue silicone boot.
- Do not handle the lead body on the clear tubing with instruments.
- When using tools, grasp the sense lead only by the blue silicone boot, blue tubing, or the suture anchors.

The respiratory sensing lead is placed between the intercostal muscle layers in the extrapleural space (Figure 13).

The potential complications can be avoided by positioning the incision as outlined in the following steps:




Figure 13. Respiratory sensing lead extrapleural placement

1. Using the inferolateral margin of the pectoralis major as the main landmark, select the target intercostal space immediately at or superior to this margin. Make a 4–6 cm (1.6–2.4 in) incision starting near the mid-axillary line, parallel to the ribs, and toward midline on the right side of the chest.

Note: The respiratory sensing lead will be tunneled approximately 3–4 cm (1.2–1.6 in) in length between the intercostal muscle layers, and therefore the incision should be approximately 3–4 cm (1.2–1.6 in) from the desired sensor location. The desired sensor location is within an intercostal space sufficiently distant from the heart's ventricles, typically around the nipple line of a non-ptotic breast.

Note: A neurovascular bundle is located inferior to each rib. Therefore, implantation of the sensor should be as close as possible to the superior surface of the inferior rib within the target space.

2. Use sharp and blunt dissection to expose the intercostal muscle layers.
 - Dissection is required to reach and identify the external intercostal muscle.
 - The sensor will be inserted between the internal intercostal muscle and the internal intercostal muscle layers.

Note: Enter the intercostal space toward the medial side of the incision in order to provide lateral space for lead anchoring within the incision.
3. Create a 5 cm (2.0 in) tunnel between the internal and external intercostal layers with the use of a narrow malleable ribbon retractor or other similar tool.
4. Insert the tip of the respiratory sensing lead between the internal and external intercostal muscle layers at a shallow angle along the superior edge of the inferior rib forming the intercostal space.
5. Insert approximately 3–4 cm (1.2–1.6 in) of the length of the distal lead between the internal and external intercostal muscle layers.
 - The sensor membrane (flat surface) is required to face inward/deep toward the pleura.
 - Raised bumps on the distal anchor are to face out/superficial, which confirms the sensor membrane is facing inward toward the thoracic cavity.




Figure 14. Anchor orientation bumps

6. Secure the respiratory sensing lead with permanent sutures on the distal anchor. It is recommended to secure the distal anchor with at least one suture around the lead body and one anchor in each of the winglets to prevent rotation (3 total sutures).
- Ensure that the sensor membrane orientation is maintained during suturing the anchors.
 - Sutures can be secured with throw-down knots to the tissue first before device placement to minimize risk of damaging product. It is recommended to secure sutures to tissue first and then to the circumferential groove anchor points of the sensor lead with a second knot.
- Note:** Both anchors are fixed/fused to the lead body and no attempt should be made to slide either anchor along the lead body.
7. Position the proximal anchor to direct the lead toward the generator pocket. The excess lead should form an omega (or loop) shape between the two lead anchors. If excess lead length is present, perform additional blunt dissection and gently loop the lead in the additional space without kinking. Secure the proximal anchor into place on robust tissues, such as the fascia of inferolateral pectoralis major. It is recommended to secure the proximal anchor to tissue with at least one suture loop, preferably through a winglet.
8. Check that the lead body exiting the intercostal muscles transitions smoothly before tunneling to the generator pocket, forming the recommended omega-shaped strain relief between the two anchors.



Caution: Do not loop the lead such that the lead body crosses and touches itself. Crossing the lead bodies can result in fibrosis at the intersection point and reduce strain relief in the lead body.

9. Tunnel the connector end of the respiratory sensing lead to the generator pocket using the tunneling tool. Refer to “Tunneling the Lead” on page 18 for instructions.

Connecting the Leads and Generator



Caution: Saline or bodily fluids in the generator connector may reduce battery longevity.

- Do not allow saline or bodily fluids to enter the generator connector ports.
 - Confirm that lead connectors are dry prior to inserting them into the generator ports.
 - Use care when inserting the hex wrench to avoid damaging the seals, inserting the hex wrench in the center of the seal while holding it perpendicular (90 degree angle) to the surface of the generator.
 - Confirm that setscrew seals fully close after securing the lead in place.
-

Connect the respiratory sensing lead to the generator

1. Wipe off any body fluids from the respiratory sensing lead connector.
2. Grasping the lead approximately 3 cm (1.2 in) from the connector end, insert the lead connector into the generator connector port marked **SENSE** (Figure 15).
 - Make sure the lead connector is fully inserted into the generator connector port by verifying that the lead connector pin is visible past the set screw block.
 - If the lead tip is hard to visualize, verify the large seals on the lead are past the insertion indicator on the connector block.




Figure 15. Insert the respiratory sensing lead connector into generator connector port marked **SENSE**

3. Use the white-handled torque limiting hex wrench provided with the generator to tighten the set screw on the **SENSE** port. Insert the wrench at a 90 degree angle from the surface of the generator and directly in the center of the seal. Tighten until resistance is felt, and then continue until audible clicking is heard from the wrench. After the set screw is tightened, pull firmly on the lead strain relief section immediately adjacent to the generator – NOT the lead body – to confirm that the set screw has secured the lead in place. Verify the lead tip is visible past the set screw connector block. After removing the hex wrench, confirm the seal covering the set screw is fully closed and undamaged.



Figure 16. Tighten the lower setscrew

Connect the stimulation lead to the generator

1. Wipe off any body fluids from the stimulation lead connector.
2. Grasping the lead approximately 3 cm (1.2 in) from the connector end, insert the lead connector into the generator connector port marked **STIM** (Figure 17).
 - Make sure the lead connector is fully inserted into the generator connector port by verifying that the lead connector pin is visible past the set screw block.
 - If the lead tip is hard to visualize, verify the large seals on the lead are past the insertion indicator on the connector block.




Figure 17. Insert stimulation lead connector into generator connector port marked **STIM**

3. Use the white-handled torque limiting hex wrench to tighten the set screw on the **STIM** port. Insert the wrench at a 90 degree angle from the surface of the generator and directly in the center of the seal. Tighten until resistance is felt, and then continue until audible clicking is heard from the wrench. After the set screw is tightened, pull firmly on the lead strain relief segment immediately adjacent to the generator – NOT the lead body – to confirm that the set screw has secured the lead in place. Verify the lead tip is visible past the set screw connector block. After removing the hex wrench, confirm the seal covering the setscrew is fully closed and undamaged.



Figure 18. Tighten the upper setscrew

Implanting the generator

When implanting the generator, consider patient lifestyle factors, such as the use of firearms, carrying backpacks, and other work or recreation-related activities.

1. Excess lead length is desirable at both lead incisions and in the generator pocket to ensure body motions do not stretch the leads. If necessary, wrap some excess lead body behind the generator (Figure 19) and position the generator and wrapped excess lead body in the pocket. Implant the generator with the logo facing up toward the skin.




Figure 19. Wrap excess lead length



Caution: When placing the generator and leads into the subcutaneous pocket:

- Do **not** coil the leads. Coiling the leads (Figure 20) can twist the lead bodies and may result in lead dislodgement.
- Do **not** grip the leads or generator with surgical instruments.
- Do not place the leads under tension. Ensure there is a substantial amount of excess lead length in the generator pocket and at the stimulation and sensor lead incisions sites to prevent body motions from stretching the leads.
- Ensure that the generator logo is facing up toward the skin.




Figure 20. Do not coil excess lead length

System Test

Perform intraoperative testing prior to closing to confirm proper lead placement and lead-generator connections.

1. Test the stimulation function as follows:
 - Place the telemetry cable into a sterile sleeve and hold the telemetry head centered over the generator.
 - Test stimulation thresholds using the Inspire programmer Record Thresholds screen (see the programming manual for instructions). It is recommended to start at 0.5 volts and increase stimulation in 0.2 volt increments. Conduct intraoperative test stimulation while observing the tongue and neck area for signs of patient muscle response to stimulation.
 - Verify that bipolar stimulation gives the appropriate response.
 - Reposition the stimulation lead cuff if necessary. During and after repositioning of the cuff, apply saline to the cuff to facilitate electrical contact of the cuff electrodes with the nerve. Continue to reposition the cuff if a stimulation response does not occur.
 - Record the functional threshold using the Inspire programmer.
2. Test the sensor function as follows:
 - Start waveform from the Adjust Sensing screen.
 - Verify sensor function by observing the sensor waveform using the programmer. Gently and firmly tap on the patient's chest directly over the location of the sensor membrane. Confirm that these taps are clearly visible on the displayed sensor waveform and that the waveform is not fuzzy or choppy.
 - Once sensor function has been verified, turn the therapy off.
3. Following the system functional check, verify that the therapy is off and the stimulation amplitude is programmed to 0 volts. It is recommended to keep the therapy off for the first month after the implant surgery to allow for healing and encapsulation of the stimulation lead.
4. If it is necessary to disconnect a lead from the generator, use care to insert the torque limiting wrench through the set screw seals. Carefully disconnect the stimulation lead connector from the generator.



Cautions:

- Use care to not loosen the generator set screws more than necessary, which can result in the set screws unseating from the connector and damaging the generator grommets.
-

Completing the Implant Procedure

After testing, complete the implant procedure:

1. Secure both permanent sutures through one of the generator suture holes, forming a V-shaped sling. Place a clamp or other tool within loose loops as the knots are secured for each so that there is a tight knot forming a loose loop that will not cause undue tension on the anchoring sutures that could cause patient pain or lead to the anchoring sutures tearing loose from their underlying anchor points.

2. Irrigate all incision sites with a generous amount of bacitracin and saline solution or equivalent before closing.
3. Close the surgical incisions.
4. At the discretion of the physician, antibiotics may also be administered postoperatively.
5. Take at least one anterior-posterior and one lateral x-ray to document the location of all system components. Both head/neck and chest x-rays may be needed to fully capture the implanted system.
6. Place pressure dressings on all three incision sites (e.g. fluffs with Hypafix™ or Medipore™.)

Postoperative Follow-up

Follow up with normal postoperative care. A 7–14 day check of surgical incision healing is recommended.

To allow for healing after surgery, the system should not be activated for about 1 month following implant. Refer to the Inspire Programmer Manual for additional information.

Regular patient follow-up should be scheduled to monitor system status and therapy effectiveness.

Physician Instructions to Patient

Give the patient information concerning the Inspire system. This should include information on the generator, the sleep remote, the stimulation lead, and the respiratory sensing lead.

Patients should be instructed as follows:

- It is normal to feel some discomfort from the incisions and to have some pain at the implant sites for 2–6 weeks.
- It is best to avoid bending, twisting, and large arm movements for several weeks after the implant procedure, as such movements could impair the healing process. This time period allows the leads and generator to fix themselves more securely in place.
- Avoid physical activities that could damage the implant site or implanted device.
- Inform personal physicians, consulting physicians, or dentists that they have an implanted stimulation system.
- Carry their Inspire Medical Systems ID card at all times.

The “Precautions” section on page 7, which includes information about cellular phones and electromagnetic interference in the home or work environment, should also be conveyed to the patient.

Patient Registration

The implanted components of the Inspire system (generator and leads) are subject to the Food and Drug Administration’s Medical Device Tracking Requirements (21 CFR 821). A device registration form must be completed by the implanting (or explanting) clinician and returned to Inspire Medical Systems. The information provided on this form is required for Inspire to meet government obligations for device tracking, product safety, effectiveness, performance and government event reporting and is a public health disclosure under Section 164.512(b)(1)(iii).

Therapy Activation

Inspire therapy should be activated approximately 4 weeks after the implant procedure to allow for healing.

Therapy Titration

At least one sleep study will be needed approximately 4–8 weeks after therapy activation to titrate stimulation settings. Additional titration sleep studies may be needed to improve therapy effectiveness and patient comfort.

Surgical Revision and Explant

Lead Repositioning

- If the stimulation or respiratory sensing lead becomes displaced, any repositioning should be attempted as soon as possible, before scar tissue builds up.
- If the lead must be repositioned (or removed) proceed with caution to avoid damage to surrounding tissue.
- Extreme forces used during removal can damage leads or result in dismantling of the leads.
- If removal is unavoidable, return the removed lead, or portion thereof, to Inspire Medical Systems.

Generator Replacement

- Use x-ray imaging to plan the surgical approach to avoid the implanted leads.
- Make incision to expose the generator. During dissection, take care not to cut the lead bodies or damage any lead body insulation.
 - CAUTION: Avoid the use of unipolar electrocautery. Unipolar electrocautery can be transmitted along the lead body leading to permanent nerve damage or damage to the generator.
 - Prior to removing the outgoing generator, carefully cut any anchoring sutures, then carefully remove the IPG enough to access the set screws. Use care during generator removal and avoid forces that may damage the leads.
- Loosen the set screws using the wrench provided with the new generator. Use gentle traction to remove the leads from the expired generator.
 - CAUTION: To avoid damaging the lead, grasp leads by the terminal connector and not the lead body during removal from the generator.
- Disrupt the fibrotic capsule to ensure that the new IPG is readily recognized by the immune system for maximum foreign body response. Use care to avoid damage to the implanted leads.
- Ensure leads are wiped clean and dry prior to inserting into the new generator.
 - Connect the respiratory sensor lead to the generator
 - Connect the stimulation lead to the generator
 - Implanting the generator
 - System Test
 - Completing the Implant Procedure
- If the new generator is smaller than the expired device then the following considerations apply:
 - Use both anchoring holes in the IPG to secure the new, smaller IPG
 - Consider placing “roof stitches” in the superior aspect of the fibrotic capsule to downsize the pocket, prior to conducting final, multi-layer closure
- Before final closing, carefully examine the generator pocket to eliminate any micro-bleeding and make sure hemostasis is established.
- Return the expired generator to Inspire Medical Systems in a biohazard container and mailer for analysis and disposal. Your local Inspire support personnel can facilitate this process as needed.

System or Generator Explant

- Extreme forces used during removal can damage the lead or result in dismantling of the lead.
- A lead that has been cut off should have the remaining lead end sealed. If the leads are left in place, the proximal connector ends of the leads should be capped to minimize tissue irritation and induced currents.
- Lead removal may not be possible due to the risk of damaging surrounding structures. The decision to remove the leads or leave them in place is made between the physician and the patient on a case-by-case basis. The implications of both options should be discussed, for example:
 - Removing the leads will extend the duration of the surgical procedure, require two additional incisions, and require the dissection of fibrotic tissue that may have formed around the leads.
 - A partially explanted system is MR Unsafe, which will prevent the patient from receiving an MRI. Furthermore, patients must be made aware that they need to notify medical personnel that they still have implanted leads even if the generator has been removed and the lead ends have been capped.
- Return all explanted components to Inspire Medical Systems for disposal.

Explant Disposition

When replacing or explanting a generator or lead, return the generator or lead to Inspire Medical Systems for analysis and disposal. See the back cover of this manual for mailing address.

Clinical Summary

Stimulation Therapy for Apnea Reduction (STAR) Clinical Trial

The Inspire Upper Airway Stimulation (UAS) system was evaluated in a multi-center trial at study centers in the United States and Europe for the indication of moderate to severe obstructive sleep apnea (OSA) in patients who were not effectively treated by continuous positive airway pressure (CPAP).

Patients Studied

The study enrolled 929 OSA patients. These patients were evaluated against patient selection criteria that included moderate to severe OSA, a BMI (body mass index) less than or equal to 32, and the absence of a complete concentric collapse at the level of the soft palate. Following the evaluation period, 126 patients met all selection criteria and proceeded to implant. All 126 implant procedures were successful, and 124 of the 126 implanted patients provided evaluable data through at least 12 months. The STAR trial was an intent-to-treat study. Therefore, the 2 patients who did not provide evaluable data through 12 and 18 months post-implant are assumed to be non-responders and were included in the evaluation as such. The patient demographics for the STAR trial are included in Table 1. The patients' baseline AHI showed a mean of 32.0 and a median of 29.3, and the baseline ODI showed a mean of 28.9 and a median of 25.4.

Table 1. STAR Trial Subject Demographics

Continuous Measures	Mean N = 126	Median
Age, year	54.5	55
Body Mass Index, kg/m ²	28.4	29.2
Neck Size, cm	41.2	41.9
Systolic BP, mmHg	128.7	128
Diastolic BP, mmHg	81.5	80.5
Male	105 (83%)	Total N = 126
Race		
Caucasian	122 (97%)	
African American	0 (0%)	
Hispanic	1 (1%)	
Asian	1 (1%)	
Others*	2 (2%)	* 1-Surinam, 1-Turkey

Study Design and Methods

The STAR trial was a multi-center, prospective trial with a 12-month single arm study and a randomized controlled therapy withdrawal study at 13 months. Following implant of the Inspire system, patients were followed at 1, 2, 3, 6, 9, 12, 13, 15, 18 months, and every 6 months thereafter. The patients' baseline AHI and ODI (oxygen desaturation index) values were the mean results from their screening (pre-implant) and 1-month (post-implant but prior to therapy activation) sleep studies. Baseline results were compared to the 12-month results to determine the percentage of patients who experienced a clinically meaningful reduction in the severity of their OSA in terms of their AHI and ODI scores. For this study, a clinically meaningful reduction in AHI and ODI was defined as (1) a 50% reduction in the AHI compared to the pre-implant screening and 1-month visit (post-implant but prior to therapy activation) and an AHI < 20 events per hour, and (2) a 25% or greater reduction in ODI at the 12-month visit compared to baseline.

Upon completion of the overnight sleep study at the 12-month visit, a randomized controlled therapy withdrawal study was conducted. The first 46 responders were randomized 1:1 to either the therapy maintenance (ON) group or the therapy withdrawal (OFF) group, resulting in 23 subjects in each group. Patients randomized to the therapy withdrawal group had Inspire therapy turned OFF for at least five days. Patients randomized to the therapy maintenance group continued their use of the Inspire system. All randomized patients participated in a sleep study at the 13-month visit. The therapy withdrawal group had the sleep study performed with Inspire therapy OFF, and the therapy maintenance group had the sleep study performed with the Inspire therapy ON. The mean change of AHI for each arm was compared to determine the extent of treatment effect from Inspire therapy.

The percentage of sleep time a patient had an oxygen saturation (SaO₂) level below 90% was recorded during the sleep studies, and two validated quality of life questionnaires were administered at follow-ups through 18 months. The quality of life questionnaire was the Epworth Sleepiness Scale (ESS), which rates a patient's daytime sleepiness, and the Functional Outcomes of Sleep Questionnaire (FOSQ), which assesses the effect of a patient's daytime sleepiness on activities of ordinary living. The hypotheses for the secondary efficacy endpoints, which included the randomized withdrawal study, FOSQ, ESS, and SaO₂, were tested according to a hierarchical strategy in order to preserve an overall Type I error rate of 5%.

Study Results

Titration

All subjects underwent polysomnography (PSG) for titration of therapy settings at 2 and 6 months. Additional titration PSG studies were performed as needed. Through 18 months, patients had an average of 3.3 (range 2–6) titration studies.

Safety

Of the 126 patients implanted with the Inspire UAS system in the STAR trial, 124 were followed through 18 months. There were no unanticipated events and only 2 events required surgical intervention. Both events consisted of an IPG migrating out of position and were resolved with a surgical procedure performed under local anesthesia to reposition the IPG.

Many of the procedure-related adverse events reported are expected with a surgical procedure. The procedure-related events are described in Table 2.

**Table 2. Procedure-Related Adverse Events
(and the probability of experiencing them in the first 18 months)**

Event	Number of Subjects with Event	Percent of Subjects (n=126)
Incision pain	35	28%
Post-operative discomfort	31	25%
Temporary tongue weakness	23	18%
Sore throat from intubation during implant	15	12%
Other post-operative symptoms (such as gastrointestinal (nausea, vomiting, abdominal pain, constipation), body pain (back, knee, wrist, hand), allergy to antibiotics, anxiety, ineffective airway clearance, loss of some taste, inability to void)	14	11%
Headache	8	6%
Mild infection	1	1%

The device-related adverse events are described in Table 3.

**Table 3. Device-Related Adverse Events
(and the probability of experiencing them in the first 18 months)**

Event	Number of Subjects with Event	Percent of Subjects (n=126)
Discomfort due to electrical stimulation	59	47%
Tongue abrasion	30	24%
Other acute symptoms (i.e., headaches, coughing, choking, dysphasia, and speech-related events)	23	17%
Mouth dryness	14	11%
Complaints related to temporary usability or functionality issues with an implanted device	13	11%
Complaints related to temporary usability or functionality issues with an external device	13	10%
Mechanical pain associated with presence of device	10	8%
Mild infection	1	1%

At the completion of the 18-month follow-up visits of all study patients, 75% of device-related events were fully resolved, primarily with either medication, device reprogramming, dental work to fix a jagged tooth, or with the aid of a lower tooth guard used during sleep to prevent tongue abrasions, or no intervention. Twenty-five percent (25%) of device-related events were unresolved at 18 months. Currently unresolved events include reports of discomfort due to stimulation, tongue abrasion and various stimulation related events including dry mouth, headaches, intermittent waking, isolated stimulation sensation events, audible buzzing, and intermittent fatigue. Despite these reported events, patients continued to report high (85%) compliance with the therapy at 18 months.

Two subjects had their devices removed, which required a surgical procedure. One chose to have the stimulator removed, and the leads were capped and left in the patient. The other had the entire system removed as a precaution due to proximity to an unrelated infection. Both explants were successfully completed without damage to the surrounding structures. There were 3 deaths over the course of the study, all were unrelated to Inspire therapy. There were 32 serious adverse events (SAE), 2 of which were related to Inspire therapy, both involving repositioning of the IPG.

Efficacy

The sleep studies, which were scored by an independent sleep scoring core lab, showed statistically significant and clinically relevant reductions in the patients' AHI and ODI scores. Table 4 reports the percentage of patients who experienced a clinically meaningful reduction in their OSA severity (i.e., responders). As this is an intent-to-treat study, these results are based on a total of 126 patients even though only 124 patients provided evaluable data through 12 and 18 months. The other 2 patients are assumed to be non-responders and are included in the evaluation as such.

Table 4. Therapy Responders at 12 Months Post-Implant

Responder	Responder Rate at 12-Month Follow-Up	Responder Rate at 18-Month Follow-Up
50% Reduction in AHI from baseline and AHI < 20	66% (83/126)	65% (80/124)
25% Reduction in ODI from baseline	75% (94/126)	80% (99/124)

The average reduction of AHI from baseline to 12 months was 68% and 70% for ODI. Baseline AHI showed a mean of 32.0. In comparison, the AHI at the 12-month PSG study showed a mean of 15.3. Baseline ODI showed a mean of 28.9. In comparison, ODI at the 12-month PSG study showed a mean of 13.9. The patients also had statistically significant improvements in terms of time with $\text{SaO}_2 < 90\%$, ESS and FOSQ scores at 12 months relative to baseline. The mean FOSQ score at baseline was 14.3, at the 12-month visit it was 17.2, and at the 18-month visit it was 17.3. The mean ESS score at baseline was 11.6, at the 12-month visit it was 7.0, and at the 18-month it was 7.0. The mean percentage of sleep time with $\text{SaO}_2 < 90$ at baseline was 8.7%, at the 12-month visit it was 5.9%, and at the 18-month visit it was 5.6%. These results through 18 months show the durability of Inspire therapy's treatment effect.

The randomized controlled therapy withdrawal study provided further evidence that improvements were attributed directly to the Inspire therapy. AHI increased significantly in the therapy withdrawal (OFF) group compared to AHI scores in the therapy maintenance (ON) group. The results from the randomized control therapy withdrawal study showing the difference between the therapy OFF arm and the therapy ON arm are provided in Table 5.

Table 5. Randomized Controlled Therapy Withdrawal Study Results in Month 13

AHI	Mean AHI		Change (13M–12M) Mean	95% CL for Mean Change	p-value
	12-Month	13-Month			
Therapy ON	7.2	8.9	1.7	(-1.1, 4.5)	< 0.0001
Therapy OFF	7.6	25.8	18.2	(11.4, 24.9)	

The randomized controlled therapy withdrawal study confirmed that the significant OSA severity reduction at 12 months is attributable to the Upper Airway Stimulation therapeutic effect. An analysis of AHI responder status relative to baseline characteristics is provided in Table 6

.Table 6. AHI Responder Analysis of Baseline Characteristics

Baseline Characteristics	Responders N = 83 Mean % (N)	Non-responders N = 43 Mean % (N)	Association of AHI Response to Baseline Characteristics p-value
Age	55.9	51.8	0.03
Gender (% Male)	82%	86% (37)	0.56
BMI	28.3	28.6	0.50
Neck Size	41.0	41.6	0.32
Baseline AHI	30.7	34.6	0.08
Baseline ODI	27.1	32.3	0.02
Prior UPPP (%)	20.5% (17)	11.6% (5)	0.22
Baseline FOSQ	14.7	13.6	0.059
Baseline ESS	11.2	12.3	0.22

While the percentage of patients with prior UPPP surgery is noted to be twice as high in the responder group as compared to the non-responder group, the observation was not statistically significant (p-value of 0.22).

Conclusion

Upper Airway Stimulation is a safe and effective treatment for patients with moderate to severe OSA who are not effectively treated by CPAP.

STAR Clinical Trial Extended Follow Up

Study Objectives

To continue following subjects from the STAR trial in the post-market environment.

Study Design

This was a prospective, single arm cohort study to evaluate the long-term safety of the device in subjects implanted with the Inspire UAS system under the STAR premarket study. The subjects were followed for five years post implant.

Study Population

See the “Patients Studied” Section above in the clinical summary of the STAR trial.

Data Source

The final data in the STAR extended follow-up study through five years post implant.

Key Study Endpoints

See the “Study Design and Methods” Section above in the clinical summary of the STAR trial.

Study Visits and Length of Follow Up

During the premarket phase of the STAR trial all subjects had completed their 18-Month follow up at a minimum. After the 18-Month follow-up visit, subjects continued to be followed long term at 6-month intervals. These long-term follow-ups included adverse event assessment, system interrogation, and therapy titration (as required).

Total Number of Enrolled Study Sites and Subjects’ Follow Up Rate

Of the 126 subjects implanted in the STAR trial, 97 (77%) subjects completed the 60-month final STAR follow-up visit at 18 sites. The final status of the subject follow-up visit compliance is presented in Table 7.

Table 7. Cumulative Implanted Subjects Accountability by Visit

Subjects	Implant	6-month visit	12-month visit	18-month visit	24-month visit	30-month visit	36-month visit	42-month visit	48-month visit	54-month visit	60-month visit
Implanted	126	126	126	126	126	126	126	126	126	126	126
Death (Unrelated)	0	0	1	1	2	3	3	3	3	4	5
Withdrawn	0	0	1	1	2	4	7	11	13	15	24
Eligible at Visit	126	126	124	124	122*	119	116	112	110	107	97
Attrition Rate	0	0	1.6%	1.6%	3.2%	5.6%	7.9%	11.1%	12.7%	15.1%	23.0%
Visit Completed at Interval	126	125	124	123	118	114	110	96	96	90	92
Missed Visit	0	1	0	1	4	5	6	16	14	17	5
Missed Visit Rate at Interval	0	0.8%	0	0.8%	3.2%	5.6%	7.9%	14.1%	12.7%	15.9%	5.2%

*One subject had a medically necessary explant unrelated to the Inspire device after the 18-month visit but continued to be followed (post-explant) for safety reasons and exited after completing the 24-month visit.

Safety

Procedure Related Adverse Events (AE)

A majority of the procedure-related events were reported within the first 12 months post-implant. However, a few late experiences of discomfort in the area of incisions or device were reported. One of the later events resulted from inadvertent severing of the stimulation lead during a standard procedure to replace the IPG and sensing lead, and thus required replacement of the entire system. All other late-reported events were minor and resolved with medication or no intervention, except for one report of a recurrent cyst at the sensor lead incision site, which eventually resolved without sequelae. Procedure-related non-serious AE's through the 60-month follow up are presented in Table 8.

Table 8. Procedure Related Adverse Events through the 60-Month Follow Up

Adverse Event	Total		Fully Resolved		Partially Resolved		Ongoing	
	# Events	% (# of Subjects)	N	% of Total	N	% of Total	N	% of Total
Events specifically related to an incision	52	30.2% (38)	47	90.4%	5	9.6%	0	0.0%
Post-operative discomfort independent of any surgical incision	42	27.0% (34)	40	95.2%	2	4.8%	0	0.0%
Acute tongue weakness	34	18.3% (23)	32	94.1%	2	5.9%	0	0.0%
Intubation Effects	18	11.9% (15)	18	100.0%	0	0.0%	0	0.0%
Headache	8	6.3% (8)	8	100.0%	0	0.0%	0	0.0%
Other post-op symptoms (such as gastrointestinal nausea, vomiting, abdominal pain, constipation), body pain (back, knee, wrist, hand), allergy to antibiotics, anxiety, ineffective airway clearance, loss of some taste, inability to void.	22	11.1% (14)	22	100.0%	0	0.0%	0	0.0%
Infection (mild or moderate)	1	0.8% (1)	1	100.0%	0	0.0%	0	0.0%

Device Related Adverse Events (AE)

Device related events include those resulting from the presence of the device, or delivery of and/or response to the therapy. Most device related events were reported in the first 1-2 years after therapy activation. During this timeframe subjects are acclimating to the therapy as optimal settings are established. Of the device-related events reported after the 36-month follow-up, 21 (34%) were classified as temporary external device usability or functionality complaints, which are intended to describe functional issues with the external component operation. These were most often issues with patient remote functioning and were generally resolved with retraining on, or replacement of, the patient programmer. There were also 12 reports (27%) of discomfort due to electrical stimulation which were generally resolved either with adjustments in programming or required no intervention. Device-related non-serious AE's through the 60-month follow up are presented in Table 9.

Table 9. Device-related Non-serious AE's through the 60-Month Follow Up

Adverse Event	Total		Fully Resolved		Partially Resolved/ Death		Fully Resolved/ Exit	
	# Events	% (# of Subjects)	N	% of Total	N	% of Total	N	% of Total
Discomfort due to electrical stimulation	142	60.3% (76)	120	84.5%	17	12.0%	5	3.5%
Tongue abrasion	49	27.0% (34)	42	85.7%	4	8.2%	3	6.1%
Mouth dryness	20	15.1% (19)	7	35.0%	9	45.0%	4	20.0%
Mechanical pain associated with presence of the device	14	11.1% (14)	9	64.3%	4	28.6%	1	7.1%
Temporary Internal Device Usability or Functionality Compliant	25	16.7% (21)	19	76.0%	5	20.0%	1	4.0%
Temporary External Device Usability or Functionality Complaint	45	26.2% (33)	45	100.0%	0	0.0%	0	0.0%
Other acute symptoms*	39	24.6% (31)	26	66.7%	12	30.8%	1	2.6%
Infection (mild or moderate)	1	0.8% (1)	1	100.0%	0	0.0%	0	0.0%

*Other acute symptoms include miscellaneous reports attributed to the system; for example instances of morning headache, face/neck ache or soreness, insomnia/fatigue, coughing/ choking, and minor dysphasia, tongue deviation or speech related events.

Serious Adverse Events (SAE)

During the five-year follow-up period in the STAR pivotal trial there were nine related serious adverse events (SAE) in eight (6%) of the patients. All SAE's involved a revision/repositioning or replacement of the Inspire system or component(s) of the Inspire system. One revision occurred during the first 12 months of follow-up, and four additional revisions were required during the 12- to 60-month follow-up period. One revision was to reposition the stimulation lead to improve therapy effectiveness, and the others were to address patient discomfort. During the last two years of follow-up four subjects required replacement of the system or a component(s) of the system to address device performance issues in order to improve or restore therapy. The revision and replacement surgeries were without complications or significant sequelae. In addition, one patient had their system explanted due to concerns over a serious infection in their sternoclavicular joint, however upon explant it was determined that there were no signs of infection around the Inspire system components.

Efficacy

Although not required by the long term follow up phase of the protocol, the collection of AHI and ODI measurements was permitted and these objective measures of OSA severity were collected on a majority of patients. The long-term results were consistent with those of the premarket pivotal trial thereby demonstrating a durable therapeutic effect. The baseline, 12-month, 36-month, and 60-month AHI and ODI results are presented in Table 10.

Table 10. Long Term AHI and ODI Results

Outcome Measure	Baseline N Mean \pm SD Median	Month 12 N Mean \pm SD Median	Month 36 N Mean \pm SD Median	Month 60 N Mean \pm SD Median
AHI	126 32.0 \pm 11.8 29.3	124 15.3 \pm 16.1 9.0	98 11.5 \pm 14.0 6.0	71 12.4 \pm 16.3 6.2
ODI (4%)	126 28.9 \pm 18.2 25.4	124 14.0 \pm 15.6 7.4	98 9.1 \pm 11.7 4.8	71 9.9 \pm 14.5 4.6

As was the case in the STAR pivotal trial, the Functional Outcomes Sleep Questionnaire (FOSQ) and Epworth Sleepiness Scale (ESS) results continued to show significant improvement in the subjects' quality of life throughout long term follow up. These validated instruments are commonly used in clinical evaluation and management of OSA. The FOSQ Scores range from 5 to 20, with higher scores indicating greater functioning. A FOSQ score greater than 17.9 is considered to be the cut point for subjects free of any sleep disorders. A change of two or more points in the FOSQ score is a clinically meaningful improvement in daily functioning. The results measured during long-term follow-up compared to baseline are presented in Table 11.

Table 11. Functional Outcomes of Sleep Questionnaire

FOSQ and Change from Baseline by Visit		
Visit	FOSQ (N) Mean \pm SD Median (Min, Max)	Change from Baseline (N) Mean \pm SD Median (Min, Max)
Month 24	(111) 17.5 \pm 2.8 18.6 (6.5, 20.0)	(111) -3.0 \pm 3.0 -2.8 (-11.9, 3.2)
Month 30	(112) 17.6 \pm 2.7 18.3 (5.0, 20.0)	(112) -3.1 \pm 3.3 -2.7 (-12.1, 9.6)
Month 36	(113) 17.3 \pm 3.5 18.8 (5.7, 20.0)	(113) -2.7 \pm 3.8 -2.6 (-12.7, 10.1)
Month 48	(94) 17.6 \pm 2.7 18.8 (6.4, 20.0)	(94) -3.0 \pm 3.6 -2.6 (-12.1, 9.0)
Month 54	(90) 17.7 \pm 3.1 18.8 (5.4, 20.0)	(90) -2.9 \pm 3.7 -3.1 (-10.7, 9.4)
Month 60	(92) 18.0 \pm 2.2 18.7 (8.5, 20.0)	(92) -3.2 \pm 2.9 -3.1 (-12.1, 3.4)

The ESS rates a subject's daytime sleepiness with scores ranging from 0 to 24, lower scores indicating greater functioning. An ESS score of less than 10 is the cut point for normal subjective sleepiness. The results measured in STAR at annual follow-ups compared to baseline are presented in Table 12.

Table 12. Epworth Sleepiness Scale

ESS and Change from Baseline by Visit		
Visit	FOSQ (N) Mean \pm SD Median (Min, Max)	Change from Baseline (N) Mean \pm SD Median (Min, Max)
Month 24	(111) 17.5 \pm 2.8 18.6 (6.5, 20.0)	(111) -3.0 \pm 3.0 -2.8 (-11.9, 3.2)
Month 30	(112) 17.6 \pm 2.7 18.3 (5.0, 20.0)	(112) -3.1 \pm 3.3 -2.7 (-12.1, 9.6)
Month 36	(113) 17.3 \pm 3.5 18.8 (5.7, 20.0)	(113) -2.7 \pm 3.8 -2.6 (-12.7, 10.1)
Month 48	(94) 17.6 \pm 2.7 18.8 (6.4, 20.0)	(94) -3.0 \pm 3.6 -2.6 (-12.1, 9.0)
Month 54	(90) 17.7 \pm 3.1 18.8 (5.4, 20.0)	(90) -2.9 \pm 3.7 -3.1 (-10.7, 9.4)
Month 60	(92) 18.0 \pm 2.2 18.7 (8.5, 20.0)	(92) -3.2 \pm 2.9 -3.1(-12.1, 3.4)

Study Strengths and Weaknesses

The follow up rate was quite high during the premarket phase of the study with 124 out of 126 patients completing the primary endpoint follow up. However, during long term follow up in the post market phase of the study the follow up rate dropped to 76.9% (i.e. 97 out of 126 patients).

Conclusion

The STAR trial exceeded all primary and secondary efficacy endpoints, providing the majority of subjects with clinically significant reductions in OSA severity and meaningful improvements in quality of life. The long-term follow up visits through 60 months further demonstrated that Inspire's therapeutic effects are durable and robust.

Pediatric Extrapolation:

Existing adult clinical data was leveraged to support Inspire therapy in the pediatric sub-population of adolescents age 18 to 21. There is no data to support use of Inspire UAS in a general pediatric population less than 18 years old. Additionally, published literature data^{1,2,3,4,5} indicates there may be a significant chance for spontaneous remission of OSA in pediatric patients who are still growing and may experience changes to their airway and disease characteristics.

References:

- 1.Spilsbury JC, Storer-Isser A, Rosen CL, Redline S. Remission and incidence of obstructive sleep apnea from middle childhood to late adolescence. *Sleep*. 2015;38(1):23 -29. Published 2015 Jan 1. doi:10.5665/sleep.4318
- 2.Chan KC, Au CT, Hui LL, Ng SK, Wing YK, Li AM (2019 Jul) How OSA evolves from childhood to young adulthood: natural history from a 10-year follow-up study. *Chest* 156(1):120 -130)
- 3.Edward O. Bixler, Julio Fernandez-Mendoza, Duanping Liao, Susan Calhoun, Sol M. Rodriguez Colon, Jordan Gaines, Fan He, Alexandros N. Vgontzas: Natural history of sleep disordered breathing in prepubertal children transitioning to adolescence. *Respiratory Journal* 2016 47: 1402-1409; DOI: 10.1183/13993003.01771-2015
- 4.Goodwin JL, Vasquez MM, Silva GE, Quan SF. Incidence and remission of sleep-disordered breathing and related symptoms in 6- to 17-year old children--the Tucson Children's Assessment of Sleep Apnea Study. *J Pediatr*. 2010;157(1):57 -61. doi:10.1016/j.jpeds.2010.01.033
- 5.Chervin RD, Ellenberg SS, Hou X, et al. Prognosis for Spontaneous Resolution of OSA in Children. *Chest*. 2015;148(5):1204 -1213. doi:10.1378/chest.14-2873

IPG Specifications

Factory Settings

Table 13. Inspire Implantable Pulse Generator (Model 3028) Factory Settings

Parameter	Value
General	
Therapy On/Off	Off
Usage	0
Start Delay	30 mins
Pause Time	15 mins
Therapy Duration	8 hrs
Stimulation	
Amplitude	0 V
Rate	33 Hz
Pulse Width	90 μ s
Electrode Configuration	Outer (+) Center (-) Case (off)
Patient Control	Off
Sensing	
Exhalation Sensitivity Threshold	-4 -1
Inhalation Sensitivity Threshold	0 +1
Refractory Hard Soft	38% 13%
Invert Signal	Off
Max Stim Time	4 secs

Configurable Settings

The parameters in Table 14 can be changed using an Inspire programmer. See the physician programmer manual for more information.

Table 14. Inspire Implantable Pulse Generator (Model 3028) Configurable Settings

Parameter	Values	Increment
Stimulation		
Start Delay	0–75 mins	5 mins
Pause Time	5–30 mins	5 mins
Therapy Duration	1–15 hrs	1 hr
Amplitude	0.0–5.0 V	0.1 V
Rate	20, 25, 30, 33, 40 Hz	
Pulse Width	60, 90, 120, 150, 180, 210 μ s	
Electrode Configuration	Outer (+) Center (-) Outer (-) Center (+) Case (+) Outer (-) Case (+) Center (-) Case (+) Outer (-) Center (-)	
Patient Amplitude Control	On, Off	
Sensing		
Exhalation Sensitivity	-4 to +3	1
Exhalation Threshold	-1, 0, +1	1
Inhalation Sensitivity	-7 to 0	1
Inhalation Threshold	0, +1	1
Hard Off Period	38, 50, 63, 75%	
Soft Off Period	13, 25%	
Invert Signal	On, Off	
Max Stim Time	2–5 secs	1.0 sec

Battery Longevity

Longevity Estimate	10.9 years average (0.6 years standard deviation) ^a
End of Service ^c	1 month after Recommended Replacement Time ^b

(a) Longevity estimate is based on STAR trial therapy settings at the 12-month endpoint. IPG longevity will vary based on usage and therapy settings. The minimum estimated longevity from the STAR trial is 7 years.

(b) Recommended Replacement Time - The sleep remote generator light turns on to indicate that replacement of the Model 3028 is recommended within 1 month.

(c) End of Service - The Model 3028 should be replaced immediately.

Physical Description

Table 15. Inspire Implantable Pulse Generator (Model 3028) Physical Description

Description	Value
Height	46mm (1.8 in)
Length	51mm (2.0 in)
Thickness	8.4mm (0.33in)
Volume	15cm ³ (0.92in ³)
Radiopaque identification	IMS1
Tissue contacting materials	Titanium, polyurethane, silicone rubber

Radiopaque identification

The IPG's radiopaque identification, IMS1 (Figure 21), can be confirmed by using fluoroscopy on the IPG.




Figure 21. Radiopaque identification

Inspire Medical Systems Limited Warranty

Summary

Inspire provides a limited warranty against defects. The warranty period for implanted products is 3 years. All other products have a warranty period of 1 year. The warranty information below is intended for doctors (referred to as physicians in the warranty), but is included here for reference. Ask your doctor if you have any questions. The information below takes precedence over the information contained in this Summary.

Inspire Medical Systems' products consist of Implantable Pulse Generators (IPG), tools to connect the IPG to implantable leads, leads, Inspire Sleep Remotes, and physician programmers.

- 1. EXCLUSION OF WARRANTIES, NO WARRANTIES FOR TOOLS.** The implied warranties of MERCHANTABILITY and fitness for a particular purpose and all other warranties, express or implied with regard to tools are EXCLUDED from any transaction and shall not apply. Inspire Medical Systems will not be liable for any damages, whether direct, consequential, or incidental caused by tool defects, failures, or malfunctions, whether such claims are based on warranty, contract, tort or otherwise. No person has any authority to bind Inspire Medical Systems to any representation or warranty with respect to tools. You may have other rights, which vary from state to state. If one or more of the provisions of this exclusion of warranties for tools shall be deemed void or unenforceable, the remaining provisions shall continue to have full force and effect.
- 2. LIMITED WARRANTY FOR PRODUCTS OTHER THAN TOOLS.** This limited warranty is available if products other than tools fail to function within normal tolerances due to defects in materials or workmanship that manifest during the specified warranty period.

During the operational life of an IPG, battery energy is consumed to monitor the patient's breathing and provide therapy. On the basis of individual patient physiology, certain patients may require more frequent therapy, thus requiring replacement of the IPG in less than the warranty period shown below. This is considered normal for those patients and not a malfunction or defect in the IPG.

If the purchaser complies with the Terms and Conditions, Inspire Medical Systems will issue a limited warranty toward the purchase of a new Inspire Medical Systems IPG product. The limited warranty credit amount will be the full purchase price of either the original unit or the replacement unit, whichever is less.

- For patient products, for example, IPG, lead, Inspire Sleep Remote, Inspire Medical Systems will issue a credit to the hospital conducting replacement surgery on behalf of the original patient. Any cost reductions extended as a result of this warranty shall be fully and accurately reflected on the patients' bill and reported to that applicable payor using the appropriate methodology.
- For physician products, for example, physician programmer, Inspire Medical Systems will issue a credit to the original purchaser of the product.

A. Terms and Conditions

1. The product labeling must indicate a limited warranty exists.
2. For implantable products, this limited warranty applies only for a product replacement in the original patient.
3. All registration materials must be completed and returned to Inspire Medical Systems within 30 days of first use.
4. The product must be replaced with an Inspire Medical Systems product.
5. If the product is implantable, it must be implanted before the product expires and implanted with other Inspire Medical Systems products.
6. The product must be returned to Inspire Medical Systems, 5500 Wayzata Blvd, Suite 1600, Golden Valley, MN 55416 within 30 days that the product first fails to function within normal tolerances. The product may be returned at no cost to you. Contact your Inspire Medical Systems representative for information on how to return the product.
7. Inspire Medical Systems will inspect the returned product and determine whether a limited warranty credit is due.
8. All products returned to Inspire Medical Systems become its property.

This limited warranty represents the entire obligation of Inspire Medical Systems for products other than tools and is made IN LIEU OF any other warranties, whether express or implied, including MERCHANTABILITY or fitness for a particular purpose.

Inspire Medical Systems will not be liable for any damages, whether direct, consequential, or incidental caused by product defects, failures, or malfunctions, whether such claims are based on warranty, contract, tort or otherwise.

No person has any authority to bind Inspire Medical Systems to any warranty or representation except those specifically contained herein.

This limited warranty gives specific legal rights, and you may also have other rights, which vary from state to state. If one or more of the provisions of this limited warranty shall be deemed void or unenforceable, the remaining provisions shall continue to have full force and effect.

B. Limited Warranty Period

The applicable limited warranty period for each product is listed and calculated as follows:

1. Three (3) years from date an IPG or lead is implanted in the patient.
2. One (1) year from the date a physician programmer or Inspire Sleep Remote is first used.



Manufacturer

Inspire Medical Systems, Inc.
5500 Wayzata Blvd, Suite 1600
Golden Valley, MN 55416
USA
Tel. 844-672-4357
763-205-7970
Fax 763-537-4310
www.inspiresleep.com